



September 24, 2019

MAKO Surgical Corp.  
Shikha Khandelwal  
Senior Staff Regulatory Affairs Specialist  
2555 Davie Road  
Fort Lauderdale, Florida 33317

Re: K191998

Trade/Device Name: Mako Total Hip Application, Mako Total Knee Application  
Regulation Number: 21 CFR 882.4560  
Regulation Name: Stereotaxic Instrument  
Regulatory Class: Class II  
Product Code: OLO  
Dated: July 23, 2019  
Received: July 26, 2019

Dear Shikha Khandelwal:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shumaya Ali, M.P.H.  
Assistant Director  
DHT6C: Division of Restorative, Repair  
and Trauma Devices  
OHT6: Office of Orthopedic Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure  
List of Cleared Devices in K191998

**List of Cleared Devices in K191998**

1. Mako Total Hip Application
2. Mako Total Knee Application

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
Food and Drug Administration

Form Approved: OMB No. 0910-0120  
Expiration Date: 06/30/2020  
See PRA Statement below.

## Indications for Use

510(k) Number (if known)

K191998

Device Name

Mako Total Knee Application

Indications for Use (Describe)

The Mako System is intended to assist the surgeon in providing software defined spatial boundaries for orientation and reference information to anatomical structures during orthopedic procedures.

The Mako System is indicated for use in surgical knee procedures in which the use of stereotactic surgery may be appropriate, and where reference to rigid anatomical bony structures can be identified relative to a CT based model of the anatomy. These procedures include:

- Total Knee Arthroplasty (TKA)

The implant systems compatible with the system:

- Triathlon Total Knee System (CR/CS/PS cemented and cementless primary)
- Triathlon Total Knee System (TS inserts cemented primary)
- Kinetis Total Knee System (CR/UC)

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
Food and Drug Administration

Form Approved: OMB No. 0910-0120  
Expiration Date: 06/30/2020  
See PRA Statement below.

## Indications for Use

510(k) Number (*if known*)

K191998

Device Name

Mako Total Hip Application

Indications for Use (*Describe*)

The Mako System is intended to assist the surgeon in providing software defined spatial boundaries for orientation and reference information to anatomical structures during orthopedic procedures. The Mako System is indicated for use in surgical knee and hip procedures in which the use of stereotactic surgery may be appropriate, and where reference to rigid anatomical bony structures can be identified relative to a CT based model of the anatomy.

These procedures include:

- Unicondylar knee replacement and/or patellofemoral knee replacement
- Total Hip Arthroplasty (THA)

Type of Use (*Select one or both, as applicable*)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*



2555 Davie Road • Ft. Lauderdale, FL 33317  
 Phone 954.927.2044 • Fax 954.927.0446  
[www.makosurgical.com](http://www.makosurgical.com)

## 510(k) SUMMARY

**Sponsor:** Mako Surgical Corp.  
 2555 Davie Road,  
 Fort Lauderdale, FL 33317

**Contact Person:** Shikha Khandelwal, PhD  
[shikha.khandelwal@stryker.com](mailto:shikha.khandelwal@stryker.com)  
 Phone: 201-831-6921

**Date Prepared:** July 23, 2019

**Proprietary Name:** Mako Total Hip Application

**Common Name:** Total Hip Application (THA)

**Regulation Name:** Stereotaxic Instrument

**Regulation Number:** 21 CFR Section 882.4560

**Device Classification:** Class II

**Product Code:** OLO

### Substantial Equivalence Claimed To:

The subject device, the Mako Total Hip Application, is substantially equivalent to the predicate device, the Mako Total Hip Application, cleared via K170593.

### Device Modification:

The following changes have been made to the Mako Total Hip Application:

- **Pre-operative Planning Process** – Treatment Design Application (TDA) is a new software tool which will be implemented into the pre-operative planning process of Mako Total Hip procedures.

The software tool is intended to accept CT imaging data and process it to generate the patient-specific 3D models and anatomic landmarks utilized for pre-operative implant planning.

**Description:**

The Mako System with the subject Total Hip Application is a stereotactic instrument that includes a robotic arm, an integrated cutting system, an optical detector, a computer, dedicated instrumentation, operating software, a planning laptop, and tools and accessories.

The system's architecture is designed to support total and partial knee procedures and total hip procedures. With application specific hardware and software, the system provides stereotactic/haptic guidance during orthopedic surgical procedures by using patient CT data to assist a surgeon with pre-surgical planning, implant placement and interpretive/intraoperative navigation of the patient's anatomy.

Once configured for a specific application, the Mako robotic-arm can serve as surgeon's "intelligent" tool holder or tool guide by passively constraining the preparation of an anatomical site for an orthopedic implant with software-defined spatial boundaries.

**Summary of Technological Characteristics Compared to Predicate Devices:**

The technological characteristics of the Mako Total Hip Application compared to the predicate device are listed below:

<b>Technological Characteristics</b>	<b>Mako Total Hip Application</b>	<b>Mako Total Hip Application - K170593</b>
Major Components	Guidance Module, robotic arm, camera stand, cutting system, preoperative planning laptop.	Guidance Module, robotic arm, camera stand, cutting system, preoperative planning laptop.
Tools/accessories	Various reusable and disposable instruments	Various reusable and disposable instruments
Image Use	CT	CT

**Intended Use** - The subject device has the same intended use as that specified in the cleared 510(k) premarket notification for the predicate device.

**Indications for Use**

The Mako System is intended to assist the surgeon in providing software defined spatial boundaries for orientation and reference information to anatomical structures during orthopedic procedures.

The Mako System is indicated for use in surgical knee and hip procedures in which the use of stereotactic surgery may be appropriate, and where reference to rigid anatomical bony structures can be identified relative to a CT based model of the anatomy. These procedures include:

- Unicondylar knee replacement and/or patellofemoral knee replacement
- Total Hip Arthroplasty (THA)

**Performance Data** – The modified Mako System with the Mako Total Hip Application has been evaluated through the following non-clinical performance testing:

- Software testing:
  - Treatment Design Application Full Software Run Through
  - Bone Registration Accuracy Verification
  - Mako Total Hip Application Full System Run Through

**Conclusions of Performance Testing:**

Performance testing has demonstrated that the characteristics of the subject Mako Total Hip Application is equivalent to the predicate device. The device is also as safe and as effective as the predicate device and does not raise different questions of safety and effectiveness. Therefore, the performance testing supports a determination of Substantial Equivalence.





2555 Davie Road • Ft. Lauderdale, FL 33317  
 Phone 954.927.2044 • Fax 954.927.0446  
[www.makosurgical.com](http://www.makosurgical.com)

## 510(k) SUMMARY

**Sponsor:** Mako Surgical Corp.  
 2555 Davie Road,  
 Fort Lauderdale, FL 33317

**Contact Person:** Shikha Khandelwal, PhD  
[shikha.khandelwal@stryker.com](mailto:shikha.khandelwal@stryker.com)  
 Phone: 201-831-6921

**Date Prepared:** July 23, 2019

**Proprietary Name:** Mako Total Knee Application

**Common Name:** Total Knee Application (TKA)

**Regulation Name:** Stereotaxic Instrument

**Regulation Number:** 21 CFR Section 882.4560

**Device Classification:** Class II

**Product Code:** OLO

### Substantial Equivalence Claimed To:

The subject device, the Mako Total Knee Application, is substantially equivalent to the predicate device, Mako Total Knee Application, cleared via K172219.

### Device Modification:

The following changes have been made to the Mako Total Knee Application:

- **Pre-operative Planning Process** – Treatment Design Application (TDA) is a new software tool which will be implemented into the pre-operative planning process of Mako Total Knee procedures.

The software tool is intended to accept CT imaging data and process it to generate the patient-specific 3D models and anatomic landmarks utilized for pre-operative implant planning.

**Description:**

The Mako System with the subject Total Knee Application is a stereotactic instrument that includes a robotic arm, an integrated cutting system, an optical detector, a computer, dedicated instrumentation, operating software, a planning laptop, and tools and accessories.

The system's architecture is designed to support total and partial knee procedures and total hip procedures. With application specific hardware and software, the system provides stereotactic/haptic guidance during orthopedic surgical procedures by using patient CT data to assist a surgeon with pre-surgical planning, implant placement and interpretive/intraoperative navigation of the patient's anatomy.

Once configured for a specific application, the Mako robotic-arm can serve as surgeon's "intelligent" tool holder or tool guide by passively constraining the preparation of an anatomical site for an orthopedic implant with software-defined spatial boundaries.

**Summary of Technological Characteristics Compared to Predicate Devices:**

The technological characteristics of the Mako Total Knee Application compared to the predicate device are listed below:

<b>Technological Characteristics</b>	<b>Mako Total Knee Application</b>	<b>Mako Total Knee Application - K172219</b>
Major Components	Guidance Module, robotic arm, camera stand, cutting system, preoperative planning laptop.	Guidance Module, robotic arm, camera stand, cutting system, preoperative planning laptop.
Tools/accessories	Various reusable and disposable instruments	Various reusable and disposable instruments
Image Use	CT	CT

**Intended Use** - The subject device has the same intended use as that specified in the cleared 510(k) premarket notification for the predicate device.

**Indications for Use**

The Mako System is intended to assist the surgeon in providing software defined spatial boundaries for orientation and reference information to anatomical structures during orthopedic procedures.

The Mako System is indicated for use in surgical knee procedures in which the use of stereotactic surgery may be appropriate, and where reference to rigid anatomical bony structures can be identified relative to a CT based model of the anatomy. These procedures include:

- Total Knee Arthroplasty (TKA)

The implant systems compatible with the system:

- Triathlon Total Knee System (CR/CS/PS cemented and cementless primary)
- Triathlon Total Knee System (TS inserts cemented primary)
- Kinetis Total Knee System (CR/UC)

**Performance Data** – The modified Mako System with the Mako Total Knee Application has been evaluated through the following non-clinical performance testing:

- Software testing:
  - Treatment Design Application Full Software Run Through
  - Bone Registration Accuracy Verification
  - Mako Total Knee Application Full System Run Through

**Conclusions of Performance Testing:**

Performance testing has demonstrated that the characteristics of the subject Mako Total Knee Application is equivalent to the predicate device. The device is also as safe and as effective as the predicate device and does not raise different questions of safety and effectiveness. Therefore, the performance testing supports a determination of Substantial Equivalence.