



September 21, 2020

Everyway Medical Instruments Co., Ltd.
Paul Hung
Official Applicant
3 FL., NO. 5, Lane 155, Section 3, Peishen Rd.,
Shenkeng District, New Taipei City, TW 22203

Re: K192589

Trade/Device Name: EV-906 & EV-906A OTC TENS/EMS Combination
Regulation Number: 21 CFR 882.5890
Regulation Name: Transcutaneous electrical nerve stimulator for pain relief
Regulatory Class: Class II
Product Code: NUH, NGX

Dear Paul Hung:

The Food and Drug Administration (FDA) is sending this letter to notify you of an administrative change related to your previous substantial equivalence (SE) determination letter dated August 19, 2020. Specifically, FDA is updating this SE Letter (e.g. incorrect trade name) as an administrative correction.

Please note that the 510(k) submission was not re-reviewed. For questions regarding this letter please contact Amber Ballard, OHT5: Office of Neurological and Physical Medicine Devices, (240) 402-9983, amber.ballard@fda.hhs.gov.

Sincerely,

Vivek J. Pinto -S

Vivek Pinto, PhD
Director
DHT5B: Division of Neuromodulation
and Physical Medicine Devices
OHT5: Office of Neurological
and Physical Medicine Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health



December 19, 2019

Everyway Medical Instruments Co., Ltd.
Paul Hung
Official Correspondent
3 FL., NO. 5, Lane 155, Section 3, Peishen Rd.,
Shenkeng District, New Taipei City, 22203 TW

Re: K192589

Trade/Device Name: Everyway OTC TENS/EMS Combination
Regulation Number: 21 CFR 882.5890
Regulation Name: Transcutaneous Electrical Nerve Stimulator For Pain Relief
Regulatory Class: Class II
Product Code: NUH, NGX
Dated: September 19, 2019
Received: September 20, 2019

Dear Paul Hung:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Amber T. Ballard -
S



For Vivek Pinto, Ph.D.

Director

DHT5B: Division of Neuromodulation
and Physical Medicine Devices

OHT5: Office of Neurological
and Physical Medicine Devices

Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K192589

Device Name

EV-906 & EV-906A OTC TENS/EMS Combination.

Indications for Use (Describe)

TENS: intended for temporary relief of pain associated with sore and aching muscles in the low back as well as upper and lower extremities (arm and/or leg) due to strain from exercise or normal household and work activities. (Choose TENS Modes B/N/M/SD1/SD2 as adjustable mode and P1 to P12 preset programs only for EV-906A).

EMS: intended to stimulate healthy muscles in order to improve or facilitate muscle performance. (Choose EMS Modes C/S/A as adjustable mode and P13 to P24 preset programs only for EV-906A).

Type of Use (Select one or both, as applicable)

☐ Prescription Use (Part 21 CFR 801 Subpart D)

☒ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."