



In2Bones USA, LLC  
Christine Scifert  
VP of Quality and Regulatory  
6000 Poplar Ave, Suite 115  
Memphis, Tennessee 38119

January 7, 2020

Re: K192754

Trade/Device Name: Avenger Radial Head System  
Regulation Number: 21 CFR 888.3170  
Regulation Name: Elbow joint radial (hemi-elbow) polymer prosthesis  
Regulatory Class: Class II  
Product Code: KWI  
Dated: December 4, 2019  
Received: December 6, 2019

Dear Christine Scifert:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Vesa Vuniqui  
Assistant Director  
DHT6A: Division of Joint Arthroplasty Devices  
OHT6: Office of Orthopedic Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K192754

Device Name

Avenger Radial Head System

Indications for Use (Describe)

The Avenger Radial Head System is intended for replacement of the proximal end of the radius:

- Replacement of the radial head for degenerative or post-traumatic disabilities presenting pain, crepitation, and decreased motion at radial-humeral and/or proximal radio-ulnar joint with:
  - a. Joint destruction and/or subluxation visible on x-ray; and/or
  - b. Resistance to conservative treatment.
- Primary replacement after fracture of the radial head.
- Symptomatic sequelae after radial head resection.
- Revision following failed radial head arthroplasty.

The components are intended for uncemented use.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRASStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**510(k) Summary**  
*Avenger Radial Head System*  
*January 6, 2020*

**Company:** In2Bones USA, LLC  
6000 Poplar Ave, Suite 115  
Memphis, TN 38119  
901-260-7931

**Primary Contact:** Christine Scifert

**Company Contact:** Rebecca Wahl

**Trade Name:** Avenger Radial Head System

**Common Name:** Modular Radial Head

**Classification:** II

**Regulation Number:** 21 CFR 888.3170, Elbow joint radial (hemi elbow) polymer prosthesis

**Panel:** 87-Orthopedic

**Product Code(s):** KWI

**Device Description:** The Avenger Radial Head System is a two-piece modular system comprised of cobalt chrome (ASTM F 1537) head and stem components. The system consists of a range of lengths and diameters for the stem as well as heads in a range of diameters and heights to accommodate the patient anatomy. The stems are offered in short and long lengths. The heads are offered in heights of standard, +2 and +4 configurations. The implant allows for replacement of the radial head.

Indications for Use

Intended Use:

The Avenger Radial Head System is intended for replacement of the proximal end of the radius:

- Replacement of the radial head for degenerative or post-traumatic disabilities presenting pain, crepitation, and decreased motion at radial-humeral and/or proximal radio-ulnar joint with:
  - a. Joint destruction and/or subluxation visible on x-ray; and/or
  - b. Resistance to conservative treatment.
- Primary replacement after fracture of the radial head.
- Symptomatic sequelae after radial head resection.
- Revision following failed radial head arthroplasty.

The components are intended for uncemented use.

**Substantial Equivalence:** The subject components were demonstrated to be substantially equivalent to the following systems previously cleared by the FDA:

Primary Predicate

- K060731 – Wright Medical Evolve® Modular Radial Head

Additional Predicates

- K991915 – Wright Medical Evolve® Modular Radial Head
- K982288 – Stryker (SBI/Avanta) Radial Head

The subject Avenger Radial Head is made of cobalt chrome and has been demonstrated to be substantially equivalent to the previously cleared devices identified above as the products are similar in indications, materials and geometry.

**Performance Testing:** Head disassembly (ASTM F2009) and fatigue testing (ASTM F2887) were performed for the Avenger Radial Head System. Bacterial endotoxin testing (LAL) was performed.

**Conclusion**

Based on the test results and the comparison to the predicate devices, the subject device is determined to be substantially equivalent to the predicate devices.