



November 14, 2019

Wilson-Cook Medical, Inc./Cook Endoscopy
Tiffany A. Thomas
Global Regulatory Affairs Specialist
4900 Bethania Station Road
Winston-Salem, NC 27105

Re: K192908
Trade/Device Name: Fusion Cytology Brush, CytoMax II Double
Lumen Cytology Brush
Regulation Number: 21 CFR 876.1500
Regulation Name: Endoscope and accessories
Regulatory Class: II
Product Code: FDX
Dated: October 10, 2019
Received: October 15, 2019

Dear Tiffany Thomas:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Angel A. Soler-Garcia, Ph.D.
Acting Assistant Director
DHT3B: Division of Reproductive,
Gynecology and Urology Devices
OHT3: Office of GastroRenal, ObGyn,
General Hospital and Urology Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K192908

Device Name
Fusion Cytology Brush
CytoMax II Double Lumen Cytology Brush

Indications for Use (Describe)
Used for collection of cells in the biliary system.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

Special 510(k): Cytology Brushes

510(k) Summary

Name: Wilson-Cook Medical, Inc. /Cook Endoscopy
Address: 4900 Bethania Station Road
Winston-Salem, North Carolina 27105
Phone: (336) 744-0157
Fax: (336) 201-5994
Contact; Tiffanny A. Thomas, Global Regulatory Affairs Specialist
Theresa de Prat, Regulatory Affairs Specialist II

Date: October 10, 2019

Trade Names: Fusion Cytology Brush
CytoMax II Double Lumen Cytology Brush

Common Name: Endoscopic Cytology Brush

Classification Name: Endoscope and Accessories 21 CFR §876.1500, FDX, Class II

Predicate Devices: Fusion Cytology Brush/ CytoMax II Double Lumen Cytology Brush cleared
December 29, 2018 (k171573)
Fusion Cytology Brush/ CytoMax II Double Lumen Cytology Brush cleared
February 8, 2019 (k181317)

Intended Use: Used for collection of cells in the biliary system.

Device Description:

The Fusion Cytology Brush and CytoMax II Double Lumen Cytology Brush (predicate devices) cleared via k171573 on 12/29/2018 as well as k181317 on 2/8/2019 have been modified from their original design as described on this submission. The subject devices have the same indications for use, methods of operation, and the fundamental technological characteristics as their cleared predicates. The modification to the subject devices is the removal of a cannula proximal to the brush.

The subject devices, the Fusion Cytology Brush and CytoMax II Double Lumen Cytology Brush, consist of a nylon brush on a stainless-steel drive wire with a stainless-steel tip, a double lumen catheter with ink markings, radiopaque markers, wire guide access via a wire guide hub or and IDE port, a detachable extension line with a luer lock for optional flushing of the wire guide hub and a pin-vise handle.

The Fusion Cytology Brush and CytoMax II Double Lumen Cytology Brushes are used by passing the device through an endoscope over a prepositioned wire guide to a target location. The cytology brush is located at the distal end (patient contacting) of the device with the pin vise handle located at the proximal end (non-patient contacting). The handle is actuated by pushing the pin vise handle forward to extend the cytology brush and then pulling backward to retract the brush.

Substantial Equivalence:

A modification was made to the currently cleared predicate devices cleared under k181317 and k171573. The modification is a device design change, removing a cannula proximal to the brush.

Performance Data:

Performance testing consisting of non-clinical bench testing demonstrates the subject devices met the performance requirements to fulfill the intended use. This testing provides reasonable assurance that the subject devices will function as intended. The subject devices do not raise new questions of safety or effectiveness as compared to the predicate devices.

Summary of non-clinical testing:

The following non-clinical testing was conducted to demonstrate the performance of the subject devices and confirms that the subject devices perform as intended.

- Tensile Strength Testing
- Tensile Strength of Handle

Conclusion:

We believe that the subject devices are substantially equivalent to the predicate devices with respect to intended use, key operating principles, materials of construction, and technological characteristics. We consider the risks associated with the modifications to the subject devices to have been adequately addressed through our Design Control Processes and do not affect the safety or effectiveness of the devices.