



December 12, 2019

Arthrex Inc.
Jessica Singelais
Regulatory Affairs Specialist
1370 Creekside Boulevard
Naples, Florida 34108-1945

Re: K193157

Trade/Device Name: Arthrex 3.9mm Corkscrew Suture Anchor
Regulation Number: 21 CFR 888.3030
Regulation Name: Single/Multiple Component Metallic Bone Fixation Appliances and Accessories
Regulatory Class: Class II
Product Code: MAI, MBI
Dated: November 7, 2019
Received: November 14, 2019

Dear Ms. Singelais:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

for

Laura Rose, Ph.D.
Acting Assistant Director
DHT6C: Division of Restorative, Repair
and Trauma Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K193157

Device Name

Arthrex 3.9mm Corkscrew Suture Anchor

Indications for Use (Describe)

The Arthrex Corkscrew is intended for fixation of suture (soft tissue) to bone in the shoulder, foot/ankle, hip, knee, hand/wrist, and elbow in the following procedures:

- Shoulder: Rotator Cuff Repairs, Bankart Repair, SLAP Lesion Repair, Biceps Tenodesis, Acromio-Clavicular Separation Repair, Deltoid Repair, Capsular Shift or Capsulolabral Reconstruction.
- Foot/Ankle: Lateral Stabilization, Medial Stabilization, Achilles Tendon Repair, Hallux Valgus Reconstruction, Midfoot reconstruction, Metatarsal Ligament Repair/Tendon Repair, Bunionectomy.
- Knee: Anterior Cruciate Ligament Repair, Medial Collateral Ligament Repair, Lateral Collateral Ligament Repair, Patellar Tendon Repair, Posterior Oblique Ligament Repair, Iliotibial Band Tenodesis.
- Hand/Wrist: Scapholunate Ligament Reconstruction, Ulnar or Radial Collateral Ligament Reconstruction, Radial Collateral Ligament Reconstruction.
- Elbow: Biceps Tendon Reattachment, Tennis Elbow Repair, Ulnar or Radial collateral Ligament Reconstruction, Lateral Epicondylitis Repair.
- Hip: Capsular Repair, acetabular labral repair.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

Date Prepared	December 5, 2019
Submitter	Arthrex Inc. 1370 Creekside Boulevard Naples, FL 34108-1945
Contact Person	Jessica L. Singelais Regulatory Affairs Specialist 1-239-598-4302, ext. 73091 Jessica.singelais@arthrex.com
Name of Device	Arthrex 3.9mm Corkscrew Suture Anchor
Common Name	Smooth or threaded metallic bone fixation fastener
Product Code	MAI, MBI
Classification Name	21 CFR 888.3030: Fastener, Fixation, Biodegradable, Soft Tissue
Regulatory Class	II
Predicate Device	K173788: Arthrex Corkscrew FT
Reference Device	K082810: Arthrex BioComposite Anchors
Purpose of Submission	This Special 510(k) premarket notification is submitted to add a line extension to the Arthrex BioComposite Anchors cleared under predicate K173788.
Device Description	The Arthrex 3.9mm Corkscrew is a partially threaded suture anchor pre-loaded with Arthrex Suture on a disposable inserter. The anchor is manufactured from either PLLA/ β TCP or PEEK and is offered sterile, single use.
Indications for Use	The Arthrex Corkscrew is intended for fixation of suture (soft tissue) to bone in the shoulder, foot/ankle, hip, knee, hand/wrist, and elbow in the following procedures: • Shoulder: Rotator Cuff Repairs, Bankart Repair, SLAP Lesion Repair, Biceps Tenodesis, Acromio-Clavicular Separation Repair, Deltoid Repair, Capsular Shift or Capsulolabral Reconstruction. • Foot/Ankle: Lateral Stabilization, Medial Stabilization, Achilles Tendon Repair, Hallux Valgus Reconstruction, Midfoot reconstruction, Metatarsal Ligament Repair/Tendon Repair, Bunionectomy. • Knee: Anterior Cruciate Ligament Repair, Medial Collateral Ligament Repair, Lateral Collateral Ligament Repair, Patellar Tendon Repair, Posterior Oblique Ligament Repair, Iliotibial Band Tenodesis. • Hand/Wrist: Scapholunate Ligament Reconstruction, Ulnar or Radial Collateral Ligament Reconstruction, Radial Collateral Ligament Reconstruction. • Elbow: Biceps Tendon Reattachment, Tennis Elbow Repair, Ulnar or Radial collateral Ligament Reconstruction, Lateral Epicondylitis Repair. • Hip: Capsular Repair, acetabular labral repair.
Comparison Summary of Technological Characteristics and Modifications Proposed	The proposed device is a line extension to the predicate device. The proposed and predicate devices (K173788) have the same basic design, intended use, indications for use, packaging and sterilization method. Proposed modifications consist of a smaller diameter, shorter length, additional material (PEEK) and minor dimensional changes. Any differences between the proposed device and the predicate device are considered minor and do not raise questions concerning safety or effectiveness.
Performance Data	Pull-out and biocompatibility testing per ISO 10993-1:2018 was conducted and compared to the predicate device to demonstrate that the proposed devices are substantially equivalent to the predicate.

Bacterial endotoxin per EP 2.6.14/USP <85> was conducted to demonstrate that the device meets pyrogen limit specifications.

Conclusion

The proposed Arthrex Corkscrew Suture Anchors are substantially equivalent to the predicate device in which the basic design features and intended uses are the same. Any differences between the proposed device and the predicate device are considered minor and do not raise questions concerning safety or effectiveness.

Based on the indications for use, technological characteristics, and the summary of data submitted, Arthrex Inc. has determined that the Arthrex Corkscrew is substantially equivalent to the currently marketed predicate device.
