

FEB 15 2002

K020122

## Section 16

# Summary of Safety and Effectiveness

K020122

Millennium Biomedical Inc.  
MB 102 Millennium Blades

December 26, 2001  
Premarket Notification

**Summary of Safety and Effectiveness**

This 510(k) summary of safety and effectiveness information is being submitted in accordance with the requirements of SMDA 1990 and CFR 807.92.

**1. Submitter's name, address, telephone number, contact person, and data summary prepared:**

a. Millennium Biomedical Inc.  
360 East Bonita Avenue  
Pomona, California 91767  
Phone: (909)-621-7646  
Fax: (909)-621-7556

b. Contact Person: Jerry Kaeni  
President

c. Date Summary Prepared: December 26, 2001

**2. Name of device, including trade name and classification name:**

a. Trade/Proprietary Name: MB 102 Millennium Blades

b. Classification Name: Keratome, AC-Powered, and/or Blades

**3. Identification of the predicate device or legally marketed device or devices to which substantial equivalence is being claimed:**

<u>Company</u>	<u>Device</u>	<u>510(k) No.</u>	<u>Date Cleared</u>
Chiron Vision (Bausch & Lomb)	Hansatome™ Microkeratome which uses Bausch & Lomb Accuglide™ blades	K972808	11-5-91

4. **A description of the device that is the subject of the 510(k), including explanation of how device function, basic scientific concepts, significant physical and performance characteristics (design, material, physical properties):**

The MB 102 Millennium Blade is a replacement blade designed to be used with the Hansatome™ Microkeratome. The MB 102 Millennium Blade is a single-use only, disposable device. The blade material is similar to that used in predicate devices (stainless steel).

5. **A statement of intended use:**

The MB 102 Millennium blade is intended to be used as a replacement blade for the Hansatome™ Microkeratome.

6. **A statement of how the technological characteristics of the device compare to those of the predicate or legally marketed devices:**

SUMMARY OF TECHNOLOGICAL CHARACTERISTICS OF DEVICE

	<u>CHARACTERISTICS</u>	<u>MILLENNIUM BLADE</u>
	B&L Accuglide™ Blade ( <u>PREDICATE DEVICE</u> )	MBI MB 102
Intended Use	Indicated for use with the Hansatome™ Microkeratome by surgeons to cut cornea in the form of a hinged flap in LASIK refractive surgery procedures.	Indicated for use as a replacement blade for the Hansatome™ Microkeratome.
Operating Principle	The blade is held in the keratome head and oscillates by means of the turbine. The keratome head adapts to the turbine by means of a threaded part. The turbine motor is gas powered.	The blade is held in the keratome head and oscillates by means of the turbine. The keratome head adapts to the turbine by means of a threaded part. The turbine motor is gas powered.
Blade Design	Single edge blade with the plastic blade holder	Single edge blade with the plastic blade holder
Blade Hardness	52 Rockwell C	52 Rockwell C
Sterilization Method	Cobalt 60 radiation	Cobalt 60 radiation
Blade Material	Stainless steel	Stainless steel
Blade Holder Material	Delrin	Delrin
Patient Contact Portion of Device	Blade cutting edge	Blade cutting edge

**7. Brief summary of clinical tests and results**

The performance of the MB 102 Millennium Blades was found to be acceptable the test results of the functional performance testing. The results showed that the functional performance of the MB 102 Millennium Blades was substantially equivalent to that of the predicate blade.



FEB 15 2002

Mr. Jerry Kaeni  
President  
Millennium Biomedical Inc.  
360 East Bonita Avenue  
Pomona, CA 91767

Re: K020122  
Trade/Device Name: MB 102 Millennium Blade  
Regulation Number: 21 CFR 886.4370  
Regulation Name: Keratome  
Regulatory Class: I  
Product Code: HNO  
Dated: December 26, 2001  
Received: January 14, 2001

Dear Mr. Kaeni:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Page 2 – Mr. Jerry Kaeni, President

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 21 CFR Part 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4613. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,



A. Ralph Rosenthal, M.D.  
Director  
Division of Ophthalmic and Ear,  
Nose and Throat Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

510(k) Number (if known): K020122

Device Name: MB 102 Millennium Blade

**Indications for Use:**

The MB 102 Millennium blade is intended to be used as a replacement blade for the Chiron Hansatome™ Microkeratome.

**(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)**

---

Concurrence of CDRH, Office of Device Evaluation (ODE)

Denis L. Mc Carthy

(Division Sign-Off)  
Division of Ophthalmic Ear,  
Nose and Throat Devices

510(k) Number K020122

Prescription Use   x  

OR Over-The-Counter Use \_\_\_\_\_

(Per 21 CFR 801.109)