



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

JUL 25 2002

Ms. Erin McGurk
Regulatory Affairs Representative
Vector Medical, Inc.
1886 Lodgepole Drive
Milton, Florida 32583

Re: K022225
Trade/Device Name: Vector Medical, Inc., Vector Targeting System
Regulation Number: 888.3020
Regulation Name: Intramedullary Fixation Rod
Regulatory Class: II
Product Code: HSB
Dated: July 7, 2002
Received: July 10, 2002

Dear Ms. McGurk:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 21 CFR Part 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4659. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,

A handwritten signature in black ink, appearing to read "Celia M. Witten". The signature is written in a cursive style with a large initial "C".

Celia M. Witten, Ph.D., M.D.

Director

Division of General, Restorative
and Neurological Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

Applicant: Vector Medical Inc.
510(k) Number (if known): K022225
Device Name: Vector Targeting System
Indications For Use:

INDICATIONS:

The Vector Medical, Inc. *Vector Targeting System* is indicated for use in locating (targeting) screw holes in an intramedullary (IM) nail in order to quickly and accurately guide the drill into the exact position to locate and drill the screw holes. The technique does not require the use of radiographic equipment.

Generally these nails are indicated for use in fractures in long bones such as the femur, tibia or humerus, which require longitudinal stabilization.

RECEIVED
JUN 10 11:05 AM '92

FDA/CDRH/ODE/DHC

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

(Per 21 CFR 801.109)
(Optional Format 1-2-96)

Steph Rhoads
(Division Sign-Off)
Division of General, Restorative
and Neurological Devices

510(k) Number K022225