



MAR 07 2003

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Mr. James A. Cochie
Regulatory Affairs Manager
Monaghan Medical Corporation
5 Latour Avenue, Suite 1600
P.O. Box 2805
Plattsburgh, New York 12901-0299

Re: K023097
Trade/Device Name: TrueZone Peak Flow Meter™
Regulation Number: 868.1860
Regulation Name: Peak-flow Meter for Spirometry
Regulatory Class: II
Product Code: BZH
Dated: January 8, 2003
Received: January 13, 2003

Dear Mr. Cochie:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (301) 594-4646. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,



Susan Runner, DDS, MA
Interim Director
Division of Anesthesiology, General Hospital,
Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

510(k) Number (if known): K023097

Device Name: TruZone™ Peak Flow Meter

Indications for Use:

The **TruZone™** Peak Flow Meter (PFM) is a portable monitoring device that measures the peak expiratory flow (PEF) generated by a patient during a forced exhalation maneuver. This measurement is useful in detecting changes in airways that could signal a worsening of symptoms or an improvement in breathing function for people with respiratory conditions such as Asthma and emphysema.

This device is packaged and labeled so it can be used as a screening device in a hospital or other clinical setting. Labeling is also included for the use of the device if dispensed by a health care provider to a patient for personal use.

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use
(Per 21 CFR 801.109)

or

Over-The-Counter Use _____



(Division Sign-Off)

Division of Anesthesiology, General Hospital,
Infection Control, Dental Devices

510(k) Number: K023097