



June 5, 2020

NeoTract, Inc.  
Brian Gall  
Regulatory Affairs Manager  
4155 Hopyard Rd.  
Pleasanton, CA 94588

Re: K200441  
Trade/Device Name: UroLift Advanced Tissue Control (ATC) System  
Regulation Number: 21 CFR§ 876.5530  
Regulation Name: Implantable Transprostatic Tissue Retractor System  
Regulatory Class: II  
Product Code: PEW  
Dated: May 8, 2020  
Received: May 11, 2020

Dear Brian Gall:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies.

You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

for Martha W. Betz, Ph.D.  
Acting Assistant Director  
DHT3B: Division of Reproductive,  
Gynecology and Urology Devices  
OHT3: Office of GastroRenal, ObGyn,  
General Hospital and Urology Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K200441

Device Name

UroLift Advanced Tissue Control (ATC) System

Indications for Use (Describe)

The UroLift System is indicated for the treatment of symptoms due to urinary outflow obstruction secondary to benign prostatic hyperplasia (BPH), including lateral and median lobe hyperplasia, in men 45 years of age or older.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

UroLift® Advanced Tissue Control (ATC®) System

03 510(k) SUMMARY

**COMPANY INFORMATION**

NeoTract, Inc.  
4155 Hopyard Road  
Pleasanton, CA 94588  
Registration Number: 3015181082

**SUBMISSION CORRESPONDENT**

Brian Gall  
Regulatory Affairs Manager, Interventional Urology  
NeoTract, Inc.  
4155 Hopyard Road  
Pleasanton, CA 94588

Telephone – 925.329.6547  
E-mail – brian.gall@teleflex.com

**DATE PREPARED**

21 February 2020

**DEVICE INFORMATION**

Trade Name:	UroLift® Advanced Tissue Control (ATC®) System
Common Name:	Implantable Transprostatic Tissue Retractor System
Regulation Name:	Implantable Transprostatic Tissue Retractor System
Product Code:	PEW
Regulation Number:	876.5530
Classification:	II
Classification Panel:	Reproductive, Gastro-Renal, Urological, General Hospital Device and Human Factors (OHT3) Reproductive and Urology Devices (DHT3B)

**DEVICE DESCRIPTION**

The UroLift Advanced Tissue Control (ATC) System is a modification of the UroLift UL400 System (last cleared in K193269). The primary difference is the addition of a wing component on the distal tip of the UL400 which provides a larger footprint. This design feature is intended to provide better mobilization of tissue when performing the UroLift System procedure.

The UroLift System (both the UL400 and UroLift ATC) is designed to access the prostatic urethra and deliver one UroLift Implant through a lobe of the prostate. The UroLift System is inserted into the urethra through the penile orifice and used to displace the urethra toward the prostatic capsule. The UroLift Implant is then deployed transversely through the prostatic tissue. Multiple implants are deployed in the UroLift System procedure. The implants secure the retracted position of the urethra, thereby maintaining an expanded urethral lumen, reducing fluid obstruction and improving lower urinary tract symptoms (LUTS). This is accomplished by holding the approximated position of the inner (urethral) tissue and the outer (capsular) tissue of the prostate with

## UroLift® Advanced Tissue Control (ATC®) System

the UroLift Implant. The procedure typically requires 2-6 implants to retract the obstruction. The UroLift System consists of two main components, the UroLift Delivery Device (single use), and the UroLift Implants (one implant per delivery device). Each Delivery Device comes pre-loaded with one UroLift Implant.

**INTENDED USE**

The UroLift System is indicated for the treatment of symptoms due to urinary outflow obstruction secondary to benign prostatic hyperplasia (BPH), including lateral and median lobe hyperplasia, in men 45 years of age or older.

**CONTRAINDICATIONS**

The UroLift System should not be used if the patient has:

- Prostate volume of >100 cc
- A urinary tract infection
- Urethra conditions that may prevent insertion of delivery system into bladder
- Urinary incontinence due to incompetent sphincter
- Current gross hematuria

**PREDICATE DEVICE**

The predicate device is the UroLift UL400 System from NeoTract (K193269).

Trade Name:	UroLift® UL400 System
Common Name:	Implantable Transprostatic Tissue Retractor System
Regulation Name:	Implantable Transprostatic Tissue Retractor System
Product Code:	PEW
Regulation Number:	876.5530
Classification:	II
Classification Panel:	Reproductive, Gastro-Renal, Urological, General Hospital Device and Human Factors (OHT3) Reproductive and Urology Devices (DHT3B)

**COMPARISON WITH THE PREDICATE DEVICE**

The UroLift ATC System device is based on the UL400 UroLift System platform cleared in K193269. The UroLift ATC System device leverages the same platform design as the UL400 UroLift system and includes a modification to the distal tip, giving the tip a larger footprint during the procedure and allowing for effective mobilization of tissue when needed.

The remainder of the device is substantially equivalent to the UL400. The implant components, including the materials, specifications and methods of manufacture are unchanged relative to the predicate device. The delivery system mechanism of action is unchanged.

**PERFORMANCE TESTING**

The design requirements for the UroLift System were reviewed and non-clinical design verification testing was required to assure that the modifications of the proposed device did not impact the safe and effective use of the device. Non-clinical testing included deployment testing, compatibility with accessories, and implant, shaft, and wing performance testing. The testing was performed on devices which had undergone worst case sterilization, accelerated aging, and transit testing. The majority of the test

## UroLift® Advanced Tissue Control (ATC®) System

methods were equivalent to the testing for the 510(k) cleared UroLift UL400 System (K193269), and all acceptance criteria were met.

### **BIOCOMPATIBILITY TESTING**

The UroLift ATC System has been tested for biocompatibility and passed the relevant tests according to ISO 10993-1: *Biological evaluation of medical devices – Part 1: Evaluation and testing within a risk management process*. The modification addressed in this 510(k) submission do introduce new materials and, therefore additional biocompatibility testing was performed.

Biocompatibility testing was performed on worst case sterilized devices and included:

- Cytotoxicity testing per *ISO 10993-5:2009 – Biological evaluation of medical devices Part 5: Tests for in vitro cytotoxicity*
- Sensitization and Intracutaneous Reactivity testing per *ISO 10993-10:2010 – Biological evaluation of medical devices Part 10: Tests for irritation and skin sensitization*
- Material Mediated Pyrogenicity and Acute Systemic Toxicity per *ISO 10993-11:2017 – Biological Evaluation of Medical Devices Part 11: Tests for Systemic Toxicity*

### **STERILIZATION AND SHELF LIFE TESTING**

The UroLift ATC System has been validated to determine the minimum gamma irradiation dose to ensure a  $10^{-6}$  Sterility Assurance Level (SAL). The modification addressed in the 510(k) submission may impact the product sterility because the modified component utilizes new materials and adds some geometric complexity to the device. These materials are manufactured, processed, and handled similarly to the predicate UroLift device.

### **CONCLUSION**

The testing demonstrated the NeoTract UroLift ATC System is as safe and effective, has the same intended use, technological characteristics and principles of operation as the predicate device. Therefore, the NeoTract UroLift ATC System is substantially equivalent to the UroLift UL400 System.