



June 15, 2020

Aziyo Biologics, Inc.  
Wendy Perreault  
Regulatory Affairs Consultant  
1100 Old Ellis Road, Suite 1200  
Roswell, Georgia 30076

Re: K201313

Trade/Device Name: CanGaroo Envelope (small), CanGaroo Envelope (medium), CanGaroo Envelope (large), CanGaroo Envelope (extra large), CanGaroo Envelope (extra extra large) (SubQ)

Regulation Number: 21 CFR 878.3300

Regulation Name: Surgical Mesh

Regulatory Class: Class II

Product Code: FTM

Dated: May 15, 2020

Received: May 18, 2020

Dear Wendy Perreault:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Hetal Odobasic  
Implantable Electrophysiology Devices Team  
Division of Cardiac Electrophysiology, Diagnostics, and  
Monitoring Devices  
Office of Cardiovascular Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
Food and Drug Administration

Special 510(k) Premarket Notification

Form Approved: OMB No. 0910-0120

Expiration Date: 06/30/2020

See PRA Statement below.

## Indications for Use

510(k) Number (if known)

Device Name

CanGaroo(R) Envelope

Indications for Use (Describe)

The CanGaroo Envelope is intended to securely hold a cardiac implantable electronic device or an implantable neurostimulator to create a stable environment when implanted in the body. The cardiac implantable electronic devices that may be used with the CanGaroo Envelope include pacemaker pulse generators, defibrillators or other cardiac implantable electronic devices. The implantable neurostimulator devices that may be used with the CanGaroo Envelope include vagus nerve stimulators, spinal cord neuromodulators, deep brain stimulators and sacral nerve stimulators.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**CONTINUE ON A SEPARATE PAGE IF NEEDED.**

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRAStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

510(k) Summary: K201313

Company Information

Company Name: Aziyo Biologics, Inc.  
Contact Name: Wendy Perreault  
Contact Title: Regulatory Consultant  
Address: 1100 Old Ellis Road, Suite 1200  
Roswell, GA 30076  
Phone (mobile): 404-542-5854  
Phone (office): 470-514-4085  
  
Date Prepared: June 15, 2020

Product Information

Trade Name: CanGaroo® Envelope  
Common Name: Surgical Mesh Envelope  
Classification Name: Surgical Mesh, 21 CFR 878.3300, Product Code FTM, Class II

Indications for Use

The CanGaroo Envelope is intended to securely hold a cardiac implantable electronic device or an implantable neurostimulator to create a stable environment when implanted in the body. The cardiac implantable electronic devices that may be used with the CanGaroo Envelope include pacemaker pulse generators, defibrillators or other cardiac implantable electronic devices. The implantable neurostimulator devices that may be used with the CanGaroo Envelope include vagus nerve stimulators, spinal cord neuromodulators, deep brain stimulators and sacral nerve stimulators.

Predicate Devices

The CanGaroo Envelope is substantially equivalent to the CanGaroo Envelope referenced in the Substantially Equivalent letter issued by FDA for 510(k) application K192616.

This Special 510(k) application describes changes in the packaging for the CanGaroo Envelope as follows:

1. Introduce a single foil-to-foil pouch configuration to improve the packaging environment for the sterilized CanGaroo Envelope during storage and shipment, and to comply with suture manufacturer's recommendations for storage;
2. Reduce the overall size of the Display Envelope; and
3. Introduce a new Five Pack Carton to replace the current Five Pack Carton.

There have been no modifications to the indications for use of the CanGaroo device and no changes to labeling are proposed, and the change does not have the potential to alter the fundamental scientific technology of the device. The operating principle(s) and mechanism of action of the device are not changing.

#### Device Description

The CanGaroo Envelope is intended to securely hold a cardiac implantable electronic device or an implantable neurostimulator to create a stable environment when implanted in the body. The cardiac implantable electronic devices that may be used with the CanGaroo Envelope include pacemaker pulse generators, defibrillators, or other cardiac implantable electronic devices. The implantable neurostimulator devices that may be used with the CanGaroo Envelope include vagus nerve stimulators, spinal cord neuromodulators, deep brain stimulators and sacral nerve stimulators.

The CanGaroo Envelope is constructed from two to four perforated, multilaminate sheets (4-ply) of decellularized, non-crosslinked, lyophilized ECM (extracellular matrix) derived from porcine small intestinal submucosa. The 3 mm perforations are spaced evenly at 10 mm apart to allow exit of any exudate. The ECM is assembled into pouch form using violet 5-0 polydioxanone suture (PDS). The device design is identical to the device cleared under K192616.

#### Substantial Equivalence

The intended use of the CanGaroo Envelope to securely hold a CIED or neurostimulator is identical to the intended use of the CanGaroo Envelope cleared under K192616. The devices are of identical design and are manufactured from the same materials. The proposed packaging functions identically to the current packaging to protect and keep the product sterile during transportation and storage.

#### Non-clinical Testing

To ensure that the new, equivalent packaging could be used to manufacture a product that continued to meet the design input requirements for the CanGaroo Envelope, packaging testing and sterilization adoption testing of product manufactured with the new packaging was repeated. The CanGaroo Envelope manufactured with the new packaging met the design input requirements for packaging and sterilization.

#### Conclusion

The CanGaroo Envelope packaged in the new packaging components is substantially equivalent to the predicate CanGaroo Envelope.