



December 15, 2020

Apollo Endosurgery, Inc.  
Natalie Allen  
Regulatory Affairs Specialist  
1120 S. Capital of Texas Hwy  
Austin, Texas 78749

Re: K201808

Trade/Device Name: X-Tack Endoscopic HeliX Tacking System  
Regulation Number: 21 CFR 876.4400  
Regulation Name: Hemorrhoidal Ligator  
Regulatory Class: Class II  
Product Code: PKL, OCW  
Dated: November 5, 2020  
Received: November 6, 2020

Dear Natalie Allen:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shanil P. Haugen, Ph.D.  
Assistant Director  
DHT3A: Division of Renal,  
Gastrointestinal, Obesity  
and Transplant Devices  
OHT3: Office of Gastrogenal, ObGyn,  
General Hospital and Urology Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K201808

Device Name

X-Tack Endoscopic HeliX Tacking System

Indications for Use (Describe)

The X-Tack™ Endoscopic HeliX Tacking System is intended for approximation of soft tissue in minimally invasive gastroenterology procedures (e.g. closure and healing of ESD/EMR sites, and closing of fistula, perforation or leaks). X-Tack is not intended for hemostasis of acute bleeding ulcers.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRASStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*



1120 S. Capital of TX Hwy, Bldg. 1, Ste. 300  
Austin, Texas 78746  
USA  
Phone: 512.279.5100  
FAX: 512.279.5105

### **510(k) Summary**

**Owner's Name & Address:** Apollo Endosurgery  
1120 S. Capital of Texas Hwy.  
Building 1, Suite 300  
Austin, TX 78746

**Contact Person:** David M. Hooper, PhD  
Vice President, Quality and Regulatory Affairs  
Phone: (512) 279-5100  
Email : david.hooper@apolloendo.com

**Date:** December 10, 2020

**Trade Name:** X-Tack™ Endoscopic HeliX Tacking System

**Common Name:** Endoscopic Tissue Approximation Device

**Product Code:** PKL, OCW

**Classification:** Class II (21 CFR 876.4400)

**Classification Name** Hemorrhoidal Ligator

**Predicate Devices:** K151802 – Resolution™ 360 Clip

### **Device Description**

#### **X-Tack™ Endoscopic HeliX Tacking System**

The X-Tack™ Endoscopic HeliX Tacking System is a sterile, single-use device that enables the user to approximate soft tissue in the gastrointestinal (GI) tract using helix tacks and a 3-0 suture through a 2.8 mm or larger working channel of an endoscope (e.g. gastroscope or colonoscope).

#### **OverStitch Suture Cinch**

The Overstitch Suture Cinch device is comprised of thermoplastic and stainless steel materials and includes an implantable PEEK Cinch component designed to secure and cut the suture once tissue approximation is complete. It is the final step of the X-Tack procedure. The device functions by squeezing the handle and deploying the PEEK components, which form a press-fit onto the tail end of the suture to maintain suture position *in situ*.

**Indications for Use:** The X-Tack™ Endoscopic HeliX Tacking System is intended for approximation of soft tissue in minimally invasive gastroenterology procedures (e.g. closure and healing of ESD/EMR sites, and closing of fistula, perforation or leaks). X-Tack is not intended for hemostasis of acute bleeding ulcers

**Technological Characteristics:**

The X-Tack™ Endoscopic Helix Tacking System shares technological characteristics similar to the predicate device. These characteristics include:

- Equivalent intended use
- Endoscopic delivery of closure device
- Manufactured from materials commonly utilized for implant devices used in the gastrointestinal (GI) tract
- Similar design concept in that a force is delivered by the implant to approximate soft tissue.
- Sterilized using an ethylene oxide (EO) process
- Equivalent performance in a randomized, controlled animal study, specifically evaluating the closure and healing of defects over the course of one month, as evaluated through direct visualization and histological analysis.

**Basis of Substantial Equivalence:**

Non-clinical testing was performed to verify anchor retention, suture tensile strength, and endoscope compatibility. Testing was performed to validate usability and human factors, packaging, shelf-life and the ability of the device to withstand distribution forces. The sterilization cycle and biocompatibility of the device were validated per recognized ISO standards. MR testing was performed to establish the conditions under which the device could be safely scanned, as well as the heating and artifact that could be expected during MRI scanning. MR testing included scanning to determine induced forces, followed by laboratory tests to verify safety under maximum scan conditions.

A randomized study involving 4 pigs and 40 defects (created in the stomach and colon), was done to compare the closure rates and healing between X-Tack and the predicate. That study demonstrated that the closure rates associated with X-Tack were equivalent to the predicate and that X-Tack enabled larger defects to be closed. Histological analysis of defects showed both devices resulted in healing of defects.

Based on the non-clinical and animal testing, the device is as safe and effective, and performs as well or better than the legally marketed predicate device.

**Summary Table of Equivalence**

	<b>X-Tack</b>	<b>Predicate</b>
Principle of Operation	<ul style="list-style-type: none"> <li>• Device is inserted into an endoscope and deployed at target site.</li> <li>• Helix tacks are deployed independently around the defect.</li> <li>• Implanted construct closes the defect through metallic anchors and suture tension.</li> </ul>	<ul style="list-style-type: none"> <li>• Device is inserted into an endoscope and deployed at target site.</li> <li>• Clips are deployed independently around the defect.</li> <li>• Implanted construct consists of metallic arms and spring tension.</li> </ul>

Endoscope compatability	Scopes with a 2.8mm working channel.	Scopes with a 2.8mm working channel.
Working length	Gastric: 155 cm Colon: 235 cm	Gastric: 160 cm Colonic: 235 cm
Respositionable prior to deployment	Yes	Yes
Physician control over placement and final locking	Yes	Yes
Expected implant duration.	Approximately 1 month or less, then implant is passed in stool.	Approximately 1 month or less, then implant is passed in stool.
Clip opening width	Not applicable. Device uses independent tacks, suture and cinch and is not limited by a clip opening dimension.	11 mm
Closure efficacy in animal study	100% closure at 4 weeks. Histologic evaluations of closure were consistent with wound healing.	100% closure at 4 weeks. Histologic evaluations of closure were consistent with wound healing.
Sterilization Method	EO	EO
Usage	Single-use	Single-use
Implanted Materials	Helix Tack: 316L Stainless Steel Cinch: VESTAKEEP i4 (PEEK) Suture: Polyproylene USP 3-0 Polypropylene Copper Phthalocyanine Blue (Below 0.5WT%) in accordance with 21 CFR 74, 3045	Capsule: 304 Stainless Steel Arms: 17-7 PH SS Tension breaker: Luran 968R (polystyrene) Yoke: F75 Cobalt Chrome
Biocompatibility	Tested per ISO 19993.	Tested per ISO 19993.
MR Compatibility	MR Conditional with 1.5 and 3 T MR scanners with spatial field gradient of 2500 Gauss/cm (extrapolated or less) and SAR of 2.0 W/kg for 15 minutes of continuous scanning.	MR Conditional with 1.5 and 3 T MR scanners with spatial field gradient of 2500 Gauss/cm (extrapolated or less) and SAR of 2.0 W/kg for 15 minutes of continuous scanning.