



September 17, 2021

Olympus Medical Systems Corp.  
% Jon Gilbert  
Regulatory Affairs Consultant for Olympus  
Olympus Corporation of the Americas  
3500 Corporate Parkway PO Box 610  
Center Valley, PA 18034-0610

Re: K201832  
Trade/Device Name: ENDOEYE FLEX DEFLECTABLE VIDEOSCOPE OLYMPUS LTF-S190-5  
Regulation Number: 21 CFR§ 884.1720  
Regulation Name: Gynecologic Laparoscope and Accessories  
Regulatory Class: II  
Product Code: HET, GCJ, NWB  
Dated: August 13, 2021  
Received: August 16, 2021

Dear Jon Gilbert:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies.

You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

**Jason Roberts -S**

Jason R. Roberts, Ph.D.

Assistant Director

DHT3B: Division of Reproductive,

Gynecology and Urology Devices

OHT3: Office of GastroRenal, ObGyn,

General Hospital and Urology Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

**Indications for Use**

510(k) Number (if known)

K201832

Device Name

ENDOYE FLEX DEFLECTABLE VIDEOSCOPE OLYMPUS LTF-S190-5

Indications for Use (Describe)

This instrument is intended to be used with a video system center, light source, documentation equipment, monitor, hand instruments, electrosurgical unit, and other ancillary equipment for endoscopy and endoscopic surgery within the thoracic and abdominal cavities including female reproductive organs.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

**CONTINUE ON A SEPARATE PAGE IF NEEDED.**

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary

### I. GENERAL INFORMATION

- 510(k) Submitter: OLYMPUS MEDICAL SYSTEMS CORP.  
2951 Ishikawa-cho, Hachioji-shi, Tokyo 192-8507,  
Japan
- Contact Person: Jon Gilbert  
c/o Olympus Corporation of the Americas  
3500 Corporate Parkway PO Box 610  
Center Valley, PA 18034-0610, USA  
Phone: 906-361-3237  
Fax: 484-896-7128  
Email: [jgilb.raca@gmail.com](mailto:jgilb.raca@gmail.com)
- Manufacturing site: Aizu Olympus Co., Ltd.,  
500 Muranishi, Niidera, Monden-machi, Aizuwakamatsu-shi,  
Fukushima 965-8520, Japan

Date Prepared: September 14, 2021

### II. DEVICE IDENTIFICATION

- Device Name ENDOEYE FLEX DEFLECTABLE VIDEOSCOPE  
OLYMPUS LTF-S190-5
- Common Name VIDEOSCOPE
- Regulation Number 21 CFR 884.1720
- Regulation Name Gynecologic laparoscope and accessories
- Regulatory Class II
- Product Code HET; Laparoscope, Gynecologic (And Accessories)  
GCJ; Laparoscope, General & Plastic Surgery  
NWB; Endoscope, Accessories, Narrow Band Spectrum
- Classification Panel Obstetrics/Gynecology

**III. PREDICATE DEVICE**

| Device name                                                        | 510(k) Submitter                 | 510(k) No. |
|--------------------------------------------------------------------|----------------------------------|------------|
| ENDOEYE FLEX 3D<br>DEFLECTABLE VIDEOSCOPE<br>OLYMPUS LTF-190-10-3D | OLYMPUS MEDICAL<br>SYSTEMS CORP. | K123365    |

The predicate device has not been subject to a design-related recall.

**IV. DEVICE DESCRIPTION****■ General Description of the subject device**

This instrument has been designed to be used with a video system center, light source, documentation equipment, monitor, hand instruments, electrosurgical unit, and other ancillary equipment for endoscopy and endoscopic surgery within the thoracic and abdominal cavities including female reproductive organs.

**■ Principle of Operation**

The LTF-S190-5 is a video endoscope consisting of three parts: the control section, the insertion section, and the connector section. The device uses an objective lens and light guide lens for capturing images/videos from the distal end of the endoscope.

**V. INDICATIONS FOR USE****ENDOEYE FLEX DEFLECTABLE VIDEOSCOPE OLYMPUS LTF-S190-5**

This instrument is intended to be used with a video system center, light source, documentation equipment, monitor, hand instruments, electrosurgical unit, and other ancillary equipment for endoscopy and endoscopic surgery within the thoracic and abdominal cavities including female reproductive organs

**VI. COMPARISON OF TECHNOLOGY CHARACTERISTICS WITH THE PREDICATE DEVICE**

The detailed comparison chart between subject device and predicate devices are as follows.

| Item                            | Subject Device (SD)                                                                                                                                                                                                                                                                                            | Predicate Device (PD)                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               |
|---------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|                                 | LTF-S190-5                                                                                                                                                                                                                                                                                                     | LTF-190-10-3D                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                       |
| Indications for Use             | This instrument is intended to be used with a video system center, light source, documentation equipment, monitor, hand instruments, electrosurgical unit, and other ancillary equipment for endoscopy and endoscopic surgery within the thoracic and abdominal cavities including female reproductive organs. | This instrument is intended to be used with Olympus video system center, light source, documentation equipment, 3D processor, monitor, hand instruments, electrosurgical unit, and other ancillary equipment for endoscopy and endoscopic surgery. This instrument is indicated for use within the thoracic and abdominal cavities including female reproductive organs. This instrument must not be used for observation or treatment of the heart and must not contact the heart or any area near the heart. In addition, this instrument must not come into contact with any device or therapeutic accessory that contacts the heart or any area near the heart. |
| Environment of use              | Healthcare facility/hospital                                                                                                                                                                                                                                                                                   | Healthcare facility/hospital                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                        |
| Single/ Repeat use              | Repeat use                                                                                                                                                                                                                                                                                                     | Repeat use                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                          |
| Sterile/ Non-sterile            | Non-sterile                                                                                                                                                                                                                                                                                                    | Non-sterile                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                         |
| Sterilization method            | Ethylene oxide, Autoclave, Hydrogen peroxide (Sterrad 100S/ NX)                                                                                                                                                                                                                                                | Ethylene oxide, Hydrogen peroxide (Sterrad NX)                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                      |
| Field of View                   | 85°                                                                                                                                                                                                                                                                                                            | 80°                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                 |
| Direction of View               | 0° (Forward viewing)                                                                                                                                                                                                                                                                                           | 0° (Forward viewing)                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                |
| Depth of Field                  | 18-100 mm                                                                                                                                                                                                                                                                                                      | 18-100 mm                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           |
| Type of CCD                     | Color CCD                                                                                                                                                                                                                                                                                                      | Color CCD                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           |
| Number of CCD Chip              | 1                                                                                                                                                                                                                                                                                                              | 2                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                   |
| Narrow Band Imaging observation | Available                                                                                                                                                                                                                                                                                                      | Available                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           |
| Insertion Tube Diameter         | Distal end size: ø 5.4 mm<br>Insertion tube outer Diameter (Maximum): ø5.4 mm                                                                                                                                                                                                                                  | Distal end size: ø 10 mm<br>Insertion tube outer Diameter (Maximum): ø 10 mm                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                        |
| Working Length                  | 370 mm                                                                                                                                                                                                                                                                                                         | 370 mm                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                              |
| Channel inner diameter          | No instrument channel                                                                                                                                                                                                                                                                                          | No instrument channel                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               |
| Angulation Range                | U/D:100°/100°<br>R/L:100°/100°                                                                                                                                                                                                                                                                                 | U/D:100°/100°<br>R/L:100°/100°                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                      |

| Item                         | Subject Device (SD)                                                                                                                                                                              | Predicate Device (PD)                                                                                                                                                                            |
|------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|                              | LTF-S190-5                                                                                                                                                                                       | LTF-190-10-3D                                                                                                                                                                                    |
| Patient contacting material  | Bending rubber: Fluoro Rubber<br>Distal end/ Distal cover/<br>Insertion tube/ Pin: Stainless steel<br>Glue: Epoxy Resin<br>Lens: Glass                                                           | Bending rubber: Fluoro Rubber<br>Distal end / Insertion tube: Stainless steel<br>Glue: Epoxy Resin<br>Lens: Glass                                                                                |
| Duration and type of contact | External communicating device in contact with mucosal membranes, damaged mucosal membranes and tissue/bone/dentin.<br>The contact duration is limited exposure (i.e. contact is up to 24 hours). | External communicating device in contact with mucosal membranes, damaged mucosal membranes and tissue/bone/dentin.<br>The contact duration is limited exposure (i.e. contact is up to 24 hours). |

The indications for use of the subject device do not raise different questions of safety and effectiveness when compared to the predicate device. The subject and predicate devices have the same intended use of providing visualization during thoracic and abdominal surgery, including the female reproductive organs. The differences in technological characteristics do not raise new questions of safety and effectiveness.

## **VII. PERFORMANCE DATA**

The following performance data were provided in support of the substantial equivalence determination.

### **1) Reprocessing validation testing**

All reprocessing methods including the Sterilizer/Sterilization process were validated in accordance with the FDA guidance “Reprocessing Medical Devices in Health Care Settings: Validation Methods and Labeling” issued on March 17, 2015. The acceptance criteria were met.

### **2) Biocompatibility testing**

Biocompatibility testing for the LTF-S190-5 final product was conducted in accordance with the FDA’s Guidance for Industry and Food and Drug Administration Staff, Use of International Standard ISO 10993-1, “Biological evaluation of medical devices – Part 1: Evaluation and testing within a risk management process”.

- Cytotoxicity Study Using the Colony Assay (ISO 10993-5:2009)
- Intracutaneous Study in Rabbits (ISO 10993-10:2010)
- Guinea Pig Maximization Sensitization Test (ISO 10993-10:2010)
- Acute Systemic Toxicity Study in Mice (ISO 10993-11:2006)

- Rabbit Pyrogen Study (USP <151>, USP <37>)

The device was shown to be non-cytotoxic, non-irritating, non-sensitizing, not systemically toxic, and non-pyrogenetic.

### **3) Software verification and validation testing**

Software verification and validation testing for the LTF-S190-5 was conducted and documentation was provided as recommended by FDA's Guidance for Industry and FDA Staff, "Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices" and "Content of Premarket Submissions for Management of Cybersecurity in Medical Devices". The device met all requirements.

### **4) Electrical safety and electromagnetic compatibility (EMC)**

Electrical safety and EMC testing were conducted on the LTF-S190-5. The system complies with the ANSI/AAMI ES 60601-1:2005/(R)2012 and A1:2012 and IEC 60601-2-18:2009 standards for safety and the IEC 60601-1-2:2014 standards for EMC. The device met all requirements.

### **5) Performance testing - Bench**

Bench testing for the LTF-S190-5 as listed below was conducted to ensure that the subject device performs as intended and meet design specifications. Device performance was assessed the design requirements, and included process verification, design verification, and design validation. The device met all acceptance criteria.

- Thermal safety test
- Composite durability test
- Color (imaging) performance
- Photobiological Safety

## **VIII. CONCLUSIONS**

The indications for use and technological characteristics of the LTF-S190-5 do not raise different questions of safety and effectiveness as compared to the predicate. The performance testing demonstrate that the subject device is as safe and effective as the predicate. Therefore, and the subject device is substantially equivalent to the predicate device.