



May 18, 2021

Zimmer, Inc.
Romil Sheth
Regulatory Affairs Project Manager
1800 W. Center Street
Warsaw, Indiana 46580

Re: K210829

Trade/Device Name: Persona® Revision Knee System Femoral Metaphyseal Cones

Regulation Number: 21 CFR 888.3560

Regulation Name: Knee Joint Patellofemorotibial Polymer/Metal/Polymer Semi-Constrained Cemented
Prosthesis

Regulatory Class: Class II

Product Code: JWH, MBH

Dated: March 16, 2021

Received: March 19, 2021

Dear Romil Sheth:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Ting Song, Ph.D., R.A.C.
Assistant Director
DHT6A: Division of Joint
Arthroplasty Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K210829

Device Name

Persona Revision Knee System Femoral Metaphyseal Cones

Indications for Use (Describe)

This device is indicated for patients with severe knee pain and disability due to:

- Rheumatoid arthritis, osteoarthritis, traumatic arthritis, polyarthritis.
- Collagen disorders, and/or avascular necrosis of the femoral condyle.
- Post-traumatic loss of joint configuration, particularly when there is patellofemoral erosion, dysfunction or prior patellectomy.
- Moderate valgus, varus, or flexion deformities.
- The salvage of previously failed surgical attempts or for a knee in which satisfactory stability in flexion cannot be obtained at the time of surgery.

Porous components may be used cemented or uncemented (biological fixation). Augments may be attached via bone cement or screw to the tibial plates and/or femoral components. Splined stem extension components are intended to be used press-fit (uncemented). All other femoral component, tibial plate, stem extension, and femoral and tibial augment components are indicated for cemented use only.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

Sponsor: Zimmer, Inc.
1800 W. Center Street
Warsaw, IN 46580
Establishment Registration Number: 1822565

Contact Person: Romil Sheth
Regulatory Affairs Project Manager
Telephone: (574-268-8196)
Romil.Sheth@zimmerbiomet.com

Date: March 16, 2021

Subject Device: **Trade Name:** Persona[®] Revision Knee System Femoral Metaphyseal Cones

Classification Name:

- JWH – Prosthesis, Knee, Patellofemorotibial, Semi-Constrained, Cemented, Polymer/Metal/Polymer (21 CFR 888.3560 – Knee joint Patellofemorotibial polymer/metal/polymer semi-constrained cemented prosthesis)
- MBH – Prosthesis, Knee, Patello/Femorotibial, Semi-Constrained, Uncemented, Porous, Coated, Polymer/Metal/Polymer (21 CFR 888.3565 – Knee joint Patellofemorotibial metal/polymer porous-coated uncemented prosthesis)

Predicate Device(s):

K181947	<i>Persona</i> Revision Knee System	Zimmer, Inc.
---------	-------------------------------------	--------------

Device Description: The purpose of this submission is to obtain clearance for the changes/proposed changes made to the *Persona* Revision Knee System femoral metaphyseal cones post their original 510(k) submission/clearance.

The *Persona* Revision Knee System femoral metaphyseal cones are intended to fill distal femoral cavitory bone defects during knee arthroplasty.

Indications for Use:

This device is indicated for patients with severe knee pain and disability due to:

- Rheumatoid arthritis, osteoarthritis, traumatic arthritis, polyarthritis.
- Collagen disorders, and/or avascular necrosis of the femoral condyle.
- Post-traumatic loss of joint configuration, particularly when there is patellofemoral erosion, dysfunction or prior patellectomy.
- Moderate valgus, varus, or flexion deformities.
- The salvage of previously failed surgical attempts or for a knee in which satisfactory stability in flexion cannot be obtained at the time of surgery.

Porous components may be used cemented or uncemented (biological fixation). Augments may be attached via bone cement or screw to the tibial plates and/or femoral components. Splined stem extension components are intended to be used press-fit (uncemented). All other femoral component, tibial plate, stem extension, and femoral and tibial augment components are indicated for cemented use only.

Summary of Technological Characteristics:

The rationale for substantial equivalence is based on consideration of the following characteristics:

- **Intended Use:** Same as the predicate device.
- **Indications for Use:** Same as the predicate device.
- **Geometry/Configurations/Sizes:** Same as the predicate device.
- **Material:** Same as the predicate device.
- **Bone Interface Finish:** Same as the predicate device.
- **Fixation Method of Bone:** Same as the predicate device.
- **Sterility:** Same as the predicate device.
- **Design Features:** Similar to the predicate device.

**Summary of Performance Data
(Nonclinical and/or Clinical)**

Finite Element Analysis (FEA) was used to predict the peak stress under physiological loading for the subject *Persona* Revision Knee System femoral metaphyseal cones.

**Substantial Equivalence
Conclusion**

Based on the information contained within this submission, it is concluded that the *Persona* Revision Knee System femoral metaphyseal cones are substantially equivalent to the identified predicate device.