



June 16, 2021

Inner Mongolia Cureguard Medical Materials Co., Ltd.  
% Boyle Wang  
Official Correspondent  
Shanghai Truthful Information Technology Co., Ltd.  
RM. 608, No. 738, Shangcheng Rd., Pudong  
Shanghai, Shanghai 200120  
China

Re: K211336

Trade/Device Name: Disposable Vinyl Examination Glove  
Regulation Number: 21 CFR 880.6250  
Regulation Name: Non-Powdered Patient Examination Glove  
Regulatory Class: Class I, reserved  
Product Code: LYZ  
Dated: April 21, 2021  
Received: May 3, 2021

Dear Boyle Wang:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Ryan Ortega, PhD  
Acting Assistant Director  
DHT4B: Division of Infection Control  
and Plastic Surgery Devices  
OHT4: Office of Surgical  
and Infection Control Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K211336

Device Name  
Disposable Vinyl Examination Glove

Indications for Use (Describe)

The Disposable Vinyl Examination Glove is a disposable device intended for medical purposes that is worn on the examiner's hands to prevent contamination between patient and examiner.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## **510(k) Summary**

### **(K211336)**

This summary of 510(k) is being submitted in accordance with 21 CFR 807.92.

#### **1.0 Submitter's Information**

Name: Inner Mongolia Cureguard Medical Materials Co., Ltd.  
Address: RM 408, Building 5, Block B of Business Science & Technology Center,  
Jining District, Ulanqab City, Inner Mongolia Autonomous Region.  
Phone Number: +86-13485097856  
Contact: Guo Hua  
Date of Preparation: 06/15/2021

#### **Designated Submission Correspondent**

Mr. Boyle Wang  
Shanghai Truthful Information Technology Co., Ltd.  
Room 608, No. 738 Shangcheng Rd., Pudong Shanghai, 200120 China  
Tel: +86-21-50313932  
Email: [Info@truthful.com.cn](mailto:Info@truthful.com.cn)

#### **2.0 Device Information**

Trade name: Disposable Vinyl Examination Glove  
Common name: Vinyl Patient Examination Glove  
Classification name: Non-powdered Patient Examination Glove  
Model(s): XS,S, M, L, XL

#### **3.0 Classification**

Production code: LYZ  
Regulation number: 21CFR880.6250  
Classification: Class I  
Panel: General Hospital

#### **4.0 Predicate Device Information**

Manufacturer: Hebei Hongtai Plastic Products Company Limited  
Device: Vinyl Patient Examination Gloves (White, Blue, Yellow)  
510(k) number: K163168

### 5.0 Indication for Use

The Disposable Vinyl Examination Glove is a disposable device intended for medical purposes that is worn on the examiner's hands to prevent contamination between patient and examiner.

### 6.0 Device Description

The subject device is powder free vinyl patient examination gloves. The subject device is clear. The design of subject device is addressing the standards as ASTM D6124, ASTM D5151, and ASTM D5250. The subject device is non-sterile.

### 7.0 Technological Characteristic Comparison Table

**Table1-General Comparison**

Item	Subject device	Predicate device	Comparison
510(k) number	K211336	K163168	-
Product Code	LYZ	LYZ	Same
Regulation No.	21CFR880.6250	21CFR880.6250	Same
Class	I	I	Same
Intended Use	The Disposable Vinyl Examination Glove is a disposable device intended for medical purposes that is worn on the examiner's hands to prevent contamination between patient and examiner.	The Vinyl Examination Glove (White, Blue, or Yellow) is a disposable device intended for medical purposes that is worn on the examiner's hands to prevent contamination between patient and examiner.	Same
Powdered or Powered free	Powdered free	Powdered free	Same
Design Feature	Ambidextrous	Ambidextrous	Same
Labeling Information	Single use, powder free, device color, device name, glove size and quantity, Vinyl Examination Gloves, Non-Sterile	Single use, powder free, device color, device name, glove size and quantity, Vinyl Examination Gloves, Non-Sterile	Similar

**Table2 Device Dimensions Comparison**

Predicate Device (K163168)	Designation	Size					Tolerance
		XS	S	M	L	XL	
	Length, mm	230	230	235	245	245	min
	Width, mm	80	85	95	105	115	±5

	Thickness, mm:						
	Finger	0.05				min	
	Palm	0.08				min	
Subject Device (K211336)	Designation	Size				Tolerance	
		XS	S	M	L		XL
	Length, mm	230	230	230	230	230	min
	Width, mm	75	85	95	105	115	±5
	Thickness, mm:						
	Finger	0.08				min	
Palm	0.08				min		
Remark	Similar						

Analysis: The physical dimensions are little different with that of the predicate, but they all meet the requirements of ASTM D5250.

**Table3 Performance Comparison**

Item			Subject device (K211336)	Predicate device (K163168)	Comparison
Colorant			Clear	White, Blue, Yellow	Analysis 1
Physical Properties	Before Aging	Tensile Strength	11MPa, min	15MPa, min	Analysis 2
		Ultimate Elongation	300%min	380%min	Analysis 2
	After Aging	Tensile Strength	11MPa, min	15MPa, min	Analysis 2
		Ultimate Elongation	300%min	380%min	Analysis 2
	Comply with ASTM D5250			Comply with ASTM D5250	Same
Freedom from Holes			Be free from holes when tested in accordance with ASTM D5151 AQL=2.5	Be free from holes when tested in accordance with ASTM D5151 AQL=2.5	Same
Powder Content			< 0.01 mg per glove. Meet the requirements of ASTM D6124	Meet the requirements of ASTM D6124	Similar

Analysis 1: The colorant of the subject device is different with that of the predicate device, the subject device was evaluated according to ISO 10993-1 standards.

Analysis 2: The tensile strength and ultimate elongation are different with that

of the predicate, but they all meet the requirements of ASTM D5250.

**Table4 Safety Comparison**

Item		Subject device	Predicated device	Comparison
Material		Vinyl	Vinyl	Same
Biocompatibility	Irritation	Comply with ISO10993-10. Under the conditions of the study, not an irritant	Comply with ISO10993-10	Same
	Sensitization	Comply with ISO10993-10. Under conditions of the study, not a sensitizer.		
	Cytotoxicity	Comply with ISO10993-10. Under conditions of the study, did not show potential toxicity to L-929 cells.	Not provided	Different
Label and Labeling		Meet FDA's Requirement	Meet FDA's Requirement	Same

**8.0 Summary of Non-clinical Testing**

Non-clinical tests were conducted to verify that the proposed device met all design specifications. The test results demonstrated that the proposed device complies with the following standards:

ISO 10993-10:2010 Biological Evaluation Of Medical Devices - Part 10: Tests For Irritation And Skin Sensitization.

ISO 10993-5:2009 Biological Evaluation Of Medical Devices - Part 5: Tests For In Vitro Cytotoxicity

ASTM D6124-06 (Reapproved 2017), Standard Test Method for Residual Powder on Medical Gloves

ASTMD5151-19, Standard Test Method for Detection of Holes in Medical Gloves.

ASTM D5250-19, Standard Specification for Poly (vinyl chloride) Gloves for Medical Application.

**Table 5 Summary of non-clinical performance testing**

Test Methodology	Purpose	Acceptance Criteria	Results
------------------	---------	---------------------	---------

ASTM D5250	Physical Dimensions Test	Length(mm): $\geq 230$ ; Width(mm): XS: $75 \pm 5$ ; S: $85 \pm 5$ ; M: $95 \pm 5$ ; L: $105 \pm 5$ ; XL: $115 \pm 5$ ;		Length: $> 230$ / Pass Width: XS: 74-77/ Pass S: 83-87 /Pass M: 96-97/ Pass L: 104-108/ Pass XL: 114-117/ Pass	
		Thickness (mm) : Finger: $\geq 0.08$ Palm: $\geq 0.08$		Finger: 0.09-0.14/Pass Palm: 0.10-0.13/Pass	
ASTM D5151	Watertightness Test for Detection of Holes	Meet the requirements of ASTM D5151 AQL 1.5		0/125 leaks / Pass	
ASTM D6124	Powder Content	Meet the requirements of ASTM D6124 $< 2.0\text{mg}$		0.01 mg/Pass	
ASTM D412	Physical properties	Before Aging	Tensile Strength	$\geq 11\text{MPa}$	15-18/Pass
			Ultimate Elongation	$\geq 300\%$	530-552/Pass
		After Aging	Tensile Strength	$\geq 11\text{MPa}$	15-18/Pass
			Ultimate Elongation	$\geq 300\%$	533-553/Pass
ISO 10993-5	Cytotoxicity	Non-cytotoxic		Under conditions of the study, did not show potential toxicity to L-929 cells./ Pass	
ISO 10993-10	Irritation	Non-irritating		Under the conditions of the study, not an irritant/ Pass	
ISO 10993-10	Sensitization	Non-sensitizing		Under conditions of the study, not a sensitizer./ Pass	



## **9.0 Summary of Clinical Testing**

Clinical testing is not needed for this device.

## **10.0 Conclusion**

The conclusions drawn from the nonclinical tests demonstrate that the proposed device, Disposable Vinyl Examination Glove is as safe, as effective, and performs as well as or better than the legally marketed predicated device under K163168.