



April 3, 2022

Protect Gloves Co., Ltd.  
% Paweena U-Thainual  
CEO  
MDR Solutions Co., Ltd.  
1435 Kanchanapisek Rd., Bang Khae Nuea  
Bangkok, Bang Khae 10160  
Thailand

Re: K213604

Trade/Device Name: ELAZ Nitrile Powder Free Examination Glove  
Regulation Number: 21 CFR 880.6250  
Regulation Name: Non-Powdered Patient Examination Glove  
Regulatory Class: Class I, reserved  
Product Code: LZA  
Dated: March 21, 2022  
Received: March 24, 2022

Dear Paweena U-Thainual:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Clarence W. Murray, III, PhD  
Assistant Director  
DHT4B: Division of Infection Control  
and Plastic Surgery Devices  
OHT4: Office of Surgical  
and Infection Control Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K213604

Device Name  
ELAZ Nitrile Powder Free Examination Glove

Indications for Use (Describe)

ELAZ Nitrile Powder Free Examination Glove is disposable device intended for medical purpose that is worn on the examiner's hand to prevent contamination between patient and examiner.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

K213604

## 510(k) Summary

### 1. General Information

Applicant/Submitter: Protect Gloves Company Limited  
 Address: 60/138 Moo 3 Mabyangporn Pluakdang  
 Rayong, 21140 THAILAND  
 Tel: +66-2-384-3049  
 Email: [nathaphon@me.com](mailto:nathaphon@me.com)

Contact Person: Paweena U-Thainual, PhD  
 Address: MDR Solutions, Co., Ltd.  
 1435 Kanchanapisek Rd., Bang Khae Nuea  
 Bang Khae, Bangkok 10160 THAILAND  
 Email: [paweena@mdrsolutions.co.th](mailto:paweena@mdrsolutions.co.th)

Preparation Date: November 8, 2021

### 2. Device Name and Code

Device Trade Name: ELAZ Nitrile Powder Free Examination Glove  
 Common Name: Nitrile Patient Examination Glove  
 Classification Name: Non-Powdered Patient Examination Glove  
 Product Code: LZA  
 Regulation Number: 21 CFR 882.6250  
 Classification: I  
 Review Panel: General Hospital

### 3. Predicate Device

Provided below is the legally marketed predicate device.

Table 1 Primary Predicate device

Applicant	Device Name	510(k) Number
Tangshan Lanhai Medical Supplies Co., Ltd.	Disposable Nitrile Examination Gloves (Powder free, Purple-Blue, Blue)	K210898

**4. Device Description**

ELAZ Nitrile Powder Free Examination Glove is non-sterile, single use only, disposable, powder free examination glove. The glove is made of Acrylonitrile-butadiene rubber. The glove is designed to meets the specification of ASTM D6319-19, Standard Specification for Nitrile Examination Gloves for Medical Application.

**5. Indications / Intended Use**

ELAZ Nitrile Powder Free Examination Glove is disposable device intended for medical purposes that is worn on the examiner’s hand to prevent contamination between patient and examiner.

**6. Technological Characteristics Comparison:**

ELAZ Nitrile Powder Free Examination Glove is compared to the legally marketed predicate device.

	<b>Proposed Subject Device</b>	<b>Predicate Device</b>	<b>Comparison</b>
<b>Manufacturer</b>	Protect Gloves Co., Ltd.	Tangshan Lanhai Medical Supplies Co., Ltd.	N/A
<b>Trade Name</b>	ELAZ Nitrile Powder Free Examination Glove	Disposable Nitrile Examination Gloves (Powder free, Purple-Blue, Blue)	N/A
<b>510(k) Number</b>	K213604	K210898	N/A
<b>Classification</b>	Class I	Class I	YES
<b>Product Code</b>	LZA	LZA	YES
<b>Regulation Number</b>	880.6250	880.6250	YES
<b>Intended Use/ Indications For use</b>	ELAZ Nitrile Powder Free Examination Glove is disposable device intended for medical purposes that is worn on the examiner’s hand to prevent contamination between patient and examiner	The DISPOSABLE NITRILE EXAMINATION GLOVES (Powder free, Blue) is a disposable device intended for medical purposes that is worn on the examiner’s hands to prevent contamination between patient and examiner.	YES
<b>Powdered Free</b>	Yes	Yes	YES

ELAZ Nitrile Powder Free Examination Glove

510(k) Summary

<b>Device Dimensions Comparison</b>										
<b>Size</b>	S, M, L, XL				XS, S, M, L, XL					YES*
<b>Color</b>	Blue				Purple-Blue, Blue					YES
<b>Sterility</b>	Non-Sterile				Non-Sterile					YES
<b>Length (mm)</b>	<b>S</b>	<b>M</b>	<b>L</b>	<b>XL</b>	<b>XS</b>	<b>S</b>	<b>M</b>	<b>L</b>	<b>XL</b>	YES
	≥ 220	≥ 230	≥ 230	≥ 230	≥ 220	≥ 220	≥ 230	≥ 230	≥ 230	
<b>Width (mm)</b>	<b>S</b>	<b>M</b>	<b>L</b>	<b>XL</b>	<b>XS</b>	<b>S</b>	<b>M</b>	<b>L</b>	<b>XL</b>	YES
	80±10	95±10	110±10	120±10	70±10	80±10	95±10	110±10	120±10	
<b>Finger Thickness (mm)</b>	≥ 0.05				≥ 0.05					YES
<b>Palm Thickness (mm)</b>	≥ 0.05				≥ 0.05					YES
<b>Performance and Physical Comparison</b>										
<b>Tensile Strength (MPa)</b>	<b>Before Aging</b>		<b>After Aging</b>		<b>Before Aging</b>		<b>After Aging</b>		YES	
	≥ 14		≥ 14		≥ 14		≥ 14			
<b>Ultimate Elongation (%)</b>	<b>Before Aging</b>		<b>After Aging</b>		<b>Before Aging</b>		<b>After Aging</b>		YES	
	≥ 500		≥ 400		≥ 500		≥ 400			
<b>Single Use</b>	Yes				Yes					YES
<b>Freedom from hole</b>	Be free from holes when tested in accordance with ASTM D5151 AQL 2.5				Be free from holes when tested in accordance with ASTM D5151 AQL 2.5					YES
<b>Powder Content</b>	Powder residue ≤ 2.0 mg				Powder residue ≤ 2.0 mg					YES
<b>Comply with ASTM D6319</b>	Yes				Yes					YES
<b>Material Used</b>	Nitrile				Nitrile					YES
<b>Shelf life</b>	3 years				N/A					N/A
<b>Safety Comparison – Biocompatibility Test</b>										
<b>Irritation</b>	Under the conditions of the study, not an irritant				Under the conditions of the study, not an irritant					YES
<b>Sensitization</b>	Under conditions of the study, not a sensitizer				Under conditions of the study, not a sensitizer					YES
<b>Cytotoxicity</b>	Under conditions of the study, it was considered as “non-cytotoxic” at 25%, 12.5%, and 6.25% and demonstrate cytotoxicity at the 50% and 100% of the test item extract.				N/A					N/A
<b>Acute Systemic Toxicity</b>	Under the conditions of the study, there was no evidence of systemic toxicity from the extract				Under the conditions of the study, there was no evidence of systemic toxicity from the extract					YES

\*When compared to the same size, the dimensions, and physical design are identical.

**7. Performance Data**

Non-clinical tests:

Bench tests were conducted according to the FDA-recognized consensus standard, to verify that the proposed device met all design specifications. The proposed device complies with the following standards:

- ASTM D6319-19, Standard Specification for Nitrile Examination Gloves for Medical Application.
- ASTM D5151-19, Standard Test Method for Detection of Holes in Medical Gloves.
- ASTM D6124-17, Standard Test Method for Residual Powder on Medical Gloves.

Biocompatibility tests were conducted according to the FDA-recognized consensus standard, to verify the safety of the device.

- ISO 10993-5:2009, Biological Evaluation of Medical Devices - Part 5: Tests for In vitro cytotoxicity.
- ISO 10993-10:2010, Biological Evaluation of Medical Devices - Part 10: Tests for irritation and skin sensitization.
- ISO 10993-11:2017 Biological evaluation of medical devices - Part 11: Tests for systemic toxicity.

Test Methodology	Purpose	Acceptance Criteria	Results
ASTM D6319-19	Physical Dimensions Test (mm)	<u>Length (mm)</u> S ≥ 220 M ≥ 230 L ≥ 230 XL ≥ 230	<u>Length (mm)</u> S ≥ 220 M ≥ 230 L ≥ 230 XL ≥ 230
		<u>Width (mm)</u> S: 80±10 M: 95±10 L: 110±10 XL: 120±10	<u>Width (mm)</u> S: 87-89 M: 99-101 L: 110-113 XL: 112-115
		Finger Thickness (mm) ≥0.05	Finger Thickness (mm) S ≥ 0.16 M ≥ 0.15 L ≥ 0.15 XL ≥ 0.13
		Palm Thickness (mm) ≥0.05	Palm Thickness (mm) S ≥ 0.11

ELAZ Nitrile Powder Free Examination Glove

510(k) Summary

			M ≥ 0.11 L ≥ 0.11 XL ≥ 0.10
ASTM D5151-19	Testing for Freedom from holes	AQL 2.5	Gloves pass AQL 2.5
ASTM D6124-17	Determine the powder residue for powder free gloves	< 2 mg/glove	Average 0.43 mg/glove
ASTM D412 ASTM D573	Testing for Physical property characteristics	<u>Before Aging</u> Tensile strength ≥14 MPa Ultimate Elongation ≥500%  <u>After Aging</u> Tensile strength ≥14 MPa Ultimate Elongation ≥400%	<u>Before Aging</u> Tensile strength ≥21 MPa Ultimate Elongation ≥502%  <u>After Aging</u> Tensile strength ≥18 MPa Ultimate Elongation ≥465%
ISO 10993-5	Tests for In vitro cytotoxicity	Under the conditions of the study non-cytotoxic.	Under conditions of the study, it was considered “non-cytotoxic” at 25%, 12.5%, and 6.25% and demonstrate cytotoxicity at the 50% and 100% of the test item extract.
ISO 10993-10	Evaluate the endpoint of irritant for biocompatibility	Under the conditions of the study, not an irritant.	Under the conditions of the study, not an irritant.
	Evaluate the endpoint of sensitization for biocompatibility	Under the conditions of the study, not a sensitizer.	Under the conditions of the study, not a sensitizer.
ISO 10993-11	Tests for systemic toxicity	Under the conditions of the study, the device extracts do not pose a systemic toxicity concern.	Under the conditions of the study, the device extracts do not pose a systemic toxicity concern.

**8. Conclusions**

The conclusions drawn from the non-clinical tests demonstrate that the subject device, ELAZ Nitrile Powder Free Examination Glove, is as safe, as effective, and performs as well as or better than the legally marketed device.