



October 17, 2022

Jiangxi Maidikang Medical Devices Co., Ltd  
Wu Huansheng  
Management  
Room 308, floor 3, building 1, Jiangxi pharmaceutical co.,  
Ltd, No 999, kelun Avenue, pharmaceutical industrial park  
Yichun, Jiangxi  
China

Re: K221308

Trade/Device Name: Safety Syringe with permanently attached needle; Disposable Syringe with permanently attached needle

Regulation Number: 21 CFR 880.5860

Regulation Name: Piston Syringe

Regulatory Class: Class II

Product Code: FMF, MEG

Dated: September 8, 2022

Received: September 14, 2022

Dear Wu Huansheng:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

CAPT Alan M. Stevens  
Assistant Director  
DHT3C: Division of Drug Delivery and  
General Hospital Devices,  
and Human Factors  
OHT3: Office of GastroRenal, ObGyn,  
General Hospital and Urology Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K221308

Device Name

Safety Syringe with permanently attached needle  
Disposable Syringe with permanently attached needle

Indications for Use (Describe)

The Safety Syringe with permanently attached needle is intended for use by health care professionals for general purpose aspiration of fluid from vials, ampoules and liquid injection below the surface of the skin. The Safety sheath of Syringe is designed to aid in the prevention of needle stick injuries and reduce the potential of syringe reuse.

The Disposable Syringe with permanently attached needle is intended for use by health care professionals for general purpose aspiration of fluid from vials, ampoules and liquid injection below the surface of the skin.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) summary

### I Submitter

Device submitter: Jiangxi Maidikang Medical Devices Co., Ltd.

Room 308, floor 3, building 1, jiangxi pharmaceutical co., Ltd, No 999,  
kelun Avenue, pharmaceutical industrial park, Yuanzhou District,  
Yichun City, Jiangxi province

Contract manufacturer: Anhui Tiangkang Medical Technology Co., Ltd.

No.228, Weiyi Road, Economic Development Zone, Tianchang  
City, Anhui, China.

510(k) Number: K221308

Contact person:

Name: Wu Huansheng

Title: Management

Phone: +86-13617915949

E-mail: [3487486481@qq.com](mailto:3487486481@qq.com)

Preparation Date: May 6th, 2022

### II Device

Trade Name of Device: Safety Syringe with permanently attached needle

Disposable Syringe with permanently attached needle

Common Name: Piston syringe

Regulation Number: 21CFR 880.5860

Regulation Name: Piston syringe

Regulatory Class: II

Product code: FMF, MEG

Review Panel: General Hospital

### III Predicate Devices

Trade name: 1ml Luer Slip or Luer Lock Syringe  
Syringe with permanently attached needle (used as Predicate  
Device)  
Safety Syringe with permanently attached needle (used as  
Predicate Device)

Common name: Piston Syringe and antistick syringe

Classification: Class II, 21 CFR 880.5860

Product Code: FMF, MEG

Premarket Notification: K192551

Manufacturer: Jiangsu Caina Medical Co., Ltd.

**IV Device description**

The proposed Syringes include Disposable Syringe with permanently attached needle and a Safety Syringe with permanently attached needle. The Disposable Syringe with permanently attached needle have one kind of product configuration, and the Safety Syringe with permanently attached needle has two kinds of product configurations. The proposed syringes are available in different combination of syringe volumes and/or needle sizes, and the syringe is sterilized by Ethylene Oxide Gas to achieve a SAL of  $10^{-6}$  and supplied sterile in packaging to maintain the sterility of the device during the shelf life of three years.

Table 1 specification of proposed device

Model	Syringe Volume	Needle Length	Needle Gauge	Wall type	Needle Bevel
Disposable Syringe with permanently attached needle -TKSPN01	0.5ml; 1ml	3/8" (10mm); 1/2" (13mm);	23G	TW	LB
		5/8" (16mm); 3/4" (20mm);	24G	RW	LB
		7/8" (22mm); 1" (25mm);	25G	RW	LB
		3/8" (10mm); 1/2" (13mm);	26G	RW	LB
		5/8" (16mm); 3/4" (20mm);	27G	RW	LB
		3/8" (10mm); 1/2" (13mm); 5/8" (16mm);	28G	RW	LB
		3/10" (8mm); 3/8" (10mm);	29G	RW	LB
		1/2" (13mm); 5/8" (16mm);	30G	RW	LB
			31G	RW	LB
Safety Syringe with permanently attached needle -TKSSPN01	0.3ml	3/8" (10mm); 1/2" (13mm);	27G	RW	LB
		5/8" (16mm);	28G	RW	LB
		3/10" (8mm); 3/8" (10mm);	29G	RW	LB
		1/2" (13mm); 5/8" (16mm);	30G	RW	LB
			31G	RW	LB

	0.5ml; 1ml	3/8" (10mm); 1/2" (13mm); 5/8" (16mm);	25G	RW	LB			
			26G	RW	LB			
			27G	RW	LB			
			28G	RW	LB			
		3/10" (8mm); 3/8" (10mm); 1/2" (13mm); 5/8" (16mm);	29G	RW	LB			
			30G	RW	LB			
			31G	RW	LB			
			Safety Syringe with permanently attached needle-TKSSPN02	1ml	3/4" (20mm);	21G	TW	LB
					7/8" (22mm);	22G	TW	LB
					1" (25mm);	23G	TW	LB
1 1/4" (32mm);	24G	RW			LB			
1 1/2" (38mm);	25G	RW			LB			
3/4" (20mm)	26G	RW			LB			
3/4" (20mm)	27G	RW			LB			

## V Indications for use

### Disposable Syringe with permanently attached needle

The Disposable Syringe with permanently attached needle is intended for use by health care professionals for general purpose aspiration of fluid from vials, ampoules and liquid injection below the surface of the skin.

### Safety Syringe with permanently attached needle

The Safety Syringe with permanently attached needle is intended for use by health care professionals for general purpose aspiration of fluid from vials, ampoules and liquid injection below the surface of the skin. The Safety sheath of Syringe is designed to aid in the prevention of needle stick injuries and reduce the potential of syringe reuse.

## VI Comparison of technological characteristics with the predicate devices

The Disposable Syringe with permanently attached needle and Safety Syringe with permanently attached needle have the same intended use, technology, design and performance specifications to the legally marketed predicate device. The main differences between subject devices and predicate devices are the specification of syringe volume, needle gauge and needle length. This difference does not affect indication for use. The difference between the subject device and the predicate does not affect the safety and effectiveness of the subject device T

Table 6-1 Substantial equivalence discussion for Disposable Syringe with permanently attached needle

Device feature	Subject Device	Predicate Device K192551	Comment
Product	Disposable Syringe with permanently attached needle	Syringe with permanently attached needle	/
Syringe type	Piston syringe	Piston syringe	Same
Indications for use	The Disposable Syringe with permanently attached needle is intended for use by health care professionals for general purpose aspiration of fluid from vials, ampoules and liquid injection below the surface of the skin.	The Syringe with permanently attached needle is intended for use by health care professionals for general purpose aspiration of fluid from vials, ampoules and liquid injection below the surface of the skin.	Same
Product code	FMF	FMF	Same
Regulation number	21 CFR 880.5860	21 CFR 880.5860	Same
Class	II	II	Same
Principle of operation	For Manual Use Only, For Single Use Only	For Manual Use Only, For Single Use Only	Same
Connector Type	Attached needle	Attached needle	Same
Needle gauge	23G, 24G, 25G, 26G, 27G, 28G, 29G, 30G, 31G	21G, 23G, 25G, 26G, 27G, 28G, 29G, 30G, 31G	Difference 1
Needle Length	8mm, 10mm, 13mm, 16mm, 20mm, 22mm, 25mm	8mm, 10mm, 13mm, 16mm, 20mm, 25mm	
Needle wall type	RW, TW	RW, TW	
Needle bevel	11°±2°	11°±2°, 15°±2°	
Syringe Volume	0.5ml, 1ml	0.3ml, 0.5ml, 1ml	

Device feature	Subject Device	Predicate Device K192551	Comment
Main structure and materials	(1) Needle cap (PP) (2) Needle (Stainless Steel 304) (3) Piston (Polyisoprene) (4) Plunger (PP) (5) Barrel (PP)	Syringe with permanently attached needle-type A (1) needle cap (PP or PE) (2) needle (Stainless Steel 304) (3) piston (Polyisoprene) (4) plunger (PP) (5) barrel (PP) (6) end cap (PP or PE) Syringe with permanently attached needle-type B (1) needle cap (PP or PE) (2) needle (Stainless Steel 304) (3) piston (Polyisoprene) (4) plunger (PP or ABS) (5) barrel (PP)	Difference 2
Performance specifications	Complies with ISO 7886-1; ISO 9626 and ISO 7864	Complies with ISO 7886-1; ISO 9626 and ISO 7864	Same
Single Use	Yes	Yes	Same
Sterilization	EO Sterilization	EO Sterilization	Same
SAL	10 <sup>-6</sup>	10 <sup>-6</sup>	Same
Biocompatibility	Contact level: blood path, indirect, limited contact (<24 hours). Conforms to the requirement of ISO 10993 series Standards	Contact level: blood path, indirect, limited contact (<24 hours). Conforms to the requirement of ISO 10993 series Standards	Same
Labeling	Complied with 21 CFR part 801	Complied with 21 CFR part 801	Same

#### Difference 1

The needle bevel of subject device ( $11^{\circ} \pm 2^{\circ}$ ) is smaller than the predicate device's bevel range of  $11^{\circ} \pm 2^{\circ}$  and  $15^{\circ} \pm 2^{\circ}$ . The subject device's 0.5ml and 1ml syringe volume is smaller than the predicated device's syringe volume range of 0.3ml, 0.5ml and 1ml. And the needle gauge and length of subject devices are similar to the predicate device. This difference does not affect intended use. In addition, differences were addressed through ISO 7886-1, ISO 7864 and ISO 9626. The differences



on syringe volume, needle bevel, needle gauge and length do not raise different question of safety and effectiveness.

**Difference 2**

The configuration of subject device is same as type B of predicate device, and compared with type A, the difference is that the subject device does not have an end cap. but the syringe with attached needle is widely used in the clinical. Whether there is an end cap or not will not affect the indication for use of the equipment itself. This difference does not raise new questions about safety and effectiveness.

The materials of subject device are similar to the predicated device. The proposed syringes' biocompatibility can be demonstrated by the biocompatibility tests. The differences on configuration and materials do not raise new questions about safety and effectiveness.

Table 6-2 Substantial equivalence discussion for Safety Syringe with permanently attached needle

<b>Device feature</b>	<b>Subject Device</b>	<b>Predicate Device K192551</b>	<b>Comment</b>
Product	Safety Syringe with permanently attached needle	Safety Syringe with permanently attached needle	/
Syringe type	Piston syringe	Piston syringe	Same
Indications for use	The Safety Syringe with permanently attached needle is intended for use by health care professionals for general purpose aspiration of fluid from vials, ampoules and liquid injection below the surface of the skin. The Safety sheath of Syringe is designed to aid in the prevention of needle stick injuries and reduce the potential or syringe reuse.	The Safety Syringe with permanently attached needle is intended for use by health care professionals for general purpose aspiration of fluid from vials, ampoules and liquid injection below the surface of the skin. The Safety sheath of Syringe is designed to aid in the prevention of needle stick injuries and reduce the potential or syringe reuse.	Same
Product code	MEG	MEG	Same
Regulation number	21 CFR 880.5860	21 CFR 880.5860	Same
Class	II	II	Same

Device feature	Subject Device	Predicate Device K192551	Comment
Principle of operation	For Manual Use Only, For Single Use Only	For Manual Use Only, For Single Use Only	Same
Connector Type	Attached needle	Attached needle	Same
Needle gauge	21G, 22G, 23G, 24G, 25G, 26G, 27G, 28G, 29G, 30G, 31G	21G, 23G, 25G, 26G, 27G, 28G, 29G, 30G, 31G	Difference 3
Needle Length	8mm, 10mm, 13mm, 16mm, 20mm, 22mm, 25mm, 32mm, 38mm	8mm, 10mm, 13mm, 16mm, 20mm, 25mm	
Needle wall type	RW, TW	RW, TW	
Needle bevel	11°±2°	11°±2°, 15°±2°	
Syringe Volume	0.3ml, 0.5ml, 1ml	0.3ml, 0.5ml, 1ml	Same
Main structure and materials	(1) Protective cap (PP) (2) Needle (Stainless Steel 304) (3) Safety mechanism- Connector base (PC) (4) Piston (Polyisoprene) (5) Safety mechanism- Sliding sleeve (PP) (6) Plunger (PP) (7) Barrel (PP)	(1) needle cap (PP or PE) (2) needle (Stainless Steel 304) (3) safety mechanism (PC) (4) piston (Polyisoprene) (5) safety mechanism (PP) (6) plunger (PP) (7) barrel (PP)	Same
Performance specifications	Complies with ISO 7886-1; ISO 9626 and ISO 7864	Complies with ISO 7886-1; ISO 9626 and ISO 7864	Same
Single Use	Yes	Yes	Same
Sterilization	EO Sterilization	EO Sterilization	Same
SAL	10 <sup>-6</sup>	10 <sup>-6</sup>	Same
Biocompatibility	Contact level: blood path, indirect, limited contact (<24 hours). Conforms to the requirement of ISO 10993 series Standards	Contact level: blood path, indirect, limited contact (<24 hours). Conforms to the requirement of ISO 10993 series Standards	Same
Labeling	Complied with 21 CFR part 801	Complied with 21 CFR part 801	Same

### Difference 3

The needle bevel of subject device ( $11^{\circ}\pm 2^{\circ}$ ) is smaller than the predicate device's bevel range of  $11^{\circ}\pm 2^{\circ}$  and  $15^{\circ}\pm 2^{\circ}$ . And the needle gauge and length of subject devices are similar to the predicate device, the difference is just in dimension. This difference does not affect intended use. In addition, differences were addressed through ISO 7864 and ISO 9626. The differences on syringe volume, needle bevel, needle gauge and length do not raise different question of safety and effectiveness.

### VII Performance data

The following performance data were provided in support of the substantial equivalence determination.

#### Biocompatibility testing

Biocompatibility of the Disposable Syringe with permanently attached needle and Safety Syringe with permanently attached needle were evaluated in accordance with ISO 10993-1:2018 for the body contact category of "External communication device – Blood path indirect" with a contact duration of "Limited (< 24 hours)". The following tests were performed, as recommended:

Cytotoxicity	ISO 10993-5: 2009
Skin sensitization	ISO 10993-10: 2010
Hemolysis	ISO 10993-4: 2017
Intracutaneous reactivity	ISO 10993-10: 2010
Acute systemic toxicity	ISO 10993-11: 2017
Pyrogenicity	ISO 10993-11: 2017
Particulate matter	USP <788>

#### Sterilization and shelf life testing

The sterilization method has been validated to ISO11135, which has thereby determined the routine control and monitoring parameters. Bacteria Endotoxin Limit is carried out according to USP42-NF37 <85> Bacterial Endotoxins Test.

The testing is performed according to the following standards:

EO residue	ISO 10993-7:2008
ECH residue	ISO 10993-7:2008
Bacteria Endotoxin	USP42-NF37 <85>

The shelf life of 3 year is determined based on stability studies which include ageing test according to FDA recognized standard ASTM F1980-16.

Package integrity testing was conducted on the final, packaged, and sterile devices after environmental conditioning and simulated transportation. All packaging deemed acceptable for protection of product and sterility maintenance.

The testing is performed according to the following standards:

Seal strength	ASTM F88/F88M-15
---------------	------------------

Blue Dye Penetration  
Seal Integrity (Visual Inspection)

ASTM F 1929-2015  
ASTM F 1886/ F 1886M-16

### **Performance testing**

Performance testing is performed according to the following standards:

- ISO 7886-1: 2017, Sterile hypodermic syringes for single use- Part 1: Syringes for manual use.
- ISO 7886-4: 2018, Sterile hypodermic syringes for single use- Part 4: Syringes with re-use prevention feature
- ISO 7864: 2016, ISO 9626:2016, Stainless Steel Needle Tubing For The Manufacture of Medical Devices.
- ISO 9626:2016, Stainless Steel Needle Tubing for The Manufacture of Medical Devices.
- ISO 23908: 2011 Sharps injury protection - Requirements and test methods – Sharps protection features for single-use hypodermic needles, introducers for catheters and needles used for blood sampling.

### **VIII Clinical Test Conclusion**

No clinical study is included in this submission.

### **IX Conclusion**

The Disposable Syringe with permanently attached needle and Safety Syringe with permanently attached needle are substantially equivalent to its predicate devices. The non-clinical testing demonstrates that the device is as safe, as effective and performs as well as the legally marketed device.