



October 27, 2022

JR Engineering & Medical Technologies (M) SDN.BHD.
% Manoj Zacharias
Consultant
Liberty Management Group Ltd.
75 Executive Dr. STE 114
Aurora, Illinois 60504

Re: K222348

Trade/Device Name: Latex Examination Powder Free Gloves
Regulation Number: 21 CFR 880.6250
Regulation Name: Non-Powdered Patient Examination Glove
Regulatory Class: Class I, reserved
Product Code: LYY
Dated: September 28, 2022
Received: September 28, 2022

Dear Manoj Zacharias:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Bifeng Qian, M.D., Ph.D.
Assistant Director
DHT4B: Division of Infection Control
and Plastic Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K222348

Device Name

Latex Examination Powder Free Gloves

Indications for Use (Describe)

Latex Examination Powder Free Gloves are disposable device intended for medical purpose that is worn on the examiner's hand to prevent contamination between patient and examiner.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510K SUMMARY

K222348

As required by 21CFR§807.92(c)

A. APPLICANT INFORMATION

Applicant	JR Engineering & Medical Technologies (M) SDN.BHD.
Address	Lot 8 &10, Jalan Zurah 3 & Lot 1&3, Jalan Zurah 3A/1, Pusat Perindustrian 2, 44200 Rasa, Hulu Selangor, Selangor Darul Ehsan, Malaysia.
Phone	+603-60572081
Fax	+603-60572181
E-mail	ganeshjrmt@gmail.com
Contact Person	Mr. Ganesan Subramaniam
Designation	Managing Director
Contact Number	+6012 224 6677
Contact Email	ganeshjrmt@gmail.com
Date Submitted	27 October 2022

B. DEVICE IDENTIFICATION

Name of the device	Latex Examination Powder Free Gloves
Product proprietary or trade name	JR MEDIC
Common or usual name	Latex Examination Powder Free Gloves
Classification name	Patient Examination Gloves
Device Classification	Class I
Product Code	LYY
Regulation Number	21 CFR 880.6250
Review Panel	General Hospital

C. PREDICATE DEVICE

Legally Marketed devices that Equivalency is claimed	JR Medic Powder free Latex Examination Gloves
510(K) Number	K192329
Regulatory Class	Class I
Product code	LYY

D. DESCRIPTION OF THE DEVICE:

Latex Examination Powder Free Gloves are manufactured to meet all the current specifications listed under the ASTM Specification D3578-2019, Standard Specification for Rubber Examination Gloves. They are made from Natural Rubber Latex. These gloves are natural in color (No color is added) and are powder free and Non-Sterile.

E. INDICATIONS FOR USE/INTENDED USE OF THE DEVICE:

Latex Examination Powder Free Gloves are disposable devices intended for medical purpose that are worn on the examiner’s hand to prevent contamination between patient and examiner.

F. TECHNOLOGICAL CHARACTERISTICS

Characteristics	Standards	Device Performance		Comparison	
		Predicate	Current		
510(K) Number	---	K192329	K222348		
Name of device	---	JR MEDIC Blue Latex Examination Powder Free Gloves	Latex Examination Powder Free Gloves	---	
Dimensions- Length	ASTM D3578-2019	Length > 230 mm	Length > 230 mm		Similar
			Size	Average	
			X-Small	305	
			Small	306	
			Medium	307	
			Large	308	
Dimensions- Width	ASTM D3578-2019	Width Min 95+/- 10 mm (for medium size)	Width Min 95+/-10 mm (for medium size)		Similar
			Size	Average	
			X-Small	76	
			Small	85	
			Medium	96	
			Large	106	
Physical Properties- Tensile Strength	ASTM D3578-2019	<u>Before Ageing</u> Tensile Strength > 18 Mpa	<u>Before Ageing</u> Tensile Strength > 18 Mpa		Similar
			Size	Actual value	
			X-Small	22.07	
			Small	22.15	
			Medium	22.22	
			Large	22.30	
		<u>After Ageing</u> Tensile Strength > 14 Mpa	<u>After Ageing</u> Tensile Strength > 14 Mpa		Similar
			Size	Actual value	
			X-Small	18.49	
			Small	18.56	
			Medium	18.67	
			Large	18.74	
X-Large	18.76				

Characteristics	Standards	Device Performance			Comparison	
		Predicate	Current			
510(K) Number		K192329	K222348			
Physical Properties- Ultimate Elongation	ASTM D3578-2019	Before Ageing Ultimate Elongation > 650%	Before Ageing Ultimate Elongation > 650%		Similar	
			Size	Actual value		
			X-Small	858		
			Small	869		
			Medium	874		
			Large	880		
		X-Large	882			
		After Ageing Ultimate Elongation >500%	After Ageing Ultimate Elongation > 500%			
			Size	Actual value		
			X-Small	841		
			Small	848		
			Medium	854		
Large	860					
X-Large	862					
Thickness	ASTM D3578-2019	Palm > 0.08 mm Finger > 0.08 mm	Palm > 0.08 mm Finger > 0.08 mm		Similar	
			Size	Palm (Actual value)		Finger (Actual value)
			X-Small	0.16		0.22
			Small	0.16		0.22
			Medium	0.16		0.22
			Large	0.16		0.22
			X-Large	0.16		0.22
Powder Free Residue	ASTM D6214	≤2 mg/glove	≤2 mg/glove		Similar	
			Size	Residual powder content (mg/glove)		
			X-Small	0.21		
			Small	0.21		
			Medium	0.22		
			Large	0.22		
X-Large	0.22					
Protein Content	ASTM D5712	Max 200 µg/ dm ²	43.19 µg/ dm ²		Similar	
Biocompatibility	Primary Skin Irritation-ISO 10993-10:2010(E)	Under the condition of study, not an irritant	Under the condition of study not an irritant		Same	
	Dermal Sensitization-ISO 10993-10:2010(E)	Under the conditions of the study, not a sensitizer	Under the conditions of the study, not a sensitizer		Same	
	In vitro cytotoxicity ISO10993-5 :2009(E)	Under the conditions of the study, cytotoxic	Under the conditions of the study, cytotoxic		Same	
	Material mediated Pyrogenicity ISO 1099311:2017(E) / USP 41<151>	Under the conditions of the study non pyrogenic	No data available		----	

Characteristics	Standards	Device Performance		Comparison
		Predicate	Current	
510(K) Number		K192329	K222348	
Biocompatibility	Acute Systemic Toxicity Test ISO 10993-11:2017(E)	Under the conditions of study the device extracts do not pose a systemic toxicity concern	Under the conditions of study the device extracts do not pose a systemic toxicity concern	same
Water Tight (1000 ml)	ASTM D5151-2019	Passes AQL-1.5	Passes AQL-1.5	Same
Intended use/ Indication for use		JR MEDIC Blue Latex Examination Powder Free Gloves are disposable devices intended for medical purpose that are worn on the examiner's hand to prevent contamination between patient and examiner.	Latex Examination Powder Free Gloves are disposable devices intended for medical purpose that are worn on the examiner's hand to prevent contamination between patient and examiner.	Similar
Material	-	Natural Latex	Natural Latex	Same
Color	-	Blue	Natural (No color is added)	different
Texture	-	Finger Texture	Finger texture	Same
Size	ASTM D3578-2019	Small, Medium, Large & X Large	X Small, Small, Medium, Large, X-Large	Similar
Single Use	Medical Glove Guidance Manual - Labeling	Single Use	Single Use	Same
Sterile/non sterile	-	Non sterile	Non sterile	Same
Powder/Powder free	-	Powder free	Powder free	Same
Label and Labeling	FDA Label requirements	Meets FDA's label and labeling requirements	Meets FDA's label and labeling requirements	Same
Manufacturer(s)	-	JR Engineering & Medical Technologies (M) SDN.BHD. Malaysia.	JR Engineering & Medical Technologies (M) SDN.BHD. Malaysia	Same

G. SUMMARY OF NON-CLINICAL PERFORMANCE DATA

Test Method	Purpose	Acceptance Criteria	Result
ASTM D3578-2019 Standard Specification for Rubber Examination Gloves	To determine the length of the gloves	Min 230 mm for all sizes	X-Small 305 mm Small 306 mm Medium 307 mm Large 308 mm X-Large 310 mm
ASTM D3578-2019 Standard Specification for Rubber Examination Gloves	To determine the width of the gloves	X-Small 70+/-10 mm Small 80+/-10mm Medium 95+/-10 mm Large 111+/-10 mm X-Large 115+/-10 mm	X-Small 76 mm Small 85 mm Medium 96 mm Large 106 mm X-Large 116 mm

Test Method	Purpose	Acceptance Criteria	Result		
ASTM D3578-2019 Standard Specification for Rubber Examination Gloves	To determine the thickness of the gloves	Palm 0.08 mm min Finger 0.08 mm min for all sizes	Size X-Small Small Medium Large X-Large	Palm 0.16mm 0.16mm 0.16mm 0.16mm 0.16mm	Finger 0.22mm 0.22mm 0.22mm 0.22mm 0.22mm
ASTM D3578-2019 Standard Specification for Rubber Examination Gloves	To Determine the physical properties-Tensile strength	Before Ageing Tensile Strength 18Mpa Min for all sizes After Ageing Tensile Strength 14Mpa Min for all sizes	Size X-Small Small Medium Large X-Large	Before ageing 22.07 Mpa 22.15 Mpa 22.22 Mpa 22.30 Mpa 22.32 Mpa	After ageing 18.49 Mpa 18.56 Mpa 18.67 Mpa 18.74 Mpa 18.76 Mpa
	To Determine the physical properties-Ultimate Elongation	Before Ageing Ultimate Elongation 650% Min for all sizes After Ageing Ultimate Elongation 500% Min for all sizes	Size X-Small Small Medium Large X-Large	Before ageing 858% 869% 874% 880% 882%	After ageing 841% 848% 854% 860% 862%
	To Determine the physical properties-stress at 500% Elongation	Before Ageing 5.5 Mpa Max for all sizes	Size X-Small Small Medium Large X-Large	Before ageing 5.1 Mpa 5.1 Mpa 5.2 Mpa 5.2 Mpa 5.2 Mpa	NA
ASTM D5151-2019 Standard Test Method for Detection of Holes in Medical Gloves	To determine the holes in the gloves	AQL 2.5	Gloves Passes AQL 1.5		
ASTM D6124-06 (Reapproved 2017) Standard Test Method for Residual Powder on Medical Gloves	To determine the residual powder in the gloves	2 Mg/Glove Max	Size X-Small Small Medium Large X-Large	Residual Powder Content 0.21 mg/glove 0.21 mg/glove 0.22 mg/glove 0.22 mg/glove 0.22 mg/glove	
ASTM D 5712-95 (Re approved 2010) Standard Test Method for the Analysis of Protein in Natural Rubber	To determine the extractable protein in the gloves.	200 µg/ dm ² Max for all sizes	Size X-Small Small Medium Large X-Large	Extractable Protein content 43.65 µg/ dm ² 43.65 µg/ dm ² 43.65 µg/ dm ² 43.65 µg/ dm ² 43.65 µg/ dm ²	

The performance test data of the non-clinical tests meet following standards:

- ASTM D 3578-2019 Standard Specification for Rubber Examination Gloves
- ASTM D 5151-2019 Standard Test Method for Detection of Holes in Medical Gloves
- ASTM D 6124-06 (Reapproved 2017) Standard Test Method for Residual Powder on Medical Gloves
- ASTM D 5712-95 (Re approved 2010) Standard Test Method for the Analysis of Protein in Natural Rubber

H. SUMMARY OF CLINICAL TESTING

Not applicable

I. CONCLUSION

The conclusion drawn from the nonclinical tests demonstrates that the subject device in 510(K) submission Latex Examination Powder Free Gloves is as safe, as effective, and performs as well as or better than the legally marketed predicate device *K192329*.