



March 29, 2023

Stryker Neurovascular
Heli Chambi
Staff Regulatory Affairs Specialist
47900 Bayside Parkway
Fremont, California 94538

Re: K223305

Trade/Device Name: Trevo NXT ProVue Retriever

Regulation Number: 21 CFR 882.5600

Regulation Name: Neurovascular Mechanical Thrombectomy Device for Acute Ischemic Stroke
Treatment

Regulatory Class: Class II

Product Code: POL, NRY

Dated: February 21, 2023

Received: February 23, 2023

Dear Heli Chambi:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Naira Muradyan -S

Naira Muradyan, Ph.D.

Assistant Director

DHT5A: Division of Neurosurgical,

Neurointerventional

and Neurodiagnostic Devices

OHT5: Office of Neurological

and Physical Medicine Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K223305

Device Name
Trevor NXT ProVue Retriever

Indications for Use (Describe)

1. The Trevo Retriever is indicated for use to restore blood flow in the neurovasculature by removing thrombus for the treatment of acute ischemic stroke to reduce disability in patients with a persistent, proximal anterior circulation, large vessel occlusion, and smaller core infarcts who have first received intravenous tissue plasminogen activator (IV t-PA). Endovascular therapy with the device should start within 6 hours of symptom onset.
2. The Trevo Retriever is intended to restore blood flow in the neurovasculature by removing thrombus in patients experiencing ischemic stroke within 8 hours of symptom onset. Patients who are ineligible for intravenous tissue plasminogen activator (IV t-PA) or who fail IV t-PA therapy are candidates for treatment.
3. The Trevo Retriever is indicated for use to restore blood flow in the neurovasculature by removing thrombus for the treatment of acute ischemic stroke to reduce disability in patients with a persistent, proximal anterior circulation, large vessel occlusion of the internal carotid artery (ICA) or middle cerebral artery (MCA)-M1 segments with smaller core infarcts (0-50cc for age <80 years, 0-20cc for age ≥80 years). Endovascular therapy with the device should start within 6-24 hours of time last seen well in patients who are ineligible for intravenous tissue plasminogen activator (IV t-PA) or who fail IV t-PA therapy.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary
K223305

Introduction:

According to the requirements of 21 CFR 807.92, the following information provides sufficient details to understand the basis for a determination of substantial equivalence.

Submitter Name, Address and Contact:

Submitter: Stryker Neurovascular
47900 Bayside Parkway
Fremont, CA 94538-6515
(FDA Registration Number: 3008853977)

Contact: **Heli Chambi**
Staff Regulatory Affairs Specialist
Phone: 239-366-6826
Fax: 510-413-2588
Email: heli.chambi@stryker.com

Date Prepared: March 29, 2023

Device Name and Classification:

Trade/Proprietary Name: Trevo NXT ProVue Retriever

Common Name: Trevo Retriever

Classification Name: Neurovascular Mechanical Thrombectomy Device for Acute Ischemic Stroke Treatment, 21 CFR 882.5600, Class II
Catheter, Thrombus Retriever, 21 CFR 870.1250 – Class II

Product Code: POL, NRY

Legally Marketed Predicate Device

Name of Predicate Device	Name of Manufacturer	510(k) Number
Trevo NXT ProVue Retriever	Stryker Neurovascular	K210502

Device Description

The Trevo Retriever consists of a flexible, tapered core wire with a shaped section at the distal end. Platinum markers at the distal end allow fluoroscopic visualization. In addition, the shaped section is also radiopaque. Retriever dimensions are indicated on product label. The Retriever delivery wire has a hydrophilic coating on the distal 101cm length to reduce friction during use. The Retriever has a shaft marker to indicate proximity of Retriever tip relative to Microcatheter tip. A Torque Device is provided with the Retriever to facilitate manipulation and retrieval. The Retriever comes preloaded in an insertion tool to introduce the Retriever into a Microcatheter.

Indications for Use

1. The Trevo Retriever is indicated for use to restore blood flow in the neurovasculature by removing thrombus for the treatment of acute ischemic stroke to reduce disability in patients with a persistent, proximal anterior circulation, large vessel occlusion, and smaller core infarcts who have first received intravenous tissue plasminogen activator (IV t-PA). Endovascular therapy with the device should start within 6 hours of symptom onset.
2. The Trevo Retriever is intended to restore blood flow in the neurovasculature by removing thrombus in patients experiencing ischemic stroke within 8 hours of symptom onset. Patients who are ineligible for intravenous tissue plasminogen activator (IV t-PA) or who fail IV t-PA therapy are candidates for treatment.
3. The Trevo Retriever is indicated for use to restore blood flow in the neurovasculature by removing thrombus for the treatment of acute ischemic stroke to reduce disability in patients with a persistent, proximal anterior circulation, large vessel occlusion of the internal carotid artery (ICA) or middle cerebral artery (MCA)-M1 segments with smaller core infarcts (0-50cc for age <80 years, 0-20cc for age ≥80 years). Endovascular therapy with the device should start within 6-24 hours of time last seen well in patients who are ineligible for intravenous tissue plasminogen activator (IV t-PA) or who fail IV t-PA therapy.

Technological Characteristics and Product Feature Comparison

There are no differences in the technological characteristics between the subject device and predicate device because they both have the same indications for use, device design, materials, manufacturing, packaging, and sterilization methods. The changes made to the subject device are solely labeling changes to update compatibility information in the Instructions for Use (IFU).


The differences between the subject and predicate devices are summarized in **Table 1** below.

Table 1. Product Feature Comparison of Subject Device to Predicate Device

Feature	<u>Predicate Device</u> Trevo NXT ProVue Retriever (K210502)	<u>Subject Device</u> Trevo NXT ProVue Retriever
Regulation Number	<ul style="list-style-type: none"> • 21 CFR 882.5600 • 21 CFR 870.1250 	Same
Regulation Name	<ul style="list-style-type: none"> • Neurovascular Mechanical Thrombectomy Device for Acute Ischemic Stroke Treatment • Percutaneous Catheter 	Same
Classification	Class II	Same
Product Code	<ul style="list-style-type: none"> • POL • NRY 	Same
Intended Use	Neurovascular mechanical thrombectomy device for acute ischemic stroke treatment used in the treatment of acute ischemic stroke to improve clinical outcomes.	Same
Indications for Use	<ol style="list-style-type: none"> 1. The Trevo Retriever is indicated for use to restore blood flow in the neurovasculature by removing thrombus for the treatment of acute ischemic stroke to reduce disability in patients with a persistent, proximal anterior circulation, large vessel occlusion, and smaller core infarcts who have first received intravenous tissue plasminogen activator (IV t-PA). Endovascular therapy with the device should start within 6 hours of symptom onset. 2. The Trevo Retriever is intended to restore blood flow in the neurovasculature by removing thrombus in patients experiencing ischemic stroke within 8 hours of symptom onset. Patients who are ineligible for intravenous tissue plasminogen activator (IV 	Same

Feature	<u>Predicate Device</u> Trevo NXT ProVue Retriever (K210502)	<u>Subject Device</u> Trevo NXT ProVue Retriever
	<p>t-PA) or who fail IV t-PA therapy are candidates for treatment.</p> <p>3. The Trevo Retriever is indicated for use to restore blood flow in the neurovasculature by removing thrombus for the treatment of acute ischemic stroke to reduce disability in patients with a persistent, proximal anterior circulation, large vessel occlusion of the internal carotid artery (ICA) or middle cerebral artery (MCA)-M1 segments with smaller core infarcts (0-50cc for age <80 years, 0-20cc for age ≥80 years). Endovascular therapy with the device should start within 6-24 hours of time last seen well in patients who are ineligible for intravenous tissue plasminogen activator (IV t-PA) or who fail IV t-PA therapy.</p>	
Target Population	Patients experiencing acute ischemic stroke	Same
Anatomical Sites	Neurovasculature	Same
Device Description	<p>The Retriever consists of a flexible, tapered core wire with a shaped section at the distal end. Platinum markers at the distal end allow fluoroscopic visualization. In addition, the shaped section is also radiopaque. Retriever dimensions are indicated on product label. The Retriever delivery wire has a hydrophilic coating on the distal 101cm to reduce friction during use. The Retriever has a shaft marker to indicate proximity of Retriever tip relative to Microcatheter tip. A torque device is provided with the Retriever to facilitate manipulation and retrieval. The Retriever comes preloaded in an insertion tool to introduce the Retriever into a Microcatheter.</p>	Same
Principle of Operation	<p>The Trevo Retriever is delivered to the thrombus using a microcatheter. The Microcatheter is then retracted to deploy the shaped section of the Retriever. If using an Aspiration Catheter, remove Microcatheter. Advance Aspiration Catheter over proximal section of Retriever while applying aspiration using a 60 mL syringe or an aspiration pump. The Retriever with Microcatheter or</p>	Same

Feature	<u>Predicate Device</u> Trevo NXT ProVue Retriever (K210502)	<u>Subject Device</u> Trevo NXT ProVue Retriever																																																							
	Aspiration Catheter are pulled back to capture the thrombus. The Retriever, thrombus, and Microcatheter or Aspiration Catheter are removed as a unit from the body.																																																								
Procedural Steps Aspiration Source	Syringe, Aspiration pump	Same																																																							
Accessory Devices	Insertion tool and torque device provided within product package	Same																																																							
Sizes	3x32mm 4x28mm 4x41mm 6x37mm	Same																																																							
Compatibility	<table border="1"> <thead> <tr> <th>Retriever Size</th> <th>Trevo Pro14 Microcatheter</th> <th>Trevo Trak 21 Microcatheter*</th> <th>Trevo Pro18 Microcatheter</th> <th>Excelsior® XT-27 Microcatheter (REF XT275081)</th> <th>Recommended Minimum Vessel ID (mm)</th> </tr> </thead> <tbody> <tr> <td>Trevo NXT 3x32</td> <td>√</td> <td>√</td> <td>√</td> <td>√</td> <td rowspan="4">2.5</td> </tr> <tr> <td>Trevo NXT 4x28</td> <td></td> <td>√</td> <td>√</td> <td>√</td> </tr> <tr> <td>Trevo NXT 4x41</td> <td></td> <td>√</td> <td>√</td> <td>√</td> </tr> <tr> <td>Trevo NXT 6x37</td> <td></td> <td>√</td> <td>√</td> <td>√</td> </tr> </tbody> </table>	Retriever Size	Trevo Pro14 Microcatheter	Trevo Trak 21 Microcatheter*	Trevo Pro18 Microcatheter	Excelsior® XT-27 Microcatheter (REF XT275081)	Recommended Minimum Vessel ID (mm)	Trevo NXT 3x32	√	√	√	√	2.5	Trevo NXT 4x28		√	√	√	Trevo NXT 4x41		√	√	√	Trevo NXT 6x37		√	√	√	<table border="1"> <thead> <tr> <th>Retriever Size</th> <th>Trevo Pro14 Microcatheter</th> <th>Trevo Trak 21 Microcatheter*</th> <th>Trevo Pro18 Microcatheter</th> <th>Excelsior® XT-27 Microcatheter (REF XT275081)</th> <th>Recommended Minimum Vessel ID (mm)</th> </tr> </thead> <tbody> <tr> <td>Trevo NXT 3x32</td> <td>√</td> <td>√</td> <td>√</td> <td>√</td> <td rowspan="3">1.5</td> </tr> <tr> <td>Trevo NXT 4x28</td> <td></td> <td>√</td> <td>√</td> <td>√</td> </tr> <tr> <td>Trevo NXT 4x41</td> <td></td> <td>√</td> <td>√</td> <td>√</td> </tr> <tr> <td>Trevo NXT 6x37</td> <td></td> <td>√</td> <td>√</td> <td>√</td> <td>2.5</td> </tr> </tbody> </table>	Retriever Size	Trevo Pro14 Microcatheter	Trevo Trak 21 Microcatheter*	Trevo Pro18 Microcatheter	Excelsior® XT-27 Microcatheter (REF XT275081)	Recommended Minimum Vessel ID (mm)	Trevo NXT 3x32	√	√	√	√	1.5	Trevo NXT 4x28		√	√	√	Trevo NXT 4x41		√	√	√	Trevo NXT 6x37		√	√	√	2.5
	Retriever Size	Trevo Pro14 Microcatheter	Trevo Trak 21 Microcatheter*	Trevo Pro18 Microcatheter	Excelsior® XT-27 Microcatheter (REF XT275081)	Recommended Minimum Vessel ID (mm)																																																			
Trevo NXT 3x32	√	√	√	√	2.5																																																				
Trevo NXT 4x28		√	√	√																																																					
Trevo NXT 4x41		√	√	√																																																					
Trevo NXT 6x37		√	√	√																																																					
Retriever Size	Trevo Pro14 Microcatheter	Trevo Trak 21 Microcatheter*	Trevo Pro18 Microcatheter	Excelsior® XT-27 Microcatheter (REF XT275081)	Recommended Minimum Vessel ID (mm)																																																				
Trevo NXT 3x32	√	√	√	√	1.5																																																				
Trevo NXT 4x28		√	√	√																																																					
Trevo NXT 4x41		√	√	√																																																					
Trevo NXT 6x37		√	√	√	2.5																																																				
	Balloon Guide Catheters and Aspiration Catheters (commercially available aspiration catheters with minimum inner diameter 0.046 inches (1.17mm)) are recommended for use during thrombus removal procedures.	Same																																																							
Core Wire	Nitinol (nickel titanium alloy)	Same																																																							
Shaped Section	Nitinol	Same																																																							
Distal Coil	Platinum/Tungsten	Same																																																							
Shaped Section Radiopaque Wire	Platinum/Tungsten	Same																																																							
Mid Coil	304 Stainless Steel	Same																																																							

Feature	<u>Predicate Device</u> Trevo NXT ProVue Retriever (K210502)	<u>Subject Device</u> Trevo NXT ProVue Retriever
Proximal Coil	Pebax	Same
Solder	Gold/Tin	Same
Hydrophilic Coating	Sodium hyaluronate mixture	Same
Dimensional Drawing	All Retriever sizes: 	Same
Overall Length (A)	200 cm	Same
Total Shaped Section Length (B)	32, 36, 40, 44 mm	Same
Full Diameter Length (C)	21, 25, 30, 35 mm	Same
Shaped Section Diameter (D)	3, 4, 6 mm	Same
Delivery Core Wire Outer Diameter (E)	0.015, 0.019 inches	Same
Cell Coverage Length (F)	28, 32, 37, 41 mm	Same
Packaging Materials and Configuration	Polyethylene Hoop, Polycarbonate Mounting Card, Tyvek/Film Pouch, Chipboard Carton	Same
Sterilization Method	100% EtO	Same
How Supplied	Sterile/Single Use	Same

Risk Assessment

Risk assessment of the Trevo NXT ProVue Retriever has been conducted in accordance with EN ISO 14971. Stryker Neurovascular has determined that the labeling changes to the Trevo NXT ProVue Retriever raise no new questions of safety or effectiveness. Results of testing are appropriate for determining that the Trevo NXT ProVue Retriever with the modified IFU is substantially equivalent to the legally marketed Predicate device.

Testing Summary

There are no changes to the device intended use or indications for use statement. Other than the proposed labeling changes regarding compatibility, there are no changes in the device design, materials, manufacturing, packaging and sterilization methods; therefore, biocompatibility data, bench performance data, sterilization and stability data from the Predicate device (**K210502**) are directly applicable. Relevant testing data supporting the subject device are summarized as follows.

Performance Data – Bench Testing

There are no changes to the device intended use or indications for use statement. Other than the labeling modification to recommend use of some Trevo retriever models in vessels with minimum diameter of 1.5 mm, there are no changes to the device design, materials, manufacturing and packaging and sterilization methods; therefore, biocompatibility data, bench performance data, sterilization and stability data from the predicate device (**K210502**) are directly applicable.

Performance Data – Clinical

To support the labeling modifications for the subject Trevo NXT ProVue Retriever 3 mm and 4 mm models, a retrospective analysis of device performance per vessel size was performed using The ASSIST Registry data. This retrospective subgroup analysis demonstrates that use of the subject Trevo NXT ProVue Retriever 3 mm and 4 mm models in vessels with diameters of 1.5 to 2.5 mm has a similar performance and safety profile using available procedural data compared to their use in vessels with diameters ≥ 2.5 mm.

Shelf-Life Testing

Shelf-life testing previously conducted for the Trevo NXT ProVue Retriever was leveraged to support the labeling changes to the device. Shelf-life testing was not performed on the subject device

since there was no impact to device material, design, or safety and efficacy as a result of the labeling changes.

Sterilization

Sterilization evaluation previously conducted for the Trevo NXT ProVue Retriever was leveraged to support the labeling changes to the device. The Trevo NXT ProVue Retrievers are sterilized with 100% Ethylene Oxide (EO) and provided sterile. A sterility assurance level (SAL) of 10^{-6} has been demonstrated. The Trevo NXT ProVue Retrievers meet EO residuals per EN ISO 10993-7 for limited contacting, externally communicated devices. The Trevo NXT ProVue Retrievers are for single use only.

Biocompatibility

Biocompatibility testing previously conducted for the Trevo NXT ProVue Retriever was leveraged to support the labeling changes to the device. The results of biocompatibility testing, and biological safety evaluation of the Trevo NXT ProVue Retrievers demonstrate that the devices meet biological safety requirements per ISO 10993-1 for externally communicating medical devices with circulating blood contact for less than 24 hours. The Trevo NXT ProVue Retrievers are considered to have no residual risk of biological hazards. Also, the devices and its packaging do not contain detectable latex. Therefore, the Trevo NXT ProVue Retriever devices, accessories, and primary packaging are considered biocompatible for their intended use.

Conclusion

Based on the conclusions drawn from the risk assessment and the analysis of clinical data from The ASSIST Registry, the subject device demonstrates substantial equivalence to the legally marketed predicate device (**K210502**).