



April 5, 2024

MMI North America, Inc.
Mauro Ercolani
Vice President, Regulatory Affairs and Quality Assurance
344 Ponte Vedra Blvd.
Ponte Vedra, Florida 32082

Re: DEN230032

Trade/Device Name: Symani Surgical System
Regulation Number: 21 CFR 878.4963
Regulation Name: Electromechanical system for open microsurgery
Regulatory Class: Class II
Product Code: SAQ
Dated: April 19, 2023
Received: April 25, 2023

Dear Mauro Ercolani:

The Center for Devices and Radiological Health (FDA) of the Food and Drug Administration (FDA) has completed its review of your FDA request for classification of the Symani Surgical System, a prescription device under 21 CFR Part 801.109 with the following indications for use:

The Symani® Surgical System is intended for soft tissue manipulation to perform anastomosis, suturing, and ligation microsurgery techniques on small blood vessels and lymphatic ducts between 0.1 and 2.5 mm in open free-flap surgery of the breast and extremities and open lymphatic surgery of the extremities.

The Symani® Surgical System is indicated for use during microsurgical procedures when use of a motion scaling function is deemed appropriate by the surgeon. The System is indicated for use in adults. It is intended to be used by trained physicians in an appropriate operating environment in accordance with the Instructions for Use

FDA concludes that this device should be classified into Class II. This order, therefore, classifies the Symani Surgical System, and substantially equivalent devices of this generic type, into Class II under the generic name electromechanical system for open microsurgery.

FDA identifies this generic type of device as:

Electromechanical system for open microsurgery. An electromechanical system for open microsurgery is a software-controlled electromechanical system with bedside human/device interfaces and without an integrated visualization system which allows a qualified user to perform surgical techniques during open microsurgical procedures using surgical instruments attached to an electromechanical arm.

Section 513(f)(2) of the Food, Drug and Cosmetic Act (the FD&C Act) was amended by section 607 of the Food and Drug Administration Safety and Innovation Act (FDASIA) on July 9, 2012. This law provides two options for FDA classification. First, any person who receives a “not substantially equivalent” (NSE) determination in response to a 510(k) for a device that has not been previously classified under the Act may request FDA to make a risk-based classification of the device under section 513(a)(1) of the Act. On December 13, 2016, the 21st Century Cures Act removed a requirement that a FDA request be submitted within 30 days of receiving an NSE determination. Alternatively, any person who determines that there is no legally marketed device upon which to base a determination of substantial equivalence may request FDA to make a risk-based classification of the device under section 513(a)(1) of the Act without first submitting a 510(k). FDA shall, within 120 days of receiving such a request, classify the device. This classification shall be the initial classification of the device. Within 30 days after the issuance of an order classifying the device, FDA must publish a notice in the Federal Register announcing the classification.

On April 25, 2023, FDA received your FDA requesting classification of the Symani Surgical System. The request was submitted under section 513(f)(2) of the FD&C Act. In order to classify the Symani Surgical System into class I or II, it is necessary that the proposed class have sufficient regulatory controls to provide reasonable assurance of the safety and effectiveness of the device for its intended use. After review of the information submitted in the FDA request, FDA has determined that, for the previously stated indications for use, the Symani Surgical System can be classified in class II with the establishment of special controls for class II. FDA believes that class II (special) controls provide reasonable assurance of the safety and effectiveness of the device type. The identified risks and mitigation measures associated with the device type are summarized in the following table:

Risks to Health	Mitigation Measures
Electrical fault, electromagnetic Disturbances, mechanical fault, or system malfunction resulting in tissue injury or prolonged procedure time	<i>In vivo</i> performance testing Electrical safety testing Electromagnetic compatibility testing Non-clinical performance testing Software verification, validation, and hazard analysis Labeling Annual reporting
Adverse tissue reaction	Biocompatibility evaluation Pyrogenicity testing
Infection	Sterilization validation Reprocessing validation Biocompatibility evaluation Pyrogenicity testing Shelf-life validation

	<i>In vivo</i> performance testing Labeling
Use error leading to patient harm or prolonged procedure time: <ul style="list-style-type: none"> • Tissue injury • Reoperation • Thrombosis • Flap failure • Necrosis • Hematoma 	<i>In vivo</i> performance testing Postmarket surveillance Training Human factors testing Labeling Annual reporting

In combination with the general controls of the FD&C Act, the electromechanical system for open microsurgery is subject to the following special controls:

- (1) Data obtained from premarket *in vivo* performance validation testing and postmarket data acquired under anticipated conditions of use must demonstrate that the device performs as intended in the intended patient population and anatomical location, unless FDA determines based on the totality of the information provided for premarket review that data from postmarket surveillance is not required. Objective performance measures (e.g., patency and rate of device related adverse events and their severity, cause, and outcomes) for the device and a clinically justified comparator must be reported with relevant descriptive or developmental performance measures.
- (2) The device manufacturer must develop, and update as necessary, a device-specific use training program that ensures proper device setup/use/shutdown, accurate control of instruments to perform the intended surgical techniques, troubleshooting and handling during unexpected events or emergencies, and safe practices to mitigate use error.
- (3) The device manufacturer may only distribute the device to facilities that implement and maintain the device-specific use training program and ensure that users of the device have completed the device-specific use training program.
- (4) Labeling must include:
 - (i) A detailed summary of *in vivo* performance testing conducted with the device, including study population, results, adverse events, and comparisons to any comparator groups identified;
 - (ii) A statement in the labeling that the safety and effectiveness of the device has not been evaluated for outcomes related to the treatment or prevention of cancer, including but not limited to risk reduction, overall survival, disease-free survival and local recurrence, unless FDA determines that it can be removed or modified based on clinical performance data submitted to FDA;
 - (iii) Identification of compatible devices;
 - (iv) Reprocessing instructions for reusable components;
 - (v) Use life for reusable components;
 - (vi) Shelf life for any sterile components;
 - (vii) A description of the device-specific use training program;
 - (viii) A statement that the device is only for distribution to facilities that implement and maintain

the device-specific use training program and ensure that users of the device have completed the device-specific use training program; and

- (ix) A detailed summary of the postmarket surveillance data collected under paragraph (1) of this section and any necessary modifications to the labeling to accurately reflect outcomes based upon the postmarket data collected under paragraph (1) of this section.
- (5) Human factors validation testing must be performed and must demonstrate that the device/user interfaces of the system support safe use in all use environments.
 - (6) Non-clinical performance testing must demonstrate that the device performs as intended under anticipated conditions of use and must include:
 - (i) Device motion accuracy and repeatability;
 - (ii) System testing;
 - (iii) Instrument reliability;
 - (iv) User-device interface performance;
 - (v) Workspace access testing; and
 - (vi) Performance testing with compatible devices.
 - (7) Software verification, validation, and hazard analysis must be performed.
 - (8) Electromagnetic compatibility and electrical, thermal, and mechanical safety testing must be performed.
 - (9) Performance data must demonstrate the sterility of all patient-contacting device components.
 - (10) Performance data must support the shelf life of the device components provided sterile by demonstrating continued sterility and package integrity over the labeled shelf life.
 - (11) Performance data must validate the reprocessing instructions for the reusable components of the device.
 - (12) Performance data must demonstrate that all patient-contacting components of the device are biocompatible.
 - (13) Performance data must demonstrate that all patient-contacting components of the device are non-pyrogenic.
 - (14) The device manufacturer must submit a report to the FDA annually on the anniversary of initial marketing authorization for the device, until such time as FDA may terminate such reporting, which comprises the following information:
 - (i) Cumulative summary, by year, of complaints and adverse events since date of initial marketing authorization; and
 - (ii) Identification and rationale for changes made to the device, labeling, or device-specific use training program, which did not require submission of a premarket notification during the reporting period.

In order to satisfy special control (1) above, FDA has determined that you must conduct two postmarket surveillance studies as outlined below.

Lymphatics

First, FDA has determined that you must collect and report postmarket surveillance data acquired under anticipated conditions of use to demonstrate long term effectiveness of lymphovenous anastomoses in a population representative of the general U.S. population. Specifically, you must conduct postmarket clinical validation performance testing of the Symani Surgical System in patients from demographic groups representative of the U.S. population, to include populations who had limited representation in the premarket study (e.g., African American, Asian, etc.) and have comorbidities similar to the general U.S. population. The comparator (i.e., cohort where all anastomoses are performed manually) must also be representative of the general U.S. population. This study is needed to understand longer term effectiveness associated with creating a lymphovenous anastomosis using your device during lymphatic surgery. This should include evaluating patency with lymphoscintigraphy at 30 days in a robotic cohort and in a cohort where all anastomoses are performed manually. Additionally, it must include novice and experienced microsurgeons evaluating proficiency as measured in part by ischemia time, intraoperative patency at first attempt, the number of robotic and manual sutures per anastomosis and anastomosis suturing time for each user group. You must also evaluate how these factors change with increased surgical experience using the Symani System. This information is needed to understand the minimum number of cases to demonstrate proficiency as prolonged operation places the patient at higher risks to anesthetic treatment and medication.

FDA expects that the postmarket clinical validation performance testing will include a clinically justified study sample size to confirm that performance of the device in postmarket use is not inferior to the performance observed in the pre-market study for the studied subgroup(s) or the overall population.

Free-Flap

Second, FDA has also determined that you must also collect and report postmarket surveillance data acquired under anticipated conditions of use to demonstrate safety in free-flap procedures of the breast and extremities in a population that is representative of the general U.S. population. Specifically, you must conduct postmarket clinical validation performance testing of the Symani Surgical System in patients from demographic groups representative of the U.S. population, to include populations who had limited representation in the premarket study (e.g., African American, Asian, etc.) and have comorbidities similar to the general U.S. population. The comparator (i.e., cohort where all anastomoses are performed manually) must also be representative of the general U.S. population. This should include determining the anastomosis specific re-operation rate and evaluating adverse events resulting in revisions of the anastomosis in a robotic cohort and in a cohort where all anastomoses are performed manually. Additionally, each anatomical location (i.e., breast and extremities) must include novice and experienced microsurgeons evaluating proficiency as measured in part by ischemia time, intraoperative patency at first attempt, the number of robotic and manual sutures per anastomosis and anastomosis suturing time for each user group. You must also evaluate how these factors change with increased surgical experience using the Symani System. This information is needed to understand the minimum number of cases to demonstrate proficiency as prolonged operation places the patient at higher risks to anesthetic treatment and medication.

FDA expects that the postmarket clinical validation performance testing will include a clinically justified study sample size to confirm that performance of the device in postmarket use is not inferior to the performance observed in the pre-market study for the studied subgroup(s) or the overall population.

Within 30 days of receipt of this order, you must submit complete and separate study protocols for the free-flap and the lymphatic studies as described above. FDA expects to work with you to approve your study protocols within 60 days of this order. Your submission should be clearly labeled as “De Novo Postmarket Study Protocols” and submitted to the Agency as specified below. Please reference the De Novo number above to facilitate processing. If there are multiple protocols being finalized after granting of this De Novo request, please submit each protocol as a separate submission, identified by their unique study name(s).

From the date of study protocol approval, you must meet the following timelines:

- First subject enrolled within 12 months
- 20% of subjects enrolled within 24 months
- 50% of subjects enrolled within 36 months
- 100% of subjects enrolled within 48 months

In addition, you must submit separate periodic reports on the progress of the study as follows:

- Postmarket surveillance progress reports every six (6) months until subject enrollment has been completed, and annually thereafter, from the date of the protocol approval letter, unless otherwise specified by FDA.
- If any enrollment milestones are not met, you must begin submitting enrollment status reports every 3 months in addition to your annual postmarket study progress reports, until FDA notifies you otherwise.
- Submit the final postmarket study report three (3) months from study completion (i.e., last subject’s last follow-up date).

Each postmarket study report should be submitted to the Agency as specified below, identified as a “De Novo Postmarket Study Report” in accordance with how the study is identified above, and bearing the applicable De Novo reference number. Be advised that failure to comply with any special control requirement, including the initiation, enrollment, completion, and reporting per the postmarket surveillance data requirements outlined above, may result in the adulteration and misbranding of your device.

Although this letter refers to your product as a device, please be aware that some granted products may instead be combination products. If you have questions on whether your product is a combination product, contact CDRHProductJurisdiction@fda.hhs.gov.

In addition, this is a prescription device and must comply with 21 CFR 801.109.

Section 510(m) of the FD&C Act provides that FDA may exempt a class II device from the premarket notification requirements under section 510(k) of the FD&C Act, if FDA determines that premarket notification is not necessary to provide reasonable assurance of the safety and effectiveness of the device type. FDA has determined premarket notification is necessary to provide reasonable assurance of the safety and effectiveness of the device type and, therefore, the device is not exempt from the premarket notification requirements of the FD&C Act. Thus, persons who intend to market this device type must submit a

premarket notification containing information on the electromechanical system for open microsurgery they intend to market prior to marketing the device.

Please be advised that FDA's decision to grant this De Novo request does not mean that FDA has made a determination that your device complies with other requirements of the FD&C Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the FD&C Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and if applicable, the electronic product radiation control provisions (Sections 531-542 of the FD&C Act; 21 CFR 1000-1050).

A notice announcing this classification order will be published in the Federal Register. A copy of this order and supporting documentation are on file in the Dockets Management Branch (HFA-305), Food and Drug Administration, 5630 Fishers Lane, Room 1061, Rockville, MD 20852 and are available for inspection between 9 a.m. and 4 p.m., Monday through Friday.

As a result of this order, you may immediately market your device as described in the De Novo request, subject to the general control provisions of the FD&C Act and the special controls identified in this order.

For comprehensive regulatory information about medical devices and radiation-emitting products, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

All required documents should be submitted, unless otherwise specified, to the address below and should reference the above De Novo number to facilitate processing.

De Novo Postmarket Surveillance
U.S. Food and Drug Administration
Center for Devices and Radiological Health
Document Control Center - WO66-G609
10903 New Hampshire Avenue
Silver Spring, MD 20993-0002

Alternatively, documents can be submitted electronically through the CDRH Portal. For more information on the CDRH Portal, please visit <https://www.fda.gov/medical-devices/industry-medical-devices/send-and-track-medical-device-premarket-submissions-online-cdrh-portal>.

If you have any questions concerning the contents of the letter, please contact Gabrielle Clark-Patterson at (240) 402-2817.

Sincerely,

Binita S. Ashar, M.D., M.B.A., F.A.C.S.
Director
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health