



April 25, 2024

LabCorp Drug Development  
Mason W. Robbins  
Associate Director, Regulatory Affairs, CDx  
100 Perimeter Park Drive, Ste C  
Morrisville, NC 27560

Re: H230005  
HUD Number: 20-0442  
Trade/Device Name: nAbCyte Anti-AAVRh74var HB-FE Assay  
Product Code: QWQ  
Filed: July 26, 2023  
Amended: August 25, 2023; December 6, 2023; March 29, 2024

Dear Mason Robbins:

The Center for Devices and Radiological Health (CDRH) of the Food and Drug Administration (FDA) has completed its review of your humanitarian device exemption (HDE) application for the nAbCyte Anti-AAVRh74var HB-FE Assay. This device is indicated for:

The nAbCyte Anti-AAVRh74var HB-FE Assay for BEQVEZ™ (fidanacogene elaparvovec) Eligibility in moderate to severe Hemophilia B, or nAbCyte Anti-AAVRh74var HB-FE Assay, is a cell-based qualitative in vitro companion diagnostic device intended for the detection of neutralizing antibodies to the AAVRh74var capsid in serum specimens from adult patients previously diagnosed with moderate to severe hemophilia B to determine eligibility for treatment with the hemophilia B gene therapy, BEQVEZ (fidanacogene elaparvovec). Patients who test positive for neutralizing antibodies to the AAVRh74var capsid are not eligible for treatment with BEQVEZ (fidanacogene elaparvovec). Patients who test negative for neutralizing antibodies to the AAVRh74var capsid are eligible for treatment with BEQVEZ (fidanacogene elaparvovec).

This diagnostic device is a single-site assay for professional use only and is to be performed only at Labcorp-Monogram Biosciences by trained and qualified laboratory personnel.

Based upon the information submitted, the HDE is approved. You may begin commercial distribution of the device in accordance with the conditions of approval described below. Although this letter refers to your product as a device, please be aware that some approved products may instead be combination products. The Premarket Approval Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfPMA/pma.cfm> identifies combination product submissions.

The sale and distribution of this device are restricted to prescription use in accordance with 21 CFR 801.109 and under section 515(d)(1)(B)(ii) of the Federal Food, Drug, and Cosmetic Act (the act). FDA has

determined that these restrictions on sale and distribution are necessary to provide reasonable assurance of the safety and probable benefit of the device. Your device is therefore a restricted device subject to the requirements in sections 502(q) and (r) of the act, in addition to the many other FDA requirements governing the manufacture, distribution, and marketing of devices.

Continued approval of the HDE is contingent upon the submission of periodic reports, required under 21 CFR 814.126, at intervals of one year (unless otherwise specified) from the date of approval of the original HDE. This report, identified as "Annual Report" and bearing the applicable HDE reference number, should be submitted to the address below. The Annual Report should indicate the beginning and ending date of the period covered by the report and should include the information required by 21 CFR 814.126.

In accordance with 21 CFR 814.124, an HDE holder is responsible for ensuring that a humanitarian use device (HUD) under an approved HDE is administered only in facilities having institutional review board (IRB) oversight. In addition, approval by an IRB or an appropriate local committee is required before the HUD can be used at a facility, with the exception of emergency use. An HDE holder is also required to maintain records of the names and addresses of the facilities to which the HUD has been shipped, correspondence with reviewing IRBs as well as any other information requested by a reviewing IRB or FDA (21 CFR 814.126(b)(2)).

This is a reminder that as of September 24, 2014, class III devices are subject to certain provisions of the final Unique Device Identification (UDI) rule. These provisions include the requirement to provide a UDI on the device label and packages (21 CFR 801.20), format dates on the device label in accordance with 21 CFR 801.18, and submit data to the Global Unique Device Identification Database (GUDID) (21 CFR 830 Subpart E). For more information on these requirements, please see the UDI website, <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-udi-system>.

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production and process controls (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Before making any change affecting the safety or probable benefit of the HDE device, you must submit an HDE supplement or an alternate submission (30-day notice) in accordance with 21 CFR 814.108 and 814.39 except a request for a new indication for use of a humanitarian use device (HUD). A request for a new indication for use for an HUD shall comply with the requirements set forth in 21 CFR 814.110 which includes obtaining a new designation of HUD status for the new indication for use and submission of an original HDE application in accordance with §814.104. The application for the new indication for use may incorporate by reference any information or data previously submitted to the agency.

You are reminded that many FDA requirements govern the manufacture, distribution, and marketing of devices. For example, in accordance with the Medical Device Reporting (MDR) regulation, 21 CFR 803.50 and 21 CFR 803.52 for devices or post-marketing safety reporting (21 CFR 4, Subpart B) for combination products, you are required to report adverse events for this device. Manufacturers of medical devices,

including in vitro diagnostic devices, are required to report to FDA no later than 30 calendar days after the day they receive or otherwise becomes aware of information, from any source, that reasonably suggests that one of their marketed devices:

1. May have caused or contributed to a death or serious injury; or
2. Has malfunctioned and such device or similar device marketed by the manufacturer would be likely to cause or contribute to a death or serious injury if the malfunction were to recur.

Additional information on MDR, including how, when, and where to report, is available at <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems> and on combination product post-marketing safety reporting is available at (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>).

In accordance with the recall requirements specified in 21 CFR 806.10 for devices or the post-marketing safety reporting requirements (21 CFR 4, Subpart B) for combination products, you are required to submit a written report to FDA of any correction or removal of this device initiated by you to: (1) reduce a risk to health posed by the device; or (2) remedy a violation of the act caused by the device which may present a risk to health, with certain exceptions specified in 21 CFR 806.10(a)(2). Additional information on recalls is available at <https://www.fda.gov/safety/recalls-market-withdrawals-safety-alerts/industry-guidance-recalls>.

CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading. CDRH will notify the public of its decision to approve your HDE by making available, among other information, a summary of the safety and probably benefit data upon which the approval is based. The information can be found on the FDA CDRH Internet Home Page located at <https://www.fda.gov/medical-devices/device-approvals-denials-and-clearances/pma-approvals>. Written requests for this information can also be made to the Food and Drug Administration, Dockets Management Branch, (HFA-305), 5630 Fishers Lane, Rm. 1061, Rockville, MD 20852. The written request should include the HDE number or docket number. Within 30 days from the date that this information is placed on the Internet, any interested person may seek review of this decision by submitting a petition for review under section 515(g) of the act and requesting either a hearing or review by an independent advisory committee. FDA may, for good cause, extend this 30-day filing period.

Failure to comply with any post-approval requirement constitutes a ground for withdrawal of approval of a HDE. The introduction or delivery for introduction into interstate commerce of a device that is not in compliance with its conditions of approval is a violation of law.

You are reminded that, as soon as possible and before commercial distribution of your device, you must submit an amendment to this HDE submission with a copy of all final labeling. Final labeling that is identical to the labeling approved in draft form will not routinely be reviewed by FDA staff when accompanied by a cover letter stating that the final labeling is identical to the labeling approved in draft form. If the final labeling is not identical, any changes from the final draft labeling should be highlighted and explained in the amendment.

All required documents should be submitted, unless otherwise specified, to the address below and should reference the above HDE number to facilitate processing.

U.S. Food and Drug Administration  
Center for Devices and Radiological Health  
Document Control Center - WO66-G609  
10903 New Hampshire Avenue  
Silver Spring, MD 20993-0002

If you have any questions concerning this approval order, please contact Yan Cai at 240-402-1094 or [Yan.Cai@fda.hhs.gov](mailto:Yan.Cai@fda.hhs.gov).

Sincerely,

Takeesha Taylor-bell -S

*for*

Lea Carrington  
Director  
Division of Immunology and Hematology Devices  
OHT7: Office of In Vitro Diagnostics  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health