



April 19, 2023

BrosMed Medical Co., Ltd.  
Crystal Lee  
Registration Affairs Manager  
15th Building, SMEs Venture Park, SongShan Lake  
Hi-Tech Industrial Development Zone  
Dongguan, GD 523808  
China

Re: K230705

Trade/Device Name: POT PTCA Balloon Dilatation Catheter  
Regulation Number: 21 CFR 870.5100  
Regulation Name: Percutaneous Transluminal Coronary Angioplasty (PTCA) Catheter  
Regulatory Class: Class II  
Product Code: LOX

Dear Crystal Lee:

The Food and Drug Administration (FDA) is sending this letter to notify you of an administrative change related to your previous substantial equivalence (SE) determination letter dated April 11, 2023. Specifically, FDA is updating this SE Letter (e.g., inaccurate device description in 510(k) summary) as an administrative correction.

Please note that the 510(k) submission was not re-reviewed. For questions regarding this letter please contact Gregory O'Connell, OHT2: Office of Cardiovascular Devices, 301-796-6075, [gregory.oconnell@fda.hhs.gov](mailto:gregory.oconnell@fda.hhs.gov).

Sincerely,

**Ariel G. Ash-**  
**shakoor -S**

Digitally signed by  
Ariel G. Ash-shakoor -S  
Date: 2023.04.19  
10:12:06 -04'00'

For

Gregory O'Connell  
Assistant Director  
DHT2C: Division of Coronary  
and Peripheral Intervention Devices  
OHT2: Office of Cardiovascular Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health



April 11, 2023

BrosMed Medical Co., Ltd.  
Crystal Lee  
Registration Affairs Manager  
15th Building, SMEs Venture Park, SongShan Lake  
Hi-Tech Industrial Development Zone  
Dongguan, GD 523808  
China

Re: K230705

Trade/Device Name: POT PTCA Balloon Dilatation Catheter  
Regulation Number: 21 CFR 870.5100  
Regulation Name: Percutaneous Transluminal Coronary Angioplasty (Ptca) Catheter  
Regulatory Class: Class II  
Product Code: LOX  
Dated: March 14, 2023  
Received: March 14, 2023

Dear Crystal Lee:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements,

including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Ariel G. Ash-  
shakoor -S

Digitally signed by Ariel G.  
Ash-shakoor -S  
Date: 2023.04.11 10:18:58  
-04'00'

For

Gregory O'Connell  
Assistant Director  
DHT2C: Division of Coronary  
and Peripheral Intervention Devices  
OHT2: Office of Cardiovascular Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K230705

Device Name  
POT PTCA Balloon Dilatation Catheter

### Indications for Use (Describe)

The POT PTCA balloon dilatation catheter is indicated for:

- balloon dilatation of the stenotic portion of a coronary artery or bypass graft stenosis for the purpose of improving myocardial perfusion
- balloon dilatation of a coronary artery occlusion for the treatment of acute myocardial infarction
- balloon dilatation of a stent after implantation (balloon models Ø2.25 mm - Ø5.00 mm only)

Note: Bench testing was conducted with the POT PTCA Balloon Dilatation Catheter and marketed balloon expandable stents. Consideration should be taken when this device is used with different manufacturers' stents due to difference in stent design.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) SUMMARY

### 1. GENERAL INFORMATION

#### 1.1 Submitter

BrosMed Medical Co., Ltd.  
15th Building, SMEs Venture Park  
SongShan Lake Hi-Tech Industrial Development Zone  
Dongguan, Guangdong, 523808, China  
Office: +86 (769) 2289 2018  
Fax: +86 (769) 2289 2016

#### 1.2 Contract person

Crystal Lee,  
Email: [crystallee@brosmed.com](mailto:crystallee@brosmed.com)  
Office: +86 (769) 2289 2018

#### 1.3 Date of Preparation

March 03, 2023

### 2. NAME OF THE DEVICE

#### 2.1.1 Trade/Proprietary Name

POT PTCA Balloon Dilatation Catheter

#### 2.1.2 Common/Usual Name

Percutaneous Transluminal Coronary Angioplasty (PTCA) Catheter

#### 2.1.3 Classification Information

|                            |  |
|----------------------------|--|
| Classification Name:       | Catheters, Transluminal Coronary Angioplasty, Percutaneous |
| Classification Regulation: | 21 CFR 870.5100  |
| Device Class:              | Class II (Special Controls)                                |
| Product Code:              | LOX  |
| Review Panel:              | Cardiovascular   |

### 3. PREDICATE DEVICE

- Apollo Balloon Dilatation Catheter (K133852 and K153742).

### 4. DESCRIPTION OF THE DEVICE

The POT PTCA balloon dilatation catheter is a sterile, flexible tube designed to be used in percutaneous transluminal coronary angioplasty (PTCA) to dilate a stenotic coronary artery by controlled inflation of a distensible Nylon balloon at its distal tip. It is typically a rapid exchange (Rx) type with a single-lumen catheter. It is a single-use device and available in various sizes.

**510(k) Summary**

**5. INDICATION FOR USE**

The indication for use / intended use statement for the Subject device is as follows:

The POT PTCA Balloon Dilatation Catheter is indicated for:

- balloon dilatation of the stenotic portion of a coronary artery or bypass graft stenosis for the purpose of improving myocardial perfusion;
- balloon dilatation of a coronary artery occlusion for the treatment of acute myocardial infarction;
- balloon dilatation of a stent after implantation (balloon models Ø2.25 mm - 5.00 mm only)

Note: Bench testing was conducted with the POT PTCA Balloon Dilatation Catheter and marketed balloon expandable stents. Consideration should be taken when this device is used with different manufacturers' stents due to difference in stent design.

**6. SUBSTANTIAL EQUIVALENCE COMPARISON**

Comparison of the POT PTCA Balloon Dilatation Catheter and Apollo Balloon Dilatation Catheter cleared in K133852 and K153742 shows that the subject device incorporates substantial equivalence general design components, material and performance specifications, manufacturing processes, sterilization process, packaging materials and design, and the same indications for use and principles of operation as predicates.

Please see **Table 1** below for the comparison between Subject device and Predicate devices.

**Table 1. Substantial Equivalence Comparison of Subject Device and Predicate Device**

| <b>Technological Characteristic</b> | <b>Subject Device POT</b>   | <b>Predicate Device Apollo (K133852 &amp; K153742)</b>   | <b>Comparison</b> |
|-------------------------------------|---|--|-------------------|
| Indications for Use                 | <p>The POT PTCA balloon dilatation catheter is indicated for:</p> <ul style="list-style-type: none"> <li>·balloon dilatation of the stenotic portion of a coronary artery or bypass graft stenosis for the purpose of improving myocardial perfusion</li> <li>·balloon dilatation of a coronary artery occlusion for the treatment of acute myocardial infarction</li> <li>·balloon dilatation of a stent after implantation (balloon models Ø2.25 mm - Ø5.00 mm only)</li> </ul> <p>Note: Bench testing was conducted with the POT PTCA Balloon Dilatation Catheter and marketed balloon expandable stents. Consideration should be taken when this device is used with different manufacturers' stents due to difference in stent design.</p> | <p>The Apollo balloon dilatation catheter is indicated for:</p> <ul style="list-style-type: none"> <li>·balloon dilatation of the stenotic portion of a coronary artery or bypass graft stenosis for the purpose of improving myocardial perfusion</li> <li>·balloon dilatation of a coronary artery occlusion for the treatment of acute myocardial infarction</li> <li>·balloon dilatation of a stent after implantation (balloon models Ø2.0 mm - Ø5.00 mm only)</li> </ul> <p>Note: Bench testing was conducted with the Apollo Balloon Dilatation Catheter and marketed balloon expandable stents. Consideration should be taken when this device is used with different manufacturers' stents due to difference in stent design.</p> | Identical         |
| Fundamental device design           | Rx type sterilized PTCA catheter  | Rx type sterilized PTCA catheter   | Identical         |

**510(k) Summary**

| <b>Technological Characteristic</b> | <b>Subject Device POT</b>                            | <b>Predicate Device Apollo (K133852 &amp; K153742)</b> | <b>Comparison</b> |
|-------------------------------------|--|--|-------------------|
| General catheter design             | Tip, balloon, body tubing, hub, 2 radiopaque markers | Tip, balloon, body tubing, hub, 2 radiopaque markers   | Identical         |
| Compatible guidewire (in)           | 0.014  | 0.014  | Identical         |
| Catheter working length (cm)        | 140  | 140  | Identical         |
| Balloon diameter range (mm)         | 2.25-5.0   | 2.0-5.0  | Equivalence       |
| Balloon length range (mm)           | 6-15   | 6-30   | Equivalence       |
| Balloon cone length                 | Shorter  | Longer   | Equivalence       |
| Nominal Pressure (atm)              | 14   | 14   | Identical         |
| Rated Burst Pressure (atm)          | 20-22  | 20-22  | Identical         |

**7. PERFORMANCE TESTING SUMMARY**

Additional bench testing were performed to support a determination of substantial equivalence between subject device and predicates in accordance with the FDA guidance entitled “Deciding When to Submit a 510(k) for a Change to an Existing Device” and “The Special 510(k) Program - Guidance for Industry and FDA Staff”. The results of all tests provide reasonable assurance that the proposed POT PTCA catheter has been designed and tested to assure conformance to the requirements of its intended use.

All additional testing were conducted on the subject device as recommended in FDA PTCA guidance:

- Dimensional verification
- Simulated Use
- Balloon Rated Burst Pressure
- Balloon Fatigue
- Balloon Compliance
- Balloon Inflation and Deflation Time
- Balloon Burst (in stent)
- Balloon Fatigue (in stent)

In vitro bench testing met all acceptance criteria and demonstrated that the subject device performed as intended and did not impact the functionality of the device.

**8. CONCLUSION**

The information presented in this special 510(k) submission demonstrates that the proposed modifications on the subject device do not raise new/different questions of safety and effectiveness as compared to the predicate device. Therefore, the POT PTCA Balloon Dilatation Catheter is substantially equivalent to the Apollo Balloon Dilatation Catheter cleared in K133852 and K153742.