

August 30, 2023

Shenzhen SINO-K Medical Technology Co.,Ltd % Yang Jie Consultant Shenzhen Choncorut Medical Consulting Co., Ltd. Room 504, Block C, No. 1029 Nanhai Avenue, Nanshan District Shenzhen, Guangdong 518067 China

Re: K231961

Trade/Device Name: Reusable NIBP Cuff, Disposable NIBP Cuff Regulation Number: 21 CFR 870.1120 Regulator Name: Blood pressure cuff Regulatory Class: Class II Product Code: DXQ Dated: June 30, 2023 Received: July 3, 2023

Dear Yang Jie:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <a href="https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm">https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm</a> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <a href="https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products">https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products</a>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <u>https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems</u>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<u>https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance</u>) and CDRH Learn (<u>https://www.fda.gov/training-and-continuing-education/cdrh-learn</u>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<u>https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice</u>) for more information or contact DICE by email (<u>DICE@fda.hhs.gov</u>) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

# Stephen C. Browning -S

LCDR Stephen Browning Asistant Director Division of Cardiac Electrophysiology Diagnostic and Monitoring Devices Office of Cardiovascular Devices Office of Product Evaluation and Quality Center for Devices and Radiological Health

Enclosure

## **Indications for Use**

510(k) Number (*if known*) K231961

Device Name

Reusable NIBP Cuff, Disposable NIBP Cuff

The reusable blood pressure cuff is an accessory used in conjunction with noninvasive blood pressure measurement systems. The cuff is non-sterile and may be reused. It is available in neonate, infant, child and adult sizes. The cuff is not designed, sold, or intended for use except as indicated.

The disposable blood pressure cuff is an accessory used in conjunction with noninvasive blood pressure measurement systems. The cuff is non-sterile and for single-patient use. It is available in neonate, infant, child and adult sizes. The cuff is not designed, sold, or intended for use except as indicated.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

## CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

#### \*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\*

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services Food and Drug Administration Office of Chief Information Officer Paperwork Reduction Act (PRA) Staff *PRAStaff@fda.hhs.gov* 

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

# 510(K) Summary

#### Prepared in accordance with the requirements of 21 CFR Part 807.92

#### **Prepared Date: 2023/06/30**

#### 1. Submission sponsor

Name: Shenzhen SINO-K Medical Technology Co., Ltd.

Room401, Bldg2, Veteran Ind. City, Gongle Community, Xixiang Street, Baoan District, Shenzhen, Guangdong, China Contact person: Lao Chengxin Title: General Manager

E-mail: boss@sino-k.com

Tel: +86 137 15333326

#### 2. Submission correspondent

Name: Shenzhen Chonconn Medical Device Consulting Co., Ltd.

Address: Room 504, Block C, No. 1029 Nanhai Avenue, Nanshan District, Shenzhen, Guangdong, P.R. China

Contact person: Yang Jie E-mail: yangjie@chonconn.com

Tel: +86-755 33941160

0		
Trade/Device Name	Reusable NIBP Cuff, Disposable NIBP Cuff	
	Reusable NIBP Cuff: SC1311, SC1511, SC1611, SC1711, SC1811.	
Model	Disposable NIBP Cuff: SC0101, SC0102, SC0103, SC0104, SC0105, SC0106,	
	SC0107, SC0108, SC0109, SC0110, SC0111, SC0112.	
Common Name	Non-invasive Blood Pressure Cuff	
Regulatory Class	Class II	
Classification	21CFR 870.1120 / Blood pressure cuff / DXQ	
Submission type	Traditional 510(K)	

#### 3. Subject Device Information

#### 4. Predicate Device

Shenzhen Caremed Medical Technology Co., Ltd. Caremed Reusable Blood Pressure Cuff, Caremed Disposable Blood Pressure Cuff under K182433.

## 5. Device Description

The proposed device is an accessory used in conjunction with noninvasive blood pressure measurement systems. It is available in neonatal, infant, child and adult sizes.

The proposed device includes disposable blood pressure cuff and reusable blood pressure cuff. Both disposable and reusable blood pressure cuff have the same structure, which contains Cuff with bladder and single-tube Air Hose.

The subject device is categorized into two types of models according to its reusability. Refer to the master list of models below.

Reusable NIBP Cuff: SC1311, SC1511, SC1611, SC1711, SC1811.

Disposable NIBP Cuff: SC0101, SC0102, SC0103, SC0104, SC0105, SC0106, SC0107, SC0108, SC0109, SC0110, SC0111, SC0112.

#### 6. Intended use & Indication for use

The reusable blood pressure cuff is an accessory used in conjunction with noninvasive blood pressure measurement systems. The cuff is non-sterile and may be reused. It is available in neonate, infant, child and adult sizes. The cuff is not designed, sold, or intended for use except as indicated.

The disposable blood pressure cuff is an accessory used in conjunction with noninvasive blood pressure measurement systems. The cuff is non-sterile and for single-patient use. It is available in neonate, infant, child and adult sizes. The cuff is not designed, sold, or intended for use except as indicated.

## 7. Comparison to the Predicate Device

The following tabulation indicates the detailed differences between the subject devices and the predicate devices.

Features	Predicate Device	Subject Device	Remark
	<b>Caremed Reusable Blood</b>	Reusable NIBP Cuff	
	Pressure Cuff	K231961	
	K182433		
Applicant	Shenzhen Caremed Medical	Shenzhen SINO-K Medical	/
	Technology Co., Ltd.	Technology Co., Ltd	
Classification	21CRF 870.1120	21CRF 870.1120	Same
Regulation			
Classification	Class II,	Class II,	Same
and Code	DXQ	DXQ	
Common	Non-invasive Blood Pressure	Non-invasive Blood Pressure	Same
name	Cuff	Cuff	
Intended use	The reusable blood pressure cuff	The reusable blood pressure cuff is	Same
	is an accessory used in	an accessory used in conjunction	
	conjunction with noninvasive	with noninvasive blood pressure	
	blood pressure measurement	measurement systems. The cuff is	
	systems. The cuff is non-sterile	non-sterile and may be reused. It is	

Comparison to Predicate Device Caremed Reusable Blood Pressure Cuff:

Features	Predicate Device	Subject Device	Remark
	<b>Caremed Reusable Blood</b>	Reusable NIBP Cuff	
	Pressure Cuff	K231961	
	K182433		
	and may be reused. It is available	available in neonate, infant, child	
	in neonate, infant, child and adult	and adult sizes. The cuff is not	
	sizes. The cuff is not designed,	designed, sold, or intended for use	
	sold, or intended for use except	except as indicated.	
	as indicated.		
Patient	Adults/Pediatrics	Adults/Pediatrics	Same
Populations			
Tube Number	One or two	One	Different
Principles of	Bladder is wrapped around the	Bladder is wrapped around the	Same
Operation	patient's limb and secured by	patient's limb and secured by hook	
	hook and loop closure Air hose is	and loop closure Air hose is	
	connected to the noninvasive	connected to the noninvasive blood	
	blood pressure measurement	pressure measurement systems	
	systems		
Limb	Neonatal (6-11cm)	Neonatal (6-11cm)	Different
Circumference	Infant (10-19cm)	Infant (10-19cm)	
(Range in cm)	Child (18-26cm)	Pediatric (18-26cm)	
	Small Adult (20-28cm)	Adult (25-35cm)	
	Adult (25-35cm)	Large Adult (33-47cm)	
	Adult Long (25-35cm)		
	Large Adult (33-47cm)		
	Adult Thigh (44-66cm)		
Pressure Range	0-300 mmHg	Cuff for neonate and infant:	Different
		180mmHg(24kPa)	
		Cuff for small child, child, small	
		adult, adult, large adult:	
		360mmHg(48kPa)	
Sterility	Non-sterile	Non-sterile	Same
Max. Leakage	< 4mm Hg/ min.	< 4mm Hg/ min.	Same
Material	Cuff (Patient contacted): PU	Cuff (Patient contacted): PU	Same
	Synthetic Leather	Synthetic Leather	
	Bladder: Transparent	Bladder: Transparent Polyurethane	
	Polyurethane (TPU Film)	(TPU Film)	
	Tubing: PVC	Tubing: PVC	
	Hook: Molded Nylon	Hook: Molded Nylon	
	Loop: Nylon	Loop: Nylon	
Biocompatibility	Comply with ISO 10993-5;	Comply with ISO 10993-5;	Same

Features	Predicate Device Caremed Reusable Blood Pressure Cuff K182433	Subject Device Reusable NIBP Cuff K231961	Remark
	Comply with ISO 10993-10	Comply with ISO 10993-10	

# Comparison to Predicate Device Caremed Disposable Blood Pressure Cuff:

Features	Predicate Device	Subject Device	Remark
	<b>Caremed Disposable Blood</b>	Disposable NIBP Cuff	
	Pressure Cuff		
Applicant	Shenzhen Caremed Medical	Shenzhen SINO-K Medical	/
	Technology Co., Ltd.	Technology Co., Ltd	
Intended use	The disposable blood pressure cuff	The disposable blood pressure	Same
	is an accessory used in conjunction	cuff is an accessory used in	
	with noninvasive blood pressure	conjunction with noninvasive	
	measurement systems. The cuff is	blood pressure measurement	
	non-sterile and for single-patient	systems. The cuff is non-sterile	
	use. It is available in neonate, infant,	and for single-patient use. It is	
	child and adult sizes. The cuff is not	available in neonate, infant,	
	designed, sold, or intended for use	child and adult sizes. The cuff is	
	except as indicated.	not designed, sold, or intended	
		for use except as indicated.	
Patient	Adults/Pediatrics	Adults/Pediatrics	Same
Populations			
Tube Number	One or two	One	Different
Limb	Neonatal 1 (3-6 cm)	Neonatal (3-6 cm)	Different
Circumference	Neonatal 2 (4-8 cm)	Neonatal (4-8 cm)	
(Range in cm)	Neonatal 3 (6-11 cm)	Neonatal (6-11 cm)	
	Neonatal 4 (7-13 cm)	Neonatal (7-13 cm)	
	Neonatal 5 (8-15 cm)	Neonatal (8-15 cm)	
	Infant (9-14.8 cm)	Neonatal (7.7-10.5 cm)	
	Child (13.8-21.5 cm)	Infant (9.8-13.3 cm)	
	Small Adult (20.5-28.5 cm)	Small Child (12.4-16.8 cm)	
	Adult (27.5-36.5 cm)	Child (15.8-21.3 cm)	
	Adult Long (27.5-36.5/46.5 cm)	Small Adult (20-27 cm)	
	Large Adult (35.5-46 cm)	Adult (25.3-34.3 cm)	
	Large Adult Long (35.5-46 cm)	Large Adult (32.1-43.4 cm)	
Pressure Range	0-300mmHg	Cuff for neonate and infant:	Different
		180mmHg(24kPa)	
		Cuff for small child, child, small	

Features	Predicate Device	Subject Device	Remark
	<b>Caremed Disposable Blood</b>	Disposable NIBP Cuff	
	Pressure Cuff		
		adult, adult, large adult:	
		360mmHg(48kPa)	
Sterility	Non-sterile	Non-sterile	Same
Material	Non-woven (Patient contacted);	Cuff (Patient contacted): Non-	Different
	nylon;	woven clad PVC	
	PVC	Bladder: PVC	
		Tubing: PVC	
		Hook: PVC	
		Loop: PVC	
Biocompatibility	Comply with ISO 10993-5;	Comply with ISO 10993-5;	Same
	Comply with ISO 10993-10	Comply with ISO 10993-10	

#### 8. Performance Data

The following performance data were provided in support of the substantial equivalence determination.

#### **Biocompatibility testing**

The biocompatibility evaluation for the proposed devices was conducted in accordance with the FDA Biocompatibility guidance, 2016 (Use of International Standard ISO 10993-1, "Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process") and International Standard ISO 10993-1 "Biological Evaluation of Medical Devices – Part 1: Evaluation and Testing Within a Risk Management Process," as recognized by FDA. The battery of testing included the following tests:

- Cytotoxicity
- Sensitization
- Irritation

The NIBP Cuffs are considered surface contacting for a duration of exceed 24 hours but not 30 days.

#### Non-clinical data

Non-clinical tests were conducted to verify that the proposed devices met all design specifications as was Substantially Equivalent (SE) to the predicate device. The test results demonstrated that the proposed device complies with the following standards:

- ISO 81060-1, Non-Invasive Sphygmomanometers Part 1: Requirements and Test Methods for Non-Automated Measurement Type, First Edition 2007.
- ISO 10993-5: 2009 Biological Evaluation of Medical Devices Part 5: Tests for In Vitro Cytotoxicity.
- ISO 10993-10: 2010 Biological Evaluation of Medical Devices Part 10: Tests for Irritation and Skin Sensitization

#### 9. Conclusion

It has been shown in the 510(k) submission that the difference between the proposed devices and the predicate devices do not raise any questions regarding safety and effectiveness. Performance testing and compliance with voluntary standards demonstrate that the proposed are substantially equivalent to the relevant aspects of the predicate devices in terms of design, components, materials, principals of operation, biocompatibility, performance characteristics, and intended use.