



January 26, 2024

Zhejiang Soudon Medical Technology Co.,Ltd
% Nick Wang
RA Specialist
Shanghai Vanhe Consulting Co., Ltd
2F, Building No.1, 3938 Huqingping Road, Qingpu District
Shanghai, 201703
China

Re: K232450

Trade/Device Name: Disposable Hemostatic Clips
Regulation Number: 21 CFR 876.4400
Regulation Name: Hemorrhoidal Ligator
Regulatory Class: Class II
Product Code: PKL
Dated: August 11, 2023
Received: August 14, 2023

Dear Nick Wang:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Sivakami Venkatachalam -S

for

Shanil P. Haugen, Ph.D.

Assistant Director

DHT3A: Division of Renal, Gastrointestinal,
Obesity and Transplant Devices

OHT3: Office of GastroRenal, ObGyn,

General Hospital and Urology Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

Submission Number (if known)

K232450

Device Name

Disposable Hemostatic Clips

Indications for Use (Describe)

Disposable Hemostatic Clips is indicated for endoscopic clip placement within the gastrointestinal tract for the purpose of:

1. Endoscopic marking
2. Hemostasis for:
 - Mucosal/sub-mucosal defects < 3cm
 - Bleeding ulcers
 - Arteries < 2mm
 - Polyps < 1.5cm in diameter
 - Diverticula in the colon
3. Anchoring to affix jejunal feeding tubes to the wall of the small bowel
4. As a supplementary method, closure for GI tract luminal perforations < 20mm that can be treated conservatively.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

This summary of 510(k) safety and effectiveness information is being submitted in accordance with requirements of 21 CFR Part 807.92.

1. Submitter

Submitted by:	Zhejiang Soudon Medical Technology Co.,Ltd Address:Room 302-1, Building 4, No.1 Nangonghe Road, Donghu Street, Linping District, Hangzhou 311100 Zhejiang, China
Contact Person:	Nick Wang RA Specialist Shanghai Vanhe Consulting Co., Ltd Address: 2F, Building No.1, 3938 Huqingping Road, Qingpu District, Shanghai, China. Phone: 0086-13585860297 Email: vanheconsulting@126.com
Date Prepared:	Jan 24, 2024

Device

Device Name:	Disposable Hemostatic Clips
Classification Name:	Hemorrhoidal Ligator
Regulatory Class:	II
Regulation Number:	21 CFR 876.4400
Regulation Name:	Hemorrhoidal Ligator
Product Code:	PKL

Predicate Device

Device Name:	Hemoclip, K172727
Manufacturer:	Hangzhou AGS MedTech CO., Ltd
Classification Name:	Hemorrhoidal Ligator
Regulatory Class:	II
Regulation Number:	21 CFR 876.4400
Regulation Name:	Hemorrhoidal Ligator
Product Code:	PKL

Reference device

Device Name:	Single Use Rotatable and Repositionable Hemoclip, K201771
Manufacturer:	Anrei Medical (Hangzhou) Co., Ltd

Classification Name:	Hemorrhoidal Ligator
Regulatory Class:	II
Regulation Number:	21 CFR 876.4400
Regulation Name:	Hemorrhoidal Ligator
Product Code:	PKL

2. Device Description

The Disposable Hemostatic Clips consist of one pre-loaded clip and delivery system for single patient use only. It is provided in sterile.

The clip is made of stainless steel with good superelasticity performance. The clip is pre-loaded in end of the Spring tube part through its deformation and it is deployed from the delivery system during use. The clip is engineered such that they can be opened and closed up to five times prior to deployment, aiding in repositioning of the clip at the lesion site.

The device is available in four clip opening widths, which are 9mm, 11mm, 13mm and 16mm, and the device effective length is available in 1650mm, 1950mm and 2300mm.

3. Indication for Use:

The Disposable Hemostatic Clips is indicated for endoscopic clip placement within the gastrointestinal tract for the purpose of:

1. Endoscopic marking
2. Hemostasis for:
 - Mucosal/sub-mucosal defects <3cm
 - Bleeding ulcers
 - Arteries<2mm
 - Polyps<1.5cm in diameter
 - Diverticula in the colon
3. Anchoring to affix jejunal feeding tubes to the wall of the small bowel
4. As a supplementary method, closure for GI tract luminal perforations<20mm that can be treated conservatively.

4. Comparison of Technological Characteristics

The Disposable Hemostatic Clips has substantially equivalent device design, configuration, packaging fundamental technology, sterilization process and intended use as those featured in the predicate device Hangzhou AGS MedTech CO., Ltd 's Hemoclip, K172727. The differences between the proposed device and the predicate devices do not raise any questions regarding its safety and effectiveness. The differences are listed in the table below.

Item	Disposable Hemostatic Clips(Proposed Device)	Hemoclip K172727	Discussion
Indication for Use	The Disposable Hemostatic Clips is indicated for	The hemoclip is indicated for endoscopic clip	Same

Item	Disposable Hemostatic Clips(Proposed Device)	Hemoclip K172727	Discussion
	endoscopic clip placement within the gastrointestinal tract for the purpose of: - Endoscopic marking - Hemostasis for: - Mucosal/sub-mucosal defects <3cm - Bleeding ulcers - Arteries<2mm - Polyps < 1.5cm in diameter - Diverticula in the colon - Anchoring to affix jejunal feeding tubes to the wall of the small bowel - As a supplementary method, closure for GI tract luminal perforations<20mm that can be treated conservatively.	placement within the gastrointestinal tract for the purpose of: - Endoscopic marking - Hemostasis for: - Mucosal/sub-mucosal defects <3cm - Bleeding ulcers - Arteries<2mm - Polyps < 1.5cm in diameter - Diverticula in the colon - Anchoring to affix jejunal feeding tubes to the wall of the small bowel - As a supplementary method, closure for GI tract luminal perforations<20mm that can be treated conservatively.	
Principle of Operation	Endoscopic accessory used to deliver metal clips to the GI tract.	Endoscopic accessory used to deliver metal clips to the GI tract.	Same
Minimum Endoscopic Working Channel	2.8mm	2.8mm	Same
Working Length	1650mm, 1950mm and 2300mm	1650mm, 1950mm, 2350mm	Similar
Clip Opening Width	9mm,11mm, 13mm, 16mm	9mm,11mm, 13mm	Similar
Outer Tube Diameter	2.5mm	2.6mm	Similar
SAL	10 ⁻⁶	10 ⁻⁶	Same
Sterilization Method	EO Sterilization	EO Sterilization	Same

5. Non-clinical Performance Data

The proposed device meets the requirements of ISO 10993 “Biological Evaluation of Medical Devices – Part 1: Evaluation and Testing”, ISO 11135-1 “Sterilization of Health Care products Ethylene Oxide - Part 1: Requirements for Development, Validation, and Routine Control of Sterilization processes for Medical Devices”, and ISO 10993-7 “Biological evaluation of medical devices - Part 7: ethylene oxide sterilization residuals” .

The following bench tests were performed on Disposable Hemostatic Clips: Appearance, Dimensional verification test, Separating force test, Clamping force test, Opening and closing force test, Rotational performance test, State of the clamping part after detachment test, and Passability of the clamping part after leaving the conveying part test. The results of all testing were passing.

6. Clinical Test Data

No Clinical Study is included in this submission.

7. Conclusion

In accordance with the Federal Food, Drug and Cosmetic Act, 21 CFR Part 807, Based on the information provided in this premarket notification, Zhejiang Soudon Medical Technology Co.,Ltd has demonstrated that proposed device Disposable Hemostatic Clips is substantially equivalent to Hangzhou AGS MedTech CO., Ltd 's currently marketed Hemoclip, K172727.