



October 13, 2023

Jordan Bauman
Vice President, Regulatory Affairs
1755 West Oak Parkway
Suite 100
Marietta, Georgia 30062

Re: K232481

Trade/Device Name: RIGEL™ 3DR Anterior Cervical Corpectomy System
Regulation Number: 21 CFR 888.3060
Regulation Name: Spinal Intervertebral Body Fixation Orthosis
Regulatory Class: Class II
Product Code: PLR
Dated: August 16, 2023
Received: August 16, 2023

Dear Mr. Bauman:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Brent Showalter -S

Brent Showalter, Ph.D.

Assistant Director

DHT6B: Division of Spinal Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K232481

Device Name

RIGEL™ 3DR Anterior Cervical Corpectomy System

Indications for Use (Describe)

The RIGEL™ 3DR Anterior Cervical Corpectomy System is indicated for vertebral body replacement in the cervical spine (C2- T1) in skeletally mature patients. The RIGEL™ 3DR Anterior Cervical Corpectomy System is intended to replace a diseased or damaged vertebral body caused by fracture, osteomyelitis, or tumor, or for reconstruction following corpectomy performed to achieve decompression of the spinal cord and neural tissues in cervical degenerative disorders. The System is intended to be used with supplemental fixation that has been cleared by the FDA for use in the cervical spine. The System is designed for use with autogenous and/or allogenic bone graft comprised of cancellous and/or corticocancellous bone graft as an adjunct to fusion. The System is also intended to restore the integrity of the spinal column even in the absence of fusion for a limited time period in patients with advanced stage tumors involving the cervical spine in whom life expectancy is of insufficient duration to permit achievement of fusion, with bone graft used at the surgeon's discretion.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

This 510(k) summary is being submitted in accordance with the requirements of 21 CFR 807.92(c).

I. SUBMITTER

MiRus™, LLC
1755 West Oak Parkway
Suite 100
Marietta, Georgia 30062
Tel: (678) 324-6272
Fax: (678) 401-5607

**II. OFFICIAL
CORRESPONDENT**

Jordan Bauman

Vice President, Regulatory Affairs
MiRus™, LLC
1755 West Oak Parkway
Suite 100
Marietta, Georgia 30062
Tel: (678) 324-6272
Fax: (678) 401-5607

III. DATE PREPARED

September 28, 2023

IV. DEVICE

Name of Device	RIGEL™ 3DR Anterior Cervical Corpectomy System
Common Name	Spinal Intervertebral Body Fixation Orthosis
Classification Name	21 CFR 882.3060
Regulatory Class	Class II
Product Codes	PLR
Submission Type	Traditional 510(k)

V. PREDICATE DEVICE

Primary Predicate
NEXXT MATRXXX® System – Nexxt Spine LLC (K193412)
Additional Predicate
C-VBR – Cardinal Spine, LLC (K152568)
Reference Devices
RIGEL™ 3DR Anterior Cervical Interbody Fusion System –
MiRus, LLC (K200685)
NGage Surgical Mesh System – Blackstone Medical
(K030744)

VI. DEVICE DESCRIPTION

The RIGEL™ 3DR Anterior Cervical Corpectomy System is a vertebral body replacement device intended to provide structural stability and mechanical support to the cervical spine following corpectomy.

The RIGEL™ 3DR Anterior Cervical Corpectomy System consists of implants additively manufactured from Titanium-6 Aluminum-4 Vanadium ELI per ASTM F3001 and instrumentation manufactured from Stainless Steel per ASTM F899. The RIGEL™ 3DR

Anterior Cervical Corpectomy System implants are offered in multiple configurations and different sizes to accommodate various patient anatomical requirements. The implants are provided sterile packed and are intended for single use only.

VII. INDICATIONS FOR USE

The RIGEL™ 3DR Anterior Cervical Corpectomy System is indicated for vertebral body replacement in the cervical spine (C2- T1) in skeletally mature patients. The RIGEL™ 3DR Anterior Cervical Corpectomy System is intended to replace a diseased or damaged vertebral body caused by fracture, osteomyelitis, or tumor, or for reconstruction following corpectomy performed to achieve decompression of the spinal cord and neural tissues in cervical degenerative disorders. The System is intended to be used with supplemental fixation that has been cleared by the FDA for use in the cervical spine. The System is designed for use with autogenous and/or allogenic bone graft comprised of cancellous and/or corticocancellous bone graft as an adjunct to fusion. The System is also intended to restore the integrity of the spinal column even in the absence of fusion for a limited time period in patients with advanced stage tumors involving the cervical spine in whom life expectancy is of insufficient duration to permit achievement of fusion, with bone graft used at the surgeon's discretion.

VIII. PREDICATE DEVICE COMPARISON

The RIGEL™ 3DR Anterior Cervical Corpectomy System has the same intended use, indications for use, labeling, and technological characteristics as the predicate systems, including the same design features, geometries, and materials, and a similar size range of footprints.

IX. PERFORMANCE DATA

The mechanical performance profile of the RIGEL™ 3DR Anterior Cervical Corpectomy System was assessed through static and dynamic testing in accordance with the following test methods:

- Static and dynamic compression testing (ASTM F2077-18)
- Static and dynamic torsion testing (ASTM F2077-18)
- Subsidence testing (ASTM F2267-04)

X. CONCLUSIONS

The RIGEL™ 3DR Anterior Cervical Corpectomy System has the same intended use, indications for use, labeling, and technological characteristics as the predicate systems, including the same design features, geometries, and materials, and a similar size range of footprints. Performance data demonstrate that the RIGEL™ 3DR Anterior Cervical Corpectomy System is substantially equivalent to legally marketed predicate systems.