



October 17, 2023

Pollogen Ltd
% Prithul Bom
Most Responsible Person
Regulatory Technology Services, LLC
1000 Westgate Drive,
Suite 510k
Saint Paul, Minnesota 55114

Re: K232903

Trade/Device Name: Legend X Platform

Regulation Number: 21 CFR 878.4400

Regulation Name: Electrosurgical cutting and coagulation device and accessories

Regulatory Class: Class II

Product Code: GEI, NGX

Dated: September 18, 2023

Received: September 18, 2023

Dear Prithul Bom:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device"

(<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Mark Trumbore -S Digitally signed by
Mark Trumbore -S
Date: 2023.10.17
15:50:52 -04'00'

Mark Trumbore, Ph.D.
Assistant Director
DHT4A: Division of General Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices

Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration

Form Approved: OMB No. 0910-0120
Expiration Date: 06/30/2023
See PRA Statement below.

Indications for Use

510(k) Number (if known)

K232903

Device Name

Legend X Platform

Indications for Use (Describe)

The Legend X Platform and its accessories are intended for dermatological procedures requiring ablation and resurfacing of the skin when using VoluDerm Energy (Applicator VO).

It is also intended for use in dermatologic and general surgical procedures for the non-invasive treatment of mild to moderate facial wrinkles and rhytides when using TriPollar RF Energy and for muscle conditioning to stimulate healthy muscles (Applicators 1-3).

Legend X Platform is not intended to be used in conjunction with therapy or treatment of medical diseases or medical conditions of any kind. Legend X Platform is intended to be operated by a trained professional who is present to monitor treatment.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

K232903

5.1. General information

510(k) Submitter	Pollogen Ltd 6 Kaufman Street Tel Aviv, IL 6801298
FDA Registration Number	3008753275
Primary Correspondent	Karen Smith Vice President, Regulatory & Quality Lumenis Be, Inc., a Pollogen Ltd Family Company
Contact Information	Email: karen.smith@lumenis.com
Secondary Correspondent	Lola Fajinmi Regulatory Affairs Consultant MedTeCheck
Contact Information	Email: lola@MedTeCheck.com
Date Prepared	September 17, 2023

5.2. Device Identification

The **Legend X Platform** consists of a Console with connected components including: a Foot Switch, Patient Controlled Manual Switch, Applicators, and Applicator accessories (disposable tips) needed to perform some of Legend X dermatologic procedures. The **Legend X Platform** can be defined as follows:

Proprietary Name	Legend X Platform
Device Classification Name:	Electrosurgical, Cutting & Coagulation & Accessories
Regulation Description:	Electrosurgical Cutting and Coagulation Device and Accessories
Regulation Number:	878.4400
Device Class:	Class II
Product Code:	GEI, NGX

5.3. Predicate Devices

Primary Predicate:

Proprietary Name	<i>Pollogen Legend+™ System</i>
Device Classification Name:	Electrosurgical, Cutting & Coagulation & Accessories
Regulation Description:	Electrosurgical Cutting and Coagulation Device and Accessories
510(k) Number:	K173503

Device: **Legend X Platform**
Section 5: 510(k) Summary

Pollogen Ltd

Traditional 510(k) Notification

Regulation Number:	878.4400
Device Class:	Class II
Product Code:	GEI

Secondary Predicate:

Proprietary Name	<i>Legend Pro DMA</i>
Device Classification Name:	Stimulator, Muscle, Powered for Muscle Conditioning
Regulation Description:	Powered Muscle Stimulator
510(k) Number:	K200545
Regulation Number:	890.5850
Device Class:	Class II
Product Code:	NGX

5.4. Device Description

The **Legend X Platform** (“Proposed Device”) is a software-controlled capital equipment platform that enables application of radiofrequency onto the skin for ablation, resurfacing, or noninvasive treatment of wrinkles and rhytides or muscle activation. The application of radiofrequency or muscle activation only occurs under the continuous and direct control of a user with direct visualization, via the user inputs into the Console, Applicator indicators and subsequent operation of the Foot Switch and Applicators.

The proposed device consists of a Console with connected Foot Switch, Patient Controlled Manual Switch, Applicators, and Applicator accessories (disposable tips) needed to perform some of Legend X dermatological and general surgical procedures. The summary descriptions of each component are outlined below.

Legend X Console: The Legend X Console (also known as the Console or Main Control Unit) contains the graphic display interface for the user that is provided by a touchscreen monitor for viewing and a computer running the Legend X software. The monitor allows for user input during initial setup and throughout the session. The console also provides power and connectivity for the applicators, foot switch, and patient manual-controlled switch.

Legend X Foot Switch: The Legend X Foot Switch (Foot Switch) is connected to the Legend X Console via a connector cord. The Foot Switch enables or disables the energy delivery through the applicators to treatment area.

Legend X Applicators: The Legend X Applicators (1, 2, 3 and VO) are handheld handpieces which deliver energy to the treatments area. Depending on the applicator and selected user inputs, the user can utilize the applicators to deliver either muscle energy onto the muscles of the face and body or RF energy for ablation and resurfacing or noninvasive treatment of wrinkles and rhytides onto the skin. Each applicator is connected to the Console via an Applicator Connector cord.

Device: **Legend X Platform**

Section 5: 510(k) Summary

Legend X Patient Controlled Manual Switch: The Legend X Patient Controlled Manual Switch is also a handheld device that serves as the patient interface that allows the patient to stop the operation of the chosen treatment mode upon the press of a button. In this situation, the system does not allow operation to continue. A Patient Controlled Manual Switch cord connects the Patient Controlled Manual Switch to the Console.

Legend X Software: The Legend X Software provides the user and patient with the ability to safely commence, drive, and stop the operation of the Applicators on the skin or muscle area of interest throughout the face and body. It receives user input from the Legend X Console, Foot Switch, and Patient Manual Controlled Switch, and computes the appropriate output to the chosen applicator connected to the Console. It provides a graphical user interface where the treatment timeline as well as status of operation is shown in real time and displays important system status information.

Legend X Disposable Tips: The Legend X Disposable Tips set is part of the Legend X Applicator VO kit. The Disposable Tips Set includes:

- gen12 disposable tip
- gen 36 disposable tip
- gen 36L disposable tip
- gen 100 disposable tip
- H7X7 disposable tip

The patient contacting portions of the disposable tips are constructed of stainless steel and gold plating. The disposable tips are connected to handheld Applicator VO for delivery of radiofrequency electrical current via an array of multi-electrode pins onto the skin surface.

5.5. Intended Use/Indications for Use

The **Legend X Platform** and its accessories are intended for dermatological procedures requiring ablation and resurfacing of the skin when using VoluDerm Energy (Applicator VO).

It is also intended for use in dermatologic and general surgical procedures for the non-invasive treatment of mild to moderate facial wrinkles and rhytides when using TriPollar RF Energy and for muscle conditioning to stimulate healthy muscles (Applicators 1-3).

Legend X Platform is not intended to be used in conjunction with therapy or treatment of medical diseases or medical conditions of any kind. Legend X Platform is intended to be operated by a trained professional who is present to monitor treatment.

5.6. Comparison of Technological Characteristics with the Predicate Device

Overall, the proposed and predicate devices are based on the following similar basic technological elements:

Device: **Legend X Platform**

Section 5: 510(k) Summary

- Device contains a central console that connects each of the main components to facilitate dermatologic and general surgical procedures and facilitates the selection of either RF, muscle stimulation, or fractional RF treatment modes via the graphic user interface
- Device contains applicators to allow energy delivery and application of the selected mode on to the area of interest on the skin or muscles
- Device contains similar components between the subject and predicate devices (i.e., console, foot switch, applicators, patient manual-controlled switch, and disposable tips)
- Device requires continuous direct control by the user to operate the Applicators
- Device mode operation is only at the command of the user
- Device mode operation may be halted at the command of either the user or the patient

The main difference between the proposed device and the predicates, Pollogen Legend+™ System (K173503) and Legend Pro DMA (K200545), is that the Legend X Platform includes two of the five disposable tips (gen100 and H7x7) as an additional option. The Legend X Platform is an upgraded version of the predicate devices. For this reason, additional testing for the disposable tips was performed and is captured as part of this submission.

Comparison of the technological characteristics with the Predicate Device for each of the **Legend X Platform** components are described below:

Proposed Device Legend X Platform	Primary Predicate Pollogen Legend+™ (K173503)	Secondary Predicate Legend Pro DMA (K200545)	Technological Characteristics Comparison
System Console/ Main Control Unit	System Console/ Main Control Unit	System Console/ Main Control Unit	Identical Characteristics: <ul style="list-style-type: none"> <input type="checkbox"/> <u>Functionality</u>: To provide central connection to all device components; as well as a user interface for user inputs during setup <input type="checkbox"/> Same specifications and materials Differences <ul style="list-style-type: none"> <input type="checkbox"/> Minor updates to console, GUI
Foot Switch	Foot Switch	Foot Switch	Same

Patient-controlled manual switch	Patient-controlled manual switch	Patient-controlled manual switch	Same
Applicators (1, 2, 3, and VO)	Applicators (1, 2, 3, and VO)	Applicators (1, 2, and 3)	Same
Disposable Tips: <input type="checkbox"/> Gen12 <input type="checkbox"/> Gen36 <input type="checkbox"/> Gen36L <input type="checkbox"/> Gen100 <input type="checkbox"/> H7x7	Disposable Tips: <input type="checkbox"/> Gen12 <input type="checkbox"/> Gen36 <input type="checkbox"/> Gen36L	N/A	Identical Characteristics: <input type="checkbox"/> Functionality: To provide delivery of fractional RF energy <input type="checkbox"/> Similar specifications and intended functionality. Differences <input type="checkbox"/> Two additional optional disposable tips (same materials with additional pin configurations for gen100 and H7x7) for applicator VO <input type="checkbox"/> Minor external color differences

5.7. Performance Data

The **Legend X Platform** was assessed for performance in accordance with internal design specification with the applicable performance standards to demonstrate safety and effectiveness. The testing identified no new issues of safety or effectiveness. The testing performed are summarized below:

Reprocessing

Cleaning: The cleaning and disinfection instructions provided in labeling for non-disposable components were validated against the following standard:

- AAMI TIR-30:2011 – A compendium of processes, materials, test methods, and acceptance criteria for cleaning reusable medical devices*

Sterility: Single- use disposable sterile (EO) devices and EO residual were validated in accordance with the following standards:

- ISO 11135:2014 – Sterilization of healthcare products – Ethylene Oxide – Requirements for development validation and routine control of a sterilization process for medical devices*
- ISO10993-7:2008/AMD-1:2019 – Biological evaluation of medical devices – Ethylene Oxide sterilization residuals*

Shelf Life and Sterile Barrier Packaging: The shelf life and sterile barrier packaging of

Device: **Legend X Platform**
Section 5: 510(k) Summary

the single-use disposable devices were evaluated per the following standards:

- ASTM F1980-16 – Standard guide for accelerated aging of sterile barrier systems for medical devices*
- ASTM F 1929-15 – Standard Test Method for Detecting Seal Leaks in Porous Medical Packaging by Dye Penetration*
- ASTM F88/F88M-15: Standard Test Method for Seal Strength of Flexible Barrier Materials*
- ISO 11607 – Packaging for terminally sterilized medical devices*

Biocompatibility

The final finished form of the subject device has been used for the biocompatibility evaluation. Biocompatibility for patient contacting components has been evaluated and validated in accordance with the provision of the following FDA Guidance document(s) and standards:

- Guidance for Industry and Food and Drug Administration Staff – Use of International Standard ISO 10993-1, “Biological evaluation of medical devices – Part 1: Evaluation and testing within a risk management process”*
- ISO 10993-1:2018 - Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process*
- ISO 10993-5:2009 - Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity*
- ISO 10993-10:2010 - Biological evaluation of medical devices - Part 10: Tests for irritation and skin sensitization*
- ISO 10993-11:2017 Biological evaluation of medical devices - Part 11: Tests for systemic toxicity*
- ISO 10993-23:2021 - Biological evaluation of medical devices - Part 23: Tests for irritation*

Electrical Safety, Performance and EMC

The proposed device has been fully evaluated for electrical safety and EMC compliance to the following standards:

- IEC 60601-1:2005 (Third edition) + CORR. 1: 2006 + CORR. 2:2007 + A1:2012 (or IEC 60601-1: 2012 reprint) – Medical electrical equipment - Part 1: General requirements for basic safety and essential performance (IEC 60601-1:2005, MOD)*
- IEC 60601-1-2:2014 – Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests*
- IEC 60601-2-10:2016 – Medical electrical equipment - Part 2-10: Particular requirements for the basic safety and essential performance of nerve and muscle stimulators*

- *IEC 60601-2-2:2017 - Medical electrical equipment - Part 2-2: Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories*

Software

Software was developed, tested, and verified per the FDA guidance document: Guidance for Industry and FDA Staff – Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices

Results of verification and validation testing confirm that the proposed device conforms to design specifications and meets the needs of the intended users.

V&V Testing

Performance testing was executed to verify the overall functionality of the proposed device to operate as specified by the design input requirements including applicator, patient-manual controlled-switch, and foot switch operation and controls, various functional safety features, and other general functionality. Requirements for safety and efficacy of the system, including but not limited to the adherence to regulatory standards were verified. Results of verification testing confirm that the proposed device conforms to design specifications and requirements and meets the needs of the intended user.

Animal Testing

Performance testing - animal was executed according to General Considerations for Animal Studies Intended to Evaluate Medical Devices issued on March 28, 2023.

5.8. Conclusion

Based on the indication for use, technological characteristic, performance testing and comparison to the predicate devices, the **Legend X Platform** raises no new questions of safety and effectiveness as compared to the predicate devices and is substantially equivalent to the predicate devices in terms of safety, efficacy, and performance.