



Hangzhou AllTest Biotech Co.,Ltd
% Joe Shia
Director
LSI International, Inc.
504 East Diamond Ave., Suite H
Gaithersburg, Maryland 20877

Re: K233417
Trade/Device Name: AllTest Fentanyl Urine Test Cassette
Regulation Number: 21 CFR 862.3650
Regulation Name: Opiate test system
Regulatory Class: Class II
Product Code: NGL
Dated: October 6, 2023
Received: October 10, 2023

Dear Joe Shia:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801 and Part 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Joseph A. Kotarek -S
Digitally signed by
Joseph A. Kotarek -S
Date: 2023.10.26
17:21:15 -04'00'

Joseph Kotarek
Branch Chief for Toxicology
Division of Chemistry
and Toxicology Devices
OHT7: Office of In Vitro Diagnostics
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

Submission Number (if known)

k233417

Device Name

AllTest Fentanyl Urine Test Cassette

Indications for Use (Describe)

AllTest Fentanyl Urine Test Cassette is a competitive binding, lateral flow immunochromatographic assay for qualitative detection of Fentanyl in human urine at the cutoff concentrations of 1 ng/mL. The test provides only preliminary test results. A more specific alternative chemical method must be used in order to obtain a confirmed analytical result. GC/MS or LC/MS is the preferred confirmatory method. For in vitro diagnostic use only.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) SUMMARY
K233417

1. Date: October 26, 2023
2. Submitter: Hangzhou Alltest Biotech Co., Ltd.
#550, Yin Hai Street
Hangzhou, Zhejiang, China 310018
3. Contact person: Joe Shia
LSI International Inc.
504 East Diamond Ave., Suite H
Gaithersburg, MD 20877
Telephone: 240-505-7880
Fax: 301-916-6213
Email: shiajl@yahoo.com
4. Device Name: AllTest Fentanyl Urine Test Cassette

Classification: Class II

Product Code	CFR #	Panel
NGL	21 CFR, 862.3650 Opiate Test System	Toxicology

5. Predicate Devices: K231698
AllTest Fentanyl Rapid Test (Urine)

6. Intended Use:

AllTest Fentanyl Urine Test Cassette is a competitive binding, lateral flow immunochromatographic assay for qualitative detection of Fentanyl in human urine at the cutoff concentrations of 1 ng/mL. The test provides only preliminary test results. A more specific alternative chemical method must be used in order to obtain a confirmed analytical result. GC/MS or LC/MS is the preferred confirmatory method. For in vitro diagnostic use only.

7. Device Description:

AllTest Fentanyl Urine Test Cassette is an immunoassay intended for the qualitative detection of fentanyl in human urine. Each AllTest Fentanyl Urine Test Cassette consists of a Test Cassette and a package insert. Each Test Cassette is sealed with sachets of desiccant in an aluminum pouch.

8. Substantial Equivalence Information

Item	Device	Predicate – K231698
Indication(s) for Use	For the qualitative determination of fentanyl in human urine.	Same
Calibrator and Cut-Off Values	Fentanyl (FTY) 1 ng/ml	Same
Methodology	Competitive binding, lateral flow immunochromatographic assays based on the principle of antigen antibody immunochemistry.	Same
Type of Test	Qualitative	Same
Specimen Type	Human Urine	Same
Intended Use	Over-The-Counter Use	For prescription use
Configurations	Cassette	Same
Storage	4-30°C	Same

9. Test Principle

AllTest Fentanyl Urine Test Cassette is a competitive and immunochromatography assay, and uses monoclonal antibody as the indicator marker to qualitatively detect fentanyl in human urine.

The test cassette contains fentanyl test strip. The nitrocellulose membrane test area (T) of the test strip is correspondingly coated with fentanyl-bovine serum albumin conjugate, and the quality control area (C) is coated with goat anti-rabbit IgG polyclonal antibody. When the concentration of fentanyl in the sample is higher than or equal to the cut-off of the product, it will compete with the corresponding conjugate coated on the test area (T) to bind to the monoclonal antibody, the test line is inhibited and the result is positive; while when the sample does not contain fentanyl or its concentration is lower than the cut-off of the product, the corresponding conjugate on the test line reacts with sufficient monoclonal antibodies, the test line will be visible and the result is negative. No matter whether the sample contains the corresponding analyte or not, the quality control area (C) will develop a colored line, which is the criteria for judging whether the chromatography process is normal or not.

10. Performance Characteristics

1. Analytical Performance Clearance of candidate device is for addition of OTC claim. See analytical performance in predicate K231698.

2. Comparison Studies See studies in predicate K231698

3. Lay-user study

A lay user study was performed at three intended user sites with 140 lay persons. They had diverse educational and professional backgrounds and ranged in age from 21 to >50 years. Urine samples were prepared at the following concentrations; -100%, +/-75%, +/-50%, +/-25% of the cut-off by spiking fentanyl into drug free-pooled urine specimens. The concentrations of the samples were confirmed by LC/MS. Each sample was aliquoted into individual containers, blind-labeled and randomized. Each participant was provided with the package insert, 1 blind labeled sample and a device. The results are summarized below:

% of Cutoff	Number of samples	Fentanyl Concentration by LC/MS (ng/mL)	Lay person results		The percentage of correct results (%)
			No. of Positive	No. of Negative	
-100% Cutoff	20	0	0	20	100
-75% Cutoff	20	0.27	0	20	100
-50% Cutoff	20	0.52	0	20	100
-25% Cutoff	20	0.74	1	19	95
+25% Cutoff	20	1.17	20	0	100
+50% Cutoff	20	1.48	20	0	100
+75% Cutoff	20	1.72	20	0	100

Lay-users were also given surveys on the ease of understanding the package insert instructions. All lay users indicated that the device instructions can be easily followed A Flesch-Kincaid reading analysis was performed on the package insert and the score revealed a reading grade level of less than 7.

4. Clinical Studies

Not applicable.

11. Conclusion

Based on the test principle and performance characteristics of the device, it's concluded that AllTest Fentanyl Urine Test Cassette is substantially equivalent to the predicate.