



December 21, 2023

Tyber Medical, LLC.
Lisa Boyle
Director, Regulatory Affairs
83 South Commerce Way
Suite 310
Bethlehem, Pennsylvania 18017

Re: K233423

Trade/Device Name: Tyber Medical Anatomical Plating System
Regulation Number: 21 CFR 888.3030
Regulation Name: Single/Multiple Component Metallic Bone Fixation Appliances And Accessories
Regulatory Class: Class II
Product Code: HRS, HWC
Dated: October 6, 2023
Received: October 10, 2023

Dear Lisa Boyle:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shumaya Ali -S

Shumaya Ali, M.P.H.

Assistant Director

DHT6C: Division of Restorative, Repair
and Trauma Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K233423

Device Name

Tyber Medical Anatomical Plating System

Indications for Use (Describe)

Mini-Frag System:

The Tyber Medical Mini-Frag System is indicated for fixation of fractures, osteotomies, nonunions, malunions, replantations, and fusions of short bones and small fragments of bone including the hand, wrist, foot, and ankle. The mini-frag system is also intended for reduction and stabilization of non-load bearing long bone fragments.

The Tyber Medical Mini-frag System is not for Spinal Use.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

Prepared on: 2023-10-10

Contact Details

[21 CFR 807.92\(a\)\(1\)](#)

Applicant Name	Tyber Medical, LLC.
Applicant Address	83 South Commerce Way Suite 310 Bethlehem PA 18107 United States
Applicant Contact Telephone	610-295-7984
Applicant Contact	Mrs. Lisa Boyle
Applicant Contact Email	lboyle@tybermed.com

Device Name

[21 CFR 807.92\(a\)\(2\)](#)

Device Trade Name	Tyber Medical Anatomical Plating System
Common Name	Single/multiple component metallic bone fixation appliances and accessories
Classification Name	Plate, Fixation, Bone
Regulation Number	888.3030
Product Code	HRS

Legally Marketed Predicate Devices

[21 CFR 807.92\(a\)\(3\)](#)

Predicate #	Predicate Trade Name (Primary Predicate is listed first)	Product Code
K150099	DePuy Synthes Variable Angle Locking Hand System	HRS
K203817/K222465	Tyber Medical, LLC.	HRS

Device Description Summary

[21 CFR 807.92\(a\)\(4\)](#)

The Tyber Medical Anatomical Plating System – Line extension to the Mini-Frag System consists of titanium plates (straight and T-plate) and locking /non-locking screws (1.3mm) and are offered in a range of configurations to accommodate patient anatomy. The plates and screws are non-sterile.

Intended Use/Indications for Use

[21 CFR 807.92\(a\)\(5\)](#)

The Tyber Medical Anatomical Plating System is intended to bridge or otherwise stabilize bone fragments to facilitate healing.

Mini-Frag System:

The Tyber Medical Mini-Frag System is indicated for fixation of fractures, osteotomies, nonunions, malunions, replantations, and fusions of short bones and small fragments of bone including the hand, wrist, foot, and ankle. The mini-frag system is also intended for reduction and stabilization of non-load bearing long bone fragments. The Tyber Medical Mini-frag System is not for Spinal Use.

Indications for Use Comparison

21 CFR 807.92(a)(5)

The indications are similar in both the subject and predicate device(s).

Technological Comparison

21 CFR 807.92(a)(6)

The subject device has the same technological characteristics (design, material, chemical composition and principle of operation) as the predicate device(s) identified above.

Non-Clinical and/or Clinical Tests Summary & Conclusions

21 CFR 807.92(b)

The mechanical testing for the subject devices were performed in conformance with the FDA currently-recognized version of ASTM F382 for metallic bone plates and ASTM F543 for bone screws.

No clinical testing was performed.

The Tyber Medical Anatomical Plating System (Mini-Frag Line Extension) is substantially equivalent to the predicate devices in material, basic design features, intended use, operation and performance. Any differences between the subject and predicate device are considered minor and do not raise different questions concerning safety, performance or effectiveness. From the evidence submitted in this 510(k), the subject devices are considered substantially equivalent to the predicate device.