



February 15, 2024

Quality Electrodynamics
JoVanna Boudreaux
Quality/Regulatory Engineer
6655 Beta Drive
Suite 100
Mayfield Heights, Ohio 44143

Re: K233656
Trade/Device Name: Contour Shoulder
Regulation Number: 21 CFR 892.1000
Regulation Name: Magnetic resonance diagnostic device
Regulatory Class: Class II
Product Code: MOS
Dated: January 18, 2024
Received: January 18, 2024

Dear JoVanna Boudreaux:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,



Daniel M. Krainak, Ph.D.
Assistant Director
DHT8C: Division of Radiological Imaging
and Radiation Therapy Devices
OHT8: Office of Radiological Health
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

Submission Number (if known)

K233656

Device Name

Contour Shoulder

Indications for Use (Describe)

The Contour Shoulder is intended for use with Siemens 0.55T MR systems to produce diagnostic images of shoulder anatomy that can be interpreted by a trained physician.

Type of Use (Select one or both, as applicable)



Prescription Use (Part 21 CFR 801 Subpart D)



Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

Contact Details

21 CFR 807.92(a)(1)

Applicant Name	Quality Electrodynamics
Applicant Address	6655 Beta Drive Ste 100 Mayfield Heights OH 44143 United States
Applicant Contact Telephone	4404842340
Applicant Contact	Mrs. JoVanna Boudreaux
Applicant Contact Email	jovanna.boudreaux@qualedyn.com

Device Name

21 CFR 807.92(a)(2)

Device Trade Name	Contour Shoulder (Q7000222)
Common Name	Magnetic resonance diagnostic device
Classification Name	Coil, Magnetic Resonance, Specialty
Regulation Number	892.1000
Product Code	MOS

Legally Marketed Predicate Devices

21 CFR 807.92(a)(3)

Predicate #	Predicate Trade Name (Primary Predicate is listed first)	Product Code
K223429	Contour Knee	MOS

Device Description Summary

21 CFR 807.92(a)(4)

The Contour Shoulder is a receive-only, 12-channel phased array coil designed for magnetic resonance imaging (MRI) using the Siemens 0.55T MR systems. The Contour Shoulder is intended to be used for imaging Shoulder anatomy.

The Contour Shoulder is a reusable, non-invasive device with limited exposure with regard to duration of contact with the body. The central coil elements are enclosed in a rigid plastic housing which is fire-rated, has impact and tensile strength, and has been tested for biocompatibility. The flexible wing coil elements are encapsulated in polycarbonate and aramid felt which is fire-rated, has impact and tensile strength, then covered with a polyurethane coated nylon fabric which has been evaluated for biocompatibility.

Intended Use/Indications for Use

21 CFR 807.92(a)(5)

The Contour Shoulder is intended for use with Siemens 0.55T MR systems to produce diagnostic images of shoulder anatomy that can be interpreted by a trained physician.

Indications for Use Comparison

21 CFR 807.92(a)(5)

The Indications for Use statement for the Contour Shoulder is not identical to that of the predicate device (Contour Knee); however, the differences do not affect the safety or effectiveness of the device relative to the predicate device. Both Indications for Use statements for the proposed Contour Shoulder and predicate Contour Knee indicate that the device is intended to be used in conjunction with a MR system to produce images of human anatomy and that the images can be interpreted by a trained physician. The indications for use statements differ only in that intended for use is to image shoulder anatomy instead of knee anatomy.

Technological Comparison

21 CFR 807.92(a)(6)

At a high level, the proposed and predicate device are based on the following same technological elements: • Receive-only phased array RF coil

- Active PIN diode switching blocking circuitry. Passive blocking circuitry.
- Materials used for flame retardancy and biocompatibility: Polycarbonate and aramid felt with a polyurethane coated nylon fabric cover

The following technological differences exist between the proposed and predicate device:

- Intended for use (shoulder anatomy (proposed device) versus knee anatomy (predicate device))
- Flexible wing coil elements are encapsulated in polycarbonate, polyimide and aramid felt (proposed device) versus flexible blanket-like enclosure for anterior coil elements, rigid plastic housing for posterior coil elements (predicate device)

Non-Clinical and/or Clinical Tests Summary & Conclusions

21 CFR 807.92(b)

The signal-to-noise ratio (SNR) and image uniformity of the Contour Shoulder were measured on a 0.55T Siemens MR System, manufactured by Siemens Healthineers. The SNR and uniformity of the Contour Shoulder were analyzed per NEMA MS-9 (using alternate method 2.5 from MS-6) and uniformity was analyzed using NEMA MS-9 (primary method from MS-6) using pre-determined acceptance criteria. See attached bench testing results.

In accordance with the FDA Guidance for Industry: Guidance for the Submission of Premarket Notifications for Magnetic Resonance Diagnostic Devices, clinical images from volunteer scanning of shoulder anatomy were obtained from a Siemens 0.55T MR system. These images were used to demonstrate that the Contour Shoulder produces diagnostic quality images of the intended anatomy. No adverse events were reported or recorded. The parameters used to obtain the clinical images for the Contour Shoulder, clinical images for the Contour Shoulder, and images also provided in the native DICOM format are attached.

The electrical safety and electromagnetic compatibility and biocompatibility data support the safety of the Contour Shoulder and the bench testing per the IEC standards and diagnostic quality sample clinical images demonstrates the performance and effectiveness of the device under the specified use conditions. This testing demonstrates that the Contour Shoulder performs as well as or better than the predicate device.