



December 20, 2023

CyDen Limited
Adam Stephens
Regulatory Engineer
Block A, Bay Studios Business Park
Fabian Way, SA1 8QB
Swansea,
United Kingdom

Re: K233782

Trade/Device Name: SmoothSkin Pure Adapt (SSG3)

Regulation Number: 21 CFR 878.4810

Regulation Name: Laser Surgical Instrument For Use In General And Plastic Surgery And In
Dermatology

Regulatory Class: Class II

Product Code: OHT

Dated: November 27, 2023

Received: November 27, 2023

Dear Adam Stephens:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shlomit
Halachmi -S

Digitally signed by
Shlomit Halachmi -S
Date: 2023.12.20 14:39:00
-05'00'

For Tanisha Hithe
Assistant Director
DHT4A: Division of General Surgery Devices

OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

Submission Number (if known)

K233782

Device Name

SmoothSkin Pure Adapt (SSG3)

Indications for Use (Describe)

The SmoothSkin Pure Adapt is intended for the removal of unwanted hair. The SmoothSkin Pure Adapt is also indicated for the permanent reduction in hair regrowth, defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9 and 12 months after the completion of a treatment regime.

Type of Use (Select one or both, as applicable)

☐

Prescription Use (Part 21 CFR 801 Subpart D)

☒

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

Prepared on: 2023-11-27

Contact Details

21 CFR 807.92(a)(1)

Applicant Name	CyDen Limited
Applicant Address	Block A, Bay Studios Business Park Fabian Way, SA1 8QB Swansea United Kingdom
Applicant Contact Telephone	+44 1792 274009
Applicant Contact	Mr. Adam Stephens
Applicant Contact Email	astephens@cyden.com

Device Name

21 CFR 807.92(a)(2)

Device Trade Name	SmoothSkin Pure Adapt (SSG3)
Common Name	Laser surgical instrument for use in general and plastic surgery and in dermatology
Classification Name	Light Based Over-The-Counter Hair Removal
Regulation Number	878.4810
Product Code	OHT

Legally Marketed Predicate Devices

21 CFR 807.92(a)(3)

Predicate #	Predicate Trade Name (Primary Predicate is listed first)	Product Code
K200184	SmoothSkin Pure / SmoothSkin Pure Fit	OHT
K143003	SmoothSkin Gold	OHT

Device Description Summary

21 CFR 807.92(a)(4)

The SmoothSkin Pure Adapt is an over the counter IPL (Intense Pulsed Light) Hair Removal device for home use.

The device is mains powered and consists of the handset, hardwired base unit and AC power cord. AC mains is fed into the AC to DC converter that is used to charge the electrical storage capacitor, the stored energy is used to produce an optical output from the flash lamp.

Light energy is transferred through the skin's surface and is absorbed by melanin present in the hair shaft. The absorbed light energy is converted to heat energy (below the surface of the skin), which reduces the hairs growing back.

Intended Use/Indications for Use

21 CFR 807.92(a)(5)

The SmoothSkin Pure Adapt is intended for the removal of unwanted hair. The SmoothSkin Pure Adapt is also indicated for the permanent reduction in hair regrowth, defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9 and 12 months after the completion of a treatment regime.

Indications for Use Comparison

21 CFR 807.92(a)(5)

Same indications for use.

Technological Comparison

21 CFR 807.92(a)(6)

The SmoothSkin Pure Adapt has the same following technical characteristics:

- The same indicated use,
- Operates with the same mode of action,
- Has the same technological characteristics,
- Has the same Materials,
- Has the same shelf-life,
- Is manufactured and packed in the same manner,
- Has similar output parameters to the predicate device,
- Has the same safety features.

The following changes have been made to the device:

- Contour Adapters: 2 spring loaded moving paddles on the front of the device for improved skin contact to the user.

Non-Clinical and/or Clinical Tests Summary & Conclusions

21 CFR 807.92(b)

The device is compliant with the following Electrical Safety and Electromagnetic Compatibility standards.

- 1) AAMI / IEC 60601-1 : Medical Electrical Equipment - Part 1: General Requirements For Basic Safety And Essential Performance
- 2) IEC 60601-1-6 : Medical Electrical Equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability
- 3) IEC 60601-2-57 : Medical Electrical Equipment : Part 2-57: Particular requirements for the basic safety and essential performance of non-laser light source equipment intended for therapeutic, diagnostic, monitoring, cosmetic and aesthetic use
- 4) IEC 60601-2-83 : Medical Electrical Equipment: Part 2-83: Particular requirement for the basic safety and essential performance of home light therapy equipment
- 5) IEC 60601-1-2 : Part 1-2: General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests
- 6) IEC 62471 : Photobiological safety of lamps and lamp systems

The device is compliant with the following Biocompatibility standards

- 1) ISO 10993-1 : Biological evaluation of medical devices Part 1: Evaluation and testing within a risk management process
- 2) ISO 10993-5 : Biological evaluation of medical devices Part 5: Tests for in vitro cytotoxicity
- 3) ISO 10993-10 : Biological evaluation of medical devices Part 10: Tests for skin sensitization