



March 11, 2024

Jeil Medical Corporation
Jinwoo Kim
RA Specialist
702, 703, 704, 705, 706, 804, 805, 807, 812, 815-ho, 55,
Digital-ro 34-gil, Guro-gu
Seoul, 08378
Korea, South

Re: K233912

Trade/Device Name: ARIX Cannulated Screw System
Regulation Number: 21 CFR 888.3040
Regulation Name: Smooth Or Threaded Metallic Bone Fixation Fastener
Regulatory Class: Class II
Product Code: HWC, NDG
Dated: December 12, 2023
Received: December 12, 2023

Dear Jinwoo Kim:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shumaya Ali -S

Shumaya Ali, MPH

Assistant Director

DHT6C: Division of Restorative, Repair
and Trauma Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K233912

Device Name
ARIX Cannulated Screw System

Indications for Use (Describe)

The ARIX Cannulated Screw System is indicated for fracture fixation, fusions, osteotomies, nonunions, and malunions of fragments of long bones of the pelvis, the femur, the tibia, the humerus, and the foot. The system is not intended for use in the spine.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

K233912
510(K) Summary
[As required by 21 CFR 807.92]

1. Date Prepared [21 CFR 807.92(a)(a)]

March 11, 2024

2. Submitter's Information [21 CFR 807.92(a)(1)]

- Name of Sponsor: Jeil Medical Corporation
 - Address: 702,703,704,705,706,804,805,807,812,815-ho, 55,
Digital-ro 34-gil, Guro-gu, Seoul, 08378, Korea

- Contact Name: Jinwoo Kim / RA Specialist
 - Telephone No. +82 2 850 3934
 - Fax No. +82 2 850 3536
 - Email Address: jinwookim@jeilmed.co.kr

- Registration Number: 3004049923

- Name of Manufacturer: Same as Sponsor
 - Address: Same as Sponsor

3. Trade Name, Common Name, Classification [21 CFR 807.92(a)(2)]

- Trade Name; ARIX Cannulated Screw System
- Common Name; Screw, Fixation, Bone
- Classification Name; Smooth or threaded metallic bone fixation fastener
- Classification Panel; Orthopedic
- Classification Regulation; 21 CFR 888.3040
- Product Code; HWC, NDG
- Device Class; II

4. Identification of Predicate Device(s) [21 CFR 807.92(a)(3)]

The legally marketed device(s) to which substantial equivalence is claimed is/are:

Primary Predicate Device	K220319 - Asnis® III Cannulated Screw System, Stryker GMBH
Reference Device	K192417 - ARIX Cannulated Screw System, Jeil Medical Corporation

There are no significant differences between the subject device and the predicate device that would adversely affect the use of the product. It is substantially equivalent to these devices in intended use and technological characteristics as internal fixation components.

5. Description of the Device [21 CFR 807.92(a)(4)]

The ARIX Cannulated Screw System consists of cannulated screws and corresponding washer. The thread diameters of cannulated screws are 6.5 and 7.3 mm. They are either fully or partially threaded. Screws and a washer are made of Titanium Alloy (Ti-6Al-4V), which meet ASTM F136, Standard Specification for Wrought Titanium-6 Aluminum-4 Vanadium ELI (Extra Low Interstitial) Alloy for Surgical Implant Applications, which are widely used for surgical implants with well-known biocompatibility. All devices in the system are provided non-sterile.

6. Indications for use [21 CFR 807.92(a)(5)]

The ARIX Cannulated Screw System is indicated for fracture fixation, fusions, osteotomies, nonunions, and malunions of fragments of long bones of the pelvis, the femur, the tibia, the humerus, and the foot. The system is not intended for use in the spine.

7. Technological Characteristics [21 CFR 807.92(a)(6)]

Subject and predicate device are cannulated screw which fully or partially threaded. Both subject and predicate device are made of Titanium Alloy (ASTM F136), and predicate device can also be made of Stainless Steel (ASTM F138). Both subject and predicate device are provided as non-sterile, and predicate device can also be provided as sterile. Compared to predicate device, subject device is not made of new materials and is not provided in different states.

Detailed design/dimension would not the same, subject device was evaluated for each mechanical test according to ASTM F543, and the test results were better than the acceptance criteria. Therefore, substantially equivalent mechanical performance of the subject device has been demonstrated.

Non-Clinical Test Summary:

Bench tests were conducted to ensure the safety and effectiveness of the device as well as to demonstrate substantial equivalence to the predicate device. The test results demonstrated that the subject device complies with the following standards:

- ASTM F543, Standard Specification and Test Methods for Metallic Medical Bone Screws
 - Torsion Test
 - Driving Torque Test
 - Axial Pull-out Test

The results of this testing indicate that the ARIX Cannulated Screw System is equivalent to the predicate device.

Clinical Test Summary:

No clinical studies were considered necessary and performed.

8. Substantial Equivalence [21 CFR 807.92(b)(1) and 807.92]

Based on the submitted information in this premarket notification, the subject device is substantially equivalent to the predicate device in terms of:

- Intended use
- Technological characteristics (Design features, Material, Surface Treatment, Sterilization methods, Biocompatibility and Performance)

The subject device has met the performance, safety, and effectiveness of the device for its intended use.

9. Conclusion [21 CFR 807.92(b)(3)]

In all respects, the ARIX Cannulated Screw System is substantially equivalent to the legally marketed device. Above all, the subject device has equivalent intended use and technological characteristics. Further, nonclinical verification and validation to determine substantial equivalence provide additional evidence that subject device is substantially equivalent to the predicate device in terms of safety, efficacy, and performance.