



March 15, 2024

Alembic, LLC  
% Prithul Bom  
Most Responsible Person  
Regulatory Technology Services  
1000 Westgate Drive, Suite #510k  
Saint Paul, Minnesota 55114

Re: K234115  
Trade/Device Name: APRO 55 Catheter  
Regulation Number: 21 CFR 870.1250  
Regulation Name: Percutaneous Catheter  
Regulatory Class: Class II  
Product Code: DQY, QJP  
Dated: March 6, 2024  
Received: March 6, 2024

Dear Prithul Bom:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

**Naira Muradyan -S**

Naira Muradyan, Ph.D.

Assistant Director

DHT5A: Division of Neurosurgical,  
Neurointerventional

and Neurodiagnostic Devices

OHT5: Office of Neurological

and Physical Medicine Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K234115

Device Name  
APRO 55 Catheter

### Indications for Use (Describe)

The APRO 55 Catheter is indicated for use in facilitating the insertion and guidance of appropriately sized interventional devices into a selected blood vessel in the peripheral and neurovascular systems. The APRO 55 Catheter is also indicated for use as a conduit for retrieval devices.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary

510(k) Number:     K234115    

This 510(k) Summary is provided in accordance with the requirements of 21 CFR §807.92.

### 1) Submitter information

**Submitter:** Alembic, LLC  
627 National Ave.  
Mountain View, CA 94043

**Contact:** Lisa Yen  
Director of Regulatory and Quality  
Telephone Number: (650) 388-5087  
Fax: (650) 390-0107  
Email: lyen@alembicllc.com

**Date Prepared:** March 13, 2024

### 2) Device Name and Classification

**Trade/Proprietary Name:** APRO<sup>®</sup> 55 Catheter

**Common Name:** Percutaneous Catheter

**Classification Name:** Percutaneous Catheter, 21 CFR 870.1250  
Catheter, Percutaneous, Neurovasculature, 21 CFR 870.1250

**Regulatory Class:** Class II

**Product Codes:** DQY, QJP

**Review Panel:** Cardiovascular, Neurology

### 3) Legally Marketed Predicate Device and Reference Device

**Predicate Device:** K151667 AXS Catalyst Distal Access Catheter  
**Reference Device:** K223545 APRO 70 Catheter and Alembic Aspiration Tubing

### 4) Device Description

The APRO 55 Catheter is a single-lumen, braid and coil reinforced catheter. The APRO 55 Catheter is designed to facilitate the insertion and guidance of interventional devices into peripheral and neuro vasculature. Using standard catheterization techniques under fluoroscopic guidance, the APRO 55 Catheter is introduced through a guide catheter or guide sheath and over a guidewire into the target vasculature. The distal segment of the catheter shaft has a hydrophilic coating to aid navigation through the vasculature. A radiopaque marker is located at the distal tip of the catheter for visualization under fluoroscopy. The APRO 55 Catheter is provided with an Introducer Sheath.

5) **Indications for Use**

The APRO 55 Catheter is indicated for use in facilitating the insertion and guidance of appropriately sized interventional devices into a selected blood vessel in the peripheral and neurovascular systems. The APRO 55 Catheter is also indicated for use as a conduit for retrieval devices.

6) **Technological Characteristics Comparison**

Alembic has demonstrated that the APRO 55 Catheter is substantially equivalent to the predicate device based on the similarity in materials, similarity in design concept, and the same fundamental operating principles. A comparison of the APRO 55 Catheter with the predicate device and reference device is summarized in **Table 1** below.

**Table 1 – APRO 55 Catheter Comparison with the Predicate Device and Reference Device**

<b>Category</b>	<b><u>Subject Device</u> APRO 55 Catheter</b>	<b><u>Predicate Device</u> AXS Catalyst Distal Access Catheter</b>	<b><u>Reference Device</u> APRO 70 Catheter and Alembic Aspiration Tubing</b>
510(k) Number	K234115	K151667	K223545
Regulatory Class	Class II, 21 CFR 870.1250, DQY, QJP	Class II, 21 CFR 870.1250, DQY	Class II, 21 CFR 870.1250, NRY
Indications for Use	The APRO 55 Catheter is indicated for use in facilitating the insertion and guidance of appropriately sized interventional devices into a selected blood vessel in the peripheral and neurovascular systems. The APRO 55 Catheter is also indicated for use as a conduit for retrieval devices.	The AXS Catalyst Distal Access Catheter is indicated for use in facilitating the insertion and guidance of appropriately sized interventional devices into a selected blood vessel in the peripheral and neurovascular systems. The AXS Catalyst Distal Access Catheter is also indicated for use as a conduit for retrieval devices.	The APRO 70 Catheter with an aspiration pump and the Alembic Aspiration Tubing is intended for use in the revascularization of patients with acute ischemic stroke secondary to intracranial large vessel occlusive disease (within the internal carotid, middle cerebral – M1 and M2 segments, basilar, and vertebral arteries) within 8 hours of symptom onset. Patients who are ineligible for intravenous tissue plasminogen activator (IV t-PA) or who fail IV t-PA therapy are candidates for treatment. The Alembic Aspiration Tubing is intended to connect the APRO 70 Catheter to the aspiration pump.
<b>Materials</b>			
Hub	Nylon	Nylon	Polycarbonate
Adhesive	Cyanoacrylate	None	Cyanoacrylate
Strain Relief	Thermoplastic elastomer	Thermoplastic rubber	Thermoplastic elastomer
Liner	Polytetrafluoroethylene (PTFE)/Tecoflex composite	PTFE	PTFE/Tecoflex composite
Shaft Coil or Braid	304V stainless steel braid 304V stainless steel coil	Stainless steel with nitinol wire and polymer fiber	304V stainless steel braid 304V stainless steel coil
Extrusions	Thermoplastic polyurethanes, thermoplastic elastomer	Pebax, Nylon	Thermoplastic polyurethanes, thermoplastic elastomer
Marker Band	Platinum/ iridium	Platinum/ iridium	Platinum/ iridium
Adhesive	None	Cyanoacrylate	None
Coating	Hydrophilic coating, identical	Hydrophilic coating	Hydrophilic coating

Category	<u>Subject Device</u> APRO 55 Catheter	<u>Predicate Device</u> AXS Catalyst Distal Access Catheter	<u>Reference Device</u> APRO 70 Catheter and Alembic Aspiration Tubing
	to reference device		
<b>Dimensions</b>			
Proximal Outer Diameter	0.066 inch	5.6 F (0.073 inch) 6.0 F (0.078 inch)	0.083 inch
Distal Outer Diameter	0.066 inch	5.3 F (0.069 inch) 5.4 F (0.070 inch)	0.083 inch
Inner Diameter	0.055 inch	0.058 inch, 0.060 inch	0.070 inch
Effective Lengths	125, 137 cm	115, 132 cm	125, 132, 135 cm
Coated Length	90, 102 cm	65, 82 cm	90, 97, 100 cm
Tip Shape	Straight	Straight	Straight
<b>Accessories</b>			
Introducer Sheath	Yes	Yes	Yes
Rotating Hemostasis Valve	None	Yes	None
<b>Packaging Materials</b>			
Pouch	Nylon/polyethylene/ Tyvek	Nylon/polyethylene/ Tyvek	Nylon/polyethylene/ Tyvek
Packaging Tube	Polyethylene	Polyethylene	Polyethylene
Packaging Card	Yes	Yes	Yes
Shelf Carton	Solid bleached sulfate paperboard	Solid bleached sulfate paperboard	Solid bleached sulfate paperboard
<b>Other</b>			
Sterilization	Ethylene oxide	Ethylene oxide	Ethylene oxide
Shelf Life	6 months	2 years	6 months
Use Conditions	Sterile, single use, disposable	Sterile, single use, disposable	Sterile, single use, disposable

## 7) Performance Data

Alembic performed non-clinical bench, sterility, shelf-life, and biocompatibility testing. The results demonstrate substantial equivalence of the APRO 55 Catheter to the legally marketed predicate device.

### A. Design Verification Testing – Non-Clinical Bench

Performance testing was conducted to support the APRO 55 Catheter submission. The results of the design verification and validation testing performed confirm that the APRO 55 Catheter conforms to the predefined specifications and met test acceptance criteria. A summary of the testing is shown in **Table 2**.

**Table 2 – Summary of Non-Clinical Bench Test Results**

Test	Acceptance Criteria	Conclusion
Visual and Dimensional Characteristics	Catheter meets the visual and dimensional specifications.	The APRO 55 Catheter met the acceptance criteria.
Particulate	Catheter meets the acceptance criteria. Subject device was evaluated with a predicate device under the same test conditions.	The APRO 55 Catheter met the acceptance criteria.
Kink Resistance	Catheter shaft shall not kink at clinically relevant radii.	The APRO 55 Catheter met the acceptance criteria.
Hub Air Leakage	Catheter does not leak air into hub	The APRO 55 Catheter met the

Test	Acceptance Criteria	Conclusion
	assembly with methods specified in ISO 10555-1, Annex D.	acceptance criteria.
Hub Compatibility	Catheter meets the requirements specified in ISO 80369-7.	The APRO 55 Catheter met the acceptance criteria.
Torque Strength	Catheter must withstand the minimum required number of rotations without breakage and without kinking compared to legally marketed devices.	The APRO 55 Catheter met the acceptance criteria.
Dynamic Burst Pressure	No damage to catheter with dynamic pressure.	The APRO 55 Catheter met the acceptance criteria.
Liquid Leakage	Catheter must withstand pressure with methods specified in ISO 10555-1, Annex C.	The APRO 55 Catheter met the acceptance criteria.
Static Burst	Catheter must withstand pressures anticipated for clinical use.	The APRO 55 Catheter met the acceptance criteria.
Hub and Shaft Tensile Strength	Catheter hub and shaft must meet tensile strength specification.	The APRO 55 Catheter met the acceptance criteria.
Tip Tensile Strength	Catheter tip must meet tip tensile strength specification.	The APRO 55 Catheter met the acceptance criteria.
Tip Buckling Force	Catheter tip buckling force must be acceptable. Forces were compared to a predicate.	The APRO 55 Catheter met the acceptance criteria.
Delivery and Retrieval Force	Catheter delivery and retrieval force must be acceptable. Forces were compared to a predicate.	The APRO 55 Catheter met the acceptance criteria.
Simulated Use	When used per the Instructions for Use with accessory devices in an anatomical neurovascular model, the Catheter must meet functionality specifications.	The APRO 55 Catheter met the acceptance criteria.

### **B. Design Verification Testing – Animal**

Substantial equivalence was established based on non-clinical bench performance data. Animal testing data were not deemed necessary.

### **C. Sterilization and Shelf-Life**

The APRO 55 Catheter is sterilized using an ethylene oxide sterilization cycle that was verified to a sterility assurance level of  $1 \times 10^{-6}$  in accordance with ISO 11135.

Aging studies for the APRO 55 Catheter have established that the subject device and packaging remain functional for the labeled expiration date. Aging studies for packaging integrity, seal strength, and the device functionality were performed and met the acceptance criteria.

### **D. Biocompatibility**

Biocompatibility testing has been completed for the APRO 55 Catheter in accordance with ISO 10993-1 and the device is deemed non-toxic (local or systemic), non-sensitizing, not locally irritating or otherwise harmful. Test results obtained were acceptable for the intended

use as shown in **Table 3**. The thrombogenicity evaluation of the APRO 55 Catheter was supplemented by the prior *in vivo* thrombogenicity evaluation of the reference device, APRO 70 Catheter, in a good laboratory practice animal study.

**Table 3 – Biocompatibility Test Results**

<b>Test</b>	<b>Results</b>	<b>Conclusions</b>
Sensitization (Guinea Pig Maximization)	The APRO 55 Catheter did not elicit a sensitization response.	Non-sensitizing
Irritation/Intracutaneous Reactivity	The APRO 55 Catheter demonstrated no evidence of irritation.	Non-irritant
Cytotoxicity (MEM Elution, L929 cells)	The APRO 55 Catheter did not elicit a cytotoxic response at 24 hours and 48 hours.	Non-cytotoxic
Hemolysis - Indirect	There were no significant differences between the test article extract and negative control article results.	Non-hemolytic
Hemolysis - Direct	There were no differences between the hemolytic index of the test article and the negative control.	Non-hemolytic
Partial Thromboplastin Time (PTT)	The average clotting time of the test article was greater than vehicle control and negative control.	Acceptable clotting times
SC5b9 Complement Activation	The Sc5b9 concentration of the test article was statistically less than the positive control and was not statistically higher than the negative control.	Acceptable
Acute Systemic Toxicity	No weight loss, mortality, or evidence of systemic toxicity from the extract exposure to the mice.	Non-toxic
Material-Mediated Pyrogenicity	All individual rabbits for both the test article and negative control showed a total rise in temperature of < 0.5 °C and were determined to be nonpyrogenic.	Non-pyrogenic

### **E. Clinical Testing**

Substantial equivalence was established based on non-clinical performance data. Human clinical data were not deemed necessary.

### **8) Conclusion**

Based on the comparison of the technological characteristics and the non-clinical testing, the subject device is found to be substantially equivalent to the predicate device. The differences in technological characteristics do not raise new questions of safety and effectiveness. Testing was conducted to demonstrate that the subject device meets the specifications and performs as intended.