



February 9, 2024

Bodycad Laboratories Inc.
Nadine Adia
Regulatory Affairs and Quality Director
2035 rue du Haut-Bord
Quebec, G1N 4R7
Canada

Re: K240066
Trade/Device Name: Fine Osteotomy™
Regulation Number: 21 CFR 888.3030
Regulation Name: Single/multiple component metallic bone fixation appliances and accessories
Regulatory Class: Class II
Product Code: PBF, HWC, HRS
Dated: November 21, 2023
Received: January 9, 2024

Dear Nadine Adia:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device"

(<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shumaya Ali -S

Shumaya Ali, M.P.H.

Assistant Director

DHT6C: Division of Restorative, Repair
and Trauma Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Indications for Use

Submission Number (if known)

K240066

Device Name

Fine Osteotomy™

Indications for Use (Describe)

Fine Osteotomy™ is a system intended for open- and closed-wedge osteotomies, treatment of bone and joint deformities, fixation of fractures and malalignment caused by injury or disease, such as osteoarthritis, of the distal femur and proximal tibia.

Fine Osteotomy™ disposable instrumentation is intended to assist in pre-operative planning and/or in guiding the marking of bone and/or guiding of surgical instruments in non-acute, non-joint replacing osteotomies around the knee.

Fine Osteotomy™ is a patient-specific device.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

Prepared on: 2024-01-09

Contact Details

[21 CFR 807.92\(a\)\(1\)](#)

Applicant Name Bodycad Laboratories Inc.

Applicant Address 2035 rue du Haut-Bord Quebec G1N 4R7 Canada

Applicant Contact Telephone 4185271388

Applicant Contact Mrs. Nadine ADIA

Applicant Contact Email nadia@bodycad.com

Device Name

[21 CFR 807.92\(a\)\(2\)](#)

Device Trade Name Fine Osteotomy™

Common Name Single/multiple component metallic bone fixation appliances and accessories

Classification Name Orthopaedic Surgical Planning And Instrument Guides

Regulation Number 21 CFR 888.3030

Product Code(s) PBF, HRS, HWC

Legally Marketed Predicate Devices

[21 CFR 807.92\(a\)\(3\)](#)

Predicate #	Predicate Trade Name (Primary Predicate is listed first)	Product Code
K231314	Fine Osteotomy™	PBF

Device Description Summary

[21 CFR 807.92\(a\)\(4\)](#)

Fine Osteotomy™ is a system for planning and performing osteotomies of the distal femur and proximal tibia, and for stabilizing the bone with bone screws and a patient-specific bone plate that fits the patient's anatomy. Fine Osteotomy™ consists of patient-specific surgical planning and instrument guides designed from images of the patient's bones, a patient-specific bone plate designed from the patient's images, compression and/or locking bone screws, and class 1 reusable manual instruments. The bone plate is a patient-specific, single-use implant; the surgical planning and instrument guides are patient-specific, single-use, and discarded after surgery. Fine Osteotomy™ is offered in three configurations: 1) as a system of patient specific implants and single use instruments for performing osteotomies and implanting hardware to stabilize the resection, 2) as patient specific single use instruments alone for performing osteotomies, and 3) as a patient specific bone plate and screws for stabilizing a bone resection or fracture.

When used as a system, Fine Osteotomy™ enables the surgeon to perform an osteotomy and stabilize the bone around the knee that matches the pre-surgical plan using the patient-specific cutting guides and bone plate. When the planning guides and resection instruments are used alone, Fine Osteotomy™ enables the surgeon to perform an osteotomy around the knee that matches the pre-surgical plan using the patient-specific cutting guides designed from the patient's CT images. When the bone plate and screws are used alone, Fine Osteotomy™ enables the surgeon to stabilize fractured or resected bone per the pre-surgical plan using the patient's CT images in design of the Bodycad plate and use of the bone models intra operatively to guide placement of the implants and alignment of bone. The Fine Osteotomy System is provided clean, not sterile to the user.

Materials: Wrought Titanium-6Aluminum-4Vanadium ELI Alloy (Ti6Al4V ELI; ASTM F136-13) for the bone plates and screws, additively manufactured Nylon-12 for patient specific, single use resection guides and models.

Intended Use/Indications for Use

[21 CFR 807.92\(a\)\(5\)](#)

Fine Osteotomy™ is a system intended for open- and closed-wedge osteotomies, treatment of bone and joint deformities, fixation of fractures and malalignment caused by injury or disease, such as osteoarthritis, of the distal femur and proximal tibia.

Fine Osteotomy™ disposable instrumentation is intended to assist in pre-operative planning and/or in guiding the marking of bone and/or guiding of surgical instruments in non-acute, non-joint replacing osteotomies around the knee.

Fine Osteotomy™ is a patient-specific device.

Indications for Use Comparison

[21 CFR 807.92\(a\)\(5\)](#)

(Identical)

Fine Osteotomy™ is a system intended for open- and closed-wedge osteotomies, treatment of bone and joint deformities, fixation of fractures and malalignment caused by injury or disease, such as osteoarthritis, of the distal femur and proximal tibia.

Fine Osteotomy™ disposable instrumentation is intended to assist in pre-operative planning and/or in guiding the marking of bone and/or guiding of surgical instruments in non-acute, non-joint replacing osteotomies around the knee.

Fine Osteotomy™ is a patient-specific device.

Technological Comparison

[21 CFR 807.92\(a\)\(6\)](#)

The subject device, Fine Osteotomy™, has the same intended use, has the same indications for use, is manufactured with the same materials and processes, and has the same technological features as the primary predicate device which is Fine Osteotomy cleared in K231314. Except for updates and additions to single use instrumentation and shelf-life, the technological characteristics of the subject and primary predicate devices are identical. Comparison of the technological characteristics and indicated use of the subject and primary predicate devices demonstrate that the subject device is substantially equivalent to the primary predicate device, both of which are Bodycad's Fine Osteotomy.

Non-Clinical and/or Clinical Tests Summary & Conclusions

[21 CFR 807.92\(b\)](#)

Verification and validation (V&V) activities included the following:

- Mechanical Engineering Analysis of updated and new single use instruments demonstrated no new risks and no new worst case scenarios, and
- Modification of Shelf-Life Referential for Fine Osteotomy Systems