



August 28, 2024

Edwards Lifesciences, LLC
Carmen Chen
Manager, Regulatory Affairs
One Edwards Way
Irvine, California 92614

Re: K240078

Trade/Device Name: Fogarty Thru-Lumen Embolectomy Catheter
Regulation Number: 21 CFR 870.5150
Regulation Name: Embolectomy catheter
Regulatory Class: Class II
Product Code: DXE, KRA
Dated: July 25, 2024
Received: July 25, 2024

Dear Carmen Chen:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device"

(<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

**Ariel G. Ash-
shakoor -S**

Digitally signed by Ariel

G. Ash-shakoor -S

Date: 2024.08.28

13:34:33 -04'00'

For

Gregory O'Connell

Assistant Director

DHT2C: Division of Coronary and
Peripheral Intervention Devices

OHT2: Office of Cardiovascular Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

Submission Number (if known)

K240078

Device Name

Fogarty Thru-Lumen Embolectomy Catheter

Indications for Use (Describe)

The Fogarty thru-lumen embolectomy catheter is indicated for adults requiring the removal of fresh, soft emboli and thrombi from vessels in the arteries of the peripheral vasculature. Additionally, it may also be used for temporary occlusion of blood vessels, infusion of fluids, and blood sampling in general surgical procedures.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

K240078 - 510(k) Summary

Fogarty Thru-Lumen Embolectomy Catheter		
510(k) Submitter	Edwards Lifesciences, LLC One Edwards Way, Irvine, CA 92614	
Contact Person	<u>Primary Contact</u> Carmen G. Chen Manager, Regulatory Affairs Edwards Lifesciences One Edwards Way Irvine, CA 92614 Telephone: (949) 250 - 5469 Email: Carmen_Chen@edwards.com	<u>Secondary Contact</u> Karen O’Leary Sr. Director, Regulatory Affairs Edwards Lifesciences One Edwards Way Irvine, CA 92614 Telephone: (949) 610-9179 Email: Karen_OLeary@edwards.com
Date Prepared	August 26, 2024	
Trade Name	Fogarty Thru-Lumen Embolectomy Catheter	
Regulation Number / Name	21 CFR 870.5150 / Embolectomy Catheter	
Product Code	DXE, KRA	
Regulation Class	Class II	
Predicate Device	Fogarty Thru-Lumen Embolectomy Catheter cleared under K892410 on July 20, 1989	
Device Description	The Fogarty Thru-Lumen Embolectomy Catheter is a double lumen catheter with a latex balloon at the distal end. One lumen is used for balloon inflation and is accessed through the gate valve at the proximal end. The thru-lumen is used for infusion of fluids, sampling of blood or guidewire passage through the distal tip. A removable stainless steel stylet(s) is provided.	
Indications for Use	The Fogarty thru-lumen embolectomy catheter is indicated for adults requiring the removal of fresh, soft emboli and thrombi from vessels in the arteries of the peripheral vasculature. Additionally, it may also be used for temporary occlusion of blood vessels, infusion of fluids, and blood sampling in general surgical procedures.	
Comparative Analysis	The subject device is identical to the predicate device in terms of the intended use, indications for use, and technological characteristics (including design, material, chemical composition, principle of operation).	
Device Testing	Biocompatibility testing was performed in accordance with ISO 10993-1: 2018 – Biological evaluation of medical devices – Part 1: Evaluation and testing within a risk management process, and FDA guidance document, Use of International Standard ISO 10993-1, “Biological evaluation of medical devices – Part 1: Evaluation and testing within a risk management process”, issued on September 4, 2020.	

K240078 - 510(k) Summary

Fogarty Thru-Lumen Embolectomy Catheter	
	Bench testing was performed in accordance with Edwards' current design requirements. In addition, shelf-life, packaging, and sterilization validations have been performed to existing specifications. All testing met the existing predetermined acceptance criteria.
Conclusion	Based on the performance testing and the technological characteristics, the Fogarty Thru-Lumen Embolectomy Catheter meets the established performance criteria and is substantially equivalent to the predicate device Fogarty Thru-Lumen Embolectomy Catheters.