



September 3, 2024

Smith & Nephew
Mandy Coe
Regulatory Affairs Specialist
1450 Brooks Road
Memphis, Tennessee 38116

Re: K240487

Trade/Device Name: EVOS Patella Plates

Regulation Number: 21 CFR 888.3030

Regulation Name: Single/Multiple Component Metallic Bone Fixation Appliances And Accessories

Regulatory Class: Class II

Product Code: HRS, HWC

Dated: June 17, 2024

Received: June 21, 2024

Dear Mandy Coe:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device"

(<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

All medical devices, including Class I and unclassified devices and combination product device constituent parts are required to be in compliance with the final Unique Device Identification System rule ("UDI Rule"). The UDI Rule requires, among other things, that a device bear a unique device identifier (UDI) on its label and package (21 CFR 801.20(a)) unless an exception or alternative applies (21 CFR 801.20(b)) and that the dates on the device label be formatted in accordance with 21 CFR 801.18. The UDI Rule (21 CFR 830.300(a) and 830.320(b)) also requires that certain information be submitted to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). For additional information on these requirements, please see the UDI System webpage at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-system-udi-system>.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See

the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shumaya Ali -S

Shumaya Ali, M.P.H.

Assistant Director

DHT6C: Division of Restorative, Repair
and Trauma Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

Submission Number (if known)

K240487

Device Name

EVOS Mini-Fragment System

Indications for Use (Describe)

The EVOS Mini Plating System, excluding EVOS Patella Plates, is indicated for patients 12 – 21 and those over 21 years of age, as well as patients with osteopenic bone. The Smith & Nephew EVOS Mini Plating System is indicated for fracture fixation or reduction of small bones and small bone fragments. This system is also indicated for non-load bearing stabilization and reduction of bone fragments in long bones.

The EVOS Patella Plates are indicated for the fixation and stabilization of patellar fractures for patients over 21 years of age.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) #: K240487

510(k) Summary

Prepared on: 2024-09-03

Contact Details

[21 CFR 807.92\(a\)\(1\)](#)

Applicant Name	Smith & Nephew
Applicant Address	1450 Brooks Road Memphis TN 38116 United States
Applicant Contact Telephone	901-949-3344
Applicant Contact	Mrs. Mandy Coe
Applicant Contact Email	mandy.coe@smith-nephew.com

Device Name

[21 CFR 807.92\(a\)\(2\)](#)

Device Trade Name	EVOS Mini-Fragment System
Common Name	Single/multiple component metallic bone fixation appliances and accessories
Classification Name	Plate, Fixation, Bone
Regulation Number	888.3030
Product Code(s)	HRS, HWC

Legally Marketed Predicate Devices

[21 CFR 807.92\(a\)\(3\)](#)

Predicate #	Predicate Trade Name (Primary Predicate is listed first)	Product Code
K140814	EVOS Mini-Fragment Plating System	HRS
K210408	Depuy Synthes Variable Angle Locking Patella Plating System	HRS
K203834	Arthrex Patella Sutureplates	HRS

Device Description Summary

[21 CFR 807.92\(a\)\(4\)](#)

The subject EVOS Patella Plates are an extension to the EVOS Mini Fragment system previously cleared under K140814 (S.E. 5/7/2014). The subject bone plates consist of a Staggered Patella Plate, a Split Patella Plate, and a Small and Large Mesh Patella Plate. The subject plates are compatible with the previously cleared EVOS 2.7mm locking and non-locking screws, and partially and fully threaded 4.0mm osteopenia screws (K140814). The proposed plates feature a variable angle locking screw hole feature, are manufactured from implant-grade 316L Stainless Steel material and will be available in a sterile packaged condition.

Intended Use/Indications for Use

[21 CFR 807.92\(a\)\(5\)](#)

The EVOS Mini Plating System, excluding EVOS Patella Plates, is indicated for patients 12 – 21 and those over 21 years of age, as well as patients with osteopenic bone. The Smith & Nephew EVOS Mini Plating System is indicated for fracture fixation or reduction of small bones and small bone fragments. This system is also indicated for non-load bearing stabilization and reduction of bone fragments in long bones.

The EVOS Patella Plates are indicated for the fixation and stabilization of patellar fractures for patients over 21 years of age.

Indications for Use Comparison

[21 CFR 807.92\(a\)\(5\)](#)

The indications for use of the subject devices are substantially equivalent to the indications for the cleared predicates.

Technological Comparison

[21 CFR 807.92\(a\)\(6\)](#)

Device comparisons described in this premarket notification demonstrated that the proposed bone plates are substantially equivalent to legally marketed predicates with respect to intended use, indications, and performance characteristics. Both the subject and predicate devices have same operating principle, material, and technological characteristics such as variable angle locking screw holes.

Non-Clinical and/or Clinical Tests Summary & Conclusions

[21 CFR 807.92\(b\)](#)

The following nonclinical tests were used to determine substantial equivalence:

- Finite element analysis (FEA) and predicate clinical data were leveraged to support device performance.
- Four Point Bend Fatigue performance evaluation was done for the subject plates.
- Magnetic resonance imaging (MRI) compatibility evaluation was done per ASTM F2052, ASTM F2213, ASTM F2182, and ASTM F2119.
- Packaging verification testing was conducted on the subject devices per ASTM F2096 and ASTM F88.

The testing detailed in this premarket notification validate that the subject EVOS Patella Plates are substantially equivalent in performance to the predicate EVOS Mini-Fragment system (K140814, S.E. 57/2014).