



Avicenna.AI  
% John J. Smith, M.D., J.D.  
Partner  
Hogan Lovells, US LLP  
Columbia Square  
555 Thirteenth Street, NW  
Washington, District of Columbia 20004

Re: K240942

September 12, 2024

Trade/Device Name: CINA-CSpine  
Regulation Number: 21 CFR 892.2080  
Regulation Name: Radiological Computer Aided Triage And Notification Software  
Regulatory Class: Class II  
Product Code: QAS  
Dated: April 5, 2024  
Received: August 16, 2024

Dear John J. Smith, M.D., J.D.:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

All medical devices, including Class I and unclassified devices and combination product device constituent parts are required to be in compliance with the final Unique Device Identification System rule ("UDI Rule"). The UDI Rule requires, among other things, that a device bear a unique device identifier (UDI) on its label and package (21 CFR 801.20(a)) unless an exception or alternative applies (21 CFR 801.20(b)) and that the dates on the device label be formatted in accordance with 21 CFR 801.18. The UDI Rule (21 CFR 830.300(a) and 830.320(b)) also requires that certain information be submitted to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). For additional information on these requirements, please see the UDI System webpage at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-system-udi-system>.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See

the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,



Jessica Lamb  
Assistant Director  
Imaging Software Team  
DHT8B: Division of Radiologic Imaging  
Devices and Electronic Products  
OHT8: Office of Radiological Health  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K240942

Device Name  
CINA-CSpine

### Indications for Use (Describe)

CINA-CSpine is a radiological computer aided triage and notification software indicated for use in the analysis of cervical spine CT images.

The device is intended to assist hospital networks and appropriately trained physician specialists by flagging and communication of suspected positive findings compatible with acute cervical spine fractures including non-displaced fracture lines and/or displaced fracture fragments.

CINA-CSpine uses an artificial intelligence algorithm to analyze images and highlight cases with detected findings on a standalone application in parallel to the ongoing standard of care image interpretation. The device is not designed to detect vertebral compression fractures.

The user is presented with notifications for cases with suspected findings. Notifications include compressed preview images that are meant for informational purposes only, and are not intended for diagnostic use beyond notification. The device does not alter the original medical image, and it is not intended to be used as a diagnostic device.

The results of CINA-CSpine are intended to be used in conjunction with other patient information and based on professional judgment to assist with triage/prioritization of medical images.

Notified clinicians are ultimately responsible for reviewing full images per the standard of care.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**510(K) SUMMARY**  
**AVICENNA.AI's CINA-CSPINE**

**Submitter**Applicant:

AVICENNA.AI  
Espace Mistral B,  
375 avenue du Mistral,  
13600 LA CIOTAT France

Contact Person:

Stephane Berger  
Regulatory Manager  
Phone: +33 6 12 12 28 13  
E-mail: stephane.berger@avicenna.ai

Date prepared: September 12, 2024

**Device Identification**

Name of Device:	CINA-CSpine
Classification Name:	Radiological computer aided triage and notification software
Regulation No:	21 CFR § 892.2080
Product Code:	QAS
Regulatory Class:	Class II
Classification Panel:	Radiology

**Predicate Device**

The CINA-CSpine device is substantially equivalent to the following predicate device with regard to indications for use, performance, and technological characteristics:

510(k):	K190896
Trade Name:	BriefCase for CSF Triage
Manufacturer:	Aidoc Medical, Ltd.
Classification Name:	Radiological computer aided triage and notification software
Regulation No:	21 CFR § 892.2080
Product Code:	QAS
Regulatory Class:	Class II

## Device Description

CINA-CSpine is a radiological computer-assisted triage and notification software device.

CINA-CSpine runs on a standard “off the shelf” server/workstation and consists of CSpine Image Processing Application, which can be integrated, deployed and used with the CINA Platform (cleared under K200855) or other medical image communications devices. CINA-CSpine receives cervical spine CT scans identified by the CINA Platform or other medical image communications device, processes them using deep learning algorithms involving execution of multiple computational steps to identify the suspected positive findings compatible with acute cervical spine fractures and generates results files to be transferred by CINA Platform or a similar medical image communications device for output to a PACS system or workstation for worklist prioritization.

To identify the suspected presence of cervical fractures, the device uses a deep learning model trained end-to-end on 1,338 cases acquired from US and France, representing a distribution of fracture presentations, locations and acquisition protocols, including multiple scanner models from Siemens, Philips, GE and Canon/Toshiba. Additional deep learning models are used to locate the individual vertebrae to exclude images that do not conform to the expected field of view.

DICOM images are received, recorded and filtered before processing. The series are processed chronologically by running algorithms on each series to detect suspected positive findings of a cervical spine fracture, then active notifications on the flagged series are sent to the worklist application.

The Worklist Application displays the active notification of new studies with suspected findings when they come in. All the cervical spine CT studies which include at least 5 visible cervical vertebrae received by CINA-CSpine device are displayed in the worklist and those on which the algorithms have detected a suspected finding are marked with an icon (i.e., active notification). In addition, a compressed, grayscale, unannotated image that is marked “not for diagnostic use” is displayed as a preview function. This compressed preview is meant for informational purposes only, does not contain any marking of the findings, and is not intended for diagnostic use beyond notification.

Presenting the trained physician specialist with notification facilitates earlier triage by allowing image prioritization in the PACS. Thus, the suspect case receives attention earlier than would have been the case in the standard of care practice alone.

The CINA platform is an example of medical image communications platform for integrating and deploying the CINA-CSpine image processing application. The medical image communications device (i.e., the technical platform) provides the necessary requirements for interoperability based on the standardized DICOM protocol and services to communicate with existing systems in the hospital radiology department such as CT modalities or other DICOM nodes (DICOM router or PACS for example). It is responsible for transferring, storing, converting formats, notifying of suspected findings and displaying medical device data such as radiological data. The medical image communications server includes the Worklist client application, which receives notifications from the CINA-CSpine Image Processing application.

## **Intended Use / Indications for Use**

CINA-CSpine is a radiological computer aided triage and notification software indicated for use in the analysis of cervical spine CT images.

The device is intended to assist hospital networks and appropriately trained physician specialists by flagging and communication of suspected positive findings compatible with acute cervical spine fractures including non-displaced fracture lines and/or displaced fracture fragments.

CINA-CSpine uses an artificial intelligence algorithm to analyze images and highlight cases with detected findings on a standalone application in parallel to the ongoing standard of care image interpretation. The device is not designed to detect vertebral compression fractures.

The user is presented with notifications for cases with suspected findings. Notifications include compressed preview images that are meant for informational purposes only, and are not intended for diagnostic use beyond notification. The device does not alter the original medical image, and it is not intended to be used as a diagnostic device.

The results of CINA-CSpine are intended to be used in conjunction with other patient information and based on professional judgment to assist with triage/prioritization of medical images. Notified clinicians are ultimately responsible for reviewing full images per the standard of care.

## **Summary of Performance Data**

The following performance data were provided in support of the substantial equivalence determination.

### *Software Verification and Validation Testing*

CINA-CSpine complies with DICOM (Digital Imaging and Communications in Medicine) – Developed by the American College of Radiology and the National Electrical Manufacturers Association. NEMA PS 3.1 - 3.20.

Avicenna.AI conducted extensive performance validation testing and software verification and validation testing of the CINA-CSpine device as standalone software. CINA-CSpine is tested against its user needs and intended use by the successful execution of planned software verification and validation testing included in this submission.

Software performance, validation and verification testing demonstrated that the CINA-CSPINE met all design requirements and specifications associated with the intended use of the software.

### *Standalone Performance Testing*

Avicenna.AI conducted a retrospective, multicenter, multinational and blinded study with the CINA-CSpine application with the primary endpoint to evaluate the software's performance (Sensitivity and Specificity) in detecting cervical spine (CSpine) fractures on non-contrast CT (NCCT) scans images, in 328 clinical anonymized cases. The dataset included 59.5% Male and 36.3% Female. Mean  $\pm$  SD age of patients included in the study was  $55.3 \pm 5.1$  yo (range: 18 - 101 yo).

Multiple subgroups of interest were considered in the analysis. In fact, the validation data was acquired from 3 different clinical sources, with one representing an U.S teleradiology company which has access to a database of medical images from various U.S. hospitals distributed throughout the U.S. territories, to account for race/ethnicity in the intended U.S. patient population. Additional scanner parameters considered were slice thickness, number of detector rows, and kVp ranges. Detailed subgroup analysis was reported in the labeling.

The data was provided from multiple US (60.4%) and OUS (39.6%) clinical sites. The independence of the standalone validation dataset from the training data was ensured using data from independent sites and different time periods. There were 155 (47.3%) positive cases (CT with CSpine) and 173 (52.7%) negative cases included in the analysis. The data was acquired primarily by 4 different scanner makers (GE-31.1%, Philips-21.6%, Siemens-28.7%, and Canon-18.3%) and 36 different scanner models.

Device Sensitivity [95% CI] and Specificity [95% CI] were computed against the ground truth established by consensus of three US-board-certified expert radiologists.

As a primary endpoint, the global Sensitivity and Specificity were found to be 90.3% [95%CI: 84.5% - 94.5%] and 91.9% [95%CI: 86.8% - 95.5%], respectively. These findings achieved the 80% performance goal of the study and are similar to the results reported for the predicate BriefCase device (Aidoc Medical) which demonstrated: Sensitivity and Specificity of 91.7% [95%CI: 82.7% - 96.9%] and 88.6% [95%CI: 81.2% - 93.8%], respectively.

Additionally, the CINA-CSpine prioritization and triage effectiveness were evaluated by the standalone per-case processing time of the device (time-to-notification), which corresponds to the time between the end of the DICOM reception (made available for CINA-CSpine image processing) and the end of processing (positive or negative identification). Finally, for better comparison with the predicate device, the statistics for only True Positive cases (i.e., identified as positive both by the ground truthers as well as the device) are also provided.

The results are presented in Table below:

**Time-to-Notification for CINA-CSpine Image Processing Application**

Time-to-Notification (minutes)	MEAN ± SD	MEDIAN	95% CI	MIN	MAX
CINA-CSpine All cases (N = 328)	2.9 ± 1.1	2.6	[2.7 - 3.0]	1.0	6.8
CINA-CSpine True Positive cases (N = 140 )	2.8 ± 1.2	2.6	[2.6 - 3.0]	1.0	6.8

The mean [95% CI] time-to-notification for all included cases (n = 328) was estimated to be 2.9 [95% CI: 2.7 - 3.0] minutes for CINA-CSpine.

When taking into account only true positive cases (n = 140), the mean [95% CI] time-to-notification was 2.8 [95%CI: 2.6 - 3.0] minutes for CINA-CSpine and 3.9 [95%CI: 3.8 - 4.1] minutes for BriefCase, the defined predicate device.



The performance testing of the CINA-CSpine device establishes that the subject device is as safe and effective as the predicate device and compatible with the same clinical use, since the performances demonstrated the clinical effectiveness of the subject device and its ability to provide effective prioritization and triage, which is substantially equivalent to that of the predicate device. This establishes that the CINA-CSpine device achieves its intended use and is substantially equivalent to the predicate device.

**Substantial Equivalence**

The subject CINA-CSpine and the predicate BriefCase for CSF Triage device are both intended to aid in prioritization and triage of radiological images of time sensitive findings for patient detection and diagnosis (i.e., cervical spine fracture) based on the analysis of medical images acquired from radiological signal acquisition systems. The CINA device provides the CINA Platform in which the subject CINA-CSpine prioritization and triage application can be integrated, deployed and used. The labeling of the subject and the predicate devices clearly states that the devices are not for diagnostic use. The subject CINA-CSpine device and the predicate BriefCase for CSF Triage are software packages with similar technological characteristics and principles of operation, and incorporate deep learning AI algorithms that process images, and software to send notifications and to display unannotated preview images. In both devices, the labeling instructs the user to further evaluate and diagnose based only on the original images in the local PACS.

The subject CINA-CSpine, the predicate BriefCase for CSF Triage device operate in parallel to the standard of care workflow in the sense that they do not change the original image, do not provide any marking on the output preview, and do not remove images from the standard of care FIFO queue; thus, not disturbing standard interpretation of the images by the attending trained physician specialists. The subject and predicate devices achieve performance of the time-to-notification metric in similar ranges of time, and thus contribute similarly to effective triage and early involvement of the trained physician specialists in evaluating suspected images of cervical spine fracture.

The standalone performance and effectiveness assessment studies demonstrated that the CINA-CSpine device performs as intended and substantially equivalent to the BriefCase for CSF Triage predicate device.

The table below compares the key features of the subject and the predicate devices.

**Comparison of key features between CINA-CSpine and predicate device (Aidoc Medical Ltd)**

	<b>Subject device: CINA-CSpine Software</b>	<b>Predicate device: BriefCase for CSF Triage Software (K190896)</b>
<i>Intended Use / Indications for Use</i>	CINA-CSpine is a radiological computer aided triage and notification software indicated for use in the analysis of cervical spine CT images.	BriefCase is a radiological computer aided triage and notification software indicated for use in the analysis of cervical spine CT images.  The device is intended to assist hospital networks and trained radiologists in workflow triage by flagging and

	<b>Subject device: CINA-CSpine Software</b>	<b>Predicate device: BriefCase for CSF Triage Software (K190896)</b>
	<p>The device is intended to assist hospital networks and appropriately trained physician specialists by flagging and communication of suspected positive findings compatible with acute cervical spine fractures including non-displaced fracture lines and/or displaced fracture fragments.</p> <p>CINA-CSpine uses an artificial intelligence algorithm to analyze images and highlight cases with detected findings on a standalone application in parallel to the ongoing standard of care image interpretation. The device is not designed to detect vertebral compression fractures.</p> <p>The user is presented with notifications for cases with suspected findings. Notifications include compressed preview images that are meant for informational purposes only, and are not intended for diagnostic use beyond notification. The device does not alter the original medical image, and it is not intended to be used as a diagnostic device.</p> <p>The results of CINA-CSpine are intended to be used in conjunction with other patient information and based on professional judgment to assist with triage/prioritization of medical images. Notified clinicians are ultimately responsible for reviewing full images per the standard of care.</p>	<p>communication of suspected positive findings of linear lucencies in the cervical spine bone in patterns compatible with fractures.</p> <p>BriefCase uses an artificial intelligence algorithm to analyse images and highlight cases with detected findings on a standalone desktop application in parallel to the ongoing standard of care image interpretation. The user is presented with notifications for cases with suspected findings.</p> <p>Notifications include compressed preview images that are meant for informational purposes only and not intended for diagnostic use beyond notification. The device does not alter the original medical image and is not intended to be used as a diagnostic device.</p> <p>The results of BriefCase are intended to be used in conjunction with other patient information an based on their professional judgment, to assist with triage/prioritization of medical images. Notified clinicians are responsible for viewing full images per the standard of care.</p>
<i>User population</i>	Trained physician specialist	Radiologist
<i>Anatomical region of interest</i>	Cervical Spine	Cervical Spine

	<b>Subject device: CINA-CSpine Software</b>	<b>Predicate device: BriefCase for CSF Triage Software (K190896)</b>
<i>Data acquisition protocol</i>	Non-contrast cervical spine CT scans	Non-contrast cervical spine CT scans
<i>View DICOM data</i>	DICOM information about the patient, study and current image	DICOM information about the patient, study and current image
<i>Segmentation of region of interest</i>	No; device does not mark, highlight, or direct users' attention to a specific location in the original image	No; device does not mark, highlight, or direct users' attention to a specific location in the original image
<i>Algorithm</i>	Artificial intelligence algorithm with database of images	Artificial intelligence algorithm with database of images
<i>Notification / Prioritization</i>	Yes	Yes
<i>Preview images</i>	Presentation of a compressed, grayscale, unannotated image that is marked "not for diagnostic use".	Presentation of a small, compressed, black and white preview image that is labeled "Not for diagnostic use".
<i>Alteration of original image</i>	No	No
<i>Removal of cases from worklist queue</i>	No. The device operates in parallel with the standard of care, which remains the default option for all cases.	No. The device operates in parallel with the standard of care, which remains the default option for all cases.
<i>Structure</i>	-CSpine image processing application - Compatibility of use with the CINA Platform (worklist and Image Viewer) or other medical image communications device.	- AHS module (orchestrator, image acquisition) - ACS module (image processing). - Aidoc Worklist application for workflow integration (worklist and non-diagnostic basic Image Viewer).

## Conclusion

CINA-CSpine and BriefCase have similar intended use and substantially similar indications, technological characteristics, and principles of operation. The following is a summary of the substantial equivalence comparison:

- The predicate device is legally marketed.
- CINA-CSpine has the same intended use as the predicate device, i.e., AIDoc Medical, Ltd.'s BriefCase, and therefore it may be found substantially equivalent.
- CINA-CSpine and the predicate device have very similar indications for use.

- CINA-CSpine has similar technological characteristics as the predicate and reference devices, i.e., the deployment of an artificial intelligence algorithm with a database of images.
- Standalone performance study demonstrates that the CINA-CSpine and the predicate BriefCase raise the same types of safety and effectiveness questions, namely, accurate detection of findings within the processed study.

Accordingly, the subject CINA-CSpine device is substantially equivalent to the predicate BriefCase device