

MiRus, LLC  
Anuradha Nagulapati  
Senior Regulatory Affairs Engineer  
1755 W. Oak Parkway  
Suite 100  
Marietta, Georgia 30062

May 17, 2024

Re: K241175

Trade/Device Name: MiRus MoRe Lumbar Plating System  
Regulation Number: 21 CFR 888.3060  
Regulation Name: Spinal Intervertebral Body Fixation Orthosis  
Regulatory Class: Class II  
Product Code: KWQ  
Dated: April 26, 2024  
Received: April 26, 2024

Dear Anuradha Nagulapati:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).


Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

**Eileen  
Cadel -  
S**  Digitally signed  
by Eileen Cadel  
-S  
Date:  
2024.05.17  
14:34:53 -04'00'

for

Colin O'Neill, M.B.E.

Assistant Director

DHT6B: Division of Spinal Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

## Indications for Use

Submission Number (if known)

K241175

Device Name

MiRus MoRe Lumbar Plating System

Indications for Use (Describe)

The MiRus MoRe Lumbar Plating System is indicated for use via lateral or anterolateral surgical approach above the bifurcation of the great vessels in the treatment of the thoracic and thoracolumbar (T1-L5) spine or via an anterior approach below the bifurcation of the great vessels in the treatment of lumbar and lumbosacral (L1-S1) spine. The system is intended to provide additional support during fusion in skeletally mature patients in the treatment of the following acute and chronic instabilities or deformities:

- Degenerative Disc Disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies);
- Pseudoarthrosis;
- Spondylolysis;
- Spondylolisthesis;
- Spinal stenosis;
- Tumors;
- Trauma (i.e. Fractures or Dislocation);
- Deformities (i.e. Scoliosis, Kyphosis or Lordosis);
- Failed Previous Fusion

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

K241175

## 510(k) Summary

Prepared on: 2024-05-15

## Contact Details

[21 CFR 807.92\(a\)\(1\)](#)

|                             |                                                               |
|-----------------------------|---------------------------------------------------------------|
| Applicant Name              | MiRus, LLC                                                    |
| Applicant Address           | 1755 W. Oak Parkway Suite 100 Marietta GA 30062 United States |
| Applicant Contact Telephone | 470-428-8684                                                  |
| Applicant Contact           | Ms. Anuradha Nagulapati                                       |
| Applicant Contact Email     | anagulapati@mirusmed.com                                      |

## Device Name

[21 CFR 807.92\(a\)\(2\)](#)

|                     |                                                 |
|---------------------|-------------------------------------------------|
| Device Trade Name   | MiRus MoRe Lumbar Plating System                |
| Common Name         | Spinal intervertebral body fixation orthosis    |
| Classification Name | Appliance, Fixation, Spinal Intervertebral Body |
| Regulation Number   | 888.3060                                        |
| Product Code(s)     | KWQ                                             |

## Legally Marketed Predicate Devices

[21 CFR 807.92\(a\)\(3\)](#)

| Predicate # | Predicate Trade Name (Primary Predicate is listed first) | Product Code |
|-------------|----------------------------------------------------------|--------------|
| K150449     | LITe™ Plate System                                       | KWQ          |
| K220441     | CYGNUS™ MoRe Anterior Cervical Plate System              | KWQ          |

## Device Description Summary

[21 CFR 807.92\(a\)\(4\)](#)

The MiRus MoRe Lumbar Plating System is a spinal intervertebral body fixation orthosis intended to provide structural stability and mechanical support to the lumbar spine following lumbar interbody fusion.

The MiRus MoRe Lumbar Plating System consists of implants manufactured from Molybdenum-47.5 Rhenium Alloy (MoRe) per ASTM F3273, screws and locking cam mechanism manufactured from Titanium-6 Aluminum-4 Vanadium ELI per ASTM F136, and instrumentation manufactured from Stainless Steel per ASTM F899. The MiRus MoRe Lumbar Plating System implants are offered in multiple configurations and different sizes to accommodate various patient anatomical requirements: ANTARES™ Anterior Lumbar Plate and MiRus Lateral Lumbar Plate.

The plates are provided sterile packed and are intended for single use only. The screws must be steam sterilized prior to use.

## Intended Use/Indications for Use

[21 CFR 807.92\(a\)\(5\)](#)

The MiRus MoRe Lumbar Plating System is indicated for use via lateral or anterolateral surgical approach above the bifurcation of the great vessels in the treatment of the thoracic and thoracolumbar (T1-L5) spine or via an anterior approach below the bifurcation of the great vessels in the treatment of lumbar and lumbosacral (L1-S1) spine. The system is intended to provide additional support during

fusion in skeletally mature patients in the treatment of the following acute and chronic instabilities or deformities:

- Degenerative Disc Disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies);
- Pseudoarthrosis;
- Spondylolysis;
- Spondylolisthesis;
- Spinal stenosis;
- Tumors;
- Trauma (i.e. Fractures or Dislocation);
- Deformities (i.e. Scoliosis, Kyphosis or Lordosis);
- Failed Previous Fusion

## Indications for Use Comparison

[21 CFR 807.92\(a\)\(5\)](#)

The indications for the use are the same for the subject device and predicate device.

## Technological Comparison

[21 CFR 807.92\(a\)\(6\)](#)

The MiRus MoRe Lumbar Plating System has the same intended use, indications for use, labeling, and technological characteristics as the predicate system, including the same design features, geometries, sizes, and materials. The predicate device components are manufactured from titanium alloy (ASTM F136). The MiRus MoRe Lumbar Plating System components are manufactured from titanium alloy (ASTM F136) and molybdenum-rhenium alloy (ASTM F3273), which are the same materials as the reference device. Performance data demonstrate that the MiRus MoRe Lumbar Plating System is substantially equivalent to other legally marketed predicate systems.

## Non-Clinical and/or Clinical Tests Summary & Conclusions

[21 CFR 807.92\(b\)](#)

The mechanical performance profile of the MiRus MoRe Lumbar Plating System was assessed through static and dynamic testing in accordance with the following test methods:

- Static compression bending testing (ASTM F1717-21)
- Static torsion testing (ASTM F1717-21)
- Dynamic compression bending testing (ASTM F1717-21)

The MiRus MoRe Lumbar Plating System was compared to the performance criteria disclosed in a recent FDA Guidance document "Spinal Plating Systems - Performance Criteria for Safety and Performance Based Pathway: Guidance for Industry and Food and Drug Administration Staff":

- Guidance for Industry and FDA Staff – Spinal Plating Systems – Performance Criteria for Safety and Performance Based Pathway, Issued on December 11, 2020

The guidance document presents acceptable values for performance testing per ASTM F1717. A comparison of the mechanical performance testing of the MiRus MoRe Lumbar Plating System to this FDA Guidance demonstrates that the MiRus MoRe Lumbar Plating System is substantially equivalent to other legally marketed systems.