



July 17, 2024

Nexus Spine, LLC
% Christine Scifert
Partner
MRC Global
9085 East Mineral Circle
Suite 110
Centennial, Colorado 80112

Re: K241467
Trade/Device Name: Stable-C Interbody System
Regulation Number: 21 CFR 888.3080
Regulation Name: Intervertebral Body Fusion Device
Regulatory Class: Class II
Product Code: OVE
Dated: May 20, 2024
Received: May 23, 2024

Dear Christine Scifert:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Brent Showalter -S

Brent Showalter, Ph.D.

Assistant Director

DHT6B: Division of Spinal Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

Submission Number (if known)

K241467

Device Name

Stable-C Interbody System

Indications for Use (Describe)

The Nexus Spine Stable-C Interbody System is an anterior cervical interbody fusion system indicated for use in skeletally mature patients with cervical disc disease (DDD) at one level from C2-T1. DDD is defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies. The Stable-C Interbody System is a stand alone system intended to be used with the bone anchors provided. The system is intended to be used with autogenous or allogenic bone graft comprised of cancellous, cortical, and/or corticocancellous bone graft to facilitate fusion. The system is to be used in patients who have had six weeks of non-operative treatment.

Type of Use (Select one or both, as applicable)



Prescription Use (Part 21 CFR 801 Subpart D)



Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

Stable-C Interbody System

July 15, 2024

Company: Nexus Spine, LLC
 2825 East Cottonwood Parkway Suite 330
 Salt Lake City, UT 84121

Primary Contact: Christine Scifert – Partner
 MRC Global
 9085 E. Mineral Cir., Suite 110
 Centennial, CO 80112
 Phone: (901) 831-8053
 Email: christine.scifert@AskMRCGlobal.com

Company/Secondary Contact: Jared Crocker
 Vice President of Quality and Regulatory Affairs
 Nexus Spine, LLC
 Phone: (801) 702-8592
 jared.crocker@nexusspine.com

Trade Name: Stable-C Interbody System

Common Name: Intervertebral Fusion Device With Integrated Fixation, Cervical

Classification: Class II

Regulation: 21 CFR 888.3080 (Intervertebral Fusion Device With Integrated Fixation, Cervical)

Panel: Orthopedic

Product Code: OVE

Primary Predicate: Nexus Spine, LLC Stable-C Interbody System – K231763
Additional Predicates: Nexus Spine, LLC Stable-C Interbody System – K232530
 Nexus Spine, LLC Tranquil-C Interbody System – K240416

Device Description:

The Stable-C Interbody System is an anterior cervical interbody device comprised of an interbody cage (lordotic angles of 0°, 6°, and 12°) made from titanium alloy (Ti-6Al-4V) per ASTM F3001 and two fixation anchors made from titanium alloy (Ti-6-Al-4V ELI) per ASTM F136. The device is offered in a variety of sizes to accommodate patient anatomy.

The purpose of this Traditional 510(k) submission is to gain clearance for modifications made to the subject device.

Indications for Use:

The Nexus Spine Stable-C Interbody System is an anterior cervical interbody fusion system indicated for use in skeletally mature patients with cervical disc disease (DDD) at one level from C2-T1. DDD is defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies. The Stable-C Interbody System is a stand alone system intended to be used with the bone anchors provided. The system is intended to be used with autogenous or allogenic bone graft comprised of cancellous, cortical, and/or corticocancellous bone graft to facilitate fusion. The system is to be used in patients who have had six weeks of non-operative treatment.

Substantial Equivalence:

The subject Nexus Spine Stable-C Interbody System is substantially equivalent to the following predicate devices:

Primary Predicate:

- Nexus Spine, LLC, Stable-C Interbody System – K231763; K232530

Additional Predicate:

- Nexus Spine, LLC, Tranquil-C Interbody System – K240416

The subject components are identical in indications to the primary predicate Stable-C Interbody System (K232530). Device sizing, geometry, and technological characteristics are similar to the predicate Stable-C (K232530) and Tranquil C devices (K240416). Materials, manufacturing, sterilization, and packaging are identical to those of the primary predicate, Stable-C (K232530). Thus, it can be concluded that the subject does not raise new questions about safety and effectiveness.

Performance Testing:

The subject changes to the Stable-C Interbody System do not introduce a new worst case, as confirmed by bench top usability testing of the device and instrumentation in a simulated implantation. Additional performance testing is not required to establish substantial equivalence.

Conclusion:

Based on the performance analysis and the comparison to the predicate devices, the subject device is determined to be substantially equivalent to the predicate devices.