



Guangzhou Wondfo Biotech Co., Ltd.
% Joe Shia
Director
LSI International Inc
504E Diamond Ave., Suite H
Gaithersburg, Maryland 20877

Re: K241741

Trade/Device Name: SAFElife™ Fentanyl Urine Home Test (Cassette); SAFElife™ Fentanyl (FTY) Urine Test Cassette; SAFElife™ T-Dip Fentanyl Urine Home Test (Dip Card); SAFElife™ T-Dip Fentanyl (FTY) Urine Test Panel

Regulation Number: 21 CFR 862.3650

Regulation Name: Opiate Test System

Regulatory Class: Class II

Product Code: NGL

Dated: June 13, 2024

Received: June 17, 2024

Dear Joe Shia:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801 and Part 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Joseph A. Kotarek -S

Digitally signed by
Joseph A. Kotarek -S
Date: 2024.07.16
11:44:38 -04'00'

Joseph Kotarek, Ph.D.

Branch Chief

Division of Chemistry

and Toxicology Devices

OHT7: Office of In Vitro Diagnostics

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

Submission Number (if known)

K241741

Device Name

SAFElife™ Fentanyl Urine Home Test (Cassette);
SAFElife™ Fentanyl (FTY) Urine Test Cassette;
SAFElife™ T-Dip Fentanyl Urine Home Test (Dip Card);
SAFElife™ T-Dip Fentanyl (FTY) Urine Test Panel

Indications for Use (Describe)

The SAFElife™ Fentanyl Urine Home Test (Cassette) is a competitive binding, lateral flow immunochromatographic assay for qualitative detection of Fentanyl in human urine at cutoff concentration of 1 ng/mL.

For in vitro diagnostic use. For Over The Counter (OTC) use.

The test provides only preliminary test results. To obtain a confirmed analytical result, a more specific alternate chemical method must be used. Chromatography/Mass Spectrometry (GC/MS) or Liquid Chromatography/Tandem Mass Spectrometry (LC/MS-MS) is the recommended confirmatory method.

The SAFElife™ Fentanyl (FTY) Urine Test Cassette is a competitive binding, lateral flow immunochromatographic assay for qualitative detection of Fentanyl (FTY) in human urine at cutoff concentration of 1 ng/mL.

For in vitro diagnostic use.

It is not intended to distinguish between prescription drug or abuse of the drug. Clinical consideration and professional judgment should be applied to the drug of abuse test result, particularly in evaluating a preliminary positive result.

The test provides only preliminary test results. To obtain a confirmed analytical result, a more specific alternate chemical method must be used. Chromatography/Mass Spectrometry (GC/MS) or Liquid Chromatography/Tandem Mass Spectrometry (LC/MS-MS) is the recommended confirmatory method.

The SAFElife™ T-Dip Fentanyl Urine Home Test (Dip Card) is a competitive binding, lateral flow immunochromatographic assay for qualitative detection of Fentanyl in human urine at cutoff concentration of 1 ng/mL.

For in vitro diagnostic use. For Over The Counter (OTC) use.

The test provides only preliminary test results. To obtain a confirmed analytical result, a more specific alternate chemical method must be used. Chromatography/Mass Spectrometry (GC/MS) or Liquid Chromatography/Tandem Mass Spectrometry (LC/MS-MS) is the recommended confirmatory method.

The SAFElife™ T-Dip Fentanyl (FTY) Urine Test Panel is a competitive binding, lateral flow immunochromatographic assay for qualitative detection of Fentanyl (FTY) in human urine at cutoff concentration of 1 ng/mL.

For in vitro diagnostic use.

It is not intended to distinguish between prescription drug or abuse of the drug. Clinical consideration and professional judgment should be applied to the drug of abuse test result, particularly in evaluating a preliminary positive result. The test provides only preliminary test results. To obtain a confirmed analytical result, a more specific alternate chemical method must be used. Chromatography/Mass Spectrometry (GC/MS) or Liquid Chromatography/Tandem Mass Spectrometry (LC/MS-MS) is the recommended confirmatory method.

Type of Use (Select one or both, as applicable)

☐ Prescription Use (Part 21 CFR 801 Subpart D)

☒ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) SUMMARY
K241741

1. Date: July 8, 2024
2. Submitter: Guangzhou Wondfo Biotech Co., Ltd.
No.8 Lizhishan Road, Science City, Huangpu District
Guangzhou, China
3. Contact person: Joe Shia
LSI International Inc.
504E Diamond Ave., Suite H
Gaithersburg, MD 20877
Telephone: 240-505-7880
Email: shiajl@yahoo.com
4. Device Names: SAFElife™ Fentanyl Urine Home Test (Cassette)
SAFElife™ Fentanyl (FTY) Urine Test Cassette
SAFElife™ T-Dip Fentanyl Urine Home Test (Dip Card)
SAFElife™ T-Dip Fentanyl (FTY) Urine Test Panel

Classification:	Class 2		
Product Code	Classification	Regulation Section	Panel
NGL	II	21 CFR § 862.3650 Opiate Test System	Toxicology (91)

5. Predicate Devices:

AllTest Fentanyl Urine Test Cassette (K233417)

6. Indications for Use

The SAFElife™ Fentanyl Urine Home Test (Cassette) is a competitive binding, lateral flow immunochromatographic assay for qualitative detection of Fentanyl in human urine at cutoff concentration of 1 ng/mL.

For in vitro diagnostic use. For Over The Counter (OTC) use.

The test provides only preliminary test results. To obtain a confirmed analytical result, a more specific alternate chemical method must be used. Chromatography/Mass Spectrometry (GC/MS) or Liquid Chromatography/Tandem Mass Spectrometry (LC/MS-MS) is the recommended confirmatory method.

The SAFElife™ Fentanyl (FTY) Urine Test Cassette is a competitive binding, lateral flow immunochromatographic assay for qualitative detection of Fentanyl (FTY) in human urine at cutoff concentration of 1 ng/mL.

For in vitro diagnostic use.

It is not intended to distinguish between prescription drug or abuse of the drug. Clinical consideration and professional judgment should be applied to the drug of abuse test result, particularly in evaluating a preliminary positive result.

The test provides only preliminary test results. To obtain a confirmed analytical result, a more specific alternate chemical method must be used. Chromatography/Mass Spectrometry (GC/MS) or Liquid Chromatography/Tandem Mass Spectrometry (LC/MS-MS) is the recommended confirmatory method.

The SAFElife™ T-Dip Fentanyl Urine Home Test (Dip Card) is a competitive binding, lateral flow immunochromatographic assay for qualitative detection of Fentanyl in human urine at cutoff concentration of 1 ng/mL.

For in vitro diagnostic use. For Over The Counter (OTC) use.

The test provides only preliminary test results. To obtain a confirmed analytical result, a more specific alternate chemical method must be used. Chromatography/Mass Spectrometry (GC/MS) or Liquid Chromatography/Tandem Mass Spectrometry (LC/MS-MS) is the recommended confirmatory method.

The SAFElife™ T-Dip Fentanyl (FTY) Urine Test Panel is a competitive binding, lateral flow immunochromatographic assay for qualitative detection of Fentanyl (FTY) in human urine at cutoff concentration of 1 ng/mL.

For in vitro diagnostic use.

It is not intended to distinguish between prescription drug or abuse of the drug. Clinical consideration and professional judgment should be applied to the drug of abuse test result, particularly in evaluating a preliminary positive result. The test provides only preliminary test results. To obtain a confirmed analytical result, a more specific alternate chemical method must be used. Chromatography/Mass Spectrometry (GC/MS) or Liquid Chromatography/Tandem Mass Spectrometry (LC/MS-MS) is the recommended confirmatory method.

7. Device Description

The Wondfo SAFElife™ Fentanyl Tests are immunoassays intended for the qualitative detection of fentanyl in human urine. Each Wondfo SAFElife™ Fentanyl Test consists of a Test Device in format of Cassette or Dip Card, and a package insert. Each Test Device is sealed with sachets of desiccant in an aluminum pouch.

8. Substantial Equivalence Information

A summary comparison of features of the Wondfo SAFElife™ Fentanyl Test and the predicate devices is provided in following table.

Table 1: Features Comparison of Wondfo SAFElife™ Fentanyl Test and the Predicate Device

Item	Device	Predicate – K233417
Indication(s) for Use	For the qualitative determination of fentanyl in human urine.	Same
Calibrator and Cut-Off Values	Fentanyl (FTY) 1 ng/ml	Same
Methodology	Competitive binding, lateral flow immunochromatographic assays based on the principle of antigen antibody immunochemistry.	Same
Type of Test	Qualitative	Same
Specimen Type	Human Urine	Same
Intended Use	For OTC use	Same
Configurations	Cassette, Dip Card	Cassette
Storage	4-30°C	Same

9. Test Principle

The Wondfo SAFElife™ Fentanyl Tests are immunoassays based on the principle of competitive binding.

During testing, a urine specimen migrates upward by capillary action. Fentanyl, if present in the urine specimen below 1 ng/mL, will not saturate the binding sites of antibody-coated particles in the test device. The antibody-coated particles will then be captured by immobilized fentanyl conjugate and a visible colored line will show up in the test line region. The colored line will not form in the test line region if the fentanyl level exceeds 1 ng/mL because it will saturate all the binding sites of anti-fentanyl antibodies.

To serve as a procedural control, a colored line will always appear at the control line region indicating that proper volume of specimen has been added and membrane wicking has occurred.

10. Performance Characteristics

1. Analytical Performance

a. Precision

Precision studies were carried out for samples with concentrations of -100% cut off, -75% cut off, -50% cut off, -25% cut off, cut off, +25% cut off, +50% cut off, +75% cut off and +100% cut off. These samples were prepared by spiking fentanyl in negative samples. Each fentanyl concentration was confirmed by LC/MS. All sample aliquots were blindly labeled by the person who prepared the samples and didn't take part in the sample testing. For each concentration, tests were performed two tests per day for 25 days per device lot in a randomized order.

Cassette

Lot Number	-100% cut off	-75% cut off	-50% cut off	-25% cutoff	cut off	+25% cut off	+50% cut off	+75% cut off	+100% cut off
Lot 1	50-/0+	50-/0+	50-/0+	48-/2+	28+/22-	50+/0-	50+/0-	50+/0-	50+/0-
Lot 2	50-/0+	50-/0+	50-/0+	46-/4+	27+/23-	50+/0-	50+/0-	50+/0-	50+/0-
Lot 3	50-/0+	50-/0+	50-/0+	48-/2+	25+/25-	50+/0-	50+/0-	50+/0-	50+/0-

Dip Card

Lot Number	-100% cut off	-75% cut off	-50% cut off	-25% cutoff	cut off	+25% cut off	+50% cut off	+75% cut off	+100% cut off
Lot 1	50-/0+	50-/0+	50-/0+	47-/3+	24+/26-	50+/0-	50+/0-	50+/0-	50+/0-
Lot 2	50-/0+	50-/0+	50-/0+	47-/3+	28+/22-	50+/0-	50+/0-	50+/0-	50+/0-
Lot 3	50-/0+	50-/0+	50-/0+	48-/2+	23+/27-	50+/0-	50+/0-	50+/0-	50+/0-

c. Stability

The devices are stable at 36-86F for 24 months based on the accelerated stability study.

d. Interference

Potential interfering substances found in human urine of physiological or pathological conditions were added to drug-free urine and target drug fentanyl urine with concentrations at 50% below and 50% above Cut-Off levels. These urine samples were tested using three batches of each device. Compounds that showed no interference at a concentration of 100µg/mL or specified concentrations are summarized in the following tables.

Acetaminophen	Doxepin	Nortriptyline
Acetone (1000 mg/dL)	Ecgonine methyl ester	Noscapine
Acetophenetidin	Ephedrine	O-Hydroxyhippuric acid

Acetylsalicylic acid	Erythromycin	Octopamine
Albumin (100 mg/dL)	Ethanol (1%)	Oxalic acid (100 mg/dL)
Albuterol	Fenoprofen	Oxazepam
Aminopyrine	Fluphenazine	Oxolinic acid
Amitriptyline	Furosemide	Oxymetazoline
Amobarbital	Galactose (10 mg/dL)	Papaverine
Amoxicillin	Gamma globulin (500 mg/dL)	Penicillin G
Ampicillin	Gentisic acid	Perphenazine
Apomorphine	Glucose (3000 mg/dL)	Phencyclidine
Ascorbic Acid	Hemoglobin	Phenelzine
Aspartame	Hydralazine	Phenobarbital
Atropine	Hydrochlorothiazide	Prednisone
Benzilic acid	Hydrocortisone	Propoxyphene (50 µg/mL)
Benzoic acid	Hydroxytyramine	Propranolol
Benzoyllecgonine	Ibuprofen	Pseudoephedrine
Bilirubin	Imipramine	Quinine
Boric acid (1% w/v)	Isoproterenol	Ranitidine
Bupropion	Isoxsuprine	Riboflavin (10 mg/dL)
Caffeine	Ketamine	Salicylic acid
Carbamazepine	Ketoprofen	Secobarbital
Chloral hydrate	Labetalol	Serotonin (5-hydroxytyramine)
Chloramphenicol	Lidocaine	Sulfamethazine
Chlorothiazide	Loperamide	Sulindac
Chlorpromazine	Maprotiline	Tetrahydrocortisone 3-(β-Dglucuronide)
Cholesterol	Meperidine (50 µg/mL)	Tetrahydrocortisone 3-acetate
Clomipramine	Meprobamate	Tetrahydrozoline
Clonidine	Methapyrilene	Thiamine
Cortisone	Methaqualone	Thioridazine
Cotinine	Methoxyphenamine	Triamterene
Creatinine	Metronidazole (300 µg/mL)	Trifluoperazine
Cyclobenzaprine	N-Acetylprocainamide	Trimethoprim
Deoxycorticosterone	NaCl (4000 mg/dL)	Tyramine
Desipramine	Nalidixic acid	Urea (2000 mg/dL)
Dextromethorphan	Naloxone	Uric acid
Diclofenac	Naltrexone	Valproic acid (250 µg/mL)
Diflunisal	Naproxen	Venlafaxine
Digoxin	Niacinamide	Verapamil
Diphenhydramine	Nicotine	Zomepirac
DL-Tryptophan	Nifedipine	β-Estradiol
DL-Tyrosine	Norethindrone	

e. Specificity

To test specificity, drug metabolites and other components that are likely to interfere in urine samples were tested using three batches of device. The lowest concentration that caused a positive result for each compound are listed below. There are no differences in the test results for different device format.

Fentanyl (Cutoff=1ng/mL)	Minimum concentration required to obtain a positive result (ng/mL)	% Cross-Reactivity
Acetyl fentanyl	16	6.25%
Acrylfentanyl	1	100.00%
ω-1-Hydroxyfentanyl	20,000	0.005%
Isobutyryl fentanyl	1	100.00%
Ocfentanil	2.3	43.48%
Butyryl fentanyl	2	50.00%
Furanyl fentanyl	1	100.00%
Valeryl fentanyl	2.5	40.00%
(±) β-hydroxythiofentanyl	2.5	40.00%
4-Fluoro-isobutyrylfentanyl	3	33.33%
Para-fluorobutyryl fentanyl	4	25.00%
Para-fluoro fentanyl	2.5	40.00%
(+)-3-cis-methyl fentanyl	50	2.00%
Carfentanil	2	50.00%
Sufentanil	15	6.67%
Alfentanil	7500	0.01%
Despropionyl fentanyl (4-ANPP)	2,000	0.05%
Remifentanil	>100 µg/mL	< 0.001%
Norfentanyl	>100 µg/mL	< 0.001%
Acetyl norfentanyl	>100 µg/mL	< 0.001%
Norcarfentanil	>100 µg/mL	< 0.001%
Trazodone	25000	0.004%

The following opioids compounds were tested at a concentration of 100ug/mL. Negative results were obtained for all these compounds. There is no cross-reactivity for these compounds using the Wondfo SAFElife™ Fentanyl Tests of all three device formats.

6-Acetyl morphine	Naloxone
Amphetamine	Naltrexone
Buprenorphine	Norbuprenorphine
Buprenorphineglucuronide	Norcodeine
Codeine	Norketamine
Dextromethorphan	Normeperidine
Dihydrocodeine	Normorphine
EDDP	Noroxycodone
EMDP	Oxycodone
Fluoxetine	Oxymorphone
Heroin	Pentazocine (Talwin)
Hydrocodone	Pipamperone
Hydromorphone	Risperidone
Ketamine	Tapentadol
Levorphanol	Thioridazine
Meperidine	Tilidine
Methadone	Tramadol
Morphine	Tramadol-O-Desmethyl
Morphine-3-glucuronide	Tramadol-N-Desmethyl

f. Effect of Urine Specific Gravity and Urine pH

To investigate the effect of urine specific gravity and urine pH, urine samples, with 1.000 to 1.035 specific gravity or urine samples with pH 4 to 9 were spiked with target fentanyl at 50% below and 50% above Cut-Off levels. These samples were tested using three lots of device. Results were all positive for samples at and above +50% Cut-Off and all negative for samples at and below -50% Cut-Off. There are no differences in the test results for different device format.

2. Comparison Studies

Method comparison studies for the Wondfo SAFElife™ Fentanyl Test were performed using three different lots of the device. Operators ran 84 (41 negative and 43 positive) unaltered clinical samples. The samples were blind labeled and compared to LC/MS results. The results are presented in the tables below.

Cassette

		Negative	Low Negative by LC/MS (less than -50%)	Near Cutoff Negative by LC/MS (Between -50% and cutoff)	Near Cutoff Positive by LC/MS (Between the cutoff and +50%)	High Positive by LC/MS (greater than +50%)
Operator 1	Positive	0	0	2	22	20
	Negative	8	13	18	1	0
Operator 2	Positive	0	0	2	23	20
	Negative	8	13	18	0	0
Operator 3	Positive	0	0	2	23	20
	Negative	8	13	18	0	0

Discordant Results

Operator	Sample ID	LC/MS Result (ng/mL)	Rapid Test Result
Operator 1, 2	39	0.825	Positive
Operator 1, 2, 3	41	0.914	Positive
Operator 3	40	0.885	Positive
Operator 1	43	1.058	Negative

Dip Card

		Negative	Low Negative by LC/MS (less than -50%)	Near Cutoff Negative by LC/MS (Between -50% and cutoff)	Near Cutoff Positive by LC/MS (Between the cutoff and +50%)	High Positive by LC/MS (greater than +50%)
Operator 1	Positive	0	0	2	22	20
	Negative	8	13	18	1	0
Operator 2	Positive	0	0	3	21	20
	Negative	8	13	17	2	0

Operator 3	Positive	0	0	1	21	20
	Negative	8	13	19	2	0

Discordant Results

Operator	Sample ID	LC/MS Result (ng/mL)	Rapid Test Result
Operator 1, 2, 3	40	0.885	Positive
Operator 1	41	0.914	Positive
Operator 2	37	0.804	Positive
Operator 2	39	0.825	Positive
Operator 1	44	1.038	Negative
Operator 2, 3	42	1.015	Negative
Operator 2, 3	45	1.077	Negative

3. Lay-user study

A lay user study was performed at three intended user sites with 140 lay persons for each device format. They had diverse educational and professional backgrounds and ranged in age from 18 to >50 years. Urine samples were prepared at the following concentrations; -100%, +/-75%, +/-50%, +/-25% of the cut-off by spiking fentanyl into drug free-pooled urine specimens. The concentrations of the samples were confirmed by LC/MS. Each sample was aliquoted into individual containers, blind-labeled and randomized. Each participant was provided with the package insert, 1 blind labeled sample and a device. The results are summarized below:

Cassette

% of Cutoff	Number of samples	Fentanyl Concentration by LC/MS (ng/mL)	Lay person results		The percentage of correct results (%)
			No. of Positive	No. of Negative	
-100% Cutoff	20	0	0	20	100
-75% Cutoff	20	0.2	0	20	100
-50% Cutoff	20	0.5	0	20	100
-25% Cutoff	20	0.7	1	19	95
+25% Cutoff	20	1.2	20	0	100
+50% Cutoff	20	1.4	20	0	100
+75% Cutoff	20	1.7	20	0	100

Dip Card

% of Cutoff	Number of samples	Fentanyl Concentration by LC/MS (ng/mL)	Lay person results		The percentage of correct results (%)
			No. of Positive	No. of Negative	
-100% Cutoff	20	0	0	20	100
-75% Cutoff	20	0.2	0	20	100
-50% Cutoff	20	0.5	0	20	100
-25% Cutoff	20	0.7	1	19	95
+25% Cutoff	20	1.2	20	0	100
+50% Cutoff	20	1.4	20	0	100
+75% Cutoff	20	1.7	20	0	100

Lay-users were also given surveys on the ease of understanding the package insert instructions. All lay users indicated that the device instructions can be easily followed. A Flesch-Kincaid

reading analysis was performed on the package insert and the score revealed a reading grade level of less than 7.

4. Clinical Studies

Not applicable.

11. Conclusion

Based on the test principle and acceptable performance characteristics including precision, cut-off, interference, specificity, method comparison and Lay-user studies of the devices, it's concluded a substantial equivalence decision.