



Asahi Intecc Co., Ltd.  
Cynthia Valenzuela  
Director, Quality Systems, Regulatory Affairs/Compliance  
Asahi Intecc USA, Inc.  
3002 Dow Avenue  
Suite 212  
Tustin, California 92780

Re: K241962  
Trade/Device Name: Crossloop  
Regulation Number: 21 CFR 870.1330  
Regulation Name: Catheter Guide Wire  
Regulatory Class: Class II  
Product Code: DQX  
Dated: February 20, 2025  
Received: February 20, 2025

Dear Cynthia Valenzuela:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

All medical devices, including Class I and unclassified devices and combination product device constituent parts are required to be in compliance with the final Unique Device Identification System rule ("UDI Rule"). The UDI Rule requires, among other things, that a device bear a unique device identifier (UDI) on its label and package (21 CFR 801.20(a)) unless an exception or alternative applies (21 CFR 801.20(b)) and that the dates on the device label be formatted in accordance with 21 CFR 801.18. The UDI Rule (21 CFR 830.300(a) and 830.320(b)) also requires that certain information be submitted to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). For additional information on these requirements, please see the UDI System webpage at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-system-udi-system>.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory->

[assistance/contact-us-division-industry-and-consumer-education-dice](#)) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Jenny R.  
Katsnelson -S

Digitally signed by Jenny R.  
Katsnelson -S  
Date: 2025.03.27 12:02:30 -04'00'

for Lydia Glaw

Assistant Director

DHT2C: Division of Coronary and

Peripheral Intervention Devices

OHT2: Office of Cardiovascular Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K241962

Device Name

CROSSLOOP

Indications for Use (Describe)

This product is intended to direct a catheter to the desired anatomical location in the peripheral vasculature (excluding coronary and cerebral vessels).

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**510(k) Summary**  
(as required by 21 CFR § 807.92)



Global Headquarters and R&D Center  
3-100 Akatsuki-cho, Seto-shi, Aichi 489-0071 Japan  
TEL : +81-561-48-5551 FAX : +81-561-48-5552  
<http://www.asahi-intecc.co.jp/>

**CROSSLOOP**

**510(k) K241962**

|                               |                                                                                                                                                                                                                                                                  |
|-------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <b>Date Prepared:</b>         | 25 JUNE 2024                                                                                                                                                                                                                                                     |
| <b>Applicant:</b>             | ASAHI INTECC CO., LTD.<br>3-100 Akatsuki-cho,<br>Seto, Aichi 489-0071 Japan                                                                                                                                                                                      |
| <b>Contact:</b>               | Mrs. Cynthia Valenzuela<br>Director, Quality and Regulatory Affairs<br>ASAHI INTECC USA, Inc.<br>3002 Dow Avenue, Suite 212<br>Tustin, CA 92780<br>Tel: (949) 745-1617<br>e-mail: <a href="mailto:cynthiav@asahi-intecc-us.com">cynthiav@asahi-intecc-us.com</a> |
| <b>Trade Name:</b>            | CROSSLOOP                                                                                                                                                                                                                                                        |
| <b>Common Name:</b>           | Guidewire                                                                                                                                                                                                                                                        |
| <b>Device Classification:</b> | Class 2 per 21 CFR § 870.1330                                                                                                                                                                                                                                    |
| <b>Classification Name:</b>   | Catheter, Guide, Wire                                                                                                                                                                                                                                            |
| <b>Product Code:</b>          | DQX – Catheter Guide Wire                                                                                                                                                                                                                                        |
| <b>Predicate Device:</b>      | Astato 30 K071721, K163426                                                                                                                                                                                                                                       |
| <b>Reference Devices:</b>     | CROSSLEAD Penetration K230377<br>SION blue K122468, K163426, K191464<br>Astato XS 40 K153443, K163426<br>V-18 Control Wire K033742                                                                                                                               |

**INTENDED USE/INDICATIONS FOR USE:**

*This product is intended to direct a catheter to the desired anatomical location in the peripheral vasculature (excluding coronary and cerebral vessels).*

**DEVICE DESCRIPTION:**

The CROSSLOOP is steerable guide wire with a maximum diameter of 0.018 inches (0.46 mm) and available in various lengths of 200 cm, 235 cm and 300 cm. The CROSSLOOP consists of a stainless-steel (SUS) core wire with a Pt-Ni alloy coil, soldered with Au-Sn solder. The distal tip is loop-structured. The coil is radiopaque to achieve visibility and can be made to bend easily to accommodate vessel tortuosity. A silicone and hydrophilic coating are applied to the distal portion of the guide wire. A hydrophobic coating (PTFE) is applied to proximal portion. The purpose of these

ASAHI INTECC CO., LTD.

surface coatings is to provide lubricity when the guide wire is passed through percutaneous catheters.

The basic structure, construction and materials of the CROSSLOOP are similar to those of the predicate Astato 30 (K071721, K163426) and reference CROSSLEAD Penetration (K230377) and SION blue (K122468, K163426, K191464).

#### **COMPARISON WITH PREDICATE AND REFERENCE DEVICES:**

Comparisons of the CROSSLOOP and predicate / reference devices show that the technological characteristics of the subject device such as the components, design, materials, sterilization method and operating principle are similar to the currently marketed predicate and reference devices. The intended use of the subject device and the predicate is similar.

| Device name                        | Subject                                                                                                                                                   | Predicate                                                                                                                                                                                          | Reference                                                                                                                                                      |                                                                                                                                                                                                                                                                                     |                                                                                                                                                                                                    |                                                                                                                                                                                                                                                                                                                                                                                                                                                    |
|------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|                                    | CROSSLOOP                                                                                                                                                 | Astato 30                                                                                                                                                                                          | CROSSLEAD Penetration                                                                                                                                          | SION blue                                                                                                                                                                                                                                                                           | Astato XS 40                                                                                                                                                                                       | V-18 Control Wire                                                                                                                                                                                                                                                                                                                                                                                                                                  |
| 510(k)                             | TBD                                                                                                                                                       | K071721,K163426                                                                                                                                                                                    | K230377                                                                                                                                                        | K122468,<br>K163426,<br>K191464                                                                                                                                                                                                                                                     | K153443,K163426                                                                                                                                                                                    | K033742                                                                                                                                                                                                                                                                                                                                                                                                                                            |
| Manufacturer                       | ASAHI INTECC CO., LTD.                                                                                                                                    |                                                                                                                                                                                                    |                                                                                                                                                                |                                                                                                                                                                                                                                                                                     |                                                                                                                                                                                                    | Boston Scientific                                                                                                                                                                                                                                                                                                                                                                                                                                  |
| Regulation Number                  | 21 CFR 870.1330                                                                                                                                           |                                                                                                                                                                                                    |                                                                                                                                                                |                                                                                                                                                                                                                                                                                     |                                                                                                                                                                                                    |                                                                                                                                                                                                                                                                                                                                                                                                                                                    |
| Regulation Name                    | Catheter Guide Wire                                                                                                                                       |                                                                                                                                                                                                    |                                                                                                                                                                |                                                                                                                                                                                                                                                                                     |                                                                                                                                                                                                    |                                                                                                                                                                                                                                                                                                                                                                                                                                                    |
| Regulatory Class                   | II                                                                                                                                                        |                                                                                                                                                                                                    |                                                                                                                                                                |                                                                                                                                                                                                                                                                                     |                                                                                                                                                                                                    |                                                                                                                                                                                                                                                                                                                                                                                                                                                    |
| Product code                       | DQX                                                                                                                                                       |                                                                                                                                                                                                    |                                                                                                                                                                |                                                                                                                                                                                                                                                                                     |                                                                                                                                                                                                    |                                                                                                                                                                                                                                                                                                                                                                                                                                                    |
| Intended and/or Indication for Use | This product is intended to direct a catheter to the desired anatomical location in the peripheral vasculature (excluding coronary and cerebral vessels). | This product is intended to facilitate the placement and exchange of diagnostic and therapeutic devices during intravascular procedures. This device is intended for peripheral vascular use only. | This product is designed to direct a catheter to the desired anatomical location in the peripheral vasculature during diagnostic or interventional procedures. | ASAHI PTCA Guide Wires are intended to facilitate the placement of balloon dilatation catheters during percutaneous transluminal coronary angioplasty (PTCA) and percutaneous transluminal angioplasty (PTA). The ASAHI PTCA Guide Wires are not to be used in the neurovasculature | This product is intended to facilitate the placement and exchange of diagnostic and therapeutic devices during intravascular procedures. This device is intended for peripheral vascular use only. | The 110cm V-18 Control Wire is intended for general intravascular use including the placement of PTA balloon catheters requiring an 0.018" guide wire in hemodialysis AV access procedures. The wire can be torqued to facilitate placement of diagnostic or therapeutic catheters. This device is intended for peripheral vascular use only. A torque device is included with each wire to facilitate directional manipulation of the guide wire. |

| Device name          | Subject                                | Predicate                                   | Reference                                          |                                              |                                             |                         |
|----------------------|----------------------------------------|---------------------------------------------|----------------------------------------------------|----------------------------------------------|---------------------------------------------|-------------------------|
|                      | CROSSLOOP                              | Astato 30                                   | CROSSLEAD Penetration                              | SION blue                                    | Astato XS 40                                | V-18 Control Wire       |
| Nominal OD           | 0.46 mm<br>(0.018 inch)                | 0.45 mm<br>(0.018 inch)<br>with tapered end | 0.36 mm<br>(0.014 inch)<br>0.46 mm<br>(0.018 inch) | 0.36 mm<br>(0.014 inch)                      | 0.36 mm<br>(0.014 inch)<br>with tapered end | 0.46 mm<br>(0.018 inch) |
| Overall Length       | 200 cm, 235 cm,<br>300 cm              | 180 cm, 300 cm                              | 200 cm, 235 cm,<br>300 cm                          | 180 cm, 190 cm,<br>300 cm                    | 180 cm, 300 cm                              | 110 cm                  |
| Coil                 | Pt-Ni alloy                            | Pt-Ni alloy                                 | Pt-Ni alloy                                        | Austenitic<br>Stainless Steel<br>Pt-Ni alloy | Pt-Ni alloy                                 | unknown                 |
| Tapered Core Wire    | Austenitic<br>Stainless Steel          | Austenitic<br>Stainless Steel               | Austenitic<br>Stainless Steel                      | Austenitic<br>Stainless Steel                | Austenitic<br>Stainless Steel               | Stainless Steel         |
| Tip Shape            | Straight, Pre-<br>shape                | Straight                                    | Straight, Pre-<br>shape                            | Straight, J-shape,<br>Pre-shape              | Straight                                    | Straight                |
| Distal Tip Structure | Round                                  | Round                                       | Micro-corn                                         | Round                                        | Round                                       | Round                   |
| Coating              | Hydrophilic<br>PTFE<br>Silicone        | Hydrophilic<br>PTFE                         | Hydrophilic<br>PTFE                                | Hydrophilic<br>PTFE<br>Silicone              | Hydrophilic<br>PTFE                         | Hydrophilic<br>PTFE     |
| Solder               | Au-Sn                                  | Ag-Sn<br>Au-Sn                              | Au-Sn                                              | Ag-Sn<br>Au-Sn                               | Ag-Sn<br>Au-Sn                              | unknown                 |
| Sterilization        | Ethylene oxide<br>SAL 10 <sup>-6</sup> | Ethylene oxide<br>SAL 10 <sup>-6</sup>      | Ethylene oxide<br>SAL 10 <sup>-6</sup>             | Ethylene oxide<br>SAL 10 <sup>-6</sup>       | Ethylene oxide<br>SAL 10 <sup>-6</sup>      | unknown                 |

The following technological differences exist between the subject and the predicate or the reference devices:

- Coil structure
- Core shaft
- Distal tip structure

## **NON-CLINICAL TESTING/PERFORMANCE DATA:**

Non-clinical laboratory testing was performed on the CROSSLOOP to determine substantial equivalence. The following testing/assessments were performed:

- Tensile Strength
- Torque Strength
- Torqueability
- Tip Flexibility
- Coating Integrity
- Catheter Compatibility
- Visual Inspection
- Corrosion Resistance
- Kink Resistance
- Radio – Detectability
- Dimensional Verification
- Coating Integrity / Acute Particulate Characterization

The *in vitro* bench tests demonstrated the CROSSLOOP met all acceptance criteria and performed similarly to the predicate and reference devices. Performance data demonstrates the device functions as intended and has a safety and effectiveness profile that is similar to the predicate device.

## **BIOCOMPATIBILITY:**

Testing was performed to assess biocompatibility of the device. The following tests were performed:

- Cytotoxicity
- Sensitization
- Intracutaneous Irritation
- Systemic Toxicity
- USP Rabbit Pyrogen, Material Mediated
- Hemolysis
- Partial Thromboplastin Time
- In Vivo Thrombogenicity
- Sc5b-9 Complement Activation

## **CONCLUSION:**

The CROSSLOOP has the same intended use and the same or similar technological characteristics such as components, design, materials, sterilization method and operating principles as the predicate and reference devices. The conclusions drawn from the nonclinical tests demonstrate that the device has a similar safety and effectiveness profile as the predicate device. Therefore, the CROSSLOOP is substantially equivalent to the predicate device.

ASAHI INTECC CO., LTD.