



August 23, 2024

Think Surgical, Inc.
Anand Patel
Manager, Regulatory Affairs
47201 Lakeview Blvd
Fremont, California 94528

Re: K242264
Trade/Device Name: TMINI Miniature Robotic System
Regulation Number: 21 CFR 882.4560
Regulation Name: Stereotaxic Instrument
Regulatory Class: Class II
Product Code: OLO
Dated: July 31, 2024
Received: August 1, 2024

Dear Anand Patel:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device"

(<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shumaya Ali -S

Shumaya Ali, M.P.H.

Assistant Director

DHT6C: Division of Restorative, Repair
and Trauma Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K242264

Device Name

TMINI® Miniature Robotic System

Indications for Use (Describe)

TMINI® Miniature Robotic System is indicated as a stereotaxic instrumentation system for total knee replacement (TKA) surgery. It is to assist the surgeon by providing software-defined spatial boundaries for orientation and reference information to identifiable anatomical structures for the accurate placement of knee implant components.

The robotic device placement is performed relative to anatomical landmarks as recorded using the system intraoperatively and based on a surgical plan determined preoperatively using CT based surgical planning tools.

The targeted population has the same characteristics as the population that is suitable for the implant(s) compatible with the TMINI® Miniature Robotic System. The TMINI® Miniature Robotic System is to be used with the following knee replacement system(s) in accordance with the indications and contraindications:

- Enovis™ EMPOWR Knee System®
- Ortho Development® BKS® and BKS TriMax® Knee System
- Total Joint Orthopedics Klassic® Knee System
- United® U2™ Knee Total Knee System
- Medacta® GMK® Sphere / SpheriKA Knee Systems
- Zimmer Biomet Anterior & Posterior Referencing Persona® Knee

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) SUMMARY

Applicant Information:

Owner Name:	THINK Surgical, Inc.
Address:	47201 Lakeview Blvd., Fremont, CA 94538
Phone number:	443-756-9392
Fax number:	510-249-2396
Establishment Registration Number:	3000719653
Contact Person:	Anand Patel
Date Prepared:	21 August 2024

Device Information:

Device Classification:	Class II
Trade Name:	TMINI® Miniature Robotic System
Common name:	Orthopedic Stereotaxic Instrument
Classification name:	Stereotaxic Instrument
Regulation number:	882.4560
Product Code:	OLO

Predicate Device:

The TMINI® Miniature Robotic System (Additional Knee System) is substantially equivalent in intended use, Indications for Use, design, materials, technology, operational principles and performance to the predicate, TMINI® Miniature Robotic System, cleared via K241031.

Device Modification:

The purpose of this submission is to add two additional FDA cleared knee implant systems that are compatible with the TMINI® Miniature Robotic System to the Enovis™ EMPOWR Knee System®, Ortho Development® BKS® and BKS TriMax® Knee System, the Total Joint Orthopedics Klassic® Knee System, and the United® U2™ Knee System that is already cleared for use with the device. The two new implant systems are: the Medacta® GMK® Sphere / SpheriKA Knee Systems, and the Zimmer Biomet Anterior & Posterior Referencing Persona® Knee. As part of this change the labeling has been modified to show that the Indications for Use of the device has been updated to include compatibility with these two additional knee implant systems.

Device Description:

Like its predicate, the TMINI® Miniature Robotic System (Additional Knee System) consists of three primary components: a three-dimensional, graphical, Preoperative Planning Workstation (TPLAN Planning Station), an Optical Tracking Navigation Console (TNav)

and a robotically controlled hand-held tool (TMINI Robot) that assists the surgeon in preparing the bone for implantation of TKA components.

The TPLAN Planning Station uses preoperative CT scans of the operative leg to create 3D surface models for case templating and intraoperative registration purposes. The Planning Workstation contains a library of 510(k) cleared knee replacement implant(s) available for use with the system. The surgeon can select an implant model from this library. The planner/surgeon can manipulate the 3D representation of the implant in relation to the bone model to optimally place the implant. The surgeon reviews and approves the case plan once the surgeon is satisfied with the implant selection, location and orientation. The data from the approved plan is written to a file that is used to guide the robotically controlled hand-held tool.

The hand-held robotic tool is optically tracked relative to optical markers placed in both the femur and tibia and articulates in two degrees-of-freedom, allowing the user to place bone pins in a planar manner in both bones. Mechanical guides are clamped to the bone pins, resulting in subsequent placement of cut slots and drill guide holes such that the distal femoral and proximal tibial cuts can be made in the pre-planned positions and orientations, and such that the implant manufacturer's multi-planer cutting block can be placed relative to drilled distal femoral pilot holes. If the surgeon needs to change the plan during surgery, it can be changed intraoperatively.

Intended Use:

Like the predicate, the TMINI[®] Miniature Robotic System (Additional Knee System) is intended to assist the surgeon in providing software defined spatial boundaries for orientation and reference information to anatomical structures during orthopedic procedures.

Indications for Use:

The differences between the indications for use of the current device and the predicate are the addition of the two additional knee systems and the removal of the sentence listing the system components. See Table-1 below.

The table below identifies the substantial equivalence of TMINI[®] Miniature Robotic System (Additional Knee System) to the predicate, TMINI[™] Miniature Robotic System cleared via K241031.

Table 1: Comparison of Intended Use and Indications for Use

Product	TMINI® Miniature Robotic System (Additional Knee System)	TMINI™ Miniature Robotic System	Conclusion
510(k) number	Subject Device	K241031	
Manufacturer	THINK Surgical, Inc	THINK Surgical, Inc	
Product Code	OLO	OLO	SAME
Regulation	21 CFR 882.4560	21 CFR 882.4560	SAME
Intended Use	Intended to assist the surgeon in providing software defined spatial boundaries for orientation and reference information to anatomical structures during orthopedic procedures.	Intended to assist the surgeon in providing software defined spatial boundaries for orientation and reference information to anatomical structures during orthopedic procedures.	SAME
Indications for Use	The TMINI® Miniature Robotic System is indicated as a stereotaxic instrumentation system for total knee replacement (TKA) surgery. It is to assist the surgeon by providing software-defined spatial boundaries for orientation and reference information to identifiable anatomical structures for the accurate placement of knee implant components. The robotic device placement is performed relative to anatomical landmarks as recorded using the system intraoperatively and based on a surgical plan-determined preoperatively using CT based surgical planning tools. The targeted population has the same characteristics as the population that is suitable for the implant(s) compatible with the TMINI® Miniature Robotic System. The TMINI® Miniature Robotic System is to be used with the following knee replacement system(s) in accordance with the indications and contraindications: - Enovis™ EMPOWR Knee System® - Ortho Development BKS® and BKS TriMax® Knee System	The TMINI™ Miniature Robotic System is indicated as a stereotaxic instrumentation system for total knee replacement (TKA) surgery. It is to assist the surgeon by providing software-defined spatial boundaries for orientation and reference information to identifiable anatomical structures for the accurate placement of knee implant components. The robotic device placement is performed relative to anatomical landmarks as recorded using the system intraoperatively and based on a surgical plan-determined preoperatively using CT based surgical planning tools. It includes a handheld robotic device, an optical sensor navigation system and accessories, software system, surgical instruments and accessories. The targeted population has the same characteristics as the population that is suitable for the implant(s) compatible with the TMINI™ Miniature Robotic System. The TMINI™ Miniature Robotic System is to be used with the following knee replacement system(s) in accordance with	Substantially Equivalent (Adds Compatibility with two additional total knee implant systems and the removal of the sentence listing the system components)

Product	TMINI® Miniature Robotic System (Additional Knee System)	TMINI™ Miniature Robotic System	Conclusion
	- Total Joint Orthopedics Klassic® Knee System - United U2™ Knee Total Knee System - Medacta® GMK® Sphere / SpheriKA Knee Systems - Zimmer Biomet Anterior & Posterior Referencing Persona® Knee	the indications and contraindications: - Enovis™ EMPOWR Knee System® - Ortho Development BKS® and BKS TriMax® Knee System - Total Joint Orthopedics Klassic® Knee System - United U2™ Knee Total Knee System	

Substantial Equivalence:

Both the TMINI® Miniature Robotic System (Additional Knee System), the subject of this submission, and the predicate device have the same intended use. Both are indicated as a stereotaxic instrumentation system for total knee replacement (TKA) surgery. It is to assist the surgeon by providing software-defined spatial boundaries for orientation and reference information to identifiable anatomical structures for the accurate placement of knee implant components during orthopedic procedures.

The robotic device placement is performed relative to anatomical landmarks as recorded using the system intraoperatively and based on a surgical plan determined preoperatively using CT based surgical planning tools. The difference between the new device and the predicate is that the new device includes compatibility with two additional knee systems. The addition of the two knee systems do not alter the intended use, Indications for Use (other than the addition of the two new implant systems), design, materials, technology, or operational principles of the TMINI® Miniature Robotic System and no new questions of safety or effectiveness resulted from the changes.

The Indications for Use of the new device and its predicate are identical except for the addition of the two new implant systems to the list of compatible implant systems and the removal of the sentence listing the system components.

Verification and validation activities were performed to ensure that the TMINI® Miniature Robotic System (Additional Knee System) fulfilled system requirements, ensure that the product design conforms to the user needs and intended uses and performs substantially equivalent to the predicate, TMINI™ Miniature Robotic System, cleared via K241031. Throughout all testing performed, test samples were representative of the production product. The verification and validation activities for the TMINI® Miniature Robotic System (Additional Knee System) followed test protocols that were previously established and reviewed in the 510(k) submissions for the predicate device. The verification and validation activities were successfully completed, and all pre-determined acceptance criteria were met.

Biocompatibility information for patient contacting materials and testing for the TMINI® Miniature Robotic System were presented in submission K232802. There are no material changes to any of the direct patient contact components of the TMINI® Miniature Robotic

System (Additional Knee System) included in this submission; therefore, no additional biocompatibility testing was required.

Substantial equivalence in technological characteristic and performance of the TMINI[®] Miniature Robotic System to the predicate device is outlined in **Table-2** below:

Table-2: Substantial Equivalence

Product	TMINI[®] Miniature Robotic System (Additional Knee System)	TMINI[™] Miniature Robotic System (TMINI 1.1)	Substantial Equivalence Conclusion
510(k) number	Subject Device	K241031	
Manufacturer	THINK Surgical Inc.	THINK Surgical Inc.	
Materials			
-Materials Used	Uses materials with a long history of use in orthopedic procedures or provided biocompatibility data consistent with ISO 10993 requirements	Uses materials with a long history of use in orthopedic procedures or provided biocompatibility data consistent with ISO 10993 requirements	SAME
Technological Characteristics			
-Major System Components	Planning and robot control software, robotic positioning device, navigation system, reusable and disposable instrumentation	Planning and robot control software, robotic positioning device, navigation system, reusable and disposable instrumentation	SAME
-Patient Imaging	CT images used to create a 3D model of the bone for surgical planning	CT images used to create a 3D model of the bone for surgical planning	SAME
-Preoperative planning workstation	TPLAN three-dimensional preoperative planning workstation	TPLAN three-dimensional preoperative planning workstation	SAME
-Surgical planning system	Technician guided surgical planning with surgeon review and approval on a desktop planning station	Technician guided surgical planning with surgeon review and approval on a desktop planning station	SAME
-Bone Marker Arrays for bone registration and tracking	Active markers on femur and tibia mounted onto the bones via an attachment assembly	Active markers on femur and tibia mounted onto the bones via an attachment assembly	SAME
-Surgical Exposure	Similar to traditional surgical exposure	Similar to traditional surgical exposure	SAME
-Patient/Robot Registration	Preoperatively determined landmarks are compared to intraoperatively identified landmarks to complete patient bone registration	Preoperatively determined landmarks are compared to intraoperatively identified landmarks to complete patient bone registration	SAME
-Camera Tracking Technology	Six camera overhead tracking with a wide-angle field of view	Six camera overhead tracking with a wide-angle field of view	SAME
-Cut guide positioning	Robotic device places bone pins in the correct plane, then cutguide or drill block is attached to the pins and bone	Robotic device places bone pins in the correct plane, then cutguide or drill block is attached to the pins and bone	SAME

Product	TMINI[®] Miniature Robotic System (Additional Knee System)	TMINI[™] Miniature Robotic System (TMINI 1.1)	Substantial Equivalence Conclusion
510(k) number	Subject Device	K241031	SAME
Manufacturer	THINK Surgical Inc.	THINK Surgical Inc.	
-Intraoperative planning changes	Implant position can be fully adjusted, allowing deviation from the intended implant positioning philosophy and implant size	Implant position can be fully adjusted, allowing deviation from the intended implant positioning philosophy and implant size	
-Bone Preparation Technique	A surgical saw is used to cut the bone through a cut guide	A surgical saw is used to cut the bone through a cut guide	SAME
-Intraoperative Anatomic Measurements	The tracked bone arrays and bone registration data are used to determine the knee flexion angle and varus/valgus laxity	The tracked bone arrays and bone registration data are used to determine the knee flexion angle and varus/valgus laxity	SAME
-Gap Balancing	Displays the maximum space in the medial and lateral compartments in millimeters with the knee in extension and in flexion allowing the surgeon to perform gap balancing, if desired.	Displays the maximum space in the medial and lateral compartments in millimeters with the knee in extension and in flexion allowing the surgeon to perform gap balancing, if desired.	SAME
-TKA Component Implantation Technique	Implants are secured to the bone, either with or without cement using standard surgical technique provided by the implant manufacturer	Implants are secured to the bone, either with or without cement using standard surgical technique provided by the implant manufacturer	SAME
-Compatible Knee Implant Systems	- Enovis [™] EMPOWR Knee System [®] -Ortho Development [®] BKS [®] and BKS TriMax [®] Knee System -Total Joint Orthopedics Klassic [®] Knee System -United [®] U2 [™] Knee System -Medacta [®] GMK [®] Sphere / SpheriKA Knee Systems -Zimmer Biomet Anterior & Posterior Referencing Persona [®] Knee	- Enovis [™] EMPOWR Knee System [®] -Ortho Development [®] BKS [®] and BKS TriMax [®] Knee System -Total Joint Orthopedics Klassic [®] Knee System -United [®] U2 [™] Knee System	Substantially Equivalent (addition of two new implant systems)
Performance Testing			
-Full System Run Through Testing	Passed	Passed	SAME

Product	TMINI® Miniature Robotic System (Additional Knee System)	TMINI™ Miniature Robotic System (TMINI 1.1)	Substantial Equivalence Conclusion
510(k) number	Subject Device	K241031	
Manufacturer	THINK Surgical Inc.	THINK Surgical Inc.	
Cutting Accuracy -Pin & Block Placement Accuracy -Cadaver Lab Validation Testing -System Gap Balance Accuracy	Passed	Passed	SAME
*Biocompatibility Testing			
-Cytotoxicity	*Passed	Passed	SAME
-Sensitization	*Passed	Passed	SAME
-Intracutaneous Reactivity	*Passed	Passed	SAME
-Acute Systemic Toxicity	*Passed	Passed	SAME
-Pyrogenicity	*Passed	Passed	SAME

* There are no material changes to any of the direct patient contact components of the TMINI® Miniature Robotic System (Additional Knee System) as a result of Additional Knee System included in this submission; therefore, no additional biocompatibility testing was required.

Risk assessment was performed on the device in accordance with ISO 14971:2019 and THINK Surgical Risk Management procedures. The risk assessment was comprised of analysis and mitigation of the risks associated with the addition of three compatible knee implant systems. The risk assessment resulted in the identification of no new clinical hazards. Each of the clinical hazards identified through this risk assessment is a previously documented hazard associated with the use of the TMINI® Miniature Robotic System. The addition of three new implant systems does not increase the likelihood or severity of these hazards; therefore, the risks associated with the use of the device remain unchanged as compared to the predicate.

Conclusion

The TMINI® Miniature Robotic System (Additional Knee System) is substantially equivalent to the predicate, TMINI™ Miniature Robotic System cleared in K241031, in the following ways: it has the same intended use, the same technological characteristics and operating principles, and incorporates the same design and materials.

Performance testing and risk analysis has demonstrated that the performance and risk profile of the TMINI® Miniature Robotic System (Additional Knee System) is substantially equivalent to that of the predicate device and does not raise any new question of safety and effectiveness.

THINK Surgical Inc. respectfully submits that this filing contains adequate information and data to demonstrate the substantial equivalence of the TMINI® Miniature Robotic System (Additional Knee System) to the legally marketed TMINI™ Miniature Robotic System cleared via K241031.