



Medtronic Sofamor Danek USA, Inc.
Wafa Mustafa
Sr. Regulatory Affairs Specialist
1800 Pyramid Place
Memphis, Tennessee 38132

October 23, 2024

Re: K243007

Trade/Device Name: CD Horizon™ Spinal System
Regulation Number: 21 CFR 888.3070
Regulation Name: Thoracolumbosacral pedicle screw system
Regulatory Class: Class II
Product Code: NKB, KWP, KWQ
Dated: September 26, 2024
Received: September 26, 2024

Dear Wafa Mustafa:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

All medical devices, including Class I and unclassified devices and combination product device constituent parts are required to be in compliance with the final Unique Device Identification System rule ("UDI Rule"). The UDI Rule requires, among other things, that a device bear a unique device identifier (UDI) on its label and package (21 CFR 801.20(a)) unless an exception or alternative applies (21 CFR 801.20(b)) and that the dates on the device label be formatted in accordance with 21 CFR 801.18. The UDI Rule (21 CFR 830.300(a) and 830.320(b)) also requires that certain information be submitted to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). For additional information on these requirements, please see the UDI System webpage at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-system-udi-system>.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Digitally signed
by Eileen Cadel
-S
Date:
2024.10.23
15:48:03 -04'00' for

Colin O'Neill, M.B.E.
Assistant Director
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

Submission Number (if known)

K243007

Device Name

CD Horizon™ Spinal System

Indications for Use (Describe)

The CD Horizon™ Spinal System with or without Sextant™ instrumentation is intended for posterior, non-cervical fixation as an adjunct to fusion for the following indications: degenerative disc disease (DDD - defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies), spondylolisthesis, trauma (i.e. fracture or dislocation), spinal stenosis, curvatures (i.e. scoliosis, kyphosis, or lordosis), tumor, pseudarthrosis, and/or failed previous fusion. Except for hooks, when used as an anterolateral thoracic/lumbar system, CD Horizon™ Spinal System titanium, cobalt chrome, and stainless steel implants may also be used for the same indications as an adjunct to fusion. With the exception of DDD, CD Horizon™ Legacy™ 3.5mm rods and associated components may be used for the aforementioned indications in skeletally mature patients as an adjunct to fusion. The 3.5mm rods may be used for the specific pediatric indications noted below.

When used for posterior non-cervical pedicle screw fixation in pediatric patients, CD Horizon™ Spinal System titanium, cobalt chrome, and stainless steel implants are indicated as an adjunct to fusion to treat progressive spinal deformities (i.e. scoliosis, kyphosis, or lordosis) including idiopathic scoliosis, neuromuscular scoliosis, and congenital scoliosis. Additionally, the CD Horizon™ Spinal System is intended to treat pediatric patients diagnosed with the following conditions: spondylolisthesis/ spondylolysis, fracture caused by tumor and/or trauma, pseudarthrosis, and/or failed previous fusion. These devices are to be used with autograft and/or allograft. Pediatric pedicle screw fixation is limited to a posterior approach.

The CD Horizon™ PEEK rods are intended to provide posterior supplemental fixation when used with an interbody fusion cage for patients diagnosed with DDD. These DDD patients may also have up to Grade 1 spondylolisthesis or retrolisthesis at the involved level. This device is intended for 1-2 level use in the lumbosacral spine (L2 – S1) in skeletally mature patients. The device is intended for use with an interbody fusion cage at the instrumented level and is not intended for stand-alone use. The CD Horizon™ Spire™ plate is a posterior, single-level, non-pedicle supplemental fixation device intended for use in the non-cervical spine (T1-S1) as an adjunct to fusion in skeletally mature patients. It is intended for plate fixation/attachment to spinous processes for the purpose of achieving supplemental fixation in the following conditions: DDD (as previously defined), spondylolisthesis, trauma, and/or tumor.

To achieve additional levels of fixation, CD Horizon™ Spinal System rods may be connected to the Vertex™ Reconstruction System with the Vertex™ rod connector. Refer to the Vertex™ Reconstruction System package insert for a list of the Vertex™ indications of use.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

MEDTRONIC
CD HORIZON™ Spinal System

September 26, 2024

I. Submitter	Medtronic Sofamor Danek, USA Inc. 1800 Pyramid Place Memphis, Tennessee 38132
Contact Person	Wafa Mustafa Sr. Regulatory Affairs Specialist Email: wafa.mustafa@medtronic.com
II. Name of Device	CD Horizon™ Spinal System
Common Name	Pedicle Screw System
Classification Name	Thoracolumbosacral Pedicle Screw System
Classification	Class II
Product Codes	NKB, KWP, and KWQ (888.3070, 888.3060, 888.3050)
III. Predicate Devices	<u>Primary Predicate:</u> 1. CD HORIZON® Spinal System (K090390, S.E. 05/15/2009) <u>Additional Predicates:</u> 2. CD Horizon™ Spinal System (K210637, S.E. 4/30/2021) 3. CD Horizon™ Spinal System (K211958, S.E. 07/21/2021) 4. CD Horizon™ Spinal System (K221244, S.E. 05/25/2022) 5. CD Horizon™ Spinal System (K233951, S.E. 03/27/2024) <i>The predicates have not been subject to a design related recall.</i>

IV. Description	<u>CD Horizon™ Spinal System</u> The CD Horizon™ Spinal System consists of a variety of shapes and sizes of rods, hooks, screws, Crosslink™ Plates, staples, and connecting components, as well as implant components from other Medtronic spinal systems which can be rigidly locked into a variety of configurations, with each construct being tailor-made for the individual case. A subset of CD Horizon™ Spinal System components may be used for posterior pedicle screw fixation in pediatric cases. These constructs may be comprised of a variety of shapes and sizes of rods (ranging in diameter from 3.5mm to 6.35mm), hooks, screws, Crosslink™ plates, and connecting components. Similar to the CD Horizon™ implants used in adult cases, these components can be rigidly locked into a variety of configurations, with each construct being tailor-made for the individual case. Certain components within the CD Horizon™ Spinal System are specifically excluded for use in pediatric patients. These include PEEK rods and Spire™ plates. Screws used in pediatric cases are only cleared for use via a posterior approach. All components used in pediatric cases are fabricated from medical grade stainless steel, medical grade titanium, titanium alloy, and medical grade cobalt-chromium-molybdenum alloy. Certain implant components from other Medtronic spinal systems can be used with the CD Horizon™ Spinal System in non-pediatric cases. These components include TSRH™ rods, hooks, screws, plates; Crosslink™ plates, connectors, staples, and washers; GDLH™ rods, hooks, and connectors; Crosslink™ bar and connectors and Medtronic multi-axial rods and screws. Note that certain components are specifically designed to connect to specific rod diameters, while other components can connect to multiple rod diameters. Care should be taken so the correct components are used in the spinal construct. CD Horizon™ hooks are intended for posterior use only. CD Horizon™ staples and associated screws are intended for anterior use only. However, for patients of smaller stature and pediatric patients, CD Horizon™ 4.5mm rods and associated components may be used posteriorly. CD Horizon™ Spinal System implant components are fabricated from medical grade stainless steel, medical grade titanium, titanium alloy, medical grade cobalt-chromium-molybdenum alloy, or medical grade PEEK Optima-LT1. Certain CD Horizon™ Spinal System components may be coated with hydroxyapatite.
-------------------------------	--

V. Indications for Use	<u>CD Horizon™ Spinal System</u> The CD Horizon™ Spinal System with or without Sextant™ instrumentation is intended for posterior, non-cervical fixation as an adjunct to fusion for the following indications: degenerative disc disease (DDD - defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies), spondylolisthesis, trauma (i.e. fracture or dislocation), spinal stenosis, curvatures (i.e. scoliosis, kyphosis, or lordosis), tumor, pseudarthrosis, and/or failed previous fusion. Except for hooks, when used as an anterolateral thoracic/lumbar system, CD Horizon™ Spinal System titanium, cobalt chrome, and stainless steel implants may also be used for the same indications as an adjunct to fusion. With the exception of DDD, CD Horizon™ Legacy™ 3.5mm rods and associated components may be used for the aforementioned indications in skeletally mature patients as an adjunct to fusion. The 3.5mm rods may be used for the specific pediatric indications noted below. When used for posterior non-cervical pedicle screw fixation in pediatric patients, CD Horizon™ Spinal System titanium, cobalt chrome, and stainless steel implants are indicated as an adjunct to fusion to treat progressive spinal deformities (i.e. scoliosis, kyphosis, or lordosis) including idiopathic scoliosis, neuromuscular scoliosis, and congenital scoliosis. Additionally, the CD Horizon™ Spinal System is intended to treat pediatric patients diagnosed with the following conditions: spondylolisthesis/ spondylolysis, fracture caused by tumor and/or trauma, pseudarthrosis, and/or failed previous fusion. These devices are to be used with autograft and/or allograft. Pediatric pedicle screw fixation is limited to a posterior approach. The CD Horizon™ PEEK rods are intended to provide posterior supplemental fixation when used with an interbody fusion cage for patients diagnosed with DDD. These DDD patients may also have up to Grade 1 spondylolisthesis or retrolisthesis at the involved level. This device is intended for 1-2 level use in the lumbosacral spine (L2 – S1) in skeletally mature patients. The device is intended for use with an interbody fusion cage at the instrumented level and is not intended for stand-alone use. The CD Horizon™ Spire™ plate is a posterior, single-level, non-pedicle supplemental fixation device intended for use in the non-
--------------------------------------	--

	cervical spine (T1-S1) as an adjunct to fusion in skeletally mature patients. It is intended for plate fixation/attachment to spinous processes for the purpose of achieving supplemental fixation in the following conditions: DDD (as previously defined), spondylolisthesis, trauma, and/or tumor. To achieve additional levels of fixation, CD Horizon™ Spinal System rods may be connected to the Vertex™ Reconstruction System with the Vertex™ rod connector. Refer to the Vertex™ Reconstruction System package insert for a list of the Vertex™ indications of use.
VI. Comparison of Technological Characteristics with the Predicate Devices	The subject devices have the same intended use, indications for use, materials, similar overall design, fundamental technology, sterilization, and surgical technique as the following CD Horizon™ Spinal System predicates: • K090390, S.E. 05/15/2009 • K210637, S.E. 4/30/2021 • K211958, S.E. 07/21/202 • K221244, S.E. 05/25/2022 • K233951, S.E. 03/27/2024 The subject and predicate implants have the same function and fundamental scientific technology.
VII. Performance Data	In accordance with the Guidance for Industry and FDA Staff – Spinal System 510(k)'s, Medtronic has evaluated the subject devices to demonstrate substantial equivalence to the predicate devices. The subject devices have been tested or rationalized based on if Medtronic believes that testing is not warranted for the subject

	devices as they do not present a new worst case when compared to the predicates. Testing and/or Rationales were completed for the following: • ASTM F1717: Static Compression, Static Torsion, Compression Fatigue • ASTM F1798: Axial Grip, Axial Torsion, Flexion Extension Static, Flexion Extension Fatigue • ASTM F2503: MRI Compatibility Evaluation For the tested subject devices, the pre-determined acceptance criteria was met for all tests. For subject devices that are rationalized, all existing predicate data previously provided in the predicate 510(k)s is still applicable. Therefore, Medtronic believes the design verification testing demonstrated that the subject devices are substantially equivalent to the predicate devices.
VIII. Conclusion	Based on the supporting evidence provided in this premarket notification, Medtronic believes the subject devices are substantially equivalent to the predicate devices.