



December 9, 2024

Smith & Nephew, Inc.
Jessica Roach
Regulatory Affairs Specialist II
1450 E Brooks Road
Memphis, Tennessee 38116

Re: K243364

Trade/Device Name: TRIGEN META-TAN Trochanteric Antegrade Nail
Regulation Number: 21 CFR 888.3020
Regulation Name: Intramedullary fixation rod
Regulatory Class: Class II
Product Code: HSB, HWC
Dated: October 29, 2024
Received: October 29, 2024

Dear Jessica Roach:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device"

(<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

All medical devices, including Class I and unclassified devices and combination product device constituent parts are required to be in compliance with the final Unique Device Identification System rule ("UDI Rule"). The UDI Rule requires, among other things, that a device bear a unique device identifier (UDI) on its label and package (21 CFR 801.20(a)) unless an exception or alternative applies (21 CFR 801.20(b)) and that the dates on the device label be formatted in accordance with 21 CFR 801.18. The UDI Rule (21 CFR 830.300(a) and 830.320(b)) also requires that certain information be submitted to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). For additional information on these requirements, please see the UDI System webpage at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-system-udi-system>.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Farzana
Sharmin -S

Digitally signed by
Farzana Sharmin -S
Date: 2024.12.09
17:54:39 -05'00'

Farzana Sharmin, PhD
Assistant Director
DHT6A: Division of Joint Arthroplasty Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

Submission Number (if known)

K243364

Device Name

TRIGEN META-TAN Trochanteric Antegrade Nail

Indications for Use (Describe)

TRIGEN META-TAN Trochanteric Antegrade Nail

The TRIGEN META-TAN Trochanteric Antegrade Nail is indicated for fixation of fractures that occur in and between the proximal third and distal fourth of the femur including simple long bone fractures, severely comminuted, spiral, and segmental fractures, and supracondylar fractures.

In addition, TRIGEN META-TAN Nails contain holes/slots proximally to accept screws that thread into the femoral head for compression and rotational stability and are indicated for the following: subtrochanteric fractures; intertrochanteric fractures; and intracapsular fractures.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) #:

510(k) Summary

Prepared on: 2024-12-09

Contact Details

[21 CFR 807.92\(a\)\(1\)](#)

Applicant Name	Smith & Nephew, Inc.
Applicant Address	1450 E Brooks Road Memphis TN 38116 United States
Applicant Contact Telephone	978-474-6318
Applicant Contact	Ms. Jessica Roach
Applicant Contact Email	jessica.roach@smith-nephew.com

Device Name

[21 CFR 807.92\(a\)\(2\)](#)

Device Trade Name	TRIGEN META-TAN Trochanteric Antegrade Nail
Common Name	Intramedullary fixation rod
Classification Name	Rod, Fixation, Intramedullary And Accessories
Regulation Number	888.3020
Product Code(s)	HSB, HWC

Legally Marketed Predicate Devices

[21 CFR 807.92\(a\)\(3\)](#)

Predicate #	Predicate Trade Name (Primary Predicate is listed first)	Product Code
K092748	SURESHOT TAN NAILS AND ACCESSORIES	HSB, HWC

Device Description Summary

[21 CFR 807.92\(a\)\(4\)](#)

The device in scope of this submission is the TRIGEN META-TAN Trochanteric Antegrade Nail.

TRIGEN META-TAN Trochanteric Antegrade Nails are intended to be used as aids to normal fracture healing of the femur. They are made from Ti-6Al-4V material and available in a nail size range of 30 to 50 cm in 2 cm increments. They are Gamma sterilized and intended for single-use only.

Intended Use/Indications for Use

[21 CFR 807.92\(a\)\(5\)](#)

TRIGEN META-TAN Trochanteric Antegrade Nail

The TRIGEN META-TAN Trochanteric Antegrade Nail is indicated for fixation of fractures that occur in and between the proximal third and distal fourth of the femur including simple long bone fractures, severely comminuted, spiral, and segmental fractures, and supracondylar fractures.

In addition, TRIGEN META-TAN Nails contain holes/slots proximally to accept screws that thread into the femoral head for compression and rotational stability and are indicated for the following: subtrochanteric fractures; intertrochanteric fractures; and intracapsular fractures.

Indications for Use Comparison

[21 CFR 807.92\(a\)\(5\)](#)

The subject TRIGEN META-TAN Trochanteric Antegrade Nail indications fall within the predicate devices' indications for use.

Technological Comparison

[21 CFR 807.92\(a\)\(6\)](#)

The overall technological characteristic including device design and material of the subject devices are similar to that of the predicate Smith & Nephew system cleared under the premarket notification SURESHOT TAN (K092748, 04/26/2010).

As result, all relevant testing makes references to existing information previously provided to the agency.

Non-Clinical and/or Clinical Tests Summary & Conclusions

[21 CFR 807.92\(b\)](#)

- MR Compatibility per ASTM F2182-19e2, ASTM F2052-15, ASTM F2213-2017, and ASTM F2119-07(2013) (META-TAN Nail)

No clinical tests were performed to support safety and efficacy of the subject device.

The performance bench tests were used as a basis for the determination of substantial equivalence. The results of the equivalence testing demonstrate acceptable outcomes. The results of these tests show that when compared to the predicate device, the subject device has met the required criteria and there are no additional risks.