



February 6, 2025

STERIS Corporation
Karly Eitzman
Regulatory Affairs Specialist
5960 Heisley Road
Mentor, Ohio 44060

Re: K250061
Trade/Device Name: Celerity Incubator
Regulation Number: 21 CFR 880.2800
Regulation Name: Sterilization Process Indicator
Regulatory Class: Class II
Product Code: FRC
Dated: January 8, 2025
Received: January 10, 2025

Dear Karly Eitzman:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device"

(<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

All medical devices, including Class I and unclassified devices and combination product device constituent parts are required to be in compliance with the final Unique Device Identification System rule ("UDI Rule"). The UDI Rule requires, among other things, that a device bear a unique device identifier (UDI) on its label and package (21 CFR 801.20(a)) unless an exception or alternative applies (21 CFR 801.20(b)) and that the dates on the device label be formatted in accordance with 21 CFR 801.18. The UDI Rule (21 CFR 830.300(a) and 830.320(b)) also requires that certain information be submitted to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). For additional information on these requirements, please see the UDI System webpage at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-system-udi-system>.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Stephen A. Anisko -S

Digitally signed by
Stephen A. Anisko -S
Date: 2025.02.06 17:39:07
-05'00'

for: Christopher K. Dugard
Division Director
DHT4C: Division of Infection
Control Devices
OHT4: Office of Surgical and
Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

Submission Number (if known)

K250061

Device Name

Celerity Incubator

Indications for Use (Describe)

Use the Celerity Incubator to incubate and automatically read STERIS Celerity Biological Indicators for Steam and Vaporized Hydrogen Peroxide sterilization at 55°C-60°C for a fluorescent result.

Type of Use (Select one or both, as applicable)

☐

Prescription Use (Part 21 CFR 801 Subpart D)

☒

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



**510(k) Summary
For
Celerity Incubator**

STERIS Corporation
5960 Heisley Road
Mentor, OH 44060
Phone: (440) 354-2600
Fax No: (440) 357-9198

Contact: Mrs. Karly Eitzman
Regulatory Affairs Specialist
Tel: 440-392-7442
karly_eitzman@steris.com

Submission Date: January 8, 2025

K Number: K250061

STERIS Special 510(k) PREMARKET NOTIFICATION
Celerity Incubator

1. Device Name

Trade Name: **Celerity Incubator**
Device Class: Class II
Common/usual Name: Reader/Incubator
Classification Name: Biological Sterilization Process Indicator
Classification Number: 21 CFR 880.2800
Product Code: FRC

2. Predicate Device

K223715 – Celerity Incubator

3. Device Description

The Celerity Incubator incubates and reads fluorescent Vaporized Hydrogen Peroxide (VHP) and Steam Biological Indicators (BIs). The BI is incubated at 59°C with an acceptable tolerance of -4°C/+1°C. During incubation, the BI is monitored for a potential fluorescence signal generated as a result of the production of α -glucosidase. When a growth response has been detected or when the required incubation time has elapsed, the incubator indicates the results to the user.

4. Indications for Use

Use the Celerity Incubator to incubate and automatically read STERIS Celerity Biological Indicators for Steam and Vaporized Hydrogen Peroxide sterilization at 55°C - 60°C for a fluorescent result.

5. Technological Characteristics Comparison Table

A comparison of technical characteristics between the proposed and predicate devices is summarized in **Table 1**.

Table 1. Technological Characteristics Comparison Table to predicate device

Feature	Proposed Device Celerity Incubator	Predicate Device Celerity Incubator (K223715)	Comparison
Indications for Use	Use the Celerity Incubator to incubate and automatically read STERIS Celerity	Use the Celerity Incubator to incubate and automatically read STERIS Celerity	Same

STERIS Special 510(k) PREMARKET NOTIFICATION
Celerity Incubator

Feature	Proposed Device Celerity Incubator	Predicate Device Celerity Incubator (K223715)	Comparison
	Biological Indicators for Steam and Vaporized Hydrogen Peroxide sterilization at 55°C - 60°C for a fluorescent result.	Biological Indicators for Steam and Vaporized Hydrogen Peroxide sterilization at 55°C - 60°C for a fluorescent result.	
Basis of Readout	Photodiode detects fluorescence produced by enzymatic activity that results from viable biological indicator organisms.	Photodiode detects fluorescence produced by enzymatic activity that results from viable biological indicator organisms.	Same
Incubation Temperature Range	55 - 60 °C	55 - 60 °C	Same
Voltage Range	100-240 VAC with 12 VDC conversion	100-240 VAC with 12 VDC conversion	Same
Test Capacity	8 wells	8 wells	Same
Calibration	Factory calibration – no calibration by customer	Factory calibration – no calibration by customer	Same
Incubation Time	5 minutes or 20 minutes, depending on Biological Indicator	5 minutes or 20 minutes, depending on the Biological Indicator	Same
Fluorescence Detection	UV LEDs are used to excite the fluorescent molecule produced by enzyme cleavage of the fluorogenic substrate contained in the BI's media. The emitted light is detected by a photodiode.	UV LEDs are used to excite the fluorescent molecule produced by enzyme cleavage of the fluorogenic substrate contained in the BI's media. The emitted light is detected by a photodiode.	Same
Indication of Results	Positive – audible alarm, visual LED lights and screen Negative – optional alarm, visual indication with LED lights and LCD screen User must acknowledge results	Positive – audible alarm, visual LED lights and screen Negative – optional alarm, visual indication with LED lights and LCD screen User must acknowledge results	Same
System Operation	The reader/incubator wells are arranged in 2 banks of 4 wells and preset to 59°C. The user scans the barcode on the label of an activated BI using	The reader/incubator wells are arranged in 2 banks of 4 wells and preset to 59°C. The user scans the barcode on the label of an activated BI using	Same

STERIS Special 510(k) PREMARKET NOTIFICATION
Celerity Incubator

Feature	Proposed Device Celerity Incubator	Predicate Device Celerity Incubator (K223715)	Comparison
	the system's barcode scanner and places it into an open well. The system detects the well the BI was placed into and begins measurement of fluorescence; a blinking yellow light indicates the incubation is in process and the read initiated. The System uses information from the barcode to apply the appropriate algorithm to the well. A "positive" reading is interpreted as an indication of a potential sterilization cycle failure. A "positive" finding is indicated to the user by red light on the front panel adjacent to the well, by an audible alarm, and by text displayed on the LCD screen. The alarm must be muted by the operator when a positive result is obtained. The LCD screen provides instructions for the user to turn off the alarm for that specific BI. Should another BI become "positive", the alarm will sound again and the above actions are repeated. If the result is not positive at the end of the incubation time, the result is negative. Negative results are identified by a green light on the front panel adjacent to the well with the "negative" BI and by text on the LCD. In addition, an optional alarm is available for negative results.	the system's barcode scanner and places it into an open well. The system detects the well the BI was placed into and begins measurement of fluorescence; a blinking yellow light indicates the incubation is in process and the read initiated. The System uses information from the barcode to apply the appropriate algorithm to the well. A "positive" reading is interpreted as an indication of a potential sterilization cycle failure. A "positive" finding is indicated to the user by red light on the front panel adjacent to the well, by an audible alarm, and by text displayed on the LCD screen. The alarm must be muted by the operator when a positive result is obtained. The LCD screen provides instructions for the user to turn off the alarm for that specific BI. Should another BI become "positive", the alarm will sound again and the above actions are repeated. If the result is not positive at the end of the incubation time, the result is negative. Negative results are identified by a green light on the front panel adjacent to the well with the "negative" BI and by text on the LCD. In addition, an optional alarm is available for negative results.	

6. Summary of Non-Clinical Performance Testing

STERIS Special 510(k) PREMARKET NOTIFICATION
Celerity Incubator

Testing was performed to evaluate performance and demonstrate substantial equivalence to the predicate as summarized in **Table 2**.

Table 2. Performance Testing

Test	Acceptance Criteria	Result
Software Validation	The software that controls the system was validated and determined to operate effectively and as designed.	Pass
Qualification Testing with intended Biological Indicators	Fluorescent Read meets $\geq 97\%$ alignment with 7-day grow out per FDA guidance on Reduced Incubation Time for intended Celerity 5 HP Biological Indicators and Celerity 20 HP Biological Indicators	Pass
Recognition of additional barcodes	Incubator reads barcodes correctly	Pass

7. Conclusion

The Celerity Incubator has met the established performance criteria. Based on the intended use, technological characteristics and non-clinical performance data, the subject device is as safe, as effective, and performs as well as the legally marketed predicate device (K223715), Class II (21 CFR 880.2800), product code FRC.