



July 18, 2025

Beckman Coulter, Inc.  
Elaine Duncan  
Staff Regulatory Affairs  
1584 Enterprise Blvd.  
West Sacramento, California 95691

Re: K250084

Trade/Device Name: MicroScan Dried Gram-Negative MIC/Combo Panels with Aztreonam (AZT)  
(0.5-64 µg/mL)

Regulation Number: 21 CFR 866.1640

Regulation Name: Antimicrobial Susceptibility Test Powder

Regulatory Class: Class II

Product Code: LTT, JWY, LRG, LTW

Dated: June 20, 2025

Received: June 20, 2025

Dear Elaine Duncan:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

FDA's substantial equivalence determination also included the review and clearance of your Predetermined Change Control Plan (PCCP). Under section 515C(b)(1) of the Act, a new premarket notification is not required for a change to a device cleared under section 510(k) of the Act, if such change is consistent with an established PCCP granted pursuant to section 515C(b)(2) of the Act. Under 21 CFR 807.81(a)(3), a new premarket notification is required if there is a major change or modification in the intended use of a device,

or if there is a change or modification in a device that could significantly affect the safety or effectiveness of the device, e.g., a significant change or modification in design, material, chemical composition, energy source, or manufacturing process. Accordingly, if deviations from the established PCCP result in a major change or modification in the intended use of the device, or result in a change or modification in the device that could significantly affect the safety or effectiveness of the device, then a new premarket notification would be required consistent with section 515C(b)(1) of the Act and 21 CFR 807.81(a)(3). Failure to submit such a premarket submission would constitute adulteration and misbranding under sections 501(f)(1)(B) and 502(o) of the Act, respectively.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801 and Part 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

All medical devices, including Class I and unclassified devices and combination product device constituent parts are required to be in compliance with the final Unique Device Identification System rule ("UDI Rule"). The UDI Rule requires, among other things, that a device bear a unique device identifier (UDI) on its label and package (21 CFR 801.20(a)) unless an exception or alternative applies (21 CFR 801.20(b)) and that the dates on the device label be formatted in accordance with 21 CFR 801.18. The UDI Rule (21 CFR 830.300(a) and 830.320(b)) also requires that certain information be submitted to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). For additional information on these requirements, please see the UDI System webpage at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-system-udi-system>.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part

803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

  
**Ribhi Shawar -S**

Ribhi Shawar, Ph.D. (ABMM)  
Branch Chief, General Bacteriology and Antimicrobial  
Susceptibility Branch  
Division of Microbiology Devices  
OHT7: Office of In Vitro Diagnostics  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K250084

Device Name

MicroScan Dried Gram-Negative MIC/Combo Panels with Aztreonam (AZT) (0.5-64 µg/mL)

Indications for Use (Describe)

The MicroScan Dried Gram-Negative MIC/Combo Panel is used to determine quantitative and qualitative antimicrobial agent susceptibility of colonies grown on solid media of rapidly growing aerobic and facultative anaerobic gram-negative bacilli. After inoculation, panels are incubated for 16-20 hours at 35°C ± 1°C in a non-CO2 incubator, and read either visually or with MicroScan instrumentation, according to the Package Insert.

This particular submission is for the addition of the antimicrobial aztreonam at concentrations of 0.5-64 µg/mL to the test panel. Testing is indicated for Enterobacterales and Pseudomonas aeruginosa, as recognized by the FDA Susceptibility Test Interpretive Criteria (STIC) webpage.

The MicroScan Dried Gram-Negative MIC/Combo Panels with Aztreonam (AZT) (0.5-64 µg/mL) has demonstrated acceptable performance with the following organisms:

Enterobacterales (Citrobacter freundii complex, Citrobacter koseri, Escherichia coli, Klebsiella oxytoca, Klebsiella pneumoniae, Proteus mirabilis, Morganella morganii, Yersinia enterocolitica)  
Pseudomonas aeruginosa

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRAStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary

### 510(k) Submission Information:

Submitter's Name: Beckman Coulter  
Contact Person: Elaine Duncan, Staff Regulatory Affairs  
Address: 1584 Enterprise Blvd.  
West Sacramento, CA 95691  
Phone: 916-318-0652  
Date prepared: July 11, 2025  
Product Name: Microdilution Minimum Inhibitory Concentration (MIC) Panels  
Trade Name: MicroScan Dried Gram-Negative MIC/Combo Panels with Aztreonam (AZT) (0.5 – 64 µg/mL)  
Classification: 21 CFR 866.1640 Antimicrobial Susceptibility Test Powder; Class II  
Product Code: LTT, JWY, LRG, LTW  
510(k) Notification: Updated antimicrobial agent - Aztreonam  
Predicate device: MicroScan Dried Gram-Negative MIC/Combo Panels Ceftazidime – (0.5 – 64 µg/mL) - (K202343)

### 510(k) Summary:

**Device Description:** MicroScan Dried Gram-Negative MIC/Combo Panels are designed for use in determining quantitative and qualitative antimicrobial agent susceptibility of colonies grown on solid media of rapidly growing aerobic and facultative anaerobic gram-negative bacilli.

The principle of MicroScan panels with antimicrobial susceptibility tests are miniaturizations of the broth dilution susceptibility test that have been diluted in broth and dehydrated. Various antimicrobial agents are diluted in broth to concentrations bridging the range of clinical interest. Panels are rehydrated with water after inoculation with a standardized suspension of the organism. After incubation in a non-CO<sub>2</sub> incubator for 16-20 hours, the minimum inhibitory concentration (MIC) for the test organism is read by determining the lowest antimicrobial concentration showing inhibition of growth.

This product is single-use and intended for laboratory professional use.

**Intended Use:** For Use with MicroScan Dried Gram-Negative MIC/Combo and Dried Gram-Negative Breakpoint Combo panels. MicroScan panels are designed for use in determining antimicrobial agent susceptibility and/or identification to the species level of aerobic and facultatively anaerobic gram-negative bacilli.

**Indications for Use:** The MicroScan Dried Gram-Negative MIC/Combo Panel is used to determine quantitative and qualitative antimicrobial agent susceptibility of colonies grown on solid media of rapidly growing aerobic and facultatively anaerobic gram-negative bacilli. After inoculation, panels are incubated for 16-20 hours at 35°C ± 1°C in a non-CO<sub>2</sub> incubator, and read either visually or with MicroScan instrumentation, according to the Package Insert.

This particular submission is for the addition of the antimicrobial aztreonam at concentrations of 0.5 - 64 µg/mL to the test panel. Testing is indicated for Enterobacterales and *Pseudomonas aeruginosa*, as recognized by the FDA Susceptibility Test Interpretive Criteria (STIC) webpage.

The MicroScan Dried Gram-Negative MIC/Combo Panels with Aztreonam (Azt) (0.5 - 64 µg/mL) has demonstrated acceptable performance with the following organisms:

Enterobacterales (*Citrobacter freundii* complex, *Citrobacter koseri*, *Escherichia coli*, *Klebsiella oxytoca*, *Klebsiella pneumoniae*, *Proteus mirabilis*, *Morganella morganii*, *Yersinia enterocolitica*)

*Pseudomonas aeruginosa*

**Substantial Equivalence Information:** The similarities and differences of the MicroScan Dried Gram Negative MIC/Combo Panels with Aztreonam (Azt) (0.5 – 64 µg/mL) compared to the predicate device, MicroScan Dried Gram-Positive MIC/Combo Panels with Ceftazidime – (K202343), are described in **Table 1** below.

**Table 1: Substantial Equivalence**

<b>Similarities</b>		
<b>Item</b>	<b>Proposed</b>	<b>Predicate</b>
	MicroScan Dried Gram-Negative MIC/Combo Panels – Aztreonam	MicroScan Dried Gram-Negative MIC/Combo Panels – Ceftazidime (K202343)
<i>Intended Use</i>	Determination of susceptibility to Aztreonam with gram-negative bacteria	Determination of susceptibility to Ceftazidime with gram-negative bacteria
<i>Technology</i>	Overnight Microdilution MIC Susceptibility Tests	Same
<i>Specimen</i>	Isolated colonies from cultures	Same
<i>Incubation Temperature</i>	35° C ± 1 ° C	Same
<i>Incubation Atmosphere</i>	Aerobic	Same
<i>Incubation Time</i>	16 – 20 hours	Same
<i>Reading Method</i>	Automated or Manual	Same
<b>Differences</b>		
<b>Item</b>	<b>Proposed</b>	<b>Predicate</b>
	MicroScan Dried Gram-Negative MIC/Combo Panels – Aztreonam	MicroScan Dried Gram-Negative MIC/Combo Panels – Ceftazidime (K202343)
<i>Antimicrobial Agent</i>	Dried Aztreonam 0.5 – 64 µg/mL	Dried Ceftazidime 0.5 – 64 µg/mL

**Performance and Conclusion:** The proposed MicroScan Dried Gram-Negative MIC/Combo Panel demonstrated substantially equivalent performance when compared with a CLSI frozen Reference Panel, as defined in the FDA document "Class II Special Controls Guidance Document: Antimicrobial Susceptibility Test (AST) Systems; Guidance for Industry and FDA", dated August 28, 2009. The Premarket Notification (510[k]) presents data in support of the MicroScan Dried Gram-Negative MIC/Combo Panel with Aztreonam.

The external evaluations were conducted with contemporary and stock Efficacy isolates and stock Challenge strains. The external evaluations were designed to confirm the acceptability of the proposed Dried Gram-Negative Panel by comparing its performance with a CLSI frozen Reference panel. The Dried Gram-Negative Panel inoculated with Prompt and read on the WalkAway instrument demonstrated acceptable performance with an Enterobacteriales Essential Agreement (EA) of 91.0% and Categorical Agreement (CA) of 93.1%, and *Pseudomonas aeruginosa* Essential Agreement (EA) of 91.2% and Categorical Agreement (CA) of 86.0%\* for Aztreonam when compared with the frozen Reference panel. \*Essential agreement of evaluable isolates 90.3% and most of the categorical discrepancies were minor errors.



Inoculum and instrument reproducibility testing demonstrated acceptable reproducibility and precision with Aztreonam, regardless of which inoculum method (i.e., Turbidity or Prompt), or instrument (autoSCAN-4 instrument or WalkAway system) was used.

Quality Control testing demonstrated acceptable results for Aztreonam.

© 2025 Beckman Coulter, Inc. All rights reserved. All other trademarks are the property of their respective owners. May be covered by one or more patent. See [www.beckmancoulter.com/patents](http://www.beckmancoulter.com/patents).

Beckman Coulter, the stylized logo, and the Beckman Coulter product and service marks mentioned herein are trademarks or registered trademarks of Beckman Coulter, Inc. in the United States and other countries.