



June 13, 2025

Abbott Diagnostics Scarborough, Inc.  
Kristen Cyr  
Regulatory Affairs Specialist  
10 Southgate Road  
Scarborough, Maine 04074

Re: K250273

Trade/Device Name: BinaxNOW COVID-19 Ag Card

Regulation Number: 21 CFR 866.3982

Regulation Name: Simple Point-Of-Care Device To Directly Detect SARS-Cov-2 Viral Targets From  
Clinical Specimens In Near-Patient Settings

Regulatory Class: Class II

Product Code: QVF

Dated: January 30, 2025

Received: January 30, 2025

Dear Kristen Cyr:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device"

(<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801 and Part 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

All medical devices, including Class I and unclassified devices and combination product device constituent parts are required to be in compliance with the final Unique Device Identification System rule ("UDI Rule"). The UDI Rule requires, among other things, that a device bear a unique device identifier (UDI) on its label and package (21 CFR 801.20(a)) unless an exception or alternative applies (21 CFR 801.20(b)) and that the dates on the device label be formatted in accordance with 21 CFR 801.18. The UDI Rule (21 CFR 830.300(a) and 830.320(b)) also requires that certain information be submitted to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). For additional information on these requirements, please see the UDI System webpage at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-system-udi-system>.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

**JOSEPH BRIGGS -S**

Joseph Briggs, Ph.D.

Deputy Division Director

Division of Microbiology Devices

OHT7: Office of In Vitro Diagnostics

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K250273

Device Name  
BinaxNOW COVID-19 Ag Card

### Indications for Use (Describe)

The BinaxNOW COVID-19 Ag Card is a lateral flow immunochromatographic assay for the rapid, qualitative detection of the SARS-CoV-2 nucleocapsid protein antigen directly in anterior nasal swab specimens from individuals with signs and symptoms of upper respiratory tract infection (i.e., symptomatic). The test is intended for use as an aid in the diagnosis of SARS-CoV-2 infections (COVID-19) in symptomatic individuals when either: tested at least twice over three days with at least 48 hours between tests; or when tested once, and negative by the BinaxNOW COVID-19 Ag Card and followed up with a molecular test.

A negative test is presumptive and does not preclude SARS-CoV-2 infection; it is recommended these results be confirmed by a molecular SARS-CoV-2 assay.

Positive results do not rule out co-infection with other bacteria or viruses and should not be used as the sole basis for diagnosis, treatment, or other patient management decisions.

Performance characteristics for SARS-CoV-2 were established from November 2020 to July 2022, when SARS-COV-2 Delta and Omicron were dominant. When other SARS-CoV-2 virus variants are emerging, performance characteristics may vary.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(K) SUMMARY

---

**Preparation Date:** June 11, 2025

### **CONTACT DETAILS**

Applicant Name: Abbott Diagnostics Scarborough, Inc.

Applicant Address: 10 Southgate Road, Scarborough, Maine 04074 USA

Applicant Contact: Ms. Kristen Cyr

Applicant Contact Email : [kristen.cyr@abbott.com](mailto:kristen.cyr@abbott.com)

Applicant Contact Phone : (207) 210-4311

### **DEVICE NAME**

Device Trade Name: BinaxNOW COVID-19 Ag Card

Common Name: BinaxNOW COVID-19

Classification Name: Simple Point-Of-Care Device To Directly Detect Sars-Cov-2 Viral Targets From Clinical Specimens In Near-Patient Settings

Regulation Number: 866.3982

Product Code: QVF

### **PREDICATE DEVICE**

K231187, Nano-Check COVID-19 Antigen Test

### **DEVICE DESCRIPTION SUMMARY**

The BinaxNOW COVID-19 Ag Card is an immunochromatographic membrane assay that uses antibodies to detect SARS-CoV-2 nucleocapsid protein from anterior nasal swab specimens. SARS-CoV-2 specific antibodies and a control antibody are immobilized onto a membrane support as two distinct lines and combined with other reagents/pads to construct a test strip. This test strip and a well to hold the swab specimen are mounted on opposite sides of a cardboard, book-shaped hinged test card.

To perform the test, an anterior nasal swab specimen is collected from the patient, 6 drops of extraction reagent from a dropper bottle are added to the top hole of the swab well. The patient sample is inserted into the test card through the bottom hole of the swab well, and firmly pushed upwards until the swab tip is visible through the top hole. The swab is rotated 3 times clockwise and the card is closed, bringing the extracted sample into contact with the test strip. Test results are interpreted visually at 15 minutes based on the presence or absence of visually detectable pink/purple colored lines. Results should not be read after 30 minutes.

### **INTENDED USE/INDICATIONS FOR USE**

The BinaxNOW™ COVID-19 Ag Card is a lateral flow immunochromatographic assay for the rapid, qualitative detection of the SARS-CoV-2 nucleocapsid protein antigen directly in anterior nasal swab specimens from individuals with signs and symptoms of upper respiratory tract infection (i.e., symptomatic). The test is intended for use as an aid in the diagnosis of SARS-CoV-2 infections (COVID-19) in symptomatic individuals when either: tested at least twice over three days with at least 48 hours between tests; or when tested once, and negative by the BinaxNOW COVID-19 Ag Card and followed up with a molecular test.

A negative test is presumptive and does not preclude SARS-CoV-2 infection; it is recommended these results be confirmed by a molecular SARS-CoV-2 assay.

Positive results do not rule out co-infection with other bacteria or viruses and should not be used as the sole basis for diagnosis, treatment, or other patient management decisions.

Performance characteristics for SARS-CoV-2 were established from November 2020 to July 2022, when SARS-CoV-2 Delta and Omicron were dominant. When other SARS-CoV-2 virus variants are emerging, performance characteristics may vary.

### **INDICATIONS FOR USE COMPARISON**

The BinaxNOW COVID-19 Ag Card and the predicate device, Nano-Check COVID-19 Antigen Test, have the same intended use. The test systems are intended for the rapid, qualitative detection of SARS-CoV-2 directly in anterior nasal swab specimens as an aid in the diagnosis of SARS-CoV-2 infection. The devices are intended for use in near-patient settings.

### **TECHNOLOGICAL COMPARISON**

The BinaxNOW COVID-19 Ag Card and the predicate device, Nano-Check COVID-19 Antigen Test, have the same technological characteristics. Both test systems are visually read lateral flow immunoassays to detect nucleocapsid protein antigen from SARS-CoV-2 in anterior nasal swab samples.

### **NON-CLINICAL AND/OR CLINICAL TESTS SUMMARY AND CONCLUSIONS ANALYTICAL STUDIES**

#### **Analytical Sensitivity (Limit of Detection)**

The BinaxNOW COVID-19 Ag Card limit of detection (LOD) was determined by evaluating different concentrations of two heat inactivated strains of SARS-CoV-2 virus. Presumed negative natural nasal swab specimens were eluted in PBS. Swab eluates were combined and mixed thoroughly to create a clinical matrix pool to be used as the diluent. Inactivated SARS-CoV-2 virus was diluted in this natural nasal swab matrix pool to generate virus dilutions for testing. Contrived nasal swab samples were prepared by absorbing 20 microliters of each virus dilution onto the swab. The contrived swab samples were tested according to the test procedure.

For each strain, the LOD was determined as the lowest virus concentration that was detected  $\geq 95\%$  of the time (i.e., concentration at which at least 19 out of 20 replicates tested positive).

The BinaxNOW COVID-19 Ag Card LOD in natural nasal swab matrix was confirmed  $3.5 \times 10^3$  TCID<sub>50</sub>/mL for USA-WA1/2020 and  $1.6 \times 10^3$  TCID<sub>50</sub>/mL for B.1.1.529 (Omicron).

#### **Limit of Detection Study Results**

<b>Strain</b>	<b>Concentration TCID<sub>50</sub>/mL</b>	<b>Concentration TCID/swab</b>
USA-WA1/2020	$3.5 \times 10^3$	70
B.1.1.529 (Omicron)	$1.6 \times 10^3$	32.06

#### **WHO Standard Testing**

A study was performed to also determine the LoD for the BinaxNOW COVID-19 Ag Card Test in nasal samples using the WHO International Standard for SARS-CoV-2 Antigen (NIBSC 21/368) as a standardized material.

As per the WHO instructions, the international standard material was reconstituted in 0.25 mL of ultra-pure water. Following reconstitution, the ampule was left at ambient temperature for 20 minutes and then mixed thoroughly, avoiding generation of excess foam. The reconstitution of the material yielded a final stock concentration equal to  $2.0 \times 10^4$  IU/mL.

For each replicate, 20 µL of virus dilution was applied to a swab and the swab was tested according to the IFU. A preliminary LoD concentration was determined by testing a series of 2-fold dilutions of the antigen spiked into negative nasal matrix (NNM) in replicates of three (3). The lowest concentration with 3 out of 3 positive replicates was considered to be the preliminary LoD. The results of the preliminary LoD study are shown in the Table below:

#### WHO SARS-CoV-2 Standard LoD Range Finding Results

Concentration (IU/ml)	BinaxNOW COVID-19 Ag Card Results	
	# Detected	% Detected
2000	3/3	100
1000	3/3	100
<b>500</b>	<b>3/3</b>	<b>100</b>
250	0/3	0
125	0/3	0

The preliminary LoD was confirmed by testing an additional seventeen (17) replicates per dilution for a total of (20) until less than 100% positive results were obtained. The results of this testing, as shown in the table below confirmed the LoD for the WHO International Standard Antigen to be 375 IU/mL (7.5 IU/swab).

#### WHO SARS-CoV-2 Standard Confirmatory LoD Results

Concentration (IU/mL)	BinaxNOW COVID-19 Ag Card Results	
	# Detected	% Detected
500	20/20	100
<b>375</b>	<b>20/20</b>	<b>100</b>
250	4/20	20

#### Analytical Reactivity (Inclusivity)

An Analytical Reactivity (Inclusivity) study was performed to determine whether the BinaxNOW COVID-19 Ag Card is able to detect a variety of SARS-CoV-2 strains.

Vendor provided stocks of SARS-CoV-2 strains were diluted in natural nasal swab matrix to generate virus dilutions for testing. Contrived swab samples were prepared by coating 20 microliters of virus dilution onto each swab.

Each dilution of virus strain was tested n = 5 replicates. A concentration level was considered “reactive/positive” in this study if all five replicates generated a positive result.

The BinaxNOW COVID-19 Ag Card detected all strains tested at the concentrations indicated in the table below:

**Analytical Reactivity Study Results**

<b>Variant</b>	<b>Concentration (TCID<sub>50</sub>/ml in the test)</b>
Alpha (B.1.1.7)	3.50 x 10 <sup>4</sup>
Beta (B.1.351)	7.00 x 10 <sup>3</sup>
Delta (B.1.617.2)	1.75 x 10 <sup>3</sup>
Gamma (P.1)	1.75 x 10 <sup>3</sup>
Iota (B.1.526)	7.00 x 10 <sup>3</sup>
Italy-INMI1	3.50 x 10 <sup>4</sup>
Kappa (B.1.617.1)	1.05 x 10 <sup>4</sup>
Zeta (P.2)	3.50 x 10 <sup>3</sup>
Omicron (BA.2.3)	2.63 x 10 <sup>3</sup>
Omicron (BA.2.12.1)	1.31 x 10 <sup>3</sup>
Omicron (BA.2.75.5)	8.75 x 10 <sup>2</sup>
Omicron (BA.4.6)	3.50 x 10 <sup>3</sup>
Omicron (BA.5)	5.60 x 10 <sup>4</sup>
Omicron (BA.5.5)	1.10 x 10 <sup>2</sup>
Omicron (BF.5)	3.50 x 10 <sup>3</sup>
Omicron (BF.7)	1.40 x 10 <sup>4</sup>
Omicron (BQ.1)	7.00 x 10 <sup>3</sup>
Omicron (BQ.1.1)	8.75 x 10 <sup>2</sup>
Omicron (XBB)	2.80 x 10 <sup>4</sup>
Omicron (JN.1)	3.63 x 10 <sup>2</sup> IFU/mL

**Analytical Specificity (Cross Reactivity) and Microbial Interference**

Cross reactivity and potential interference of BinaxNOW COVID-19 Ag Card was evaluated by testing 28 commensal and pathogenic microorganisms (9 bacteria, 17 viruses, 1 yeast and pooled human nasal wash) that may be present in the nasal cavity. Each organism, virus, and yeast was tested n=5 in the absence or presence of heat inactivated SARS-CoV-2 virus at a concentration of 3X LOD (210 TCID<sub>50</sub>/swab). No cross-reactivity or interference was seen with the microorganisms in the table below when tested at 1 x 10<sup>6</sup> CFU/mL for bacteria and yeast and at 1 x 10<sup>5</sup> TCID<sub>50</sub>/mL for viruses.

<b>Type</b>	<b>Panel Member</b>
<b>Viruses</b>	Human Adenovirus 1
	Human Coronavirus 229E
	Human Coronavirus NL63
	Human Coronavirus OC43
	Enterovirus 70
	Human Metapneumovirus (hMPV)
	Human Parainfluenza virus 1
	Human Parainfluenza virus 2
	Human Parainfluenza virus 3
	Human Parainfluenza virus 4
	RSV A
	Rhinovirus 1A

Type	Panel Member
	MERS-coronavirus
	Human Influenza A/California/07/09
	Human Influenza A/New Caledonia/20/99
	Human Influenza A/Brisbane/02/18
	Human Influenza B/Wisconsin/1/10
<b>Bacteria</b>	<i>Bordetella pertussis</i>
	<i>Chlamydia pneumoniae</i>
	<i>Haemophilus influenzae</i>
	<i>Legionella pneumophila</i>
	<i>Mycoplasma pneumoniae</i>
	<i>Staphylococcus aureus</i>
	<i>Staphylococcus epidermidis</i>
	<i>Streptococcus pneumoniae</i>
<i>Streptococcus pyogenes</i>	
<b>Yeast</b>	<i>Candida albicans</i>
<b>Other</b>	Pooled Human Nasal Wash

Additionally, four (4) clinical specimens containing Coronavirus HKU1 were tested n=5 in the absence or presence of heat inactivated SARS-CoV-2 virus at a concentration of 3X LOD (210 TCID<sub>50</sub>/swab). No cross-reactivity or interference was seen with the Coronavirus HKU1 clinical specimens.

To estimate the likelihood of cross-reactivity with SARS-CoV-2 virus in the presence of organisms that were not available for wet testing, *In silico* analysis using the Basic Local Alignment Search Tool (BLAST) managed by the National Center for Biotechnology Information (NCBI) was used to assess the degree of protein sequence homology.

- For *P. jirovecii*, the potential for cross-reactivity is very low. No significant sequence homology was found.
- BinaxNOW COVID-19 Ag Card may be susceptible to cross reactivity with SARS-CoV. However, there have been no reported cases of SARS since 2004, reducing the likelihood of a false positive result due to cross reactivity.

### High Dose Hook Effect

No high dose hook effect was observed when tested with up to a concentration of 1.4 x 10<sup>6</sup> TCID<sub>50</sub>/mL of inactivated SARS-CoV-2 virus with the BinaxNOW COVID-19 Ag Card.

### Interfering Substances

The following substances, naturally present in respiratory specimens or that may be artificially introduced into the upper respiratory tract, were evaluated with the BinaxNOW COVID-19 Ag Card at the concentrations listed below and were found not to affect test performance.

Substance	Active Ingredient	Concentration
Throat Lozenge	Menthol, Benzocaine	3 mg/mL
Sore Throat Spray	Phenol	5% w/v
OTC Nasal Spray 1	Mometasone Furoate	15% v/v

Substance	Active Ingredient	Concentration
OTC Nasal Spray 2	Triamcinolone	15% v/v
OTC Nasal Spray 3	Budesonide	15% v/v
OTC Nasal Spray 4	Fluticasone	15% v/v
OTC nasal gel	Sodium Chloride with Preservatives	15% v/v
OTC Nasal Spray 5	Phenylephrine	15% v/v
OTC Nasal Spray 6	Oxymetazoline	15% v/v
OTC Nasal Spray 7	Cromolyn	15% v/v
OTC Homeopathic Nasal Spray	<i>Zicam (Galphimia glauca, Histaminum hydrochloricum, Luffa operculate, sulfur)</i>	15% v/v
OTC Homeopathic Nasal Wash	Alkalol	15% v/v
Hand Sanitizer	Ethyl Alcohol 62%	1% w/v
Hand Soap		1% w/v
Endogenous	Whole Blood	2.5% v/v
Endogenous	Mucin	2.5 mg/mL
Endogenous	Leukocytes	5 x 10 <sup>6</sup> cells/mL
Antibiotic, Nasal Ointment	Mupirocin	10 mg/mL
Nasal Corticosteroid 1	Beclomethasone	15% v/v
Nasal Corticosteroid 2	Dexamethasone	15% v/v
Nasal Corticosteroid 3	Flunisolide	15% v/v
Anti-Viral Drug 1	Tamiflu (Oseltamivir Phosphate)	5 mg/mL
Anti-Viral Drug 2	Remdesivir	5 mg/mL
Anti-Viral Drug 3	Molnupiravir	5 mg/mL
Antibiotic	Tobramycin	1.44 mg/mL
Anti-Viral Drug	Zanamivir	281.5 ng/mL

### Reproducibility/Near the Cut Off

A reproducibility/near the cut off study of BinaxNOW COVID-19 Ag Card was conducted by operators from three sites using panels of blind coded specimens containing negative, low positive (near the limit of detection), moderate positive (above the limit of detection), and high negative SARS-CoV-2 samples.

Participants tested each sample multiple times on five different days. The percent agreement relative to the expected results for the moderate positive samples was 100% (135/135). The percent agreement relative to the expected results for the low positive samples was 94.1% (127/135). The percent agreement relative to the expected results for the high negative samples was 99.2% (132/133) and the true negative samples were 99.3% (134/135).

The Reproducibility Study site-to-site qualitative results (agreements relative to the expected results) are presented in the table below:

Sample	Site 1	Site 2	Site 3	Overall Agreement and 95% CI	
Moderate Positive	100% (45/45)	100% (45/45)	100% (45/45)	100% (135/135)	97.2% - 100.0%
Low Positive	93.3% (42/45)	97.8% (44/45)	91.1% (41/45)	94.1% (127/135)	88.7% - 97.0%
High Negative	100% (45/45)	100% (45/45)	97.7% (42/43)	99.2% (132/133)	95.9% - 99.9%
True Negative	100% (45/45)	100% (45/45)	97.8% (44/45)	99.3% (134/135)	95.9% - 99.9%

## CLINICAL STUDIES

Clinical performance characteristics of the BinaxNOW COVID-19 Ag Card were evaluated in two prospective studies conducted in symptomatic individuals within the United States. In the first study, conducted across five (5) investigational sites from November, 2020 through March, 2021, each subject either self-collected one (1) nasal swab (from both nostrils) or had one sample collected from him/her by another individual and performed the BinaxNOW COVID-19 Ag Card. A total of 397 individuals were enrolled in the study of which 102 were excluded for not meeting inclusions criteria or were unevaluable resulting in a total of 295 evaluable subjects.

In the second study, an additional all comers, real world, prospective clinical study was performed at a high-volume COVID community testing site from February 2022 to July 2022 when Omicron and its variants were prevalent in the United States. This study was led by Johns Hopkins Medicine in collaboration with the University of Maryland Medical Center and Maryland Department of Health. In this study, participants independently performed the BinaxNOW COVID-19 Ag Card, from nasal swab collection to result interpretation. A total of 333 individuals were enrolled in the study of which 24 were excluded for not meeting inclusion criteria or were unevaluable resulting in a total of 309 evaluable subjects.

In each study, a matched anterior nasal swab sample was taken from each study subject by a healthcare professional for testing on an FDA Emergency Use Authorized real-time Polymerase Chain Reaction (RT-PCR) assay for the detection of SARS-CoV-2 as the comparator method in each study.

Across both studies, the performance of BinaxNOW COVID-19 Ag Card was established with a total of 604 nasal swabs from symptomatic patients (within 5 days of symptom onset) who were suspected of COVID-19. Performance is shown in the tables below combined for both studies and individual by study.

### BinaxNOW™ COVID-19 Ag Card Performance within 5 days of symptom onset against the Comparator Method – Overall/Combined

BinaxNOW COVID-19 Ag Card	Comparator Method		
	Positive	Negative	Total
Positive	186	6	192
Negative	28	384	412
<b>Total</b>	214	390	604
Positive Agreement: 186/214 86.9% (95% CI: 81.7, 90.8)			

Negative Agreement: 384/390 98.5% (95% CI: 96.7, 99.3)
--

Note: Five (5) samples generated an invalid BinaxNOW COVID-19 Ag Card result and are not included in the analysis. The invalid rate is 5/730, or 0.68% (95% CI from 0.29% to 1.59%). The denominator for the invalid rate is based on total study enrollment.

### BinaxNOW COVID-19 Ag Card Performance within 5 days of symptom onset against the Comparator Method – Original Study (November 2020 – March 2021)

BinaxNOW COVID-19 Ag Card	Comparator Method		
	Positive	Negative	Total
Positive	71	3	74
Negative	16	205	221
Total	87	208	295
Positive Agreement: 71/87	81.6% (95% CI: 72.2, 88.4)		
Negative Agreement: 205/208	98.6% (95% CI: 95.8, 99.5)		

Note: Three (3) samples generated an invalid BinaxNOW COVID-19 Ag Card result and are not included in the analysis. The invalid rate is 3/397, or 0.76% (95% CI from 0.26% to 2.20%). The denominator for the invalid rate is based on total study enrollment.

### BinaxNOW COVID-19 Ag Card Performance within 5 days of symptom onset against the Comparator Method – Omicron Study (February 2022 – July 2022)

BinaxNOW COVID-19 Ag Card	Comparator Method		
	Positive	Negative	Total
Positive	115	3	118
Negative	12	179	191
Total	127	182	309
Positive Agreement: 115/127	90.6% (95% CI: 84.2, 94.5)		
Negative Agreement: 179/182	98.4% (95% CI: 95.3, 99.4)		

Note: Two (2) samples generated an invalid BinaxNOW COVID-19 Ag Card result and are not included in the analysis. The invalid rate is 2/333, or 0.60% (95% CI from 0.16% to 2.10%). The denominator for the invalid rate is based on total study enrollment.

### BinaxNOW COVID-19 Ag Card Positive Agreement (PPA) against the Comparator Method – Stratified by Days Post Symptom Onset (DPSO)

DPSO	PPA - Original Study (November 2020-March 2021)	PPA - Omicron Study (February 2022-July 2022)
Day 0	N/A (0/0)	69.23% (9/13)
Day 1	94.12% (16/17)	88.24% (45/51)
Day 2	73.33% (22/30)	97.22% (35/36)
Day 3	76.00% (19/25)	100.00% (20/20)
Day 4	88.89% (8/9)	66.67% (2/3)
Day 5	100.00% (6/6)	100.00% (4/4)

### Serial Testing

A prospective clinical study was also conducted between January 2021 and May 2022 as a component of the Rapid Acceleration of Diagnostics (RADx) initiative from the National Institutes of Health (NIH)<sup>1,2</sup>. A total of 7,361 individuals were enrolled via a decentralized clinical study design, with a broad geographical representation of the United States. Per inclusion criteria, all individuals were asymptomatic upon enrollment

<sup>1</sup> <https://www.medrxiv.org/content/10.1101/2022.08.04.22278274v1>

<sup>2</sup> <https://www.medrxiv.org/content/10.1101/2022.08.05.22278466v1>

in the study and at least 14 days prior to it and did not have a SARS-CoV-2 infection in the three months prior to enrollment. Participants were assigned to one of three EUA authorized SARS-CoV-2 OTC rapid antigen tests to conduct serial testing (every 48 hours) for 15 days. If an antigen test was positive, the serial-antigen testing result is considered positive.

At each rapid antigen testing time point, study subjects also collected a nasal swab for comparator testing using a home collection kit (using a 15-minute normalization window between swabs). SARS-CoV-2 infection status was determined by a composite comparator method on the day of the first antigen test, using at least two highly sensitive EUA RT-PCRs. If results of the first two molecular tests were discordant a third highly sensitive EUA RT-PCR test was performed, and the final test result was based upon the majority rule.

Study participants reported symptom status throughout the study using the MyDataHelps app. Two-day serial antigen testing is defined as performing two antigen tests 36 – 48 hours apart. Three-day serial antigen testing is defined as performing three antigen tests over five days with at least 48 hours between each test.

Out of the 7,361 participants enrolled in the study, 5,609 were eligible for analysis. Among eligible participants, 154 tested positive for SARS-CoV-2 infection based on RT-PCR, of which 97 (62%) were asymptomatic on the first day of their infection, whereas 57 (39%) reported symptoms on the first day of infection.

Performance of the antigen test with serial testing in individuals is described in the table below.

**Data establishing PPA of COVID-19 antigen serial testing compared to the molecular comparator single day testing throughout the course of infection with serial testing. Data is from all antigen tests in study combined.**

Days After First PCR Positive Test Result	Symptomatic On First Day Of Testing		
	Ag Positive / PCR Positive (Antigen Test Performance % PPA)		
	1 Test	2 Tests	3 Tests
0	34/57 (59.6%)	47/51 (92.2%)	44/47 (93.6%)
2	58/62 (93.5%)	59/60 (98.3%)	43/43 (100%)
4	55/58 (94.8%)	53/54 (98.1%)	39/40 (97.5%)
6	27/34 (79.4%)	26/33 (78.8%)	22/27 (81.5%)
8	12/17 (70.6%)	12/17 (70.6%)	7/11 (63.6%)
10	4/9 (44.4%)	3/7 (42.9%)	

1 Test = one (1) test performed on the noted days after the first PCR positive test result. Day 0 is the first day of documented infection with SARS-CoV-2.  
2 Tests = two (2) tests performed an average of 48 hours apart. The first test performed on the indicated day and the second test performed 48 hours later.  
3 Tests = three (3) tests performed an average of 48 hours apart. The first test performed on the indicated day, the second test performed 48 hours later, and a final test performed 48 hours after the second test.

---

**Flex Studies**

Using risk analysis as a guide, analytical flex studies were conducted on BinaxNOW COVID-19 Ag Card. The testing evaluated various sources of potential human errors and environmental factors that could affect the accuracy of results, including those related to sample handling, reagent handling, and extremes of operational conditions. The studies demonstrated that the test is robust to usage variation and environmental factors that may be encountered. It was observed that exposure of the device test strip directly to wet cleaning solutions such as bleach, ethanol or isopropyl alcohol or excessive glove powder may cause erroneous results. Therefore, the Instructions for Use include a statement cautioning users to ensure the test card is placed on a clean, dry surface to initiate the test.

**Conclusion**

The data presented in this 510(k) premarket notification demonstrate that the subject device, BinaxNOW COVID-19 Ag Card, is substantially equivalent to the predicate device (Nano-Check COVID-19 Antigen Test, K231187). The differences in the BinaxNOW COVID-19 Ag Card (proposed device) and the Nano-Check COVID-19 Antigen Test (predicate device, K231187) are limited to the Intended Use population (individuals with symptoms within five (5) days of symptom onset vs. individuals with symptoms within four (4) days of symptom onset). This difference does not affect the overall substantial equivalence of the proposed device to the predicate device in terms of the technological similarity, intended use, safety, and effectiveness. Furthermore, the information contained within this notification demonstrates BinaxNOW COVID-19 Ag Card compliance with the special controls applicable to a simple point-of-care device to directly detect SARS-CoV-2 viral target from clinical specimens in near-patient settings.

There is no known potential adverse effect to the operator when using this in vitro device according to the BinaxNOW COVID-19 Ag Card package insert.