



March 21, 2025

Tenon Medical
% Michael Coladonato
MRCA, LLC
803 7th Street NW
Third Floor
Washington, District of Columbia 20001

Re: K250403

Trade/Device Name: CATAMARAN™ SI Joint Fusion System
Regulation Number: 21 CFR 888.3040
Regulation Name: Smooth Or Threaded Metallic Bone Fixation Fastener
Regulatory Class: Class II
Product Code: OUR
Dated: March 14, 2025
Received: March 14, 2025

Dear Michael Coladonato:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

All medical devices, including Class I and unclassified devices and combination product device constituent parts are required to be in compliance with the final Unique Device Identification System rule ("UDI Rule"). The UDI Rule requires, among other things, that a device bear a unique device identifier (UDI) on its label and package (21 CFR 801.20(a)) unless an exception or alternative applies (21 CFR 801.20(b)) and that the dates on the device label be formatted in accordance with 21 CFR 801.18. The UDI Rule (21 CFR 830.300(a) and 830.320(b)) also requires that certain information be submitted to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). For additional information on these requirements, please see the UDI System webpage at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-system-udi-system>.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory->

[assistance/contact-us-division-industry-and-consumer-education-dice](#)) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Colin
O'Neill -S 

Colin O'Neill, M.B.E.

Assistant Director

DHT6B: Division of Spinal Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

Submission Number (if known)

K250403

Device Name

CATAMARAN SI Joint Fusion System

Indications for Use (Describe)

The Tenon Medical Catamaran SI Joint Fusion System is intended for sacroiliac joint fusion for conditions including:

- Sacroiliac joint disruptions and degenerative sacroiliitis
- To augment immobilization and stabilization of the sacroiliac joint in skeletally mature patients undergoing sacropelvic fixation as part of a lumbar or thoracolumbar fusion.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

Submitter:

Tenon Medical
104 Cooper Ct
Los Gatos, CA 95032

Official Correspondent:

Michael Coladonato
Associate Director, Regulatory Affairs
MCRA, LLC
803 7th Street NW, Third Floor
Washington, DC 20001
Phone: (202) 552-5800
mcoladonato@mcra.com

Date Prepared:

March 14, 2025

Trade Name:

CATAMARAN™ SI Joint Fusion System

Common Name:

Sacroiliac Joint Fixation

Classification:

21 CFR 888.3040 - Smooth or threaded bone fixation fastener

Product Code:

OUR

Predicate Devices:

Primary: CATAMARAN™ Sacroiliac Joint Fixation System,
Tenon Medical (K231944)

Additional: SI-BONE iFuse Implant System®, SI-Bone, Inc.
(K193524)

Device Description:

The Catamaran SI Joint Fusion System is intended for sacroiliac joint fusion for conditions including sacroiliac joint disruptions and degenerative sacroiliitis. The Catamaran SI Joint Fusion System was developed as a less invasive alternative to traditional open posterior surgical SI Joint fusion. The system consists of the Catamaran SI Joint Fixation Implant (Ti6Al-4V ELI Titanium alloy / ASTM F136) included in a reusable System Tray containing: Access, Drill and Delivery, Bone Graft Packing, and Extraction Instruments. The Catamaran SI Joint Fixation Implants are designed in various widths and lengths, and allow autologous bone graft material placement in the barrels of the implant to support SI Joint fixation and fusion. The Instrument Set includes Class II single use and reusable surgical instruments designed to facilitate placement of the implant within the sacroiliac joint using an inferior-posterior surgical approach. The implants and associated instruments are provided clean and non-sterile and are designed for steam sterilization prior to use.

Indications for Use:

The Tenon Medical Catamaran SI Joint Fusion System is intended for sacroiliac joint fusion for conditions including:

- Sacroiliac joint disruptions and degenerative sacroiliitis
- To augment immobilization and stabilization of the sacroiliac joint in skeletally mature patients undergoing sacropelvic fixation as part of a lumbar or thoracolumbar fusion.

Substantial Equivalence:

The Catamaran SI Joint Fusion System has the same technological characteristics and principles of operation as compared to the predicate device; the Catamaran Sacroiliac Joint Fixation System cleared in K231944 and K180818. The sole purpose of this submission is to modify the indications for use. The new indications for use do not raise any new types of safety or effectiveness questions between the Catamaran SI Joint Fusion System and the predicate devices.

Performance Testing:

No clinical or non-clinical performance testing was necessary to support the change in indications for use proposed.

Conclusion:

The subject Catamaran SI Joint Fusion System is substantially equivalent to the cited predicate devices.