



September 9, 2025

Shenzhen Yongquan Medical Devices Co., Ltd.
% Tulin Li
Medical Device Consultant
Huide Medical Technology Service Group Co., Ltd
Room 703, Building 16, South Bank Plaza, Exhibition Bay
Zhancheng Community, Fuhai Street, Bao'an District
ShenZhen, 518053
China

Re: K251773
Trade/Device Name: Water-soluble Personal Lubricant (Satin Gel Vitamin E flavor lubricant; Satin Gel Caramel flavor lubricant; Satin Gel Peach flavor lubricant; Satin Gel Strawberry flavor lubricant.)
Regulation Number: 21 CFR§ 884.5300
Regulation Name: Condom
Regulatory Class: II
Product Code: NUC
Dated: June 8, 2025
Received: June 10, 2025

Dear Tulin Li:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

All medical devices, including Class I and unclassified devices and combination product device constituent parts are required to be in compliance with the final Unique Device Identification System rule ("UDI Rule"). The UDI Rule requires, among other things, that a device bear a unique device identifier (UDI) on its label and package (21 CFR 801.20(a)) unless an exception or alternative applies (21 CFR 801.20(b)) and that the dates on the device label be formatted in accordance with 21 CFR 801.18. The UDI Rule (21 CFR 830.300(a) and 830.320(b)) also requires that certain information be submitted to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). For additional information on these requirements, please see the UDI System webpage at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-system-udi-system>.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,


Monica D. Garcia -S

Monica D. Garcia, Ph.D.
Assistant Director
DHT3B: Division of Reproductive,
Gynecology and Urology Devices
OHT3: Office of GastroRenal, ObGyn,
General Hospital and Urology Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K251773

Device Name

Water-soluble Personal Lubricant (Satin Gel Vitamin E flavor lubricant; Satin Gel Caramel flavor lubricant; Satin Gel Peach flavor lubricant; Satin Gel Strawberry flavor lubricant).

Indications for Use (Describe)

Water-soluble Personal Lubricant is a water based personal lubricant, for penile and/or vaginal application, intended to lubricate and moisturize, to enhance the ease and comfort of intimate sexual activity and supplement the body's natural lubrication. This product is not compatible with natural rubber latex, polyurethane, and polyisoprene condoms.

Type of Use (Select one or both, as applicable)

☐ Prescription Use (Part 21 CFR 801 Subpart D)

☒ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary – K251773

I. General Information on Submitter

Applicant: Shenzhen Yongquan Medical Devices Co., Ltd.
Address: Room 309, Baiwangda Smart Park, Bantian, Longgang District Shenzhen, China
Telephone: +86-18344000056
Contact Person: Mr. Amos Zou
Contact Title: Medical Device Consultant Huide Medical Technology Service Group Co., Ltd
Email: amos.zou@139.com
Date Prepared: September 8, 2025

II. General Information on Device

Proprietary Name: Water-soluble Personal Lubricant (Satin Gel Vitamin E flavor lubricant; Satin Gel Caramel flavor lubricant; Satin Gel Peach flavor lubricant; Satin Gel Strawberry flavor lubricant).
Common Name: Personal Lubricant
Regulation Name: Condom
Regulation Number 21 CFR 884.5300
Regulatory Class: II
Product Code: NUC (Lubricant, Personal)

III. Predicate Device

Predicate Device	510(k) Number
WLX & Bio Human Lubricant, Jiangxi Wulinxiang Biological Technology Co., Ltd	K240081

This predicate device has not been subject to a device-related recall.

IV. Description of Device

Water-soluble Personal Lubricant is a non-sterile, water-based personal lubricant for penile and/or vaginal application, intended to lubricate and moisturize, to enhance the ease and comfort of intimate sexual activity and supplement the body's

natural lubrication. Its formulation consists of pure water, glycerin, hydroxyethyl cellulose, potassium sorbate, sodium benzoate, sodium lactate, flavors, ethylenediaminetetraacetic acid, etc. The subject device includes four different flavors of water-soluble personal lubricants: Vitamin E flavor, Caramel flavor, Peach flavor, and Strawberry flavor. This device is not compatible with natural rubber latex, polyisoprene and polyurethane condoms.

The lubricant is packaged in PE bottles of 100gm. The specifications for Water-soluble Personal Lubricant are described in **Table 1**.

Table 1. Device Specifications

Parameter	Test Method	Specification
Appearance/Color	Visual	Semi-Viscous liquid/Clear
Odor	Olfactory	Four different Odors (Vitamin E, Caramel, Peach, Strawberry)
pH	USP<791>	4.5-6.0
Viscosity	USP<912>	4,500-6,500 cps
Osmolality	USP<785>	800-1,350 mOsm/kg
Antimicrobial Effectiveness	USP <51>	Meets USP <51> acceptance criteria for Category 2 products. Category 2, bacteria should show not be less than 2.0 log reduction at 14 days and no increase from 14-day Count at the 28 day count. Yeasts and molds should show no increase from the initial calculated count at 14 and 28 days.
Total Aerobic Microbial Count (TAMC)	USP<61>	Less than 100 cfu/mL
Total Yeast and Mold Count (TYMC)	USP<61>	Less than 10 cfu/mL
Absence of Pathogenic Organisms (<i>Candida albicans</i> , <i>Pseudomonas aeruginosa</i> , <i>Staphylococcus aureus</i>)	USP<62>	Absent

V. Indications for Use

Water-soluble Personal Lubricant is a water based personal lubricant, for penile and/or vaginal application, intended to lubricate and moisturize, to enhance the ease and comfort of intimate sexual activity and supplement the body's natural

lubrication. This product is not compatible with natural rubber latex, polyurethane, and polyisoprene condoms.

VI. Substantial Equivalence Discussion

The following table compares the intended use and key technological characteristics of the subject and predicate device:

Characteristic / Feature	Water-soluble Personal Lubricant (subject device)	WLX & Bio Human Lubricant (predicate device) – K240081	Comparison
Indication for use	Water-soluble Personal Lubricant is a water based personal lubricant, for penile and/or vaginal application, intended to lubricate and moisturize, to enhance the ease and comfort of intimate sexual activity and supplement the body's natural lubrication. This product is not compatible with natural rubber latex, polyurethane, and polyisoprene condoms.	WLX & Bio Human Lubricant is a water based personal lubricant, for penile and/or vaginal application, intended to lubricate and moisturize, to enhance the ease and comfort of intimate sexual activity and supplement the body's natural lubrication. This product is not compatible with natural rubber latex, polyurethane, and polyisoprene condoms.	Same
Water-Based Lubricant	Yes	Yes	Same
Over the Counter	Yes	Yes	Same
Not a contraceptive or Spermicide	Yes	Yes	Same
Non-sterile	Yes	Yes	Same
Primary Ingredients	Water, Glycerin, Hydroxyethyl cellulose, Potassium sorbate, Sodium benzoate, Sodium lactate, flavors, ethylenediaminetetraacetic acid	Water, Glycerin, Propylene Glycol, Carbomer, Hydroxyethylcellulose, Triethanolamine, and Phenoxyethanol	Different
Microbial Limits	Total mold/yeast count <10 cfu/mL Total aerobic microbial count <100 cfu/mL Absence of pathogens (<i>Candida albicans</i> , <i>Pseudomonas aeruginosa</i> , <i>Staphylococcus aureus</i>)	Total mold/yeast count <10 cfu/mL Total aerobic microbial count <100 cfu/mL Absence of pathogens (<i>Candida albicans</i> , <i>Pseudomonas aeruginosa</i> , <i>Staphylococcus aureus</i>)	Same

Viscosity	4,500-6,500 cps	100-300 cps	Different
Osmolality	800-1,350 mOsm/kg	750-950 mOsm/kg	Different
pH	4.5-6.0	5.5-7.5	Different
Condom Compatibility	Not compatible with natural rubber latex, polyurethane, and polyisoprene condoms.	Not compatible with natural rubber latex, polyurethane, and polyisoprene condoms.	Same

The subject and predicate devices have similar indications for use and have the same intended use – to provide lubrication during intimate sexual activity. The subject and predicate devices have different technological characteristics, including different formulations and device specifications. The different technological characteristics do not raise different questions of safety and effectiveness.

VII. Summary of Non-Clinical Performance Testing

Biocompatibility

Biocompatibility testing on the subject lubricant was performed in accordance with the 2020 FDA guidance document *Use of International Standard ISO 10993-1, “Biological Evaluation of Medical Devices – Part 1: Evaluation and testing within a risk management process”* and ISO 10993-1:2009 as follows:

- Cytotoxicity (ISO 10993-5:2009/(R)2014)
- Guinea Pig Maximization Sensitization Test (ISO 10993-10: 2021)
- Vaginal Irritation (ISO 10993-23: 2021)
- Acute Systemic Toxicity (ISO 10993-11:2017)

The results of this testing demonstrate that the subject lubricant is non-cytotoxic, non-irritating, non-sensitizing, and not systemically toxic.

Shelf Life

The subject device is a non-sterile personal lubricant packaged in PE bottles with a 3-year shelf-life in accordance with the results of real time aging study. The device specifications listed in **Table 1** were tested across the device shelf-life and the subject device met the specifications at all time points.

Condom Compatibility

Condom compatibility testing was not conducted for the subject device. The subject

device is not compatible with natural rubber latex, polyisoprene, and polyurethane condoms.

VIII. Conclusion

The results of the testing described above demonstrate that Water-soluble Personal Lubricant is as safe and effective as the predicate device and supports a determination of substantial equivalence.