



June 18, 2026

Avalign Technologies, Inc.
Jennifer Staunton
Director Regulatory Affairs
8727 Clinton Park Dr.
Fort Wayne, Indiana 46825

Re: K252982

Trade/Device Name: CONDUIT™ ZERO-P™ VA Secured Spacer System
Regulation Number: 21 CFR 888.3080
Regulation Name: Intervertebral Body Fusion Device
Regulatory Class: Class II
Product Code: OVE
Dated: May 20, 2026
Received: May 20, 2026

Dear Ms. Staunton:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device"

(<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality Management System Regulation (QMSR) (21 CFR Part 820), which includes, but is not limited to, ISO 13485 clause 7.3 (Design controls), ISO 13485 clause 8.3 (Nonconforming product), ISO 13485 clause 8.5.2 (Corrective action), and ISO 13485 clause 8.5.3 (Preventative action). Please note that regardless of whether a change requires premarket review, the QMSR requires device manufacturers to review and approve changes to device design and production (ISO 13485 clause 7.3 and ISO 13485 clause 7.5) and document changes and approvals in the Medical Device File (ISO 13485 clause 4.2.3).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the Quality Management System Regulation (QMSR) (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

All medical devices, including Class I and unclassified devices and combination product device constituent parts are required to be in compliance with the final Unique Device Identification System rule ("UDI Rule"). The UDI Rule requires, among other things, that a device bear a unique device identifier (UDI) on its label and package (21 CFR 801.20(a)) unless an exception or alternative applies (21 CFR 801.20(b)) and that the dates on the device label be formatted in accordance with 21 CFR 801.18. The UDI Rule (21 CFR 830.300(a) and 830.320(b)) also requires that certain information be submitted to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). For additional information on these requirements, please see the UDI System webpage at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-system-udi-system>.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

BRENT SHOWALTER -S

Brent Showalter, Ph.D.

Assistant Director

DHT6B: Division of Spinal Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K252982

Device Name
CONDUITTM ZERO-P VA Secured Spacer System

Indications for Use (Describe)

The CONDUIT ZERO-P VA Spacer is a stand-alone anterior cervical interbody fusion device with a microscopic roughened surface and micro and nano-scale features indicated for use at one or two levels of the cervical spine (C2-T1) in skeletally mature patients with cervical disc disease, instability, trauma including fractures, deformity defined as kyphosis, lordosis, or scoliosis, cervical spondylotic myelopathy, spinal stenosis, and failed previous fusion.

Cervical disc disease is defined as intractable radiculopathy and/or myelopathy with herniated disc and/or osteophyte formation on posterior vertebral endplates producing symptomatic nerve root and/or spinal cord compression confirmed by radiographic studies. These patients should be skeletally mature and have had at least six (6) weeks of non-operative treatment.

The interior of the spacer component should be packed with autogenous and/or allogeneic bone graft comprised of cancellous, and/or corticocancellous bone graft and implanted via an anterior approach.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

Submitter: Avalign Technologies, Inc.
8727 Clinton Park Drive
Fort Wayne, IN 46825 USA

Contact Person: Jennifer Staunton
Director Regulatory Affairs
Telephone: 219-718-1152

Date Prepared: September 17, 2025

Trade Name: CONDUIT™ ZERO-P™ VA Secured Spacer System

Device Class: Class II

Product Code: OVE

Common Name: Intervertebral Fusion Device with Integrated Fixation, Cervical

Classification Name: Intervertebral Body Fusion Device

Regulation Number: 21 CFR 888.3080

Classification Panel: Orthopaedic and Rehabilitation Devices Panel (87)

Primary Predicate: Synthes Zero-P Variable Angle (K112068)

Additional Predicates: EIT Cellular Titanium® Cervical Cage (K170503)
EIT Cellular Titanium® Cervical Cage (K201605)
Globus Medical Inc. HEDRON™ Cervical Spacers (K191243)
CONDUIT™ SYNFIX™ Evolution Secured Spacer System (K250072)

Device Description: The CONDUIT ZERO-P VA Secured Spacers are intervertebral body fusion devices intended for anterior cervical interbody fusion. Two screws are inserted at convergent and variable angles into adjacent vertebral bodies with a one step blocking mechanism. The CONDUIT ZERO-P VA Secured Spacer System is available in various heights and geometries to suit individual pathology and anatomical conditions.

The CONDUIT ZERO-P VA Cages are made from Ti-6Al-4V ELI conforming to ASTM F3001 with an additive manufacturing process (Selective Laser Melting). The design contains solid structures and porous structures. The hollow geometry of the implants allows them to be packed with autogenous and/or allogeneic bone graft comprised of cancellous, and/or corticocancellous bone graft. The 3D Printed Conduit Cellular Titanium Cages have a microscopic

roughened surface with micro and nano-scale features. The micro and nano features are on all surfaces of the Cage, including the superior, inferior, and peripheral surfaces, as well as each member of the internal cell structure.

Indications:

The CONDUIT™ ZERO-P™ VA Spacer is a stand-alone anterior cervical interbody fusion device with a microscopic roughened surface and micro and nano-scale features indicated for use at one or two levels of the cervical spine (C2-T1) in skeletally mature patients with cervical disc disease, instability, trauma including fractures, deformity defined as kyphosis, lordosis, or scoliosis, cervical spondylotic myelopathy, spinal stenosis, and failed previous fusion.

Cervical disc disease is defined as intractable radiculopathy and/or myelopathy with herniated disc and/or osteophyte formation on posterior vertebral endplates producing symptomatic nerve root and/or spinal cord compression confirmed by radiographic studies. These patients should be skeletally mature and have had at least six (6) weeks of non-operative treatment.

The interior of the spacer component should be packed with autogenous and/or allogeneic bone graft comprised of cancellous, and/or corticocancellous bone graft and implanted via an anterior approach.

Materials:

The components are manufactured from medical grade Titanium. The Cage from standard specification for Ti-6Al-4V ELI using full-melt powder bed fusion in conformance with ASTM F3001. The pre-assembled blocking mechanism, consisting of Qty 2ea Screw Catch and Locking Pins are comprised of Ti-6Al-4V per ASTM F136, while the Springs are comprised of Phynox® per ASTM 1058. The Bone Screws are supplied by DePuy Synthes and manufactured from Ti-6Al-7Nb (ISO 5832).

Comparison to Predicate Device:

The substantial equivalence of the subject device to the predicates identified above is based upon the equivalence of intended use, design (fundamental scientific technology), performance, sterility, and biocompatibility.

Performance Data:

Mechanical performance, including static and dynamic axial compression per ASTM F2077-24, static and dynamic compression shear per ASTM F2077-24, static and dynamic torsion per ASTM F2077-24 and subsidence per ASTM F2267-24, was assessed to support a substantial equivalence determination by characterizing the properties and functionality of the system and allowing comparison to established acceptance criteria.

Additionally, the subject device was evaluated for magnetically

induced displacement force per ASTM F2052-21, magnetically induced torque per ASTM F2213-17, MR image artifact per ASTM F2119-24 and RF-induced heating per ASTM F2182-19e2 to support the MR Conditional labeling.

**Clinical Test
Summary:**

No clinical data was necessary to demonstrate substantial equivalence, nor safety and effectiveness of this system.

Conclusion:

Based on the predicate comparison of intended use, indications, technological characteristics, and device performance, the CONDUIT™ ZERO-P™ VA Spacer has demonstrated substantial equivalence to the identified predicate devices.