



February 23, 2026

Xenix Medical
Teresa Cherry
Vice President of Regulatory Affairs
149 W. Michigan St.
Orlando, California 32806

Re: K253583

Trade/Device Name: LUX Expandable Lumbar Interbody System
Regulation Number: 21 CFR 888.3080
Regulation Name: Intervertebral Body Fusion Device
Regulatory Class: Class II
Product Code: MAX
Dated: January 23, 2026
Received: January 23, 2026

Dear Teresa Cherry:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device"

(<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality Management System Regulation (QMSR) (21 CFR Part 820), which includes, but is not limited to, ISO 13485 clause 7.3 (Design controls), ISO 13484 clause 8.3 (Nonconforming product), and ISO 13485 clause 8.5 (Corrective and preventative action). Please note that regardless of whether a change requires premarket review, the QMSR requires device manufacturers to review and approve changes to device design and production (ISO 13485 clause 7.3 and 21 CFR 820.70) and document changes and approvals in the Medical Device File (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the Quality Management System Regulation (QMSR) (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

All medical devices, including Class I and unclassified devices and combination product device constituent parts are required to be in compliance with the final Unique Device Identification System rule ("UDI Rule"). The UDI Rule requires, among other things, that a device bear a unique device identifier (UDI) on its label and package (21 CFR 801.20(a)) unless an exception or alternative applies (21 CFR 801.20(b)) and that the dates on the device label be formatted in accordance with 21 CFR 801.18. The UDI Rule (21 CFR 830.300(a) and 830.320(b)) also requires that certain information be submitted to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). For additional information on these requirements, please see the UDI System webpage at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-system-udi-system>.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

KATHERINE D. KAVLOCK -S

for

Brent Showalter, Ph.D.

Assistant Director

DHT6B: Division of Spinal Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

Please type in the marketing application/submission number, if it is known. This textbox will be left blank for original applications/submissions.

K253583

?

Please provide the device trade name(s).

?

LUX Expandable Lumbar Interbody System

Please provide your Indications for Use below.

?

The LUX Expandable Lumbar Interbody System is indicated for intervertebral body fusion procedures in skeletally mature patients with degenerative disc disease (DDD) of the lumbar spine at one or two contiguous levels from L2-S1. Degenerative disc disease is defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies. These DDD patients may have up to Grade 1 spondylolisthesis or retrolisthesis at the involved level(s). LUX Expandable Lumbar Interbody System implants are to be used with autogenous bone graft and/or allograft comprised of cancellous and/or corticocancellous bone graft implanted via a transforaminal approach or an open posterior approach. The LUX Expandable Lumbar Interbody System implants are to be used with supplemental fixation. Patients should have at least (6) months of non-operative treatment prior to treatment with an intervertebral cage. Patients with previous non-fusion spinal surgery at involved level may be treated with the device.

Please select the types of uses (select one or both, as applicable).

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

?

510(k) Summary

Device Trade Name: LUX™ Expandable Lumbar Interbody System

Manufacturer: Xenix Medical
149 W Michigan St.
Orlando, FL 32806

Contact: Teresa Cherry
Vice President of Regulatory Affairs
Xenix Medical

Date Prepared: October 11, 2025

Classifications: 21 CFR §888.3080 Intervertebral Body Fusion Device

Class: II

Product Codes: MAX

Indications for Use:

The LUX™ Expandable Lumbar Interbody System is indicated for intervertebral body fusion procedures in skeletally mature patients with degenerative disc disease (DDD) of the lumbar spine at one or two contiguous levels from L2-S1. Degenerative disc disease is defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies. These DDD patients may have up to Grade 1 spondylolisthesis or retrolisthesis at the involved level(s). LUX™ Expandable Lumbar Interbody System implants are to be used with autogenous bone graft and/or allograft comprised of cancellous and/or corticocancellous bone graft implanted via a transforaminal approach or an open posterior approach. The LUX™ Expandable Lumbar Interbody System implants are to be used with supplemental fixation. Patients should have at least (6) months of non-operative treatment prior to treatment with an intervertebral cage. Patients with previous non-fusion spinal surgery at involved level may be treated with the device.

Device Description:

The Xenix Medical LUX™ Expandable Lumbar Interbody System implants are manufactured using Titanium Alloy (Ti-6AL-4V ELI per ASTM F136 and F3001). Implants are available in a

variety of footprints and geometric options to fit the anatomical variations of patients. The implants are provided sterile. The reusable instrumentation is provided non-sterile in a steam sterilization instrument tray.

Devices incorporate Xenix Medical’s proprietary NANOACTIV™ micro and nano-roughened surface designed to improve fixation to adjacent bone. The Xenix Medical implant surfaces have been engineered with surface features at a nanometer (10⁻⁹) level, which have demonstrated the ability to elicit an endogenous cellular and biochemical response as represented by mineralization in human mesenchymal stem cells *in vitro*. The implant surface demonstrates elements to be considered a nanotechnology as outlined in the FDA nanotechnology guidance document. *In vitro* results may not be representative of clinical performance. Note: Supporting Study Data will be provided upon request.

Predicate Devices:

Device Name(s)	Manufacturer	K-Number
<i>Primary Predicate</i>		
Rise Spacer	Globus	K113447
<i>Additional Predicate Devices</i>		
neoWave LS Lumbar Straight	Xenix Medical	K222988
ProLift Expandable System	Life Spine Inc	K193258

Performance Testing Summary:

The Xenix Medical LUX™ Expandable Lumbar Interbody System demonstrated substantially equivalent mechanical performance to the predicate devices through static and dynamic mechanical testing with reference to ASTM F2077, ASTM F2267, ASTM F1877.

Substantial Equivalence:

The subject device was demonstrated to be substantially equivalent to the predicate cited in the passage above with respect to indications, design, materials, function, manufacturing, and performance.

Conclusion:

The Xenix Medical LUX™ Expandable Lumbar Interbody System is substantially equivalent to the cited predicate devices with respect to intended use, indications for use, design, function, materials, and performance. The differences in the technological characteristics between the subject device and the predicate devices do not raise new or different questions of safety and effectiveness.