



December 22, 2025

Think Surgical, Inc.
Meliha Mulalic
Vice President, QA/RA
47201 Lakeview Blvd.
Fremont, California 94538

Re: K253661

Trade/Device Name: TMINI Miniature Robotic System
Regulation Number: 21 CFR 882.4560
Regulation Name: Stereotaxic Instrument
Regulatory Class: Class II
Product Code: OLO
Dated: November 20, 2025
Received: November 20, 2025

Dear Meliha Mulalic:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

All medical devices, including Class I and unclassified devices and combination product device constituent parts are required to be in compliance with the final Unique Device Identification System rule ("UDI Rule"). The UDI Rule requires, among other things, that a device bear a unique device identifier (UDI) on its label and package (21 CFR 801.20(a)) unless an exception or alternative applies (21 CFR 801.20(b)) and that the dates on the device label be formatted in accordance with 21 CFR 801.18. The UDI Rule (21 CFR 830.300(a) and 830.320(b)) also requires that certain information be submitted to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). For additional information on these requirements, please see the UDI System webpage at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-system-udi-system>.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory->

[assistance/contact-us-division-industry-and-consumer-education-dice](#)) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shumaya Ali -S

Shumaya Ali, M.P.H.

Assistant Director

DHT6C: Division of Restorative,
Repair, and Trauma Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K253661

Device Name

TMINI® Miniature Robotic System

Indications for Use (Describe)

TMINI® Miniature Robotic System is indicated as a stereotaxic instrumentation system for total knee replacement (TKA) surgery. It is to assist the surgeon by providing software-defined spatial boundaries for orientation and reference information to identifiable anatomical structures for the accurate placement of knee implant components.

The robotic device placement is performed relative to anatomical landmarks as recorded using the system intraoperatively and based on a surgical plan determined preoperatively using CT based surgical planning tools.

The targeted population has the same characteristics as the population that is suitable for the implant(s) compatible with the TMINI® Miniature Robotic System. The TMINI® Miniature Robotic System is compatible with and to be used with the following knee replacement systems in accordance with the indications and contraindications:

- Enovis™ EMPOWR Knee System®
- Ortho Development® BKS® and BKS TriMax® Knee System
- Total Joint Orthopedics Klassic® Knee System
- United® U2™ Total Knee System
- Medacta® GMK® Sphere / SpheriKA Knee Systems
- Zimmer Biomet Anterior & Posterior Referencing Persona® Knee
- b-ONE MOBIO® Total Knee System
- Maxx Orthopedics Freedom® Total & Titan Knee
- LINK® LinkSymphoKnee System
- Stryker® Triathlon® Knee System

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



**TMINI® Miniature Robotic System
Special 510(k) Submission**

510(k) SUMMARY

Applicant Information:

Owner Name: THINK Surgical, Inc.
Address: 47201 Lakeview Blvd., Fremont, CA 94538
Phone number: 408-921-5648
Fax number: 510-249-2396
Establishment Registration Number: 3000719653
Contact Person: Meliha Mulalic
Date Prepared: 20 November 2025

Device Information:

Device Classification: Class II
Trade Name: TMINI® Miniature Robotic System
Common name: Orthopedic Stereotaxic Instrument
Classification name: Stereotaxic Instrument
Regulation number: 882.4560
Product Code: OLO

Predicate Device:

The TMINI® Miniature Robotic System (AIM 3-2) is substantially equivalent in intended use, Indications for Use, design, materials, technology, operational principles and performance to the predicate, TMINI® Miniature Robotic System, cleared via K252544.

Device Modification:

The purpose of this submission is to add another compatible TKA implant system to the already cleared compatible implant systems for use with the TMINI® Miniature Robotic System. There have been no changes to the technology, operating principles, patient contacting materials or user interface of the TMINI® Miniature Robotic System associated with this 510(k) submission.

Design verification and validation testing was conducted on the implant module and implant system to evaluate performance during use with the TMINI Miniature Robotic System. All testing met the acceptance criteria of the testing protocols and demonstrated substantial equivalence to the predicate. The modification described in this submission had no impact on the intended use, Indications for Use, design,



TMINI® Miniature Robotic System Special 510(k) Submission

materials, technology, or operational principles of the device and no new questions of safety or effectiveness resulted from the changes. Therefore, the TMINI® Miniature Robotic System (AIM 3-2) is demonstrated to be substantially equivalent to the predicate device cleared via K252544.

Device Description:

The TMINI® Miniature Robotic System (AIM 3-2) like its predicate, the TMINI® Miniature Robotic System, consists of three primary components: a three-dimensional, graphical, Preoperative Planning Workstation with the TCM web based plan review, approval and download component, an Optical Tracking Navigation Console (TNav) and a robotically controlled hand-held tool (TMINI Robot) that assists the surgeon in preparing the bone for implantation of TKA components.

Like the predicate device, the TPLAN Planning Station uses preoperative CT scans of the operative leg to create 3D surface models for case templating and intraoperative registration purposes. The Planning Workstation contains a library of 510(k) cleared knee replacement implant(s) available for use with the system. The THINK Surgical planner can select a surgeon's intended implant model from this library. The planner/surgeon can manipulate the 3D representation of the implant in relation to the bone model to optimally place the implant. The surgeon reviews and approves the case plan using either TPLAN or the TCM web-based application once the surgeon is satisfied with the implant selection, location and orientation. The data from the approved plan is written to a file that is used to guide the robotically controlled hand-held tool.

The hand-held robotic tool is optically tracked relative to optical markers placed in both the femur and tibia and articulates in two degrees-of-freedom, allowing the user to place bone pins in a planar manner in both bones. Mechanical guides are clamped to the bone pins, resulting in subsequent placement of cut slots and drill guide holes such that the distal femoral and proximal tibial cuts can be made in the pre-planned positions and orientations, and such that the implant manufacturer's multi-planer cutting block can be placed relative to drilled distal femoral pilot holes. If the surgeon needs to change the plan during surgery, it can be changed intraoperatively on the TNav navigation console

Intended Use:

The TMINI System (AIM 3-2) like the predicate TMINI Miniature Robotic System is intended to assist the surgeon in providing software defined spatial boundaries for orientation and reference information to anatomical structures during orthopedic procedures.

Indications for Use:

The Indications for Use of the TMINI® Miniature Robotic System (AIM 3-2) and the predicate are the same with the exception of the new implant system added to the list of compatible implants (See **Table 1**). Both describe the use of a system that



**TMINI® Miniature Robotic System
Special 510(k) Submission**

uses diagnostic imaging of the patient to allow a surgeon to plan and execute orthopedic total knee procedures. The only difference is that the device, that is the subject of this 510(k) submission, adds another compatible TKA implant system to already cleared compatible implant systems for use with the TMINI® Miniature Robotic System.

None of the changes made to the predicate, nor the aggregate of all changes, have altered the indications for use (except adding the new implant system to the list of compatible implants in the Indications for Use), nor have they raised new types of safety or effectiveness questions.

Tables 1 and 2, below identify the substantial equivalence of the TMINI® Miniature Robotic System (AIM 3-2) to the predicate, TMINI® Miniature Robotic System cleared via K252544.

Table 1: Comparison of Intended Use and Indications for Use

Product	TMINI® Miniature Robotic System (AIM 3-2)	TMINI® Miniature Robotic System (AIM 3-0)	Conclusion
510(k) number	Subject Device	K252544	
Manufacturer	THINK Surgical, Inc	THINK Surgical, Inc	
Product Code	OLO	OLO	SAME
Regulation	21 CFR 882.4560	21 CFR 882.4560	SAME
Intended Use	Intended to assist the surgeon in providing software defined spatial boundaries for orientation and reference information to anatomical structures during orthopedic procedures.	Intended to assist the surgeon in providing software defined spatial boundaries for orientation and reference information to anatomical structures during orthopedic procedures.	SAME
Indications for Use	<p>The TMINI® Miniature Robotic System is indicated as a stereotaxic instrumentation system for total knee replacement (TKA) surgery. It is to assist the surgeon by providing software-defined spatial boundaries for orientation and reference information to identifiable anatomical structures for the accurate placement of knee implant components.</p> <p>The robotic device placement is performed relative to anatomical landmarks as recorded using the system intraoperatively and based on a surgical plan-determined preoperatively using CT based surgical planning tools.</p>	<p>The TMINI® Miniature Robotic System is indicated as a stereotaxic instrumentation system for total knee replacement (TKA) surgery. It is to assist the surgeon by providing software-defined spatial boundaries for orientation and reference information to identifiable anatomical structures for the accurate placement of knee implant components.</p> <p>The robotic device placement is performed relative to anatomical landmarks as recorded using the system intraoperatively and based on a surgical plan-determined preoperatively using CT based surgical planning tools.</p>	Substantially Equivalent

Product	TMINI® Miniature Robotic System (AIM 3-2)	TMINI® Miniature Robotic System (AIM 3-0)	Conclusion
	<p>The targeted population has the same characteristics as the population that is suitable for the implant(s) compatible with the TMINI® Miniature Robotic System.</p> <p>The TMINI® Miniature Robotic System is compatible with and to be used with the following knee replacement system(s) in accordance with the indications and contraindications:</p> <ul style="list-style-type: none"> - Enovis™ EMPOWR Knee System® - Ortho Development BKS® and BKS TriMax® Knee System - Total Joint Orthopedics Klassic® Knee System - United U2™ Knee Total Knee System - Medacta® GMK® Sphere / SpheriKA Knee Systems - Zimmer Biomet Anterior & Posterior Referencing Persona® Knee - b-ONE MOBIO® Total Knee System - Maxx Orthopedics Freedom® Total & Titan Knee - LINK® LinkSymphoKnee System - Stryker® Triathlon® Knee System 	<p>The targeted population has the same characteristics as the population that is suitable for the implant(s) compatible with the TMINI® Miniature Robotic System.</p> <p>The TMINI® Miniature Robotic System is compatible with and to be used with the following knee replacement system(s) in accordance with the indications and contraindications:</p> <ul style="list-style-type: none"> - Enovis™ EMPOWR Knee System® - Ortho Development BKS® and BKS TriMax® Knee System - Total Joint Orthopedics Klassic® Knee System - United U2™ Knee Total Knee System - Medacta® GMK® Sphere / SpheriKA Knee Systems - Zimmer Biomet Anterior & Posterior Referencing Persona® Knee - b-ONE MOBIO® Total Knee System - Maxx Orthopedics Freedom® Total & Titan Knee - LINK® LinkSymphoKnee System 	

Substantial Equivalence:

Both the TMINI® Miniature Robotic System (AIM 3-2), the subject of this submission, and the predicate device have the same intended use. Both are indicated as a stereotaxic instrumentation system for total knee replacement (TKA) surgery. It is to assist the surgeon by providing software-defined spatial boundaries for orientation and reference information to identifiable anatomical structures for the accurate placement of knee implant components during orthopedic procedures.

The robotic device placement is performed relative to anatomical landmarks as recorded using the system intraoperatively and based on a surgical plan determined preoperatively using CT based surgical planning tools. The difference between the new device and the predicate is that the new device introduces another TKA implant system. The implant module for this new implant system was created using the proprietary method of generation, verification and validation cleared in K252544. The modification, described in this submission, does not alter the intended use, Indications for Use (except adding the new implant system to the



**TMINI® Miniature Robotic System
Special 510(k) Submission**

list of compatible implants in the Indications for Use), design, materials, technology, or operational principles of the TMINI® Miniature Robotic System and no new questions of safety or effectiveness resulted from the changes.

The Indications for Use of the TMINI® Miniature Robotic System (AIM 3-2) and the predicate are the same with the exception of the new implant system added to the list of compatible implants (See Table 1). Both describe the use of a system that uses diagnostic imaging of the patient to allow a surgeon to plan and execute orthopedic total knee procedures. The only difference is that the device, that is the subject of the 510(k) submission, introduces another compatible TKA implant system to the already cleared compatible implant systems for use with the TMINI® Miniature Robotic System.

Design verification and validation testing was conducted to evaluate performance of the device. All testing met the acceptance criteria of the testing protocols and demonstrated substantial equivalence to the predicate. The modifications made had no impact on the intended use, Indications for Use, design, materials, technology, or operational principles of the device and no new questions of safety or effectiveness resulted from the changes.

Biocompatibility information for patient contacting materials and testing for the TMINI® Miniature Robotic System were presented in the predicate device submission K252544. There are no material changes to any of the direct patient contact components of the TMINI® Miniature Robotic System (AIM 3-2); therefore, no additional biocompatibility testing was required.

Substantial equivalence in technological characteristic and performance of the TMINI® Miniature Robotic System (AIM 3-2) to the predicate device is outlined in **Table-2** below:

Table-2: Substantial Equivalence

Product	TMINI® Miniature Robotic System (AIM 3-2)	TMINI® Miniature Robotic System (AIM 3-0)	Conclusion
510(k) number	Subject Device	K252544	
Manufacturer	THINK Surgical Inc.	THINK Surgical Inc.	
Materials			
<ul style="list-style-type: none"> Materials Used 	Uses materials with a long history of use in orthopedic procedures or provided biocompatibility data consistent with ISO 10993 requirements	Uses materials with a long history of use in orthopedic procedures or provided biocompatibility data consistent with ISO 10993 requirements	SAME
Technological Characteristics			
<ul style="list-style-type: none"> Major System Components 	Planning, web-based plan review, and robot control software, robotic positioning device, navigation system, reusable and disposable instrumentation	Planning, web-based plan review, and robot control software, robotic positioning device, navigation system, reusable and disposable instrumentation	SAME



**TMINI® Miniature Robotic System
Special 510(k) Submission**

Product	TMINI® Miniature Robotic System (AIM 3-2)	TMINI® Miniature Robotic System (AIM 3-0)	Conclusion
510(k) number	Subject Device	K252544	
Manufacturer	THINK Surgical Inc.	THINK Surgical Inc.	
<ul style="list-style-type: none"> • Patient Imaging 	CT images used to create a 3D model of the bone for surgical planning	CT images used to create a 3D model of the bone for surgical planning	SAME
<ul style="list-style-type: none"> • Preoperative planning workstation 	TPLAN three-dimensional preoperative planning workstation	TPLAN three-dimensional preoperative planning workstation	SAME
<ul style="list-style-type: none"> • Implant Module creation, verification and validation for each compatible implant system 	Proprietary method of generation, verification and validation of new implant modules	Proprietary method of generation, verification and validation of new implant modules	SAME
<ul style="list-style-type: none"> • Surgical planning system 	Technician guided surgical planning with surgeon review and approval on a desktop planning station	Technician guided surgical planning with surgeon review and approval on a desktop planning station	SAME
<ul style="list-style-type: none"> • Surgical planning review, approval, and approved plan export or download 	Performed by surgeon on either the TPLAN Planning Station or the THINK Case Manager (TCM)	Performed by surgeon on either the TPLAN Planning Station or the THINK Case Manager (TCM)	SAME
<ul style="list-style-type: none"> • Bone Marker Arrays for bone registration and tracking 	Active markers on femur and tibia mounted onto the bones via an attachment assembly	Active markers on femur and tibia mounted onto the bones via an attachment assembly	SAME
<ul style="list-style-type: none"> • Surgical Exposure 	Similar to traditional surgical exposure	Similar to traditional surgical exposure	SAME
<ul style="list-style-type: none"> • Patient/Robot Registration 	Preoperatively determined landmarks are compared to intraoperatively identified landmarks to complete patient bone registration	Preoperatively determined landmarks are compared to intraoperatively identified landmarks to complete patient bone registration	SAME
<ul style="list-style-type: none"> • Camera Tracking Technology 	Six camera overhead tracking with a wide-angle field of view	Six camera overhead tracking with a wide-angle field of view	SAME
<ul style="list-style-type: none"> • Cut guide positioning 	Robotic device places bone pins in the correct plane, then cutguide or drill block is attached to the pins and bone	Robotic device places bone pins in the correct plane, then cutguide or drill block is attached to the pins and bone	SAME
<ul style="list-style-type: none"> • Intraoperative planning changes 	Implant position can be fully adjusted, allowing deviation from the intended implant positioning philosophy and implant size	Implant position can be fully adjusted, allowing deviation from the intended implant positioning philosophy and implant size	SAME
<ul style="list-style-type: none"> • Bone Preparation Technique 	A surgical saw is used to cut the bone through a cut guide	A surgical saw is used to cut the bone through a cut guide	SAME
<ul style="list-style-type: none"> • Intraoperative Anatomic Measurements 	The tracked bone arrays and bone registration data are used to determine the knee flexion angle and varus/valgus laxity	The tracked bone arrays and bone registration data are used to determine the knee flexion angle and varus/valgus laxity	SAME
<ul style="list-style-type: none"> • Gap Balancing 	Displays the maximum space in the medial and lateral compartments in millimeters with the knee in extension and in flexion allowing the surgeon to perform gap balancing, if desired	Displays the maximum space in the medial and lateral compartments in millimeters with the knee in extension and in flexion allowing the surgeon to perform gap balancing, if desired	SAME



**TMINI® Miniature Robotic System
Special 510(k) Submission**

Product	TMINI® Miniature Robotic System (AIM 3-2)	TMINI® Miniature Robotic System (AIM 3-0)	Conclusion
510(k) number	Subject Device	K252544	
Manufacturer	THINK Surgical Inc.	THINK Surgical Inc.	
<ul style="list-style-type: none"> TKA Component Implantation Technique 	Implants are secured to the bone, either with or without cement using standard surgical technique provided by the implant manufacturer	Implants are secured to the bone, either with or without cement using standard surgical technique provided by the implant manufacturer	SAME
<ul style="list-style-type: none"> Compatible Knee Implant Systems 	<ul style="list-style-type: none"> Enovis™ EMPOWR Knee System® Ortho Development® BKS® and BKS TriMax® Knee System Total Joint Orthopedics Classic® Knee System United® U2™ Knee System Medacta® GMK® Sphere / SpheriKA Knee Systems Zimmer Biomet Anterior & Posterior Referencing Persona® Knee b-ONE MOBIO® Total Knee System Maxx Orthopedics Freedom® Total & Titan Knee LINK® LinkSymphoKnee System Stryker® Triathlon® Knee System 	<ul style="list-style-type: none"> Enovis™ EMPOWR Knee System® Ortho Development® BKS® and BKS TriMax® Knee System Total Joint Orthopedics Classic® Knee System United® U2™ Knee System Medacta® GMK® Sphere / SpheriKA Knee Systems Zimmer Biomet Anterior & Posterior Referencing Persona® Knee b-ONE MOBIO® Total Knee System Maxx Orthopedics Freedom® Total & Titan Knee LINK® LinkSymphoKnee System 	Substantially Equivalent
Performance Testing			
Full System Run Through Testing	Passed	Passed	SAME
Cutting Accuracy			SAME
<ul style="list-style-type: none"> Dimensional Accuracy Pin & Block Placement Accuracy Cadaver Lab Validation Testing System Gap Balance Accuracy 	Passed Unchanged Passed Unchanged	Passed Unchanged Passed Unchanged	
User Needs Validation Testing	Unchanged	Passed	SAME
Usability Testing	Unchanged	Passed	SAME
System Software Testing	Passed	Passed	SAME
Implant System Module Generation, Verification & Validation Testing	Passed	Passed	SAME
*Biocompatibility Testing			
<ul style="list-style-type: none"> Cytotoxicity Sensitization Intracutaneous Reactivity Acute Systemic Toxicity Pyrogenicity 	*Unchanged *Unchanged *Unchanged *Unchanged *Unchanged	Passed Passed Passed Passed Passed	SAME SAME SAME SAME SAME



TMINI® Miniature Robotic System Special 510(k) Submission

* There are no material changes to any of the direct patient contact components of the TMINI® Miniature Robotic System as a result of TMINI System (AIM 3-2) modifications included in this submission; therefore, no additional biocompatibility testing was required.

Risk assessment was performed on the device in accordance with ISO 14971:2019 and THINK Surgical Risk Management procedures. Changes were assessed separately and in the aggregate through risk analysis and appropriate performance testing to evaluate the impact of the change. Risk analysis resulted in the identification of new instances of risk; however, no new clinical hazards were identified, and no new questions of safety or effectiveness were identified as a result of these changes. The risks identified have been mitigated to acceptable levels and there is no change in the overall risk profile of the device compared to the predicate cleared in K252544.

Conclusion

The TMINI® Miniature Robotic System (AIM 3-2) is substantially equivalent to the predicate, TMINI® Miniature Robotic System (K252544), in the following ways:

- it has the same intended use,
- it has the same Indication for Use (except for the addition of one more compatible implant system to the list of compatible implants it has the same technological characteristics and operating principles and incorporates the same design and materials.

Performance testing and risk analysis has demonstrated that the performance and risk profile of the TMINI® Miniature Robotic System (AIM 3-2) is substantially equivalent to that of the predicate device and does not raise any new question of safety and effectiveness.

THINK Surgical Inc. respectfully submits that this filing contains adequate information and data to demonstrate the substantial equivalence of the TMINI® Miniature Robotic System (AIM 3-2) to the legally marketed TMINI® Miniature Robotic System cleared via K252544.